

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
 Data File : BP018685.D
 Acq On : 07 Dec 2023 21:59
 Operator : MA/JU
 Sample : 05459-15
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AK1

Quant Time: Dec 07 22:29:32 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	254907	20.000	ng/u1	-0.01
20) Naphthalene-d8	10.640	136	928793	20.000	ng/u1	-0.01
38) Acenaphthene-d10	14.475	164	553011	20.000	ng/u1	-0.01
64) Phenanthrene-d10	17.222	188	1325506	20.000	ng/u1	0.00
79) Chrysene-d12	21.310	240	1542377	20.000	ng/u1	0.00
88) Perylene-d12	23.680	264	1812134	20.000	ng/u1	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	30475	4.074	ng/uL	0.00
4) Pyridine-d5	3.681	84	424144	21.169	ng/u1	0.00
7) Phenol-d5	7.011	99	518095	20.253	ng/u1	-0.01
9) Bis-(2-Chloroethyl)eth...	7.181	67	339286	22.378	ng/u1	-0.01
11) 2-Chlorophenol-d4	7.375	132	432356	21.660	ng/u1	-0.01
15) 4-Methylphenol-d8	8.557	113	438883	21.200	ng/u1	-0.01
21) Nitrobenzene-d5	9.005	128	185180	22.401	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.728	143	193650	22.595	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.263	165	413406	23.590	ng/u1	-0.01
31) 4-Chloroaniline-d4	10.781	131	585655	25.571	ng/u1	-0.01
46) Dimethylphthalate-d6	13.893	166	1349363	27.339	ng/u1	-0.01
49) Acenaphthylene-d8	14.169	160	1487782	27.375	ng/u1	-0.01
54) 4-Nitrophenol-d4	14.681	143	56525	6.977	ng/u1	0.00
60) Fluorene-d10	15.469	176	1219745	29.441	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.587	200	37444	4.711	ng/u1	0.00
73) Anthracene-d10	17.322	188	2085945	29.934	ng/u1	0.00
81) Pyrene-d10	19.557	212	2859932	23.821	ng/u1	-0.01
92) Benzo(a)pyrene-d12	23.527	264	3271111	30.959	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120723\
Data File : BP018685.D
Acq On : 07 Dec 2023 21:59
Operator : MA/JU
Sample : 05459-15
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AK1

Quant Time: Dec 07 22:29:32 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 04 21:39:40 2023
Response via : Initial Calibration

