

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120921\
 Data File : BP008318.D
 Acq On : 09 Dec 2021 18:52
 Operator : CG/JU
 Sample : M4965-32
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 FLOORS-COMP

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/10/2021
 Supervised By :mohammad ahmed 12/15/2021

Quant Time: Dec 10 04:09:40 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 09 01:43:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	72401	20.000	ng	0.00
21) Naphthalene-d8	10.563	136	250205	20.000	ng	0.00
39) Acenaphthene-d10	14.422	164	147674	20.000	ng	0.00
64) Phenanthrene-d10	17.181	188	297856	20.000	ng	-0.01
76) Chrysene-d12	21.292	240	335295	20.000	ng	0.00
86) Perylene-d12	23.680	264	394409	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.369	112	601690	133.807	ng	0.00
7) Phenol-d6	6.958	99	667301	109.930	ng	0.00
23) Nitrobenzene-d5	8.928	82	582423	105.852	ng	-0.01
42) 2,4,6-Tribromophenol	15.922	330	285233	151.115	ng	0.00
45) 2-Fluorobiphenyl	13.040	172	1074214	105.661	ng	0.00
79) Terphenyl-d14	19.757	244	1621583	101.075	ng	-0.01
Target Compounds						
32) Benzoic acid	9.940	122	11637m	4.132	ng	Qvalue
84) Bis(2-ethylhexyl)phtha...	21.204	149	92091	6.353	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP120921\
Data File : BP008318.D
Acq On : 09 Dec 2021 18:52
Operator : CG/JU
Sample : M4965-32
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
FLOORS-COMP

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/10/2021
Supervised By :mohammad ahmed 12/15/2021

Quant Time: Dec 10 04:09:40 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 09 01:43:36 2021
Response via : Initial Calibration

