

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121021\
 Data File : BP008354.D
 Acq On : 10 Dec 2021 22:48
 Operator : CG/JU
 Sample : M4964-01
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 LAUNDRYROOM-PIT-BOTTOM

Quant Time: Dec 11 04:02:17 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Dec 11 03:53:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.757	152	67422	20.000	ng	0.00
21) Naphthalene-d8	10.557	136	211729	20.000	ng	0.00
39) Acenaphthene-d10	14.410	164	126814	20.000	ng	0.00
64) Phenanthrene-d10	17.175	188	276049	20.000	ng	0.00
76) Chrysene-d12	21.286	240	366345	20.000	ng	0.00
86) Perylene-d12	23.663	264	384751	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	412232	98.444	ng	0.00
7) Phenol-d6	6.952	99	515843	91.255	ng	0.00
23) Nitrobenzene-d5	8.922	82	358956	77.093	ng	0.00
42) 2,4,6-Tribromophenol	15.916	330	173087	106.785	ng	0.00
45) 2-Fluorobiphenyl	13.028	172	638553	73.140	ng	0.00
79) Terphenyl-d14	19.751	244	1098446	62.664	ng	0.00
Target Compounds						
84) Bis(2-ethylhexyl)phtha...	21.198	149	101676	6.419	ng	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121021\
Data File : BP008354.D
Acq On : 10 Dec 2021 22:48
Operator : CG/JU
Sample : M4964-01
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
LAUNDRYROOM-PIT-BOTTOM

Quant Time: Dec 11 04:02:17 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 03:53:01 2021
Response via : Initial Calibration

