



2-Nitrophenol



Response = 1.784e-001 * Amt - 2.354e-002
 Coef of Det (r^2) = 0.993220 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M
 Calibration Table Last Updated: Tue Dec 15 10:27:21 2020


Benzoic acid



Response = $2.052e-001 * \text{Amt} - 7.447e-002$
 Coef of Det (r^2) = 0.991829 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M
 Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

2,4,6-Trichlorophenol

Response Ratio



$$\text{Response} = 4.539\text{e-}001 * \text{Amt} - 3.374\text{e-}002$$


Coef of Det (r^2) = 0.991470 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M

Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

2,4,5-Trichlorophenol

Response Ratio



$$\text{Response} = 4.761\text{e-}001 * \text{Amt} - 2.183\text{e-}002$$


Coef of Det (r^2) = 0.992221 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M

Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

2-Nitroaniline

Response Ratio




$$\text{Response} = 3.932\text{e-}001 * \text{Amt} - 2.137\text{e-}002$$

Coef of Det (r^2) = 0.991143 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M


Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

2,4-Dinitrophenol



Response = $1.726 \times 10^{-1} \times \text{Amt} - 4.554 \times 10^{-2}$
 Coef of Det (r^2) = 0.993515 Curve Fit: Linear
 Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M
 Calibration Table Last Updated: Tue Dec 15 10:27:21 2020


4,6-Dinitro-2-methylphenol



Response = $1.079 \times 10^{-1} \times \text{Amt} - 2.246 \times 10^{-2}$
Coef of Det (r^2) = 0.991840 Curve Fit: Linear
Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M
Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

Butylbenzylphthalate

Response Ratio



$$\text{Response} = 5.530\text{e-}001 * \text{Amt} - 9.849\text{e-}002$$


Coef of Det (r^2) = 0.992397 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M

Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

3,3'-Dichlorobenzidine

Response Ratio



$$\text{Response} = 4.874\text{e-}001 * \text{Amt} - 7.620\text{e-}002$$


Coef of Det (r^2) = 0.998618 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M

Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 8.567\text{e-}001 * \text{Amt} - 1.462\text{e-}001$$


Coef of Det (r^2) = 0.998475 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M

Calibration Table Last Updated: Tue Dec 15 10:27:21 2020

Di-n-octyl phthalate

Response Ratio



$$\text{Response} = 1.431\text{e}+000 * \text{Amt} - 3.217\text{e}-001$$

Coef of Det (r^2) = 0.995910 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270E-BP121120.M

Calibration Table Last Updated: Tue Dec 15 10:27:21 2020