

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121120\
 Data File : BP004228.D
 Acq On : 11 Dec 2020 14:48
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP121120

Quant Time: Dec 15 10:35:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP121120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 10:27:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	111	0.00
2	1,4-Dioxane	40.000	39.799	0.5	109	0.00
3	Pyridine	40.000	41.994	-5.0	112	0.00
4	n-Nitrosodimethylamine	40.000	42.289	-5.7	111	0.00
5 S	2-Fluorophenol	80.000	85.922	-7.4	114	0.00
6	Aniline	40.000	43.471	-8.7	114	0.01
7 S	Phenol-d6	80.000	87.087	-8.9	114	0.01
8	2-Chlorophenol	40.000	42.893	-7.2	114	0.01
9	Benzaldehyde	40.000	37.255	6.9	112	0.01
10 C	Phenol	40.000	43.039	-7.6	113	0.01
11	bis(2-Chloroethyl)ether	40.000	41.571	-3.9	111	0.01
12	1,3-Dichlorobenzene	40.000	41.013	-2.5	112	0.01
13 C	1,4-Dichlorobenzene	40.000	41.265	-3.2	113	0.00
14	1,2-Dichlorobenzene	40.000	40.880	-2.2	111	0.01
15	Benzyl Alcohol	40.000	45.800	-14.5	117	0.01
16	2,2'-oxybis(1-Chloropropane	40.000	40.887	-2.2	110	0.01
17	2-Methylphenol	40.000	43.913	-9.8	114	0.01
18	Hexachloroethane	40.000	41.159	-2.9	112	0.01
19 P	n-Nitroso-di-n-propylamine	40.000	43.178	-7.9	113	0.01
20	3+4-Methylphenols	40.000	44.786	-12.0	114	0.01
21 I	Naphthalene-d8	20.000	20.000	0.0	113	0.00
22	Acetophenone	40.000	41.349	-3.4	112	0.01
23 S	Nitrobenzene-d5	80.000	83.402	-4.3	113	0.01
24	Nitrobenzene	40.000	41.776	-4.4	113	0.01
25	Isophorone	40.000	43.444	-8.6	114	0.01
26 C	2-Nitrophenol	40.000	42.871	-7.2	121	0.01
27	2,4-Dimethylphenol	40.000	43.676	-9.2	114	0.01
28	bis(2-Chloroethoxy)methane	40.000	41.834	-4.6	113	0.01
29 C	2,4-Dichlorophenol	40.000	45.270	-13.2	116	0.01
30	1,2,4-Trichlorobenzene	40.000	41.321	-3.3	113	0.01
31	Naphthalene	40.000	40.957	-2.4	112	0.01
32	Benzoic acid	40.000	41.910	-4.8	123	0.02
33	4-Chloroaniline	40.000	43.014	-7.5	113	0.01
34 C	Hexachlorobutadiene	40.000	41.233	-3.1	113	0.01
35	Caprolactam	40.000	44.958	-12.4	119	0.02
36 C	4-Chloro-3-methylphenol	40.000	45.193	-13.0	116	0.01
37	2-Methylnaphthalene	40.000	41.344	-3.4	112	0.00
38	1-Methylnaphthalene	40.000	41.299	-3.2	112	0.01
39 I	Acenaphthene-d10	20.000	20.000	0.0	113	0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	41.461	-3.7	113	0.01
41 P	Hexachlorocyclopentadiene	40.000	43.428	-8.6	115	0.01
42 S	2,4,6-Tribromophenol	80.000	86.985	-8.7	117	0.01
43 C	2,4,6-Trichlorophenol	40.000	42.847	-7.1	118	0.01
44	2,4,5-Trichlorophenol	40.000	42.233	-5.6	116	0.01
45 S	2-Fluorobiphenyl	80.000	81.489	-1.9	112	0.00
46	1,1'-Biphenyl	40.000	40.937	-2.3	113	0.01
47	2-Chloronaphthalene	40.000	41.135	-2.8	113	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121120\
 Data File : BP004228.D
 Acq On : 11 Dec 2020 14:48
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP121120

Quant Time: Dec 15 10:35:15 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP121120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 15 10:27:21 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	42.609	-6.5	116	0.01
49	Acenaphthylene	40.000	41.916	-4.8	113	0.01
50	Dimethylphthalate	40.000	42.149	-5.4	114	0.01
51	2,6-Dinitrotoluene	40.000	43.635	-9.1	114	0.01
52 C	Acenaphthene	40.000	41.417	-3.5	113	0.01
53	3-Nitroaniline	40.000	44.964	-12.4	115	0.01
54 P	2,4-Dinitrophenol	40.000	42.429	-6.1	124	0.01
55	Dibenzofuran	40.000	40.796	-2.0	112	0.01
56 P	4-Nitrophenol	40.000	44.493	-11.2	118	0.01
57	2,4-Dinitrotoluene	40.000	42.575	-6.4	115	0.01
58	Fluorene	40.000	41.170	-2.9	112	0.01
59	2,3,4,6-Tetrachlorophenol	40.000	45.479	-13.7	117	0.01
60	Diethylphthalate	40.000	42.770	-6.9	114	0.01
61	4-Chlorophenyl-phenylether	40.000	41.301	-3.3	113	0.00
62	4-Nitroaniline	40.000	45.000	-12.5	114	0.00
63	Azobenzene	40.000	42.071	-5.2	112	0.01
64 I	Phenanthrene-d10	20.000	20.000	0.0	113	0.01
65	4,6-Dinitro-2-methylphenol	40.000	43.149	-7.9	124	0.01
66 c	n-Nitrosodiphenylamine	40.000	42.409	-6.0	114	0.01
67	4-Bromophenyl-phenylether	40.000	42.242	-5.6	114	0.01
68	Hexachlorobenzene	40.000	41.407	-3.5	114	0.01
69	Atrazine	40.000	44.105	-10.3	113	0.01
70 C	Pentachlorophenol	40.000	45.364	-13.4	121	0.00
71	Phenanthrene	40.000	41.018	-2.5	113	0.01
72	Anthracene	40.000	41.876	-4.7	112	0.01
73	Carbazole	40.000	42.419	-6.0	112	0.01
74	Di-n-butylphthalate	40.000	45.640	-14.1	115	0.01
75 C	Fluoranthene	40.000	41.968	-4.9	112	0.01
76 I	Chrysene-d12	20.000	20.000	0.0	111	0.01
77	Benzydine	40.000	56.603	-41.5#	115	0.00
78	Pyrene	40.000	42.594	-6.5	111	0.01
79 S	Terphenyl-d14	80.000	82.367	-3.0	110	0.01
80	Butylbenzylphthalate	40.000	43.265	-8.2	120	0.00
81	Benzo(a)anthracene	40.000	41.825	-4.6	111	0.02
82	3,3'-Dichlorobenzidine	40.000	42.125	-5.3	114	0.01
83	Chrysene	40.000	41.464	-3.7	111	0.01
84	Bis(2-ethylhexyl)phthalate	40.000	41.698	-4.2	115	0.01
85 c	Di-n-octyl phthalate	40.000	41.635	-4.1	118	0.01
86 I	Perylene-d12	20.000	20.000	0.0	113	0.02
87	Indeno(1,2,3-cd)pyrene	40.000	42.853	-7.1	114	0.02
88	Benzo(b)fluoranthene	40.000	41.312	-3.3	109	0.01
89	Benzo(k)fluoranthene	40.000	41.997	-5.0	116	0.01
90 C	Benzo(a)pyrene	40.000	42.944	-7.4	113	0.02
91	Dibenzo(a,h)anthracene	40.000	42.484	-6.2	113	0.03
92	Benzo(g,h,i)perylene	40.000	42.357	-5.9	113	0.03

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121120\
Data File : BP004228.D
Acq On : 11 Dec 2020 14:48
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP121120

Quant Time: Dec 15 10:35:15 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270E-BP121120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 15 10:27:21 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0