

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
 Data File : BP018807.D
 Acq On : 13 Dec 2023 14:21
 Operator : MA/JU
 Sample : PB157666BS
 Misc :
 ALS Vial : 87 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS666

Quant Time: Dec 13 21:36:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 11 22:02:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	242306	20.000	ng/u1	0.00
20) Naphthalene-d8	10.628	136	940824	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	596532	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	1361975	20.000	ng/u1	-0.01
79) Chrysene-d12	21.304	240	1389879	20.000	ng/u1	0.00
88) Perylene-d12	23.669	264	1559609	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	39168	5.508	ng/uL	0.00
4) Pyridine-d5	3.681	84	505378	26.535	ng/u1	0.00
7) Phenol-d5	7.005	99	619475	25.476	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.175	67	361100	25.056	ng/u1	0.00
11) 2-Chlorophenol-d4	7.369	132	477785	25.181	ng/u1	0.00
15) 4-Methylphenol-d8	8.552	113	473394	24.056	ng/u1	0.00
21) Nitrobenzene-d5	8.999	128	220763	26.364	ng/u1	0.00
24) 2-Nitrophenol-d4	9.716	143	233152	26.856	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.252	165	482460	27.178	ng/u1	0.00
31) 4-Chloroaniline-d4	10.775	131	593404	25.578	ng/u1	0.00
46) Dimethylphthalate-d6	13.881	166	1394256	26.188	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1579521	26.943	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	253555	29.013	ng/u1	0.00
60) Fluorene-d10	15.457	176	1221743	27.338	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.581	200	217939	26.687	ng/u1	0.00
73) Anthracene-d10	17.310	188	1994228	27.852	ng/u1	-0.01
81) Pyrene-d10	19.551	212	2450628	22.651	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.516	264	2504627	27.543	ng/u1	-0.01
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.299	88	82034	10.521	ng/uL	97
5) Pyridine	3.699	79	549155	28.552	ng/u1	97
6) Benzaldehyde	6.975	77	312858	24.920	ng/u1	98
8) Phenol	7.034	94	661841	27.755	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.269	93	530725	27.100	ng/u1	100
12) 2-Chlorophenol	7.399	128	517609	26.955	ng/u1	99
13) 2-Methylphenol	8.281	108	485917	26.617	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.363	45	625914	25.189	ng/u1	97
16) Acetophenone	8.663	105	771133	26.294	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.652	70	367574	22.783	ng/u1	98
18) 4-Methylphenol	8.610	108	523384	26.255	ng/u1	99
19) Hexachloroethane	8.910	117	211633	26.883	ng/u1	92
22) Nitrobenzene	9.040	77	575179	28.252	ng/u1	98
23) Isophorone	9.569	82	1042914	25.124	ng/u1	98
25) 2-Nitrophenol	9.746	139	271823	29.243	ng/u1	98
26) 2,4-Dimethylphenol	9.810	107	491141	27.016	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.052	93	686514	26.196	ng/u1	99
29) 2,4-Dichlorophenol	10.281	162	502607	29.394	ng/u1	98
30) Naphthalene	10.681	128	1609627	28.203	ng/u1	100
32) 4-Chloroaniline	10.799	127	520077	23.521	ng/u1	100
33) Hexachlorobutadiene	10.963	225	357165	31.189	ng/u1	99
34) Caprolactam	11.575	113	156201	29.792	ng/u1	90
35) 4-Chloro-3-methylphenol	11.922	107	511151	26.685	ng/u1	99
36) 2-Methylnaphthalene	12.293	142	1094068	27.058	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
 Data File : BP018807.D
 Acq On : 13 Dec 2023 14:21
 Operator : MA/JU
 Sample : PB157666BS
 Misc :
 ALS Vial : 87 Sample Multiplier: 1


Instrument :
 BNA_P
 ClientSampleId :
 SLCS666


Quant Time: Dec 13 21:36:57 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 11 22:02:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.510	142	1077486	26.388	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.657	216	663425	30.653	ng/ul	99
40) Hexachlorocyclopentadiene	12.634	237	332796	25.942	ng/ul	98
41) 2,4,6-Trichlorophenol	12.904	196	402712	28.925	ng/ul	97
42) 2,4,5-Trichlorophenol	12.969	196	448418	29.708	ng/ul	100
43) 1,1'-Biphenyl	13.304	154	1452689	28.157	ng/ul	99
44) 2-Chloronaphthalene	13.346	162	1181349	28.992	ng/ul	98
45) 2-Nitroaniline	13.551	65	309205	27.992	ng/ul	98
47) Dimethylphthalate	13.934	163	1486841	28.370	ng/ul	100
48) 2,6-Dinitrotoluene	14.051	165	293899	29.682	ng/ul	94
50) Acenaphthylene	14.187	152	1895691	28.760	ng/ul	100
51) 3-Nitroaniline	14.375	138	291728	29.515	ng/ul	99
52) Acenaphthene	14.528	153	1254748	28.791	ng/ul	99
53) 2,4-Dinitrophenol	14.581	184	137306	26.165	ng/ul	95
55) 4-Nitrophenol	14.687	109	207227	31.386	ng/ul	97
56) Dibenzofuran	14.863	168	1769023	29.206	ng/ul	98
57) 2,4-Dinitrotoluene	14.834	165	441363	31.689	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.092	232	379728	29.992	ng/ul	92
59) Diethylphthalate	15.292	149	1458841	28.446	ng/ul	99
61) Fluorene	15.516	166	1441565	30.077	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.510	204	764594	30.417	ng/ul	95
63) 4-Nitroaniline	15.540	138	320921	34.082	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.598	198	251867	28.847	ng/ul#	90
67) N-Nitrosodiphenylamine	15.728	169	1240622	27.489	ng/ul	99
68) 4-Bromophenyl-phenylether	16.404	248	489978	28.496	ng/ul	96
69) Hexachlorobenzene	16.516	284	579985	30.212	ng/ul	96
70) Atrazine	16.681	200	318450	23.733	ng/ul	99
71) Pentachlorophenol	16.863	266	336015	29.430	ng/ul	99
72) Phenanthrene	17.257	178	2418722	29.404	ng/ul	100
74) Anthracene	17.345	178	2447452	29.683	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.263	216	664733	28.227	ng/uL	98
76) Pentachlorobenzene	14.781	250	661412	28.592	ng/uL	100
77) Carbazole	17.622	167	2209976	33.630	ng/ul	99
78) Di-n-butylphthalate	18.175	149	2524147	28.847	ng/ul	100
80) Fluoranthene	19.228	202	2996163	23.428	ng/ul	97
82) Pyrene	19.581	202	3121153	24.192	ng/ul	95
83) Butylbenzylphthalate	20.457	149	1164312	24.651	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.227	252	1018645	30.159	ng/ul	99
85) Benzo(a)anthracene	21.286	228	3279208	29.067	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.216	149	1685969	25.971	ng/ul	100
87) Chrysene	21.339	228	3034693	29.211	ng/ul	100
89) Di-n-octyl phthalate	22.127	149	2962202	26.014	ng/ul	100
90) Benzo(b)fluoranthene	22.945	252	3249912	28.262	ng/ul	99
91) Benzo(k)fluoranthene	22.992	252	3264677	29.342	ng/ul	99
93) Benzo(a)pyrene	23.563	252	3088368	29.393	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.121	276	3768931	30.022	ng/ul	96
95) Dibenzo(a,h)anthracene	26.145	278	3142475	30.091	ng/ul	97
96) Benzo(g,h,i)perylene	26.874	276	3026008	29.952	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Dec 13 21:36:57 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 11 22:02:13 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 10.521 ng/uL
RT: 3.299 min Scan# 30
Delta R.T. 0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:	88	Resp:	82034
Ion	Ratio	Lower	Upper
88	100		
43	25.7	21.8	32.6
58	67.8	56.5	84.7



#5
Pyridine
Concen: 28.552 ng/uL
RT: 3.699 min Scan# 104
Delta R.T. 0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:	79	Resp:	549155
Ion	Ratio	Lower	Upper
79	100		
52	68.3	57.4	86.0
51	31.6	26.2	39.4









#6
Benzaldehyde
Concen: 24.920 ng/ul
RT: 6.975 min Scan# 601
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion: 77 Resp: 312858
Ion Ratio Lower Upper
77 100
105 99.4 77.9 116.9
106 94.4 74.6 111.8



#8
Phenol
Concen: 27.755 ng/ul
RT: 7.034 min Scan# 671
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion: 94 Resp: 661841
Ion Ratio Lower Upper
94 100
65 25.9 21.0 31.6
66 35.7 28.6 43.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 27.100 ng/ul
 RT: 7.269 min Scan# 711
 Delta R.T. 0.006 min
 Lab File: BP018807.D
 Acq: 13 Dec 2023 14:21




Instrument :
 BNA_P
 ClientSampleId :
 SLCS666


Tgt Ion: 93 Resp: 530725
 Ion Ratio Lower Upper
 93 100
 63 72.4 58.3 87.5
 95 32.3 25.9 38.9



#12
 2-Chlorophenol
 Concen: 26.955 ng/ul
 RT: 7.399 min Scan# 733
 Delta R.T. 0.000 min
 Lab File: BP018807.D
 Acq: 13 Dec 2023 14:21

Tgt Ion:128 Resp: 517609
 Ion Ratio Lower Upper
 128 100
 64 45.4 37.0 55.4
 130 32.6 26.8 40.2





#13

2-Methylphenol

Concen: 26.617 ng/ul

RT: 8.281 min Scan# 884

Delta R.T. 0.000 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :





SLCS666

Tgt Ion:108 Resp: 485917

Ion Ratio Lower Upper

108 100

107 88.5 70.6 105.8



#14

2,2'-oxybis(1-Chloropropane)

Concen: 25.189 ng/ul

RT: 8.363 min Scan# 897

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion: 45 Resp: 625914


Ion Ratio Lower Upper

45 100

77 17.6 12.7 19.1

79 12.3 10.5 15.7









#16
Acetophenone
Concen: 26.294 ng/ul
RT: 8.663 min Scan# 948
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:105 Resp: 771133
Ion Ratio Lower Upper
105 100
77 72.8 59.4 89.0
51 23.2 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 22.783 ng/ul
RT: 8.652 min Scan# 946
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion: 70 Resp: 367574
Ion Ratio Lower Upper
70 100
42 49.9 41.0 61.4
101 10.4 7.8 11.8
130 21.5 16.9 25.3









#18
4-Methylphenol
Concen: 26.255 ng/ul
RT: 8.610 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:108 Resp: 523384
Ion Ratio Lower Upper
108 100
107 112.4 90.6 136.0



#19
Hexachloroethane
Concen: 26.883 ng/ul
RT: 8.910 min Scan# 990
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:117 Resp: 211633
Ion Ratio Lower Upper
117 100
201 114.3 83.7 125.5
199 70.5 52.3 78.5









#22
Nitrobenzene
Concen: 28.252 ng/ul
RT: 9.040 min Scan# 1012
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion: 77 Resp: 575179
Ion Ratio Lower Upper
77 100
123 46.4 35.8 53.8
65 13.1 10.5 15.7



#23
Isophorone
Concen: 25.124 ng/ul
RT: 9.569 min Scan# 1102
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion: 82 Resp: 1042914
Ion Ratio Lower Upper
82 100
95 7.0 5.5 8.3
138 17.8 13.5 20.3









#25
2-Nitrophenol
Concen: 29.243 ng/ul
RT: 9.746 min Scan# 1133
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:139 Resp: 271823
Ion Ratio Lower Upper
139 100
65 47.9 39.1 58.7
109 31.3 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 27.016 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:107 Resp: 491141
Ion Ratio Lower Upper
107 100
121 50.5 40.0 60.0
122 85.1 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 26.196 ng/ul
 RT: 10.052 min Scan# 1184
 Delta R.T. -0.006 min
 Lab File: BP018807.D
 Acq: 13 Dec 2023 14:21




Instrument :
 BNA_P
 ClientSampleId :
 SLCS666


Tgt Ion: 93 Resp: 686514
 Ion Ratio Lower Upper
 93 100
 95 31.8 26.1 39.1
 123 10.9 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 29.394 ng/ul
 RT: 10.281 min Scan# 1223
 Delta R.T. 0.000 min
 Lab File: BP018807.D
 Acq: 13 Dec 2023 14:21

Tgt Ion:162 Resp: 502607
 Ion Ratio Lower Upper
 162 100
 164 64.9 51.7 77.5
 98 33.8 29.2 43.8









#30
Naphthalene
Concen: 28.203 ng/ul
RT: 10.681 min Scan# 1291
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:128 Resp: 1609627
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.1 10.5 15.7



#32
4-Chloroaniline
Concen: 23.521 ng/ul
RT: 10.799 min Scan# 1311
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:127 Resp: 520077
Ion Ratio Lower Upper
127 100
129 31.9 25.6 38.4









#33
Hexachlorobutadiene
Concen: 31.189 ng/ul
RT: 10.963 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:225 Resp: 357165
Ion Ratio Lower Upper
225 100
223 60.7 49.5 74.3
227 64.0 50.9 76.3



#34
Caprolactam
Concen: 29.792 ng/ul
RT: 11.575 min Scan# 1443
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:113 Resp: 156201
Ion Ratio Lower Upper
113 100
55 152.3 133.7 200.5
56 112.8 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 26.685 ng/ul
RT: 11.922 min Scan# 1502
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:107 Resp: 511151
Ion Ratio Lower Upper
107 100
144 26.1 21.0 31.6
142 80.7 63.6 95.4



#36
2-Methylnaphthalene
Concen: 27.058 ng/ul
RT: 12.293 min Scan# 1565
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:142 Resp: 1094068
Ion Ratio Lower Upper
142 100
141 89.4 71.4 107.0





#37

1-Methylnaphthalene

Concen: 26.388 ng/ul

RT: 12.510 min Scan# 1602

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:142 Resp: 1077486

Ion Ratio Lower Upper

142 100

141 91.9 73.5 110.3

116 3.7 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 30.653 ng/ul

RT: 12.657 min Scan# 1627

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Tgt Ion:216 Resp: 663425




Ion Ratio Lower Upper


216 100

214 78.3 63.9 95.9

179 20.0 16.3 24.5

108 15.7 12.6 19.0





#40

Hexachlorocyclopentadiene

Concen: 25.942 ng/ul

RT: 12.634 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:237 Resp: 332796

Ion Ratio Lower Upper

237 100

235 63.8 49.4 74.0

272 12.4 9.8 14.8



#41

2,4,6-Trichlorophenol

Concen: 28.925 ng/ul

RT: 12.904 min Scan# 1669

Delta R.T. 0.000 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:196 Resp: 402712


Ion Ratio Lower Upper

196 100

198 98.9 76.9 115.3

200 32.1 25.0 37.6





#42

2,4,5-Trichlorophenol

Concen: 29.708 ng/ul

RT: 12.969 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:196 Resp: 448418

Ion Ratio Lower Upper

196 100

198 97.3 77.9 116.9

200 30.9 25.2 37.8



#43

1,1'-Biphenyl

Concen: 28.157 ng/ul

RT: 13.304 min Scan# 1737

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:154 Resp: 1452689


Ion Ratio Lower Upper

154 100

153 41.1 32.8 49.2

76 12.7 10.7 16.1





#44

2-Chloronaphthalene

Concen: 28.992 ng/ul

RT: 13.346 min Scan# 1744

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:162 Resp: 1181349

Ion Ratio Lower Upper

162 100

164 32.9 25.7 38.5

127 36.0 30.3 45.5



#45

2-Nitroaniline

Concen: 27.992 ng/ul

RT: 13.551 min Scan# 1779

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion: 65 Resp: 309205


Ion Ratio Lower Upper

65 100

92 74.9 56.9 85.3

138 111.3 88.2 132.2









#47
Dimethylphthalate
Concen: 28.370 ng/ul
RT: 13.934 min Scan# 1844
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:163 Resp: 1486841
Ion Ratio Lower Upper
163 100
194 4.2 3.2 4.8
164 10.0 8.0 12.0



#48
2,6-Dinitrotoluene
Concen: 29.682 ng/ul
RT: 14.051 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:165 Resp: 293899
Ion Ratio Lower Upper
165 100
89 52.8 46.6 70.0
121 20.3 17.3 25.9









#50
Acenaphthylene
Concen: 28.760 ng/ul
RT: 14.187 min Scan# 1887
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:152 Resp: 1895691
Ion Ratio Lower Upper
152 100
151 19.7 15.9 23.9
153 13.4 10.6 16.0



#51
3-Nitroaniline
Concen: 29.515 ng/ul
RT: 14.375 min Scan# 1919
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:138 Resp: 291728
Ion Ratio Lower Upper
138 100
108 11.1 9.5 14.3
92 117.6 93.1 139.7









#52
Acenaphthene
Concen: 28.791 ng/ul
RT: 14.528 min Scan# 1945
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:153 Resp: 1254748
Ion Ratio Lower Upper
153 100
152 47.6 38.1 57.1
154 87.6 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 26.165 ng/ul
RT: 14.581 min Scan# 1954
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:184 Resp: 137306
Ion Ratio Lower Upper
184 100
63 54.9 48.2 72.4
154 58.5 49.1 73.7









#55
4-Nitrophenol
Concen: 31.386 ng/ul
RT: 14.687 min Scan# 1972
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:109 Resp: 207227
Ion Ratio Lower Upper
109 100
139 137.9 109.0 163.4
65 126.9 106.2 159.4



#56
Dibenzofuran
Concen: 29.206 ng/ul
RT: 14.863 min Scan# 2002
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:168 Resp: 1769023
Ion Ratio Lower Upper
168 100
139 35.4 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 31.689 ng/ul
RT: 14.834 min Scan# 1997
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:165 Resp: 441363
Ion Ratio Lower Upper
165 100
63 38.4 32.7 49.1
182 2.8 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 29.992 ng/ul
RT: 15.092 min Scan# 2041
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:232 Resp: 379728
Ion Ratio Lower Upper
232 100
131 36.1 33.8 50.8
130 2.1 1.8 2.8
166 27.0 24.7 37.1









#59
Diethylphthalate
Concen: 28.446 ng/ul
RT: 15.292 min Scan# 2076
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:149 Resp: 1458841
Ion Ratio Lower Upper
149 100
177 23.1 18.2 27.4
150 12.5 9.9 14.9



#61
Fluorene
Concen: 30.077 ng/ul
RT: 15.516 min Scan# 2113
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:166 Resp: 1441565
Ion Ratio Lower Upper
166 100
165 97.0 78.0 117.0
167 13.2 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 30.417 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:204 Resp: 764594
Ion Ratio Lower Upper
204 100
206 33.4 27.1 40.7
141 51.7 45.8 68.8



#63
4-Nitroaniline
Concen: 34.082 ng/ul
RT: 15.540 min Scan# 2117
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:138 Resp: 320921
Ion Ratio Lower Upper
138 100
92 52.4 42.3 63.5
108 80.1 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 28.847 ng/ul
RT: 15.598 min Scan# 2127
Delta R.T. 0.000 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:198 Resp: 251867
Ion Ratio Lower Upper
198 100
182 4.4 3.4 5.2
77 19.9 20.6 30.8#



#67
N-Nitrosodiphenylamine
Concen: 27.489 ng/ul
RT: 15.728 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:169 Resp: 1240622
Ion Ratio Lower Upper
169 100
168 67.5 53.1 79.7
167 37.2 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 28.496 ng/ul
RT: 16.404 min Scan# 21
Delta R.T. -0.012 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:248 Resp: 489978
Ion Ratio Lower Upper
248 100
250 97.6 79.3 118.9
141 64.8 57.2 85.8



#69
Hexachlorobenzene
Concen: 30.212 ng/ul
RT: 16.516 min Scan# 2283
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:284 Resp: 579985
Ion Ratio Lower Upper
284 100
142 28.5 25.8 38.6
249 30.6 24.9 37.3





#70

Atrazine

Concen: 23.733 ng/ul

RT: 16.681 min Scan# 2311

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:200 Resp: 318450

Ion Ratio Lower Upper

200 100

173 25.2 21.0 31.4

215 48.4 38.2 57.2



#71

Pentachlorophenol

Concen: 29.430 ng/ul

RT: 16.863 min Scan# 2342

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:266 Resp: 336015


Ion Ratio Lower Upper

266 100

264 61.8 50.2 75.2

268 63.3 50.5 75.7





#72

Phenanthrene

Concen: 29.404 ng/ul

RT: 17.257 min Scan# 2409

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:178 Resp: 2418722

Ion Ratio Lower Upper

178 100

179 15.3 12.3 18.5

176 19.4 15.4 23.2



#74

Anthracene

Concen: 29.683 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:178 Resp: 2447452


Ion Ratio Lower Upper

178 100

179 15.5 12.5 18.7

176 18.4 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 28.227 ng/uL
RT: 13.263 min Scan# 1730
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:216 Resp: 664733
Ion Ratio Lower Upper
216 100
214 79.1 61.4 92.2
218 48.3 39.2 58.8



#76
Pentachlorobenzene
Concen: 28.592 ng/uL
RT: 14.781 min Scan# 1988
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:250 Resp: 661412
Ion Ratio Lower Upper
250 100
248 63.4 50.6 75.8
252 64.4 51.3 76.9





#77

Carbazole

Concen: 33.630 ng/ul

RT: 17.622 min Scan# 2471

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:167 Resp: 2209976

Ion Ratio Lower Upper

167 100

166 21.8 17.1 25.7

139 12.4 10.2 15.4



#78

Di-n-butylphthalate

Concen: 28.847 ng/ul

RT: 18.175 min Scan# 2565

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:149 Resp: 2524147


Ion Ratio Lower Upper

149 100

150 9.4 7.4 11.2

104 5.0 3.9 5.9









#80
Fluoranthene
Concen: 23.428 ng/ul
RT: 19.228 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21




Instrument :
BNA_P
ClientSampleId :
SLCS666


Tgt Ion:202 Resp: 2996163
Ion Ratio Lower Upper
202 100
101 10.2 9.1 13.7
100 7.6 7.1 10.7



#82
Pyrene
Concen: 24.192 ng/ul
RT: 19.581 min Scan# 2804
Delta R.T. -0.006 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:202 Resp: 3121153
Ion Ratio Lower Upper
202 100
101 11.7 11.0 16.4
100 9.7 9.1 13.7





#83

Butylbenzylphthalate

Concen: 24.651 ng/ul

RT: 20.457 min Scan# 2953

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:149 Resp: 1164312

Ion Ratio Lower Upper

149 100

91 68.4 56.6 85.0

206 22.1 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 30.159 ng/ul

RT: 21.227 min Scan# 3084

Delta R.T. -0.006 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:252 Resp: 1018645


Ion Ratio Lower Upper

252 100

254 64.5 51.0 76.4

126 9.1 7.8 11.6





#85

Benzo(a)anthracene

Concen: 29.067 ng/ul

RT: 21.286 min Scan# 3094

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:228 Resp: 3279208

Ion Ratio Lower Upper

228 100

229 19.9 15.7 23.5

226 27.3 21.9 32.9



#86

Bis(2-ethylhexyl)phthalate

Concen: 25.971 ng/ul

RT: 21.216 min Scan# 3082

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:149 Resp: 1685969


Ion Ratio Lower Upper

149 100

167 27.8 22.2 33.2

279 4.4 3.3 4.9





#87

Chrysene

Concen: 29.211 ng/ul

RT: 21.339 min Scan# 3103

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:228 Resp: 3034693

Ion Ratio Lower Upper

228 100

226 30.0 23.9 35.9

229 19.5 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 26.014 ng/ul




RT: 22.127 min Scan# 3237


Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Tgt Ion:149 Resp: 2962202





#90

Benzo(b)fluoranthene

Concen: 28.262 ng/ul

RT: 22.945 min Scan# 3376

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:252 Resp: 3249912

Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 9.2 7.8 11.6



#91

Benzo(k)fluoranthene

Concen: 29.342 ng/ul

RT: 22.992 min Scan# 3384

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:252 Resp: 3264677


Ion Ratio Lower Upper

252 100

253 22.3 17.7 26.5

125 9.3 8.0 12.0





#93

Benzo(a)pyrene

Concen: 29.393 ng/ul

RT: 23.563 min Scan# 3481

Delta R.T. -0.012 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21

Instrument :

BNA_P

ClientSampleId :

SLCS666





Tgt Ion:252 Resp: 3088368

Ion Ratio Lower Upper

252 100

253 22.1 17.8 26.8

125 9.9 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 30.022 ng/ul

RT: 26.121 min Scan# 3916

Delta R.T. -0.018 min

Lab File: BP018807.D

Acq: 13 Dec 2023 14:21




Tgt Ion:276 Resp: 3768931


Ion Ratio Lower Upper

276 100

138 20.5 19.4 29.2

277 24.7 19.8 29.6









#95
Dibenzo(a,h)anthracene
Concen: 30.091 ng/ul
RT: 26.145 min Scan# 3920
Delta R.T. -0.018 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Instrument :
BNA_P
ClientSampleId :
SLCS666

Tgt Ion:278 Resp: 3142475
Ion Ratio Lower Upper
278 100
139 14.3 13.0 19.6
279 24.0 18.7 28.1



#96
Benzo(g,h,i)perylene
Concen: 29.952 ng/ul
RT: 26.874 min Scan# 4044
Delta R.T. -0.029 min
Lab File: BP018807.D
Acq: 13 Dec 2023 14:21

Tgt Ion:276 Resp: 3026008
Ion Ratio Lower Upper
276 100
138 19.6 18.1 27.1
277 23.9 19.4 29.0

