

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
 Data File : BP018728.D
 Acq On : 11 Dec 2023 15:12
 Operator : MA/JU
 Sample : 05459-04
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0AA2

Quant Time: Dec 11 15:48:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 04 21:39:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.840	152	187959	20.000	ng/ul	-0.02
20) Naphthalene-d8	10.634	136	709341	20.000	ng/ul	-0.02
38) Acenaphthene-d10	14.469	164	440901	20.000	ng/ul	-0.02
64) Phenanthrene-d10	17.216	188	1045225	20.000	ng/ul	-0.01
79) Chrysene-d12	21.310	240	1095598	20.000	ng/ul	0.00
88) Perylene-d12	23.674	264	1198160	20.000	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	4375	0.793	ng/uL	0.00
4) Pyridine-d5	3.687	84	44191	2.991	ng/ul	0.00
7) Phenol-d5	7.011	99	154063	8.168	ng/ul	-0.01
9) Bis-(2-Chloroethyl)eth...	7.175	67	98550	8.815	ng/ul	-0.02
11) 2-Chlorophenol-d4	7.369	132	119574	8.124	ng/ul	-0.02
15) 4-Methylphenol-d8	8.552	113	93749	6.141	ng/ul	-0.02
21) Nitrobenzene-d5	8.999	128	53229	8.431	ng/ul	-0.02
24) 2-Nitrophenol-d4	9.722	143	49512	7.564	ng/ul	-0.02
28) 2,4-Dichlorophenol-d3	10.258	165	110502	8.256	ng/ul	-0.02
31) 4-Chloroaniline-d4	10.781	131	126383	7.225	ng/ul	-0.01
46) Dimethylphthalate-d6	13.887	166	371627	9.444	ng/ul	-0.02
49) Acenaphthylene-d8	14.163	160	392198	9.051	ng/ul	-0.02
54) 4-Nitrophenol-d4	14.675	143	34517	5.344	ng/ul	-0.01
60) Fluorene-d10	15.463	176	338115	10.236	ng/ul	-0.01
65) 4,6-Dinitro-2-methylph...	15.581	200	27480	4.385	ng/ul	-0.01
73) Anthracene-d10	17.316	188	519412	9.453	ng/ul	-0.01
81) Pyrene-d10	19.557	212	695772	8.158	ng/ul	-0.01
92) Benzo(a)pyrene-d12	23.522	264	677559	9.699	ng/ul	-0.01
Target Compounds						
16) Acetophenone	8.669	105	86155	3.787	ng/ul	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
Data File : BP018728.D
Acq On : 11 Dec 2023 15:12
Operator : MA/JU
Sample : 05459-04
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0AA2

Quant Time: Dec 11 15:48:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 04 21:39:40 2023
Response via : Initial Calibration

