

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
 Data File : BP018805.D
 Acq On : 13 Dec 2023 13:13
 Operator : MA/JU
 Sample : 05626-22
 Misc :
 ALS Vial : 85 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCLG7

Quant Time: Dec 13 21:54:31 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 11 22:02:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	212367	20.000	ng/u1	0.00
20) Naphthalene-d8	10.628	136	905574	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	567081	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	1336931	20.000	ng/u1	-0.01
79) Chrysene-d12	21.298	240	1216997	20.000	ng/u1	-0.01
88) Perylene-d12	23.663	264	1529531	20.000	ng/u1	-0.02

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	30528	4.899	ng/uL	0.00
4) Pyridine-d5	3.681	84	193686	11.603	ng/u1	0.00
7) Phenol-d5	7.005	99	95786	4.495	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	303954	24.064	ng/u1	0.00
11) 2-Chlorophenol-d4	7.364	132	303809	18.269	ng/u1	0.00
15) 4-Methylphenol-d8	8.546	113	207291	12.019	ng/u1	0.00
21) Nitrobenzene-d5	8.999	128	207608	25.758	ng/u1	0.00
24) 2-Nitrophenol-d4	9.716	143	204135	24.429	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.252	165	393949	23.056	ng/u1	0.00
31) 4-Chloroaniline-d4	10.769	131	550214	24.640	ng/u1	0.00
46) Dimethylphthalate-d6	13.881	166	1368642	27.042	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1503723	26.982	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	31497	3.791	ng/u1	0.00
60) Fluorene-d10	15.457	176	1258409	29.621	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	149373	18.634	ng/u1	-0.01
73) Anthracene-d10	17.310	188	2276990	32.397	ng/u1	-0.01
81) Pyrene-d10	19.551	212	2778822	29.333	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.516	264	2835262	31.792	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
Data File : BP018805.D
Acq On : 13 Dec 2023 13:13
Operator : MA/JU
Sample : 05626-22
Misc :
ALS Vial : 85 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCLG7

Quant Time: Dec 13 21:54:31 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 11 22:02:13 2023
Response via : Initial Calibration

