

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
 Data File : BP018806.D
 Acq On : 13 Dec 2023 13:47
 Operator : MA/JU
 Sample : 05626-23
 Misc :
 ALS Vial : 86 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCLG8

Quant Time: Dec 13 21:36:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 11 22:02:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.834	152	231826	20.000	ng/u1	0.00
20) Naphthalene-d8	10.628	136	901993	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.463	164	568424	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.216	188	1151729	20.000	ng/u1	0.00
79) Chrysene-d12	21.304	240	1345719	20.000	ng/u1	0.00
88) Perylene-d12	23.663	264	1482760	20.000	ng/u1	-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.264	96	29433	4.326	ng/uL	0.00
4) Pyridine-d5	3.687	84	131509	7.217	ng/u1	0.01
7) Phenol-d5	7.005	99	132405	5.691	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	379851	27.548	ng/u1	0.00
11) 2-Chlorophenol-d4	7.363	132	375840	20.704	ng/u1	0.00
15) 4-Methylphenol-d8	8.546	113	233033	12.377	ng/u1	0.00
21) Nitrobenzene-d5	8.999	128	232888	29.009	ng/u1	0.00
24) 2-Nitrophenol-d4	9.716	143	222468	26.729	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.251	165	436004	25.618	ng/u1	0.00
31) 4-Chloroaniline-d4	10.775	131	529244	23.795	ng/u1	0.00
46) Dimethylphthalate-d6	13.881	166	1573099	31.008	ng/u1	0.00
49) Acenaphthylene-d8	14.157	160	1686133	30.183	ng/u1	0.00
54) 4-Nitrophenol-d4	14.675	143	36586	4.393	ng/u1	0.00
60) Fluorene-d10	15.457	176	1434168	33.678	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	166573	24.121	ng/u1	-0.01
73) Anthracene-d10	17.310	188	2228342	36.803	ng/u1	-0.01
81) Pyrene-d10	19.551	212	3185206	30.407	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.515	264	3181021	36.794	ng/u1	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121123\
Data File : BP018806.D
Acq On : 13 Dec 2023 13:47
Operator : MA/JU
Sample : 05626-23
Misc :
ALS Vial : 86 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCLG8

Quant Time: Dec 13 21:36:49 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 11 22:02:13 2023
Response via : Initial Calibration

