

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121125\
 Data File : BP026299.D
 Acq On : 11 Dec 2025 19:19
 Operator : RC/JU
 Sample : Q3826-08
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 TP-11

Quant Time: Dec 11 23:42:05 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Dec 05 14:52:44 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.649	152	239941	20.000	ng	0.00
21) Naphthalene-d8	10.407	136	895211	20.000	ng	-0.01
39) Acenaphthene-d10	14.284	164	548114	20.000	ng	0.00
64) Phenanthrene-d10	17.089	188	1088637	20.000	ng	0.00
76) Chrysene-d12	21.530	240	1262375	20.000	ng	0.00
86) Perylene-d12	24.824	264	1400638	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.278	112	1657077	111.030	ng	0.00
7) Phenol-d6	6.831	99	1800392	95.567	ng	0.00
23) Nitrobenzene-d5	8.778	82	1289840	82.012	ng	-0.01
42) 2,4,6-Tribromophenol	15.795	330	959159	134.561	ng	-0.01
45) 2-Fluorobiphenyl	12.884	172	3114562	82.614	ng	-0.02
79) Terphenyl-d14	19.836	244	5407154	87.548	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121125\
Data File : BP026299.D
Acq On : 11 Dec 2025 19:19
Operator : RC/JU
Sample : Q3826-08
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
TP-11

Quant Time: Dec 11 23:42:05 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Dec 05 14:52:44 2025
Response via : Initial Calibration

