




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121321\
Data File : BP008374.D
Acq On : 13 Dec 2021 10:10
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Dec 13 15:31:18 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP112521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Dec 11 03:53:01 2021
Response via : Initial Calibration



TIC: BP008374.D\data.ms

(70) Pentachlorophenol (C)

16.839min (-0.006) 748043.84 ng

response 116319

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.60	67.57
264.00	61.90	62.48
0.00	0.00	0.00