

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121323\
 Data File : BP018872.D
 Acq On : 15 Dec 2023 08:15
 Operator : MA/JU
 Sample : PB157656BS
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS656

Quant Time: Dec 15 10:07:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	207679	20.000	ng/u1	0.00
20) Naphthalene-d8	10.622	136	729967	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.457	164	408182	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.210	188	708517	20.000	ng/u1	0.00
79) Chrysene-d12	21.298	240	1047071	20.000	ng/u1	0.00
88) Perylene-d12	23.663	264	1226262	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.258	96	35950	5.899	ng/uL	0.00
4) Pyridine-d5	3.676	84	457375	28.018	ng/u1	0.00
7) Phenol-d5	7.005	99	517563	24.833	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	309157	25.028	ng/u1	0.00
11) 2-Chlorophenol-d4	7.364	132	415745	25.565	ng/u1	0.00
15) 4-Methylphenol-d8	8.546	113	382372	22.670	ng/u1	0.00
21) Nitrobenzene-d5	8.993	128	156099	24.027	ng/u1	0.00
24) 2-Nitrophenol-d4	9.710	143	195368	29.004	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.246	165	397128	28.833	ng/u1	0.00
31) 4-Chloroaniline-d4	10.769	131	470288	26.127	ng/u1	0.00
46) Dimethylphthalate-d6	13.875	166	985973	27.064	ng/u1	0.00
49) Acenaphthylene-d8	14.151	160	1190693	29.682	ng/u1	0.00
54) 4-Nitrophenol-d4	14.669	143	177308	29.650	ng/u1	0.00
60) Fluorene-d10	15.451	176	880582	28.796	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.575	200	155315	36.559	ng/u1	0.00
73) Anthracene-d10	17.310	188	1101513	29.572	ng/u1	0.00
81) Pyrene-d10	19.545	212	1839101	22.564	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.510	264	2130225	29.793	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.293	88	76247	11.409	ng/uL	98
5) Pyridine	3.693	79	483687	29.341	ng/u1	98
6) Benzaldehyde	6.975	77	272211	25.297	ng/u1	96
8) Phenol	7.028	94	557027	27.254	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.264	93	459992	27.404	ng/u1	100
12) 2-Chlorophenol	7.393	128	453088	27.529	ng/u1	99
13) 2-Methylphenol	8.281	108	399906	25.558	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.358	45	530364	24.902	ng/u1	98
16) Acetophenone	8.658	105	637134	25.347	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.646	70	295857	21.395	ng/u1	97
18) 4-Methylphenol	8.611	108	425566	24.907	ng/u1	97
19) Hexachloroethane	8.905	117	176742	26.194	ng/u1	90
22) Nitrobenzene	9.034	77	408121	25.837	ng/u1	99
23) Isophorone	9.563	82	755090	23.445	ng/u1	98
25) 2-Nitrophenol	9.740	139	225684	31.293	ng/u1	98
26) 2,4-Dimethylphenol	9.810	107	387530	27.474	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.046	93	551619	27.129	ng/u1	100
29) 2,4-Dichlorophenol	10.275	162	401020	30.228	ng/u1	97
30) Naphthalene	10.675	128	1347983	30.441	ng/u1	100
32) 4-Chloroaniline	10.793	127	400284	23.332	ng/u1	100
33) Hexachlorobutadiene	10.957	225	316722	35.647	ng/u1	99
34) Caprolactam	11.569	113	106204	26.107	ng/u1	91
35) 4-Chloro-3-methylphenol	11.916	107	375133	25.241	ng/u1	98
36) 2-Methylnaphthalene	12.287	142	871182	27.770	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121323\
 Data File : BP018872.D
 Acq On : 15 Dec 2023 08:15
 Operator : MA/JU
 Sample : PB157656BS
 Misc :
 ALS Vial : 67 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS656

Quant Time: Dec 15 10:07:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120123.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 13 23:34:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.504	142	861770	27.201	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.652	216	532879	35.982	ng/ul	98
40) Hexachlorocyclopentadiene	12.628	237	284812	32.446	ng/ul	98
41) 2,4,6-Trichlorophenol	12.899	196	300834	31.578	ng/ul	98
42) 2,4,5-Trichlorophenol	12.969	196	326718	31.633	ng/ul	97
43) 1,1'-Biphenyl	13.299	154	1125378	31.879	ng/ul	98
44) 2-Chloronaphthalene	13.340	162	908145	32.571	ng/ul	98
45) 2-Nitroaniline	13.546	65	214462	28.373	ng/ul	99
47) Dimethylphthalate	13.928	163	1059308	29.540	ng/ul	100
48) 2,6-Dinitrotoluene	14.046	165	205009	30.259	ng/ul	94
50) Acenaphthylene	14.181	152	1414547	31.363	ng/ul	100
51) 3-Nitroaniline	14.375	138	203462	30.083	ng/ul	98
52) Acenaphthene	14.522	153	932351	31.265	ng/ul	100
53) 2,4-Dinitrophenol	14.575	184	97272	27.089	ng/ul	98
55) 4-Nitrophenol	14.681	109	149130	33.009	ng/ul	97
56) Dibenzofuran	14.857	168	1298563	31.331	ng/ul	97
57) 2,4-Dinitrotoluene	14.828	165	303973	31.895	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.087	232	268199	30.957	ng/ul	94
59) Diethylphthalate	15.287	149	1010076	28.784	ng/ul	100
61) Fluorene	15.510	166	1044915	31.861	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.504	204	550912	32.030	ng/ul	97
63) 4-Nitroaniline	15.534	138	223334	34.663	ng/ul	99
66) 4,6-Dinitro-2-methylph...	15.593	198	180203	39.675	ng/ul	93
67) N-Nitrosodiphenylamine	15.722	169	881738	37.556	ng/ul	100
68) 4-Bromophenyl-phenylether	16.398	248	333185	37.249	ng/ul	96
69) Hexachlorobenzene	16.510	284	391799	39.232	ng/ul	96
70) Atrazine	16.675	200	185169	26.528	ng/ul	99
71) Pentachlorophenol	16.857	266	216964	36.529	ng/ul	99
72) Phenanthrene	17.251	178	1338472	31.279	ng/ul	100
74) Anthracene	17.345	178	1359878	31.704	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.257	216	519147	42.376	ng/ul	99
76) Pentachlorobenzene	14.775	250	491791	40.867	ng/ul	99
77) Carbazole	17.616	167	1277094	37.358	ng/ul	100
78) Di-n-butylphthalate	18.175	149	1768215	38.845	ng/ul	99
80) Fluoranthene	19.222	202	2239987	23.250	ng/ul	96
82) Pyrene	19.575	202	2368388	24.368	ng/ul	95
83) Butylbenzylphthalate	20.451	149	864422	24.294	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.222	252	849658	33.392	ng/ul	100
85) Benzo(a)anthracene	21.286	228	2639939	31.062	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.216	149	1283610	26.247	ng/ul	99
87) Chrysene	21.333	228	2444779	31.237	ng/ul	99
89) Di-n-octyl phthalate	22.122	149	2297211	25.658	ng/ul	100
90) Benzo(b)fluoranthene	22.939	252	2764493	30.576	ng/ul	99
91) Benzo(k)fluoranthene	22.986	252	2702775	30.896	ng/ul	99
93) Benzo(a)pyrene	23.557	252	2608690	31.577	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.116	276	3285785	33.289	ng/ul	97
95) Dibenzo(a,h)anthracene	26.133	278	2753627	33.536	ng/ul	97
96) Benzo(g,h,i)perylene	26.863	276	2642578	33.267	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.409 ng/uL
RT: 3.293 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion: 88	Resp: 76247
Ion Ratio	Lower Upper
88 100	
43 25.0	21.8 32.6
58 71.5	56.5 84.7



#5
Pyridine
Concen: 29.341 ng/uL
RT: 3.693 min Scan# 103
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion: 79	Resp: 483687
Ion Ratio	Lower Upper
79 100	
52 69.1	57.4 86.0
51 32.2	26.2 39.4









#6
Benzaldehyde
Concen: 25.297 ng/ul
RT: 6.975 min Scan# 60
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion: 77 Resp: 272211
Ion Ratio Lower Upper
77 100
105 102.1 77.9 116.9
106 96.0 74.6 111.8



#8
Phenol
Concen: 27.254 ng/ul
RT: 7.028 min Scan# 670
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion: 94 Resp: 557027
Ion Ratio Lower Upper
94 100
65 26.0 21.0 31.6
66 35.0 28.6 43.0









#10
 Bis(2-Chloroethyl)ether
 Concen: 27.404 ng/ul
 RT: 7.264 min Scan# 710
 Delta R.T. 0.000 min
 Lab File: BP018872.D
 Acq: 15 Dec 2023 08:15

Instrument :
 BNA_P
 ClientSampleId :
 SLCS656

Tgt Ion: 93 Resp: 459992




Ion	Ratio	Lower	Upper
93	100		
63	73.1	58.3	87.5
95	32.3	25.9	38.9




#12
 2-Chlorophenol
 Concen: 27.529 ng/ul
 RT: 7.393 min Scan# 732
 Delta R.T. 0.000 min
 Lab File: BP018872.D
 Acq: 15 Dec 2023 08:15

Tgt Ion: 128 Resp: 453088

Ion	Ratio	Lower	Upper
128	100		
64	45.2	37.0	55.4
130	32.7	26.8	40.2









#13
2-Methylphenol
Concen: 25.558 ng/ul
RT: 8.281 min Scan# 883
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:108 Resp: 399906
Ion Ratio Lower Upper
108 100
107 89.3 70.6 105.8



#14
2,2'-oxybis(1-Chloropropane)
Concen: 24.902 ng/ul
RT: 8.358 min Scan# 896
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion: 45 Resp: 530364
Ion Ratio Lower Upper
45 100
77 17.2 12.7 19.1
79 12.7 10.5 15.7









#16
Acetophenone
Concen: 25.347 ng/ul
RT: 8.658 min Scan# 947
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:105 Resp: 637134
Ion Ratio Lower Upper
105 100
77 74.0 59.4 89.0
51 22.9 19.4 29.0



#17
N-Nitroso-di-n-propylamine
Concen: 21.395 ng/ul
RT: 8.646 min Scan# 945
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion: 70 Resp: 295857
Ion Ratio Lower Upper
70 100
42 48.5 41.0 61.4
101 10.3 7.8 11.8
130 20.5 16.9 25.3









#18
4-Methylphenol
Concen: 24.907 ng/ul
RT: 8.611 min Scan# 911
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:108 Resp: 425566
Ion Ratio Lower Upper
108 100
107 110.1 90.6 136.0



#19
Hexachloroethane
Concen: 26.194 ng/ul
RT: 8.905 min Scan# 989
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:117 Resp: 176742
Ion Ratio Lower Upper
117 100
201 116.9 83.7 125.5
199 71.7 52.3 78.5









#22
Nitrobenzene
Concen: 25.837 ng/ul
RT: 9.034 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion: 77 Resp: 408121
Ion Ratio Lower Upper
77 100
123 46.0 35.8 53.8
65 13.1 10.5 15.7



#23
Isophorone
Concen: 23.445 ng/ul
RT: 9.563 min Scan# 1101
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion: 82 Resp: 755090
Ion Ratio Lower Upper
82 100
95 7.3 5.5 8.3
138 17.9 13.5 20.3









#25
2-Nitrophenol
Concen: 31.293 ng/ul
RT: 9.740 min Scan# 1132
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:139 Resp: 225684
Ion Ratio Lower Upper
139 100
65 46.8 39.1 58.7
109 29.6 23.8 35.6



#26
2,4-Dimethylphenol
Concen: 27.474 ng/ul
RT: 9.810 min Scan# 1143
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:107 Resp: 387530
Ion Ratio Lower Upper
107 100
121 51.2 40.0 60.0
122 88.0 69.1 103.7









#27
 Bis(2-Chloroethoxy)methane
 Concen: 27.129 ng/ul
 RT: 10.046 min Scan# 1183
 Delta R.T. 0.000 min
 Lab File: BP018872.D
 Acq: 15 Dec 2023 08:15




Instrument :
 BNA_P
 ClientSampleId :
 SLCS656


Tgt Ion: 93 Resp: 551619
 Ion Ratio Lower Upper
 93 100
 95 32.7 26.1 39.1
 123 11.2 8.9 13.3



#29
 2,4-Dichlorophenol
 Concen: 30.228 ng/ul
 RT: 10.275 min Scan# 1222
 Delta R.T. 0.000 min
 Lab File: BP018872.D
 Acq: 15 Dec 2023 08:15

Tgt Ion: 162 Resp: 401020
 Ion Ratio Lower Upper
 162 100
 164 66.4 51.7 77.5
 98 34.7 29.2 43.8









#30
Naphthalene
Concen: 30.441 ng/ul
RT: 10.675 min Scan# 11
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:128 Resp: 1347983
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 13.0 10.5 15.7



#32
4-Chloroaniline
Concen: 23.332 ng/ul
RT: 10.793 min Scan# 1310
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:127 Resp: 400284
Ion Ratio Lower Upper
127 100
129 32.0 25.6 38.4









#33
Hexachlorobutadiene
Concen: 35.647 ng/ul
RT: 10.957 min Scan# 1338
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:225 Resp: 316722
Ion Ratio Lower Upper
225 100
223 63.0 49.5 74.3
227 63.3 50.9 76.3



#34
Caprolactam
Concen: 26.107 ng/ul
RT: 11.569 min Scan# 1442
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:113 Resp: 106204
Ion Ratio Lower Upper
113 100
55 154.1 133.7 200.5
56 113.4 97.4 146.2









#35
4-Chloro-3-methylphenol
Concen: 25.241 ng/ul
RT: 11.916 min Scan# 1501
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:107 Resp: 375133
Ion Ratio Lower Upper
107 100
144 26.6 21.0 31.6
142 81.2 63.6 95.4



#36
2-Methylnaphthalene
Concen: 27.770 ng/ul
RT: 12.287 min Scan# 1564
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:142 Resp: 871182
Ion Ratio Lower Upper
142 100
141 90.2 71.4 107.0





#37

1-Methylnaphthalene

Concen: 27.201 ng/ul

RT: 12.504 min Scan# 1601

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:142 Resp: 861770

Ion Ratio Lower Upper

142 100

141 91.7 73.5 110.3

116 3.8 2.8 4.2



#39

1,2,4,5-Tetrachlorobenzene

Concen: 35.982 ng/ul

RT: 12.652 min Scan# 1626

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Tgt Ion:216 Resp: 532879




Ion Ratio Lower Upper


216 100

214 78.8 63.9 95.9

179 19.6 16.3 24.5

108 14.5 12.6 19.0









#40
Hexachlorocyclopentadiene
Concen: 32.446 ng/ul
RT: 12.628 min Scan# 1622
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:237 Resp: 284812
Ion Ratio Lower Upper
237 100
235 63.0 49.4 74.0
272 12.9 9.8 14.8



#41
2,4,6-Trichlorophenol
Concen: 31.578 ng/ul
RT: 12.899 min Scan# 1668
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:196 Resp: 300834
Ion Ratio Lower Upper
196 100
198 98.9 76.9 115.3
200 31.3 25.0 37.6









#42
2,4,5-Trichlorophenol
Concen: 31.633 ng/ul
RT: 12.969 min Scan# 1679
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:196 Resp: 326718
Ion Ratio Lower Upper
196 100
198 94.5 77.9 116.9
200 30.2 25.2 37.8



#43
1,1'-Biphenyl
Concen: 31.879 ng/ul
RT: 13.299 min Scan# 1736
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:154 Resp: 1125378
Ion Ratio Lower Upper
154 100
153 39.8 32.8 49.2
76 12.3 10.7 16.1









#44
2-Chloronaphthalene
Concen: 32.571 ng/ul
RT: 13.340 min Scan# 1742
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:162 Resp: 908145
Ion Ratio Lower Upper
162 100
164 32.9 25.7 38.5
127 36.2 30.3 45.5



#45
2-Nitroaniline
Concen: 28.373 ng/ul
RT: 13.546 min Scan# 1778
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion: 65 Resp: 214462
Ion Ratio Lower Upper
65 100
92 72.9 56.9 85.3
138 110.6 88.2 132.2





#47

Dimethylphthalate

Concen: 29.540 ng/ul

RT: 13.928 min Scan# 1842

Delta R.T. 0.000 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:163 Resp: 1059308

Ion Ratio Lower Upper

163 100

194 4.1 3.2 4.8

164 10.2 8.0 12.0



#48

2,6-Dinitrotoluene

Concen: 30.259 ng/ul

RT: 14.046 min Scan# 1863

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15




Tgt Ion:165 Resp: 205009


Ion Ratio Lower Upper

165 100

89 53.3 46.6 70.0

121 19.7 17.3 25.9





#50

Acenaphthylene

Concen: 31.363 ng/ul

RT: 14.181 min Scan# 1886

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:152 Resp: 1414547

Ion Ratio Lower Upper

152 100

151 20.2 15.9 23.9

153 13.4 10.6 16.0



#51

3-Nitroaniline

Concen: 30.083 ng/ul

RT: 14.375 min Scan# 1919

Delta R.T. 0.000 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15




Tgt Ion:138 Resp: 203462


Ion Ratio Lower Upper

138 100

108 11.0 9.5 14.3

92 113.8 93.1 139.7









#52
Acenaphthene
Concen: 31.265 ng/ul
RT: 14.522 min Scan# 1944
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:153 Resp: 932351
Ion Ratio Lower Upper
153 100
152 47.8 38.1 57.1
154 89.0 71.7 107.5



#53
2,4-Dinitrophenol
Concen: 27.089 ng/ul
RT: 14.575 min Scan# 1953
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:184 Resp: 97272
Ion Ratio Lower Upper
184 100
63 58.1 48.2 72.4
154 60.9 49.1 73.7









#55
4-Nitrophenol
Concen: 33.009 ng/ul
RT: 14.681 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:109 Resp: 149130
Ion Ratio Lower Upper
109 100
139 138.0 109.0 163.4
65 127.1 106.2 159.4



#56
Dibenzofuran
Concen: 31.331 ng/ul
RT: 14.857 min Scan# 2001
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:168 Resp: 1298563
Ion Ratio Lower Upper
168 100
139 34.7 29.1 43.7









#57
2,4-Dinitrotoluene
Concen: 31.895 ng/ul
RT: 14.828 min Scan# 1996
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:165 Resp: 303973
Ion Ratio Lower Upper
165 100
63 38.8 32.7 49.1
182 2.7 2.1 3.1



#58
2,3,4,6-Tetrachlorophenol
Concen: 30.957 ng/ul
RT: 15.087 min Scan# 2040
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:232 Resp: 268199
Ion Ratio Lower Upper
232 100
131 37.4 33.8 50.8
130 2.1 1.8 2.8
166 28.7 24.7 37.1









#59
Diethylphthalate
Concen: 28.784 ng/ul
RT: 15.287 min Scan# 2074
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:149 Resp: 1010076
Ion Ratio Lower Upper
149 100
177 22.9 18.2 27.4
150 12.4 9.9 14.9



#61
Fluorene
Concen: 31.861 ng/ul
RT: 15.510 min Scan# 2112
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:166 Resp: 1044915
Ion Ratio Lower Upper
166 100
165 97.8 78.0 117.0
167 13.4 10.8 16.2









#62
4-Chlorophenyl-phenylether
Concen: 32.030 ng/ul
RT: 15.504 min Scan# 2111
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:204 Resp: 550912
Ion Ratio Lower Upper
204 100
206 32.9 27.1 40.7
141 54.2 45.8 68.8



#63
4-Nitroaniline
Concen: 34.663 ng/ul
RT: 15.534 min Scan# 2116
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:138 Resp: 223334
Ion Ratio Lower Upper
138 100
92 51.5 42.3 63.5
108 76.6 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 39.675 ng/ul
RT: 15.593 min Scan# 2126
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:198 Resp: 180203
Ion Ratio Lower Upper
198 100
182 3.8 3.4 5.2
77 21.6 20.6 30.8



#67
N-Nitrosodiphenylamine
Concen: 37.556 ng/ul
RT: 15.722 min Scan# 2148
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:169 Resp: 881738
Ion Ratio Lower Upper
169 100
168 66.4 53.1 79.7
167 36.5 29.1 43.7









#68
4-Bromophenyl-phenylether
Concen: 37.249 ng/ul
RT: 16.398 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:248 Resp: 333185
Ion Ratio Lower Upper
248 100
250 96.6 79.3 118.9
141 66.0 57.2 85.8



#69
Hexachlorobenzene
Concen: 39.232 ng/ul
RT: 16.510 min Scan# 2282
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:284 Resp: 391799
Ion Ratio Lower Upper
284 100
142 27.7 25.8 38.6
249 30.6 24.9 37.3





#70

Atrazine

Concen: 26.528 ng/ul

RT: 16.675 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:200 Resp: 185169

Ion Ratio Lower Upper

200 100

173 25.1 21.0 31.4

215 48.5 38.2 57.2



#71

Pentachlorophenol

Concen: 36.529 ng/ul

RT: 16.857 min Scan# 2341

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15




Tgt Ion:266 Resp: 216964


Ion Ratio Lower Upper

266 100

264 62.9 50.2 75.2

268 64.0 50.5 75.7





#72

Phenanthrene

Concen: 31.279 ng/ul

RT: 17.251 min Scan# 2408

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:178 Resp: 1338472

Ion Ratio Lower Upper

178 100

179 15.6 12.3 18.5

176 19.4 15.4 23.2



#74

Anthracene

Concen: 31.704 ng/ul

RT: 17.345 min Scan# 2424

Delta R.T. 0.000 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15




Tgt Ion:178 Resp: 1359878


Ion Ratio Lower Upper

178 100

179 15.0 12.5 18.7

176 18.3 15.0 22.6









#75
1,2,3,4-Tetrachlorobenzene
Concen: 42.376 ng/uL
RT: 13.257 min Scan# 1729
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:216 Resp: 519147
Ion Ratio Lower Upper
216 100
214 78.1 61.4 92.2
218 48.2 39.2 58.8



#76
Pentachlorobenzene
Concen: 40.867 ng/uL
RT: 14.775 min Scan# 1987
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:250 Resp: 491791
Ion Ratio Lower Upper
250 100
248 62.7 50.6 75.8
252 63.8 51.3 76.9





#77

Carbazole

Concen: 37.358 ng/ul

RT: 17.616 min Scan# 2470

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :





BNA_P

ClientSampleId :

SLCS656

Tgt Ion:167 Resp: 1277094

Ion	Ratio	Lower	Upper
167	100		
166	21.5	17.1	25.7
139	12.5	10.2	15.4



#78

Di-n-butylphthalate

Concen: 38.845 ng/ul

RT: 18.175 min Scan# 2565




Delta R.T. -0.006 min


Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Tgt Ion:149 Resp: 1768215

Ion	Ratio	Lower	Upper
149	100		
150	9.0	7.4	11.2
104	4.9	3.9	5.9









#80
Fluoranthene
Concen: 23.250 ng/u1
RT: 19.222 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:202 Resp: 2239987
Ion Ratio Lower Upper
202 100
101 9.9 9.1 13.7
100 7.5 7.1 10.7



#82
Pyrene
Concen: 24.368 ng/u1
RT: 19.575 min Scan# 2803
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:202 Resp: 2368388
Ion Ratio Lower Upper
202 100
101 11.9 11.0 16.4
100 9.6 9.1 13.7





#83

Butylbenzylphthalate

Concen: 24.294 ng/ul

RT: 20.451 min Scan# 2952

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:149 Resp: 864422

Ion Ratio Lower Upper

149 100

91 71.2 56.6 85.0

206 22.1 17.4 26.0



#84

3,3'-Dichlorobenzidine

Concen: 33.392 ng/ul

RT: 21.222 min Scan# 3083

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15




Tgt Ion:252 Resp: 849658


Ion Ratio Lower Upper

252 100

254 64.0 51.0 76.4

126 10.1 7.8 11.6









#85
Benzo(a)anthracene
Concen: 31.062 ng/ul
RT: 21.286 min Scan# 3094
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:228 Resp: 2639939
Ion Ratio Lower Upper
228 100
229 19.2 15.7 23.5
226 27.0 21.9 32.9



#86
Bis(2-ethylhexyl)phthalate
Concen: 26.247 ng/ul
RT: 21.216 min Scan# 3082
Delta R.T. 0.000 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:149 Resp: 1283610
Ion Ratio Lower Upper
149 100
167 28.0 22.2 33.2
279 4.9 3.3 4.9





#87

Chrysene

Concen: 31.237 ng/ul

RT: 21.333 min Scan# 3102

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:228 Resp: 2444779

Ion Ratio Lower Upper

228 100

226 30.1 23.9 35.9

229 19.8 15.6 23.4



#89

Di-n-octyl phthalate

Concen: 25.658 ng/ul




RT: 22.122 min Scan# 3236


Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Tgt Ion:149 Resp: 2297211









#90
Benzo(b)fluoranthene
Concen: 30.576 ng/ul
RT: 22.939 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15




Instrument :
BNA_P
ClientSampleId :
SLCS656


Tgt Ion:252 Resp: 2764493
Ion Ratio Lower Upper
252 100
253 21.5 17.3 25.9
125 8.7 7.8 11.6



#91
Benzo(k)fluoranthene
Concen: 30.896 ng/ul
RT: 22.986 min Scan# 3383
Delta R.T. -0.006 min
Lab File: BP018872.D
Acq: 15 Dec 2023 08:15

Tgt Ion:252 Resp: 2702775
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 8.8 8.0 12.0





#93

Benzo(a)pyrene

Concen: 31.577 ng/ul

RT: 23.557 min Scan# 3479

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15

Instrument :

BNA_P

ClientSampleId :

SLCS656





Tgt Ion:252 Resp: 2608690

Ion Ratio Lower Upper

252 100

253 21.7 17.8 26.8

125 9.8 9.2 13.8



#94

Indeno(1,2,3-cd)pyrene

Concen: 33.289 ng/ul

RT: 26.116 min Scan# 3915

Delta R.T. -0.006 min

Lab File: BP018872.D

Acq: 15 Dec 2023 08:15




Tgt Ion:276 Resp: 3285785


Ion Ratio Lower Upper

276 100

138 21.4 19.4 29.2

277 24.4 19.8 29.6









#95
 Dibenzo(a,h)anthracene
 Concen: 33.536 ng/ul
 RT: 26.133 min Scan# 3918
 Delta R.T. -0.006 min
 Lab File: BP018872.D
 Acq: 15 Dec 2023 08:15

Instrument :
 BNA_P
 ClientSampleId :
 SLCS656

Tgt Ion:278 Resp: 2753627
 Ion Ratio Lower Upper
 278 100
 139 13.7 13.0 19.6
 279 23.9 18.7 28.1



#96
 Benzo(g,h,i)perylene
 Concen: 33.267 ng/ul
 RT: 26.863 min Scan# 4042
 Delta R.T. -0.012 min
 Lab File: BP018872.D
 Acq: 15 Dec 2023 08:15

Tgt Ion:276 Resp: 2642578
 Ion Ratio Lower Upper
 276 100
 138 19.4 18.1 27.1
 277 23.7 19.4 29.0

