

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121422\
 Data File : BP013104.D
 Acq On : 14 Dec 2022 14:18
 Operator : CG/JU
 Sample : PB149062BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 PB149062BL

Quant Time: Dec 15 03:50:03 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 13 03:29:24 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.958	152	489985	20.000	ng	0.00
21) Naphthalene-d8	10.775	136	1858028	20.000	ng	0.00
39) Acenaphthene-d10	14.598	164	1072649	20.000	ng	0.00
64) Phenanthrene-d10	17.357	188	2059960	20.000	ng	0.00
76) Chrysene-d12	21.439	240	1945672	20.000	ng	0.00
86) Perylene-d12	23.921	264	2090053	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.511	112	4095168	152.707	ng	0.00
7) Phenol-d6	7.122	99	5051558	144.242	ng	0.00
23) Nitrobenzene-d5	9.128	82	2991381	88.263	ng	0.00
42) 2,4,6-Tribromophenol	16.092	330	1822051	140.886	ng	0.00
45) 2-Fluorobiphenyl	13.222	172	6983335	88.397	ng	0.00
79) Terphenyl-d14	19.910	244	9361896	94.928	ng	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121422\
Data File : BP013104.D
Acq On : 14 Dec 2022 14:18
Operator : CG/JU
Sample : PB149062BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
PB149062BL

Quant Time: Dec 15 03:50:03 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121322.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 13 03:29:24 2022
Response via : Initial Calibration

