

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121525\
 Data File : BP026321.D
 Acq On : 15 Dec 2025 12:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 16 04:22:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 16 04:20:55 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	89	0.00
2	1,4-Dioxane	40.000	39.959	0.1	92	0.00
3	Pyridine	40.000	38.545	3.6	87	0.00
4	n-Nitrosodimethylamine	40.000	37.368	6.6	84	0.00
5 S	2-Fluorophenol	80.000	79.043	1.2	88	0.00
6	Aniline	40.000	35.622	10.9	78	0.00
7 S	Phenol-d6	80.000	76.035	5.0	83	0.00
8	2-Chlorophenol	40.000	40.288	-0.7	89	0.00
9	Benzaldehyde	40.000	46.885	-17.2	94	0.00
10 C	Phenol	40.000	37.619	6.0	82	0.00
11	bis(2-Chloroethyl)ether	40.000	38.052	4.9	83	0.00
12	1,3-Dichlorobenzene	40.000	41.825	-4.6	93	0.00
13 C	1,4-Dichlorobenzene	40.000	41.224	-3.1	92	0.00
14	1,2-Dichlorobenzene	40.000	41.160	-2.9	92	0.00
15	Benzyl Alcohol	40.000	38.669	3.3	84	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	36.815	8.0	80	0.00
17	2-Methylphenol	40.000	38.653	3.4	84	0.00
18	Hexachloroethane	40.000	41.912	-4.8	93	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.358	4.1	82	0.00
20	3+4-Methylphenols	40.000	37.883	5.3	84	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	82	0.00
22	Acetophenone	40.000	40.897	-2.2	84	0.00
23 S	Nitrobenzene-d5	80.000	82.542	-3.2	84	0.00
24	Nitrobenzene	40.000	41.259	-3.1	84	0.00
25	Isophorone	40.000	40.999	-2.5	82	0.00
26 C	2-Nitrophenol	40.000	45.102	-12.8	88	0.00
27	2,4-Dimethylphenol	40.000	37.094	7.3	75	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.060	-0.2	82	0.00
29 C	2,4-Dichlorophenol	40.000	43.155	-7.9	87	0.00
30	1,2,4-Trichlorobenzene	40.000	44.310	-10.8	91	0.00
31	Naphthalene	40.000	41.574	-3.9	86	0.00
32	Benzoic acid	40.000	48.701	-21.8	99	0.00
33	4-Chloroaniline	40.000	38.928	2.7	78	0.00
34 C	Hexachlorobutadiene	40.000	47.385	-18.5	97	0.00
35	Caprolactam	40.000	40.702	-1.8	82	0.00
36 C	4-Chloro-3-methylphenol	40.000	42.296	-5.7	84	0.00
37	2-Methylnaphthalene	40.000	41.932	-4.8	85	0.00
38	1-Methylnaphthalene	40.000	41.820	-4.6	84	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	81	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	44.895	-12.2	92	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.227	-3.1	83	0.00
42 S	2,4,6-Tribromophenol	80.000	95.102	-18.9	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	44.103	-10.3	88	0.00
44	2,4,5-Trichlorophenol	40.000	44.143	-10.4	88	0.00
45 S	2-Fluorobiphenyl	80.000	86.067	-7.6	89	0.00
46	1,1'-Biphenyl	40.000	41.880	-4.7	85	0.00
47	2-Chloronaphthalene	40.000	41.470	-3.7	85	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121525\
 Data File : BP026321.D
 Acq On : 15 Dec 2025 12:03
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Dec 16 04:22:21 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 16 04:20:55 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.187	-3.0	81	0.00
49	Acenaphthylene	40.000	41.824	-4.6	85	0.00
50	Dimethylphthalate	40.000	43.118	-7.8	87	0.00
51	2,6-Dinitrotoluene	40.000	45.125	-12.8	85	0.00
52 C	Acenaphthene	40.000	40.923	-2.3	83	0.00
53	3-Nitroaniline	40.000	40.329	-0.8	78	0.00
54 P	2,4-Dinitrophenol	40.000	60.766	-51.9#	118	0.00
55	Dibenzofuran	40.000	41.850	-4.6	85	0.00
56 P	4-Nitrophenol	40.000	32.842	17.9	66	0.00
57	2,4-Dinitrotoluene	40.000	44.816	-12.0	86	0.00
58	Fluorene	40.000	42.659	-6.6	87	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.124	-10.3	87	0.00
60	Diethylphthalate	40.000	42.936	-7.3	86	0.00
61	4-Chlorophenyl-phenylether	40.000	44.056	-10.1	89	0.00
62	4-Nitroaniline	40.000	39.519	1.2	79	0.00
63	Azobenzene	40.000	39.601	1.0	79	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	82	0.00
65	4,6-Dinitro-2-methylphenol	40.000	49.448	-23.6	96	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.840	-2.1	83	0.00
67	4-Bromophenyl-phenylether	40.000	44.345	-10.9	88	0.00
68	Hexachlorobenzene	40.000	45.283	-13.2	91	0.00
69	Atrazine	40.000	38.759	3.1	78	0.00
70 C	Pentachlorophenol	40.000	42.036	-5.1	85	0.00
71	Phenanthrene	40.000	41.382	-3.5	86	0.00
72	Anthracene	40.000	41.263	-3.2	83	0.00
73	Carbazole	40.000	39.119	2.2	80	0.00
74	Di-n-butylphthalate	40.000	41.973	-4.9	82	0.00
75 C	Fluoranthene	40.000	40.608	-1.5	84	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	74	0.00
77	Benzydine	40.000	34.612	13.5	66	0.00
78	Pyrene	40.000	45.676	-14.2	83	0.00
79 S	Terphenyl-d14	80.000	93.899	-17.4	86	0.00
80	Butylbenzylphthalate	40.000	44.602	-11.5	79	0.00
81	Benzo(a)anthracene	40.000	41.137	-2.8	76	0.00
82	3,3'-Dichlorobenzidine	40.000	38.585	3.5	71	0.00
83	Chrysene	40.000	41.272	-3.2	77	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.916	-4.8	72	0.00
85 c	Di-n-octyl phthalate	40.000	39.312	1.7	70	0.00
86 I	Perylene-d12	20.000	20.000	0.0	65	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.531	-3.8	67	0.00
88	Benzo(b)fluoranthene	40.000	42.821	-7.1	69	0.00
89	Benzo(k)fluoranthene	40.000	43.728	-9.3	72	0.00
90 C	Benzo(a)pyrene	40.000	42.380	-6.0	68	0.00
91	Dibenzo(a,h)anthracene	40.000	41.333	-3.3	67	0.00
92	Benzo(g,h,i)perylene	40.000	41.174	-2.9	67	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121525\
Data File : BP026321.D
Acq On : 15 Dec 2025 12:03
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Dec 16 04:22:21 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP111825.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 16 04:20:55 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0