

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121621\
 Data File : BP008454.D
 Acq On : 16 Dec 2021 15:58
 Operator : CG/JU
 Sample : M5074-05
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CS-3

Quant Time: Dec 16 16:52:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121421.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Dec 14 17:59:47 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.734	152	50751	20.000	ng	-0.01
21) Naphthalene-d8	10.534	136	172036	20.000	ng	0.00
39) Acenaphthene-d10	14.392	164	101730	20.000	ng	0.00
64) Phenanthrene-d10	17.157	188	199078	20.000	ng	0.00
76) Chrysene-d12	21.268	240	205007	20.000	ng	# 0.00
86) Perylene-d12	23.645	264	248937	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.340	112	380956	118.544	ng	0.00
7) Phenol-d6	6.934	99	410418	95.090	ng	0.00
23) Nitrobenzene-d5	8.898	82	359128	88.437	ng	-0.01
42) 2,4,6-Tribromophenol	15.904	330	191249	127.926	ng	0.00
45) 2-Fluorobiphenyl	13.010	172	683460	85.404	ng	0.00
79) Terphenyl-d14	19.733	244	984602	83.587	ng	0.00
Target Compounds						
50) Dimethylphthalate	13.869	163	36810	4.779	ng	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121621\
Data File : BP008454.D
Acq On : 16 Dec 2021 15:58
Operator : CG/JU
Sample : M5074-05
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CS-3

Quant Time: Dec 16 16:52:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP121421.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Dec 14 17:59:47 2021
Response via : Initial Calibration

