

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
 Data File : BP013214.D
 Acq On : 21 Dec 2022 18:37
 Operator : CG/JU
 Sample : N6032-01
 Misc :
 ALS Vial : 96 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 COL84

Quant Time: Dec 21 21:54:41 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 19 21:43:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	555320	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	2382521	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.593	164	1539928	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.345	188	3137279	20.000	ng/u1	-0.01
79) Chrysene-d12	21.433	240	2099140	20.000	ng/u1	0.00
88) Perylene-d12	23.910	264	2398476	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	57946	4.462	ng/uL	0.00
4) Pyridine-d5	3.781	84	52903	1.434	ng/u1	0.02
7) Phenol-d5	7.111	99	233699	5.167	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	531715	19.487	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	514628	14.453	ng/u1	0.00
15) 4-Methylphenol-d8	8.658	113	368750	9.950	ng/u1	-0.01
21) Nitrobenzene-d5	9.116	128	325871	18.616	ng/u1	-0.01
24) 2-Nitrophenol-d4	9.840	143	328276	16.658	ng/u1	-0.01
28) 2,4-Dichlorophenol-d3	10.381	165	610245	15.942	ng/u1	0.00
31) 4-Chloroaniline-d4	10.904	131	307233	5.685	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	2173361	18.364	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	2251375	17.312	ng/u1	0.00
54) 4-Nitrophenol-d4	14.793	143	51185	2.404	ng/u1	0.00
60) Fluorene-d10	15.587	176	1912397	19.100	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	195590	9.688	ng/u1	0.00
73) Anthracene-d10	17.445	188	3225139	22.787	ng/u1	-0.01
81) Pyrene-d10	19.680	212	3108811	25.801	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	2699739	22.583	ng/u1	-0.01

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
Data File : BP013214.D
Acq On : 21 Dec 2022 18:37
Operator : CG/JU
Sample : N6032-01
Misc :
ALS Vial : 96 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
COL84

Quant Time: Dec 21 21:54:41 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 19 21:43:11 2022
Response via : Initial Calibration

