

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
 Data File : BP013222.D
 Acq On : 21 Dec 2022 23:45
 Operator : CG/JU
 Sample : PB149681BL
 Misc :
 ALS Vial : 105 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK681

Quant Time: Dec 22 01:47:09 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Dec 21 23:38:37 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)
Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.951	152	491082	20.000 ng/u1	0.00
20) Naphthalene-d8	10.763	136	2006865	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.592	164	1138089	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.345	188	2085373	20.000 ng/u1	0.00
79) Chrysene-d12	21.433	240	1972981	20.000 ng/u1	0.00
88) Perylene-d12	23.909	264	2378708	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.334	96	67374	5.866 ng/uL	0.00
4) Pyridine-d5	3.763	84	954391	29.252 ng/u1	0.00
7) Phenol-d5	7.104	99	1135032	28.376 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	729253	30.223 ng/u1	0.00
11) 2-Chlorophenol-d4	7.475	132	930698	29.557 ng/u1	0.00
15) 4-Methylphenol-d8	8.657	113	890058	27.157 ng/u1	0.00
21) Nitrobenzene-d5	9.116	128	440631	29.883 ng/u1	0.00
24) 2-Nitrophenol-d4	9.840	143	492444	29.665 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	894776	27.751 ng/u1	0.00
31) 4-Chloroaniline-d4	10.898	131	1221272	26.830 ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	2496351	28.540 ng/u1	0.00
49) Acenaphthylene-d8	14.286	160	2840050	29.549 ng/u1	0.00
54) 4-Nitrophenol-d4	14.786	143	289821	18.420 ng/u1	0.00
60) Fluorene-d10	15.586	176	2076193	28.058 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	251054	18.708 ng/u1	0.00
73) Anthracene-d10	17.445	188	2952629	31.384 ng/u1	0.00
81) Pyrene-d10	19.674	212	3404000	30.058 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.751	264	3797932	32.033 ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
Data File : BP013222.D
Acq On : 21 Dec 2022 23:45
Operator : CG/JU
Sample : PB149681BL
Misc :
ALS Vial : 105 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK681

Quant Time: Dec 22 01:47:09 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Dec 21 23:38:37 2022
Response via : Initial Calibration

