

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
 Data File : BP013181.D
 Acq On : 20 Dec 2022 23:20
 Operator : CG/JU
 Sample : N5984-09
 Misc :
 ALS Vial : 62 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 COLD4

Quant Time: Dec 20 23:49:12 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 19 21:43:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	556687	20.000	ng/u1	0.00
20) Naphthalene-d8	10.769	136	2271407	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.593	164	1318268	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	2460602	20.000	ng/u1	0.00
79) Chrysene-d12	21.433	240	2341657	20.000	ng/u1	0.00
88) Perylene-d12	23.916	264	2857370	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.340	96	52636	4.043	ng/uL	0.00
4) Pyridine-d5	3.764	84	761446	20.588	ng/u1	0.00
7) Phenol-d5	7.111	99	996260	21.971	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.281	67	628564	22.980	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	808590	22.653	ng/u1	0.00
15) 4-Methylphenol-d8	8.664	113	799135	21.509	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	393433	23.575	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	452187	24.068	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	805307	22.067	ng/u1	0.00
31) 4-Chloroaniline-d4	10.905	131	1026126	19.917	ng/u1	0.00
46) Dimethylphthalate-d6	13.999	166	2210253	21.816	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	2580735	23.181	ng/u1	0.00
54) 4-Nitrophenol-d4	14.793	143	237646	13.040	ng/u1	0.00
60) Fluorene-d10	15.587	176	1878466	21.916	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	210965	13.323	ng/u1	0.00
73) Anthracene-d10	17.451	188	2632831	23.717	ng/u1	0.00
81) Pyrene-d10	19.681	212	2956687	21.997	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.757	264	3370880	23.668	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
Data File : BP013181.D
Acq On : 20 Dec 2022 23:20
Operator : CG/JU
Sample : N5984-09
Misc :
ALS Vial : 62 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
COLD4

Quant Time: Dec 20 23:49:12 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 19 21:43:11 2022
Response via : Initial Calibration

