

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
 Data File : BP013187.D
 Acq On : 21 Dec 2022 03:18
 Operator : CG/JU
 Sample : N6009-04
 Misc :
 ALS Vial : 69 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DBYB6

Quant Time: Dec 21 04:22:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 19 21:43:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc Units	Dev(Min)

Internal Standards					
1) 1,4-Dichlorobenzene-d4	7.952	152	443921	20.000 ng/u1	0.00
20) Naphthalene-d8	10.769	136	1778532	20.000 ng/u1	0.00
38) Acenaphthene-d10	14.592	164	1057504	20.000 ng/u1	0.00
64) Phenanthrene-d10	17.351	188	2217882	20.000 ng/u1	0.00
79) Chrysene-d12	21.433	240	2307554	20.000 ng/u1	0.00
88) Perylene-d12	23.915	264	2734487	20.000 ng/u1	0.00
System Monitoring Compounds					
3) 1,4-Dioxane-d8	3.340	96	39700	3.824 ng/uL	0.00
4) Pyridine-d5	3.781	84	51978	1.762 ng/u1	0.02
7) Phenol-d5	7.110	99	673139	18.616 ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	510808	23.419 ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	606840	21.319 ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	212527	7.173 ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	303463	23.223 ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	341318	23.201 ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.387	165	573903	20.084 ng/u1	0.00
31) 4-Chloroaniline-d4	10.904	131	486527	12.061 ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	1890055	23.255 ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	2088180	23.382 ng/u1	0.00
54) 4-Nitrophenol-d4	14.792	143	193983	13.269 ng/u1	0.00
60) Fluorene-d10	15.586	176	1654100	24.057 ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	204408	14.322 ng/u1	0.00
73) Anthracene-d10	17.451	188	2373644	23.722 ng/u1	0.00
81) Pyrene-d10	19.680	212	3151654	23.794 ng/u1	0.00
92) Benzo(a)pyrene-d12	23.757	264	3344296	24.537 ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
Data File : BP013187.D
Acq On : 21 Dec 2022 03:18
Operator : CG/JU
Sample : N6009-04
Misc :
ALS Vial : 69 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DBYB6

Quant Time: Dec 21 04:22:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 19 21:43:11 2022
Response via : Initial Calibration

