

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
 Data File : BP013191.D
 Acq On : 21 Dec 2022 05:34
 Operator : CG/JU
 Sample : N5984-04
 Misc :
 ALS Vial : 73 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0LC8

Quant Time: Dec 21 06:33:28 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Dec 19 21:43:11 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.952	152	437945	20.000	ng/u1	0.00
20) Naphthalene-d8	10.763	136	1733733	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.592	164	988337	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.351	188	1981347	20.000	ng/u1	0.00
79) Chrysene-d12	21.433	240	2087372	20.000	ng/u1	0.00
88) Perylene-d12	23.916	264	2485758	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.334	96	91169	8.901	ng/uL	0.00
4) Pyridine-d5	3.764	84	677116	23.272	ng/u1	0.00
7) Phenol-d5	7.111	99	1786575	50.084	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.275	67	1123218	52.198	ng/u1	0.00
11) 2-Chlorophenol-d4	7.481	132	1434929	51.099	ng/u1	0.00
15) 4-Methylphenol-d8	8.663	113	1406007	48.104	ng/u1	0.00
21) Nitrobenzene-d5	9.122	128	698881	54.865	ng/u1	0.00
24) 2-Nitrophenol-d4	9.846	143	806923	56.268	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.381	165	1413747	50.754	ng/u1	0.00
31) 4-Chloroaniline-d4	10.904	131	1695909	43.127	ng/u1	0.00
46) Dimethylphthalate-d6	13.998	166	3940100	51.872	ng/u1	0.00
49) Acenaphthylene-d8	14.287	160	4629116	55.462	ng/u1	0.00
54) 4-Nitrophenol-d4	14.787	143	579013	42.377	ng/u1	0.00
60) Fluorene-d10	15.587	176	3412365	53.102	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.704	200	504531	39.570	ng/u1	0.00
73) Anthracene-d10	17.451	188	5089108	56.933	ng/u1	0.00
81) Pyrene-d10	19.680	212	6191111	51.672	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.757	264	6994631	56.454	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP121922\
Data File : BP013191.D
Acq On : 21 Dec 2022 05:34
Operator : CG/JU
Sample : N5984-04
Misc :
ALS Vial : 73 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
COLC8

Quant Time: Dec 21 06:33:28 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP120722.M
Quant Title : SVOA CALIBRATION
QLast Update : Mon Dec 19 21:43:11 2022
Response via : Initial Calibration

