

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
 Data File : BP008522.D
 Acq On : 23 Dec 2021 12:13
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Quant Time: Dec 23 14:48:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	426991	20.000	ng/ul	0.00
20) Naphthalene-d8	10.710	136	1704531	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.534	164	1088059	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.281	188	2210610	20.000	ng/ul	0.00
79) Chrysene-d12	21.363	240	2243441	20.000	ng/ul	0.00
88) Perylene-d12	23.792	264	2337611	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	68950	7.279	ng/uL	0.00
4) Pyridine-d5	3.758	84	486980	18.369	ng/ul	0.00
7) Phenol-d5	7.064	99	601082	17.646	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	366554	18.946	ng/ul	0.00
11) 2-Chlorophenol-d4	7.434	132	500227	18.460	ng/ul	0.00
15) 4-Methylphenol-d8	8.605	113	469909	17.196	ng/ul	0.00
21) Nitrobenzene-d5	9.058	128	221346	18.352	ng/ul	0.00
24) 2-Nitrophenol-d4	9.787	143	213018	19.262	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	515057	18.973	ng/ul	0.00
31) 4-Chloroaniline-d4	10.834	131	669428	17.864	ng/ul	0.00
46) Dimethylphthalate-d6	13.946	166	1453873	18.085	ng/ul	0.00
49) Acenaphthylene-d8	14.228	160	1893751	18.839	ng/ul	0.00
54) 4-Nitrophenol-d4	14.710	143	225178	17.670	ng/ul	0.00
60) Fluorene-d10	15.528	176	1244317	17.827	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	152162	15.637	ng/ul	0.00
73) Anthracene-d10	17.381	188	1866378	17.817	ng/ul	0.00
81) Pyrene-d10	19.610	212	2066079	15.369	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.633	264	2294655	19.073	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.387	88	69812	6.816	ng/uL	97
5) Pyridine	3.776	79	400769	14.734	ng/ul	98
6) Benzaldehyde	7.040	77	378040	19.471	ng/ul	97
8) Phenol	7.087	94	610041	17.430	ng/ul	99
10) Bis(2-Chloroethyl)ether	7.328	93	484426	17.293	ng/ul	98
12) 2-Chlorophenol	7.469	128	510461	18.049	ng/ul	98
13) 2-Methylphenol	8.346	108	467905	17.254	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.440	45	566379	17.553	ng/ul	99
16) Acetophenone	8.722	105	761988	18.026	ng/ul	97
17) N-Nitroso-di-n-propyla...	8.716	70	365475	17.371	ng/ul	99
18) 4-Methylphenol	8.669	108	501404	17.270	ng/ul	98
19) Hexachloroethane	8.993	117	202498	17.589	ng/ul	98
22) Nitrobenzene	9.105	77	520653	18.708	ng/ul	98
23) Isophorone	9.634	82	974333	16.694	ng/ul	98
25) 2-Nitrophenol	9.816	139	245523	19.291	ng/ul	99
26) 2,4-Dimethylphenol	9.881	107	539952	17.952	ng/ul	96
27) Bis(2-Chloroethoxy)met...	10.116	93	630508	17.283	ng/ul	98
29) 2,4-Dichlorophenol	10.352	162	499612	18.348	ng/ul	98
30) Naphthalene	10.758	128	1644496	18.227	ng/ul	99
32) 4-Chloroaniline	10.858	127	696454	18.328	ng/ul	99
33) Hexachlorobutadiene	11.057	225	346228	18.338	ng/ul	99
34) Caprolactam	11.610	113	135172	17.422	ng/ul	96
35) 4-Chloro-3-methylphenol	11.981	107	471874	17.676	ng/ul	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
 Data File : BP008522.D
 Acq On : 23 Dec 2021 12:13
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Quant Time: Dec 23 14:48:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
36) 2-Methylnaphthalene	12.369	142	1212795	18.996	ng/ul	99
37) 1-Methylnaphthalene	12.587	142	1115675	17.165	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.734	216	652712	18.585	ng/ul	99
40) Hexachlorocyclopentadiene	12.722	237	421787	17.945	ng/ul	100
41) 2,4,6-Trichlorophenol	12.969	196	380889	19.099	ng/ul	98
42) 2,4,5-Trichlorophenol	13.034	196	408361	18.816	ng/ul	99
43) 1,1'-Biphenyl	13.375	154	1473950	17.475	ng/ul	100
44) 2-Chloronaphthalene	13.410	162	1198983	18.437	ng/ul	98
45) 2-Nitroaniline	13.604	65	245831	18.808	ng/ul	100
47) Dimethylphthalate	13.993	163	1443315	17.995	ng/ul	100
48) 2,6-Dinitrotoluene	14.104	165	295646	20.737	ng/ul	93
50) Acenaphthylene	14.257	152	1852668	18.012	ng/ul	99
51) 3-Nitroaniline	14.428	138	296319	20.169	ng/ul#	97
52) Acenaphthene	14.598	153	1231582	18.337	ng/ul	99
53) 2,4-Dinitrophenol	14.628	184	95241	15.826	ng/ul	94
55) 4-Nitrophenol	14.728	109	158702	17.912	ng/ul	99
56) Dibenzofuran	14.934	168	1787045	18.196	ng/ul	100
57) 2,4-Dinitrotoluene	14.887	165	374783	18.999	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.157	232	326244	19.352	ng/ul	98
59) Diethylphthalate	15.357	149	1386315	17.696	ng/ul	99
61) Fluorene	15.581	166	1437147	18.367	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.581	204	741666	17.870	ng/ul	100
63) 4-Nitroaniline	15.587	138	295173	21.920	ng/ul	98
66) 4,6-Dinitro-2-methylph...	15.651	198	164071	15.935	ng/ul	99
67) N-Nitrosodiphenylamine	15.793	169	1239155	18.330	ng/ul	99
68) 4-Bromophenyl-phenylether	16.475	248	454261	17.917	ng/ul	99
69) Hexachlorobenzene	16.593	284	523534	17.736	ng/ul	97
70) Atrazine	16.745	200	435822	16.989	ng/ul	99
71) Pentachlorophenol	16.934	266	250743	16.619	ng/ul	99
72) Phenanthrene	17.328	178	2207885	18.134	ng/ul	99
74) Anthracene	17.416	178	2254170	18.464	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.340	216	633086	16.883	ng/uL	99
76) Pentachlorobenzene	14.857	250	606477	17.617	ng/uL	100
77) Carbazole	17.681	167	1981140	19.043	ng/ul	99
78) Di-n-butylphthalate	18.245	149	2297733	18.916	ng/ul	100
80) Fluoranthene	19.286	202	2548250	15.943	ng/ul	98
82) Pyrene	19.639	202	2645002	16.423	ng/ul	97
83) Butylbenzylphthalate	20.522	149	941532	17.248	ng/ul	94
84) 3,3'-Dichlorobenzidine	21.280	252	1006544	21.183	ng/ul	100
85) Benzo(a)anthracene	21.345	228	2607520	18.004	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.286	149	1554870	18.766	ng/ul	100
87) Chrysene	21.398	228	2493131	17.829	ng/ul	100
89) Di-n-octyl phthalate	22.227	149	2506551	16.839	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	2687792	17.352	ng/ul	100
91) Benzo(k)fluoranthene	23.104	252	2681300	17.935	ng/ul	99
93) Benzo(a)pyrene	23.686	252	2645800	17.844	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.327	276	3187242	19.493	ng/ul	98
95) Dibenzo(a,h)anthracene	26.345	278	2713105	19.123	ng/ul	99
96) Benzo(g,h,i)perylene	27.098	276	2614104	19.343	ng/ul	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed


Instrument :
BNA_P
ClientSampleId :
SICV649

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
 Data File : BP008522.D
 Acq On : 23 Dec 2021 12:13
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Quant Time: Dec 23 14:48:45 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration





1,4-Dioxane

Concen: 6.816 ng/uL

RT: 3.387 min Scan#

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649





Tgt Ion: 88 Resp: 69812

Ion Ratio Lower Upper

88 100

43 24.1 20.2 30.2

58 59.8 49.6 74.4



Pyridine

Concen: 14.734 ng/uL

RT: 3.776 min Scan# 134

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13




Tgt Ion: 79 Resp: 400769


Ion Ratio Lower Upper

79 100

52 66.1 51.6 77.4

51 32.4 25.2 37.8









Benzaldehyde
 Concen: 19.471 ng/ul
 RT: 7.040 min Scan# 689
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument : BNA_P
 ClientSampleId : SICV649


Tgt Ion:	77	Resp:	378040
Ion	Ratio	Lower	Upper
77	100		
105	102.4	78.8	118.2
106	98.0	76.4	114.6



Phenol
 Concen: 17.430 ng/ul
 RT: 7.087 min Scan# 697
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	94	Resp:	610041
Ion	Ratio	Lower	Upper
94	100		
65	25.4	20.6	31.0
66	33.2	26.3	39.5









Bis(2-Chloroethyl)ether
 Concen: 17.293 ng/ul
 RT: 7.328 min Scan#
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Tgt Ion:	93	Resp:	484426
Ion	Ratio	Lower	Upper
93	100		
63	67.8	56.3	84.5
95	33.1	26.5	39.7



2-Chlorophenol
 Concen: 18.049 ng/ul
 RT: 7.469 min Scan# 762
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	128	Resp:	510461
Ion	Ratio	Lower	Upper
128	100		
64	40.4	34.2	51.4
130	33.0	26.2	39.2



Abundance Scan 911 (8.346 min): BP008518.D\data.ms (-900) #13

2-Methylphenol

Concen: 17.254 ng/ul

RT: 8.346 min Scan# 911

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649

Tgt Ion:108 Resp: 467905

Ion Ratio Lower Upper

108 100

107 89.8 72.3 108.5

Abundance Scan 911 (8.346 min): BP008522.D\data.ms

Ref

108.0

39.0

203.6269.6336.8 417.3 491.0

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 911 (8.346 min): BP008522.D\data.ms (-877) (-)

Sub

107.0

39.0

230.8296.3 398.2465.5530.3

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

8.346

250000

200000

150000

100000

50000

0

Time-->

8.25 8.30 8.35 8.40

Abundance Scan 927 (8.440 min): BP008518.D\data.ms (-917) #14

2,2'-oxybis(1-Chloropropane)

Concen: 17.553 ng/ul

RT: 8.440 min Scan# 927

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Tgt Ion: 45 Resp: 566379

Ion Ratio Lower Upper

45 100

77 17.1 13.5 20.3

79 13.2 10.3 15.5

Abundance Scan 927 (8.440 min): BP008522.D\data.ms

Ref

45.0

121.0

207.0 295.3 370.8437.5506.6

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance Scan 927 (8.440 min): BP008522.D\data.ms (-893) (-)

Sub

45.0

121.0

188.5 269.0338.3 449.9517.7

m/z-->

50 100 150 200 250 300 350 400 450 500

Abundance

8.440

200000

150000


100000

50000

0

Time-->

8.35 8.40 8.45 8.50



Acetophenone

Concen: 18.026 ng/ul

RT: 8.722 min Scan# 975

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649





Tgt Ion: 105 Resp: 761988

Ion Ratio Lower Upper

105 100

77 73.1 60.8 91.2

51 21.8 18.2 27.4



N-Nitroso-di-n-propylamine

Concen: 17.371 ng/ul

RT: 8.716 min Scan# 974

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Tgt Ion: 70 Resp: 365475




Ion Ratio Lower Upper


70 100

42 49.8 39.9 59.9

101 10.1 8.2 12.4

130 24.8 19.0 28.4





4-Methylphenol

Concen: 17.270 ng/ul

RT: 8.669 min Scan# 966

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :





SICV649

Tgt Ion:108 Resp: 501404

Ion Ratio Lower Upper

108 100

107 112.2 88.3 132.5



Hexachloroethane

Concen: 17.589 ng/ul

RT: 8.993 min Scan# 1021

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13




Tgt Ion:117 Resp: 202498

Ion Ratio Lower Upper


117 100

201 115.9 91.0 136.4

199 71.5 56.7 85.1



Abundance Scan 1040 (9.104 min): BP008518.D\data.ms (-1024)



Nitrobenzene

Concen: 18.708 ng/ul

RT: 9.105 min Scan#

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649

Tgt Ion: 77 Resp: 520653



Ion Ratio Lower Upper

77 100


123 49.0 37.8 56.8

65 12.6 9.8 14.6


Abundance Scan 1040 (9.105 min): BP008522.D\data.ms



Abundance Scan 1040 (9.105 min): BP008522.D\data.ms (-1006) (-)



Abundance Scan 1130 (9.634 min): BP008518.D\data.ms (-1123)



Isophorone

Concen: 16.694 ng/ul

RT: 9.634 min Scan# 1130

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Tgt Ion: 82 Resp: 974333



Ion Ratio Lower Upper

82 100


95 7.2 5.5 8.3


138 20.0 15.3 22.9

Abundance Scan 1130 (9.634 min): BP008522.D\data.ms



Abundance Scan 1130 (9.634 min): BP008522.D\data.ms (-1096) (-)





2-Nitrophenol

Concen: 19.291 ng/ul

RT: 9.816 min Scan# 1161

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649





Tgt Ion:139 Resp: 245523

Ion Ratio Lower Upper

139 100

65 42.5 34.9 52.3

109 26.3 21.3 31.9



2,4-Dimethylphenol

Concen: 17.952 ng/ul

RT: 9.881 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13




Tgt Ion:107 Resp: 539952


Ion Ratio Lower Upper

107 100

121 53.6 41.6 62.4

122 89.9 68.7 103.1









Bis(2-Chloroethoxy)methane
 Concen: 17.283 ng/ul
 RT: 10.116 min Scan# 1212
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	93	Resp:	630508
Ion	Ratio	Lower	Upper
93	100		
95	32.0	26.4	39.6
123	11.9	9.8	14.8



2,4-Dichlorophenol
 Concen: 18.348 ng/ul
 RT: 10.352 min Scan# 1252
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	162	Resp:	499612
Ion	Ratio	Lower	Upper
162	100		
164	66.8	52.5	78.7
98	33.1	27.3	40.9









Scan 1321 (10.757 min): BP008518.D\data.ms (-1330) (-)
 Naphthalene
 Concen: 18.227 ng/ul
 RT: 10.758 min Scan# 1321
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 Client Sample Id :
 SICV649


Tgt Ion:	128	Resp:	1644496
Ion	Ratio	Lower	Upper
128	100		
129	10.5	8.8	13.2
127	13.0	10.0	15.0



Scan 1338 (10.857 min): BP008518.D\data.ms (-1330) (-)
 4-Chloroaniline
 Concen: 18.328 ng/ul
 RT: 10.858 min Scan# 1338
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	127	Resp:	696454
Ion	Ratio	Lower	Upper
127	100		
129	32.5	26.6	39.8









Hexachlorobutadiene
Concen: 18.338 ng/ul
RT: 11.057 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13




Instrument :
BNA_P
ClientSampleId :
SICV649


Tgt Ion:	225	Resp:	346228
Ion	Ratio	Lower	Upper
225	100		
223	62.5	51.0	76.4
227	64.5	51.4	77.0



Caprolactam
Concen: 17.422 ng/ul
RT: 11.610 min Scan# 1466
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:	113	Resp:	135172
Ion	Ratio	Lower	Upper
113	100		
55	133.5	111.3	166.9
56	101.7	84.6	127.0









4-Chloro-3-methylphenol
 Concen: 17.676 ng/ul
 RT: 11.981 min Scan# 1529
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	107	Resp:	471874
Ion	Ratio	Lower	Upper
107	100		
144	28.9	21.9	32.9
142	87.1	68.7	103.1



2-Methylnaphthalene
 Concen: 18.996 ng/ul
 RT: 12.369 min Scan# 1595
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	142	Resp:	1212795
Ion	Ratio	Lower	Upper
142	100		
141	88.4	71.8	107.8









Scan 1632 (12.587 min): BP008518.D\data.ms (-1627) (-)
 1-Methylnaphthalene
 Concen: 17.165 ng/u1
 RT: 12.587 min Scan# 1632
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Tgt Ion:142 Resp: 1115675




Ion	Ratio	Lower	Upper
142	100		
141	92.6	73.6	110.4
116	3.3	3.0	4.4




Scan 1657 (12.734 min): BP008518.D\data.ms (-1648) (-)
 1,2,4,5-Tetrachlorobenzene
 Concen: 18.585 ng/u1
 RT: 12.734 min Scan# 1657
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:216 Resp: 652712

Ion	Ratio	Lower	Upper
216	100		
214	78.5	63.3	94.9
179	19.9	16.6	24.8
108	15.4	12.3	18.5









Scan 1655 (12.722 min): BP008518.D\data.ms (-1648) (-)
 Hexachlorocyclopentadiene
 Concen: 17.945 ng/ul
 RT: 12.722 min Scan# 1648
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	237	Resp:	421787
Ion	Ratio	Lower	Upper
237	100		
235	63.2	50.6	76.0
272	13.1	10.8	16.2



Scan 1697 (12.969 min): BP008518.D\data.ms (-1690) (-)
 2,4,6-Trichlorophenol
 Concen: 19.099 ng/ul
 RT: 12.969 min Scan# 1697
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	196	Resp:	380889
Ion	Ratio	Lower	Upper
196	100		
198	98.2	76.4	114.6
200	31.1	24.4	36.6









2,4,5-Trichlorophenol
Concen: 18.816 ng/ul
RT: 13.034 min Scan# 1708
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13




Instrument :
BNA_P
ClientSampleId :
SICV649


Tgt Ion:	196	Resp:	408361
Ion	Ratio	Lower	Upper
196	100		
198	96.1	77.4	116.0
200	30.3	24.9	37.3



1,1'-Biphenyl
Concen: 17.475 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:	154	Resp:	1473950
Ion	Ratio	Lower	Upper
154	100		
153	40.0	32.1	48.1
76	11.6	9.7	14.5





2-Chloronaphthalene

Concen: 18.437 ng/ul

RT: 13.410 min Scan# 1772

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649





Tgt Ion:162 Resp: 1198983

Ion Ratio Lower Upper

162 100

164 32.7 26.2 39.4

127 34.9 29.3 43.9



2-Nitroaniline

Concen: 18.808 ng/ul

RT: 13.604 min Scan# 1805

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13




Tgt Ion: 65 Resp: 245831


Ion Ratio Lower Upper

65 100

92 74.2 60.1 90.1

138 123.3 98.5 147.7









Dimethylphthalate
 Concen: 17.995 ng/ul
 RT: 13.993 min Scan# 1871
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	163	Resp:	1443315
Ion	Ratio	Lower	Upper
163	100		
194	4.3	3.5	5.3
164	10.2	8.1	12.1



2,6-Dinitrotoluene
 Concen: 20.737 ng/ul
 RT: 14.104 min Scan# 1890
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	165	Resp:	295646
Ion	Ratio	Lower	Upper
165	100		
89	47.2	42.1	63.1
121	19.1	17.4	26.0





Acenaphthylene

Concen: 18.012 ng/ul

RT: 14.257 min Scan# 1916

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649





Tgt Ion:152 Resp: 1852668

Ion Ratio Lower Upper

152 100

151 20.0 16.0 24.0

153 12.9 10.8 16.2



3-Nitroaniline

Concen: 20.169 ng/ul

RT: 14.428 min Scan# 1945

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13




Tgt Ion:138 Resp: 296319


Ion Ratio Lower Upper

138 100

108 11.2 12.4 18.6#

92 104.2 85.0 127.4









Scan 1974 (14.598 min): BP008518.D\data.ms (-1952) (-)
 Acenaphthene
 Concen: 18.337 ng/ul
 RT: 14.598 min Scan# 1952
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	153	Resp:	1231582
Ion	Ratio	Lower	Upper
153	100		
152	47.6	38.3	57.5
154	90.0	71.3	106.9



Scan 1979 (14.628 min): BP008518.D\data.ms (-1953) (-)
 2,4-Dinitrophenol
 Concen: 15.826 ng/ul
 RT: 14.628 min Scan# 1979
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	184	Resp:	95241
Ion	Ratio	Lower	Upper
184	100		
63	53.5	41.8	62.8
154	71.7	52.0	78.0









Scan 1996 (14.728 min): BP008518.D\data.ms (-1955) (-)
 4-Nitrophenol
 Concen: 17.912 ng/ul
 RT: 14.728 min Scan# 1996
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:109 Resp: 158702
 Ion Ratio Lower Upper
 109 100
 139 157.0 123.5 185.3
 65 126.7 101.5 152.3



Scan 2031 (14.934 min): BP008518.D\data.ms (-2056) (-)
 Dibenzofuran
 Concen: 18.196 ng/ul
 RT: 14.934 min Scan# 2031
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:168 Resp: 1787045
 Ion Ratio Lower Upper
 168 100
 139 34.7 28.0 42.0









2,4-Dinitrotoluene
 Concen: 18.999 ng/ul
 RT: 14.887 min Scan# 2023
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	165	Resp:	374783
Ion	Ratio	Lower	Upper
165	100		
63	33.3	28.3	42.5
182	3.0	2.4	3.6



2,3,4,6-Tetrachlorophenol
 Concen: 19.352 ng/ul
 RT: 15.157 min Scan# 2069
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	232	Resp:	326244
Ion	Ratio	Lower	Upper
232	100		
131	39.1	32.3	48.5
130	2.0	1.7	2.5
166	30.2	23.8	35.6









Diethylphthalate
 Concen: 17.696 ng/ul
 RT: 15.357 min Scan# 2103
 Delta R.T. -0.006 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	149	Resp:	1386315
Ion	Ratio	Lower	Upper
149	100		
177	23.2	18.8	28.2
150	12.6	9.8	14.8



Fluorene
 Concen: 18.367 ng/ul
 RT: 15.581 min Scan# 2141
 Delta R.T. -0.006 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	166	Resp:	1437147
Ion	Ratio	Lower	Upper
166	100		
165	96.0	77.4	116.0
167	13.5	11.0	16.4









Scan 2141 (15.581 min): BP008518.D\data.ms (-2163) (-)
 4-Chlorophenyl-phenylether
 Concen: 17.870 ng/ul
 RT: 15.581 min Scan# 2141
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Tgt Ion:	204	Resp:	741666
Ion	Ratio	Lower	Upper
204	100		
206	32.9	26.6	39.8
141	52.8	42.5	63.7




Scan 2142 (15.586 min): BP008518.D\data.ms (-2163) (-)
 4-Nitroaniline
 Concen: 21.920 ng/ul
 RT: 15.587 min Scan# 2142
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	138	Resp:	295173
Ion	Ratio	Lower	Upper
138	100		
92	46.9	39.0	58.4
108	67.3	54.9	82.3



Abundance Scan 2153 (15.651 min): BP008518.D\data.ms (-2119) (-)





4,6-Dinitro-2-methylphenol
Concen: 15.935 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13


Instrument :
BNA_P
ClientSampleId :
SICV649

Tgt Ion:198 Resp: 164071
Ion Ratio Lower Upper
198 100
182 4.5 3.4 5.0
77 20.1 16.5 24.7


Abundance Scan 2153 (15.651 min): BP008522.D\data.ms



Abundance Scan 2153 (15.651 min): BP008522.D\data.ms (-2119) (-)





Abundance Scan 2177 (15.792 min): BP008518.D\data.ms (-2119) (-)




N-Nitrosodiphenylamine
Concen: 18.330 ng/ul
RT: 15.793 min Scan# 2177
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:169 Resp: 1239155
Ion Ratio Lower Upper
169 100
168 65.5 53.3 79.9
167 36.1 28.7 43.1

Abundance Scan 2177 (15.793 min): BP008522.D\data.ms



Abundance Scan 2177 (15.793 min): BP008522.D\data.ms (-2143) (-)



Abundance Scan 2293 (16.475 min): BP008518.D\data.ms (-2293) (-)

4-Bromophenyl-phenylether

Concen: 17.917 ng/ul

RT: 16.475 min Scan# 2293

Delta R.T. 0.000 min

Lab File: BP008522.D


Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649






Tgt Ion:248 Resp: 454261

Ion Ratio Lower Upper

248 100

250 97.5 78.6 118.0

141 64.6 52.9 79.3



Abundance Scan 2313 (16.593 min): BP008518.D\data.ms (-2313) (-)

Hexachlorobenzene

Concen: 17.736 ng/ul

RT: 16.593 min Scan# 2313

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13




Tgt Ion:284 Resp: 523534


Ion Ratio Lower Upper

284 100

142 28.3 24.6 37.0

249 29.8 24.1 36.1









Atrazine
Concen: 16.989 ng/ul
RT: 16.745 min Scan# 2339
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13




Instrument :
BNA_P
ClientSampleId :
SICV649


Tgt Ion:	200	Resp:	435822
Ion Ratio	Lower	Upper	
200	100		
173	26.0	20.9	31.3
215	49.5	38.8	58.2



Pentachlorophenol
Concen: 16.619 ng/ul
RT: 16.934 min Scan# 2371
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:	266	Resp:	250743
Ion Ratio	Lower	Upper	
266	100		
264	62.0	49.1	73.7
268	63.7	49.8	74.6









Scan 2438 (17.328 min): BP008518.D\data.ms (-2438)
 Phenanthrene
 Concen: 18.134 ng/ul
 RT: 17.328 min Scan# 2438
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument : BNA_P
 ClientSampleId : SICV649

Tgt Ion:	178	Resp:	2207885
Ion	Ratio	Lower	Upper
178	100		
179	15.2	12.3	18.5
176	19.0	15.4	23.2



Scan 2453 (17.416 min): BP008518.D\data.ms (-2453)
 Anthracene
 Concen: 18.464 ng/ul
 RT: 17.416 min Scan# 2453
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	178	Resp:	2254170
Ion	Ratio	Lower	Upper
178	100		
179	15.8	12.4	18.6
176	18.8	14.8	22.2



Abundance Scan 1760 (13.340 min): BP008518.D\data.ms (-1725) (-)

1,2,3,4-Tetrachlorobenzene

Concen: 16.883 ng/uL

RT: 13.340 min Scan# 1760

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Instrument :

BNA_P

ClientSampleId :

SICV649

Tgt Ion:216 Resp: 633086

Ion Ratio Lower Upper

216 100

214 77.5 63.4 95.0

218 47.4 38.2 57.2

Abundance Scan 1760 (13.340 min): BP008522.D\data.ms

Ref 50

215.9

37.0 108.0 281.0 371.9436.7 508.8

m/z--> 50 100 150 200 250 300 350 400 450 500

Raw 50

37.0 107.9 301.7366.3433.9501.7

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance Scan 1760 (13.340 min): BP008522.D\data.ms (-1726) (-)

Sub 50

215.8

37.0 107.9 282.9 366.3433.9504.7

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

400000

300000

200000

100000

0

13.340

Time--> 13.30 13.40

Abundance Scan 1818 (14.857 min): BP008518.D\data.ms (-2076) (-)

Pentachlorobenzene

Concen: 17.617 ng/uL

RT: 14.857 min Scan# 1818

Delta R.T. 0.000 min

Lab File: BP008522.D

Acq: 23 Dec 2021 12:13

Tgt Ion:250 Resp: 606477

Ion Ratio Lower Upper

250 100

248 61.9 49.5 74.3

252 63.0 50.5 75.7

Abundance Scan 1818 (14.857 min): BP008522.D\data.ms

Ref 50

249.8

37.0 108.0 177.9 348.5415.6483.9548.4

m/z--> 50 100 150 200 250 300 350 400 450 500

Raw 50

37.0 108.0 179.9 318.5386.1450.6518.3

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance Scan 1818 (14.857 min): BP008522.D\data.ms (-1984) (-)

Sub 50

249.8

37.0 108.0 179.9 362.5431.6500.8

m/z--> 50 100 150 200 250 300 350 400 450 500

Abundance

400000

300000


200000

100000

0

14.857

Time--> 14.80 14.90







Carbazole
 Concen: 19.043 ng/ul
 RT: 17.681 min Scan# 2498
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Instrument :
 BNA_P
 ClientSampleId :
 SICV649

Tgt Ion:167 Resp: 1981140




Ion	Ratio	Lower	Upper
167	100		
166	21.4	17.4	26.0
139	12.0	10.0	15.0




Di-n-butylphthalate
 Concen: 18.916 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:149 Resp: 2297733

Ion	Ratio	Lower	Upper
149	100		
150	9.3	7.4	11.2
104	4.5	3.6	5.4






Abundance Scan 2771 (19.286 min): BP008518.D\data.ms (-2737) (-)




Fluoranthene
Concen: 15.943 ng/ul
RT: 19.286 min Scan# 2771
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Instrument :
BNA_P
ClientSampleId :
SICV649

Tgt Ion:202 Resp: 2548250
Ion Ratio Lower Upper
202 100
101 10.4 9.0 13.4
100 7.9 6.6 10.0







Abundance Scan 2831 (19.639 min): BP008518.D\data.ms (-2737) (-)



Pyrene
Concen: 16.423 ng/ul
RT: 19.639 min Scan# 2831
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:202 Resp: 2645002
Ion Ratio Lower Upper
202 100
101 11.6 10.3 15.5
100 9.7 8.6 13.0









Butylbenzylphthalate
 Concen: 17.248 ng/ul
 RT: 20.522 min Scan# 2981
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	149	Resp:	941532
Ion	Ratio	Lower	Upper
149	100		
91	61.2	53.4	80.2
206	23.3	18.2	27.2



3,3'-Dichlorobenzidine
 Concen: 21.183 ng/ul
 RT: 21.280 min Scan# 3110
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	252	Resp:	1006544
Ion	Ratio	Lower	Upper
252	100		
254	64.0	51.3	76.9
126	9.6	8.1	12.1









Benzo(a)anthracene
 Concen: 18.004 ng/u1
 RT: 21.345 min Scan# 3122
 Delta R.T. -0.006 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	228	Resp:	2607520
Ion	Ratio	Lower	Upper
228	100		
229	19.7	16.2	24.2
226	27.7	22.2	33.2



Bis(2-ethylhexyl)phthalate
 Concen: 18.766 ng/u1
 RT: 21.286 min Scan# 3111
 Delta R.T. -0.006 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	149	Resp:	1554870
Ion	Ratio	Lower	Upper
149	100		
167	28.9	23.0	34.4
279	4.7	3.7	5.5









Chrysene
Concen: 17.829 ng/u1
RT: 21.398 min Scan# 3130
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Instrument :
BNA_P
ClientSampleId :
SICV649




Tgt Ion:228 Resp: 2493131


Ion	Ratio	Lower	Upper
228	100		
226	30.2	24.2	36.4
229	19.7	16.0	24.0



Di-n-octyl phthalate
Concen: 16.839 ng/u1
RT: 22.227 min Scan# 3271
Delta R.T. 0.000 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:149 Resp: 2506551









Benzo(b)fluoranthene
 Concen: 17.352 ng/u1
 RT: 23.051 min Scan# 3411
 Delta R.T. -0.006 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	252	Resp:	2687792
Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.7	26.5
125	9.6	7.4	11.2



Benzo(k)fluoranthene
 Concen: 17.935 ng/u1
 RT: 23.104 min Scan# 3420
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	252	Resp:	2681300
Ion	Ratio	Lower	Upper
252	100		
253	21.6	17.7	26.5
125	8.7	7.4	11.0









Benzo(a)pyrene
 Concen: 17.844 ng/u1
 RT: 23.686 min Scan# 3519
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13




Instrument :
 BNA_P
 ClientSampleId :
 SICV649


Tgt Ion:	252	Resp:	2645800
Ion	Ratio	Lower	Upper
252	100		
253	21.9	17.4	26.0
125	10.0	8.0	12.0



Indeno(1,2,3-cd)pyrene
 Concen: 19.493 ng/u1
 RT: 26.327 min Scan# 3968
 Delta R.T. 0.000 min
 Lab File: BP008522.D
 Acq: 23 Dec 2021 12:13

Tgt Ion:	276	Resp:	3187242
Ion	Ratio	Lower	Upper
276	100		
138	22.1	19.1	28.7
277	24.8	20.3	30.5









Dibenzo(a,h)anthracene
Concen: 19.123 ng/ul
RT: 26.345 min Scan# 4095
Delta R.T. -0.006 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Instrument :
BNA_P
ClientSampleId :
SICV649

Tgt Ion:	278	Resp:	2713105
Ion	Ratio	Lower	Upper
278	100		
139	15.4	12.0	18.0
279	24.1	19.4	29.0



Benzo(g,h,i)perylene
Concen: 19.343 ng/ul
RT: 27.098 min Scan# 4099
Delta R.T. -0.006 min
Lab File: BP008522.D
Acq: 23 Dec 2021 12:13

Tgt Ion:	276	Resp:	2614104
Ion	Ratio	Lower	Upper
276	100		
138	20.1	16.2	24.4
277	24.0	19.1	28.7

