

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
 Data File : BP008529.D
 Acq On : 23 Dec 2021 17:42
 Operator : CG/JU
 Sample : PB141621BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS621

Quant Time: Dec 24 04:12:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Jagrut Upadhyay 12/27/2021
 Supervised By :mohammad ahmed 12/31/2021

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	533362	20.000	ng/u1	0.00
20) Naphthalene-d8	10.710	136	2209334	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	1478213	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.280	188	2966525	20.000	ng/u1	0.00
79) Chrysene-d12	21.363	240	2081272	20.000	ng/u1	0.00
88) Perylene-d12	23.792	264	1970984	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	73588	6.219	ng/uL	0.00
4) Pyridine-d5	3.758	84	850455	25.682	ng/u1	0.00
7) Phenol-d5	7.063	99	1411144	33.165	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.234	67	756159	31.289	ng/u1	0.00
11) 2-Chlorophenol-d4	7.434	132	1150729	33.997	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	1147272	33.612	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	555620	35.540	ng/u1	0.00
24) 2-Nitrophenol-d4	9.787	143	569789	39.750	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.328	165	1248905	35.493	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	1228163	25.286	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	3735957	34.206	ng/u1	0.00
49) Acenaphthylene-d8	14.228	160	4630862	33.909	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	620062	35.815	ng/u1	0.00
60) Fluorene-d10	15.528	176	3205693	33.806	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	442623	33.895	ng/u1	0.00
73) Anthracene-d10	17.380	188	4786589	34.050	ng/u1	0.00
81) Pyrene-d10	19.610	212	4884713	39.168	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.639	264	3563458	35.128	ng/u1	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.387	88	144048	11.260	ng/uL	99
5) Pyridine	3.775	79	981580	28.889	ng/u1	95
6) Benzaldehyde	7.040	77	808192	33.324	ng/u1	95
8) Phenol	7.093	94	1394689	31.901	ng/u1	98
10) Bis(2-Chloroethyl)ether	7.328	93	1074099	30.697	ng/u1	98
12) 2-Chlorophenol	7.469	128	1147113	32.472	ng/u1	96
13) 2-Methylphenol	8.346	108	1082198	31.948	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.440	45	1194296	29.632	ng/u1	99
16) Acetophenone	8.728	105	1666148	31.555	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.716	70	837737	31.876	ng/u1	98
18) 4-Methylphenol	8.675	108	1171751	32.310	ng/u1	100
19) Hexachloroethane	8.993	117	442745	30.787	ng/u1	98
22) Nitrobenzene	9.104	77	1163276	32.248	ng/u1	97
23) Isophorone	9.634	82	2448854	32.371	ng/u1	99
25) 2-Nitrophenol	9.816	139	611529	37.070	ng/u1	98
26) 2,4-Dimethylphenol	9.881	107	1237331	31.738	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.116	93	1481409	31.330	ng/u1	99
29) 2,4-Dichlorophenol	10.351	162	1178883	33.401	ng/u1	99
30) Naphthalene	10.757	128	3645608	31.174	ng/u1	99
32) 4-Chloroaniline	10.863	127	1533579	31.137	ng/u1	98
33) Hexachlorobutadiene	11.057	225	775289	31.681	ng/u1	99
34) Caprolactam	11.616	113	376221m	37.411	ng/u1	
35) 4-Chloro-3-methylphenol	11.987	107	1192987	34.477	ng/u1	98
36) 2-Methylnaphthalene	12.369	142	2640009	31.903	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
 Data File : BP008529.D
 Acq On : 23 Dec 2021 17:42
 Operator : CG/JU
 Sample : PB141621BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS621


Quant Time: Dec 24 04:12:12 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 12/27/2021
 Supervised By :mohammad ahmed 12/31/2021


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.587	142	2697948	32.025	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.734	216	1528430	32.034	ng/ul	99
40) Hexachlorocyclopentadiene	12.722	237	900258	28.193	ng/ul	100
41) 2,4,6-Trichlorophenol	12.969	196	971820	35.869	ng/ul	98
42) 2,4,5-Trichlorophenol	13.040	196	1067521	36.206	ng/ul	100
43) 1,1'-Biphenyl	13.375	154	3614576	31.544	ng/ul	100
44) 2-Chloronaphthalene	13.410	162	2811125	31.818	ng/ul	99
45) 2-Nitroaniline	13.604	65	649008	36.548	ng/ul	95
47) Dimethylphthalate	13.992	163	3586004	32.910	ng/ul	99
48) 2,6-Dinitrotoluene	14.104	165	728480	37.610	ng/ul	92
50) Acenaphthylene	14.257	152	4502413	32.221	ng/ul	100
51) 3-Nitroaniline	14.428	138	723813	36.263	ng/ul#	99
52) Acenaphthene	14.598	153	2903133	31.817	ng/ul	100
53) 2,4-Dinitrophenol	14.634	184	264212	32.317	ng/ul	90
55) 4-Nitrophenol	14.734	109	398624	33.116	ng/ul	97
56) Dibenzofuran	14.934	168	4248718	31.843	ng/ul	99
57) 2,4-Dinitrotoluene	14.886	165	1018743	38.013	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.157	232	848715	37.056	ng/ul	97
59) Diethylphthalate	15.363	149	3506449	32.945	ng/ul	99
61) Fluorene	15.586	166	3432242	32.288	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.581	204	1837616	32.590	ng/ul	98
63) 4-Nitroaniline	15.592	138	724434	39.598	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.651	198	457171	33.087	ng/ul	98
67) N-Nitrosodiphenylamine	15.792	169	3049650	33.616	ng/ul	99
68) 4-Bromophenyl-phenylether	16.475	248	1163643	34.202	ng/ul	97
69) Hexachlorobenzene	16.592	284	1341381	33.863	ng/ul	97
70) Atrazine	16.745	200	1047541	30.430	ng/ul	99
71) Pentachlorophenol	16.933	266	684675	33.816	ng/ul	97
72) Phenanthrene	17.328	178	5384162	32.953	ng/ul	100
74) Anthracene	17.416	178	5300578	32.353	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.340	216	1563741	31.074	ng/ul	98
76) Pentachlorobenzene	14.857	250	1525996	33.033	ng/ul	98
77) Carbazole	17.680	167	4672814	33.470	ng/ul	100
78) Di-n-butylphthalate	18.245	149	5689570	34.903	ng/ul	100
80) Fluoranthene	19.286	202	5807649	39.167	ng/ul	99
82) Pyrene	19.639	202	5683500	38.038	ng/ul	97
83) Butylbenzylphthalate	20.516	149	2030503	40.095	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.280	252	1430432	32.449	ng/ul	99
85) Benzo(a)anthracene	21.351	228	4536381	33.762	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.286	149	2857642	37.177	ng/ul	99
87) Chrysene	21.398	228	4336835	33.430	ng/ul	99
89) Di-n-octyl phthalate	22.221	149	4393081	35.002	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	4224840	32.348	ng/ul	100
91) Benzo(k)fluoranthene	23.104	252	4125884	32.732	ng/ul	99
93) Benzo(a)pyrene	23.686	252	4176487	33.407	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.333	276	5164585	37.461	ng/ul	99
95) Dibenzo(a,h)anthracene	26.351	278	4424248	36.984	ng/ul	99
96) Benzo(g,h,i)perylene	27.103	276	4406093	38.668	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 11.260 ng/uL
RT: 3.387 min Scan# 68
Delta R.T. 0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion: 88 Resp: 144048
Ion Ratio Lower Upper
88 100
43 25.8 20.2 30.2
58 62.4 49.6 74.4

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#5
Pyridine
Concen: 28.889 ng/uL
RT: 3.775 min Scan# 134
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion: 79 Resp: 981580
Ion Ratio Lower Upper
79 100
52 60.1 51.6 77.4
51 29.6 25.2 37.8





#6
Benzaldehyde
Concen: 33.324 ng/ul
RT: 7.040 min Scan# 689
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion: 77 Resp: 808192
Ion Ratio Lower Upper
77 100
105 103.5 78.8 118.2
106 100.1 76.4 114.6

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#8
Phenol
Concen: 31.901 ng/ul
RT: 7.093 min Scan# 698
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion: 94 Resp: 1394689
Ion Ratio Lower Upper
94 100
65 24.6 20.6 31.0
66 32.0 26.3 39.5





#10
Bis(2-Chloroethyl)ether
Concen: 30.697 ng/ul
RT: 7.328 min Scan# 738
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion: 93 Resp: 1074099
Ion Ratio Lower Upper
93 100
63 67.8 56.3 84.5
95 32.8 26.5 39.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#12
2-Chlorophenol
Concen: 32.472 ng/ul
RT: 7.469 min Scan# 762
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:128 Resp: 1147113
Ion Ratio Lower Upper
128 100
64 39.1 34.2 51.4
130 32.4 26.2 39.2





#13
2-Methylphenol
Concen: 31.948 ng/ul
RT: 8.346 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion: 108 Resp: 1082198
Ion Ratio Lower Upper
108 100
107 90.5 72.3 108.5

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#14
2,2'-oxybis(1-Chloropropane)
Concen: 29.632 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion: 45 Resp: 1194296
Ion Ratio Lower Upper
45 100
77 17.3 13.5 20.3
79 13.3 10.3 15.5





#16
Acetophenone
Concen: 31.555 ng/ul
RT: 8.728 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:105 Resp: 166614
Ion Ratio Lower Upper
105 100
77 72.0 60.8 91.2
51 21.0 18.2 27.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#17
N-Nitroso-di-n-propylamine
Concen: 31.876 ng/ul
RT: 8.716 min Scan# 974
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion: 70 Resp: 837737
Ion Ratio Lower Upper
70 100
42 48.8 39.9 59.9
101 10.7 8.2 12.4
130 24.8 19.0 28.4





#18
4-Methylphenol
Concen: 32.310 ng/ul
RT: 8.675 min Scan# 967
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:108 Resp: 117175
Ion Ratio Lower Upper
108 100
107 110.9 88.3 132.5




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#19
Hexachloroethane
Concen: 30.787 ng/ul
RT: 8.993 min Scan# 1021
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:117 Resp: 442745
Ion Ratio Lower Upper
117 100
201 115.4 91.0 136.4
199 73.3 56.7 85.1





#22
Nitrobenzene
Concen: 32.248 ng/ul
RT: 9.104 min Scan# 1040
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion: 77 Resp: 1163270
Ion Ratio Lower Upper
77 100
123 49.7 37.8 56.8
65 12.7 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#23
Isophorone
Concen: 32.371 ng/ul
RT: 9.634 min Scan# 1130
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42

Tgt Ion: 82 Resp: 2448854
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 19.9 15.3 22.9





#25
2-Nitrophenol
Concen: 37.070 ng/ul
RT: 9.816 min Scan# 1161
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42

Instrument :

BNA_P

ClientSampleId :





SLCS621

Tgt Ion:139 Resp: 611529
Ion Ratio Lower Upper
139 100
65 41.9 34.9 52.3
109 26.1 21.3 31.9

Manual Integrations




APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#26
2,4-Dimethylphenol
Concen: 31.738 ng/ul
RT: 9.881 min Scan# 1172
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:107 Resp: 1237331
Ion Ratio Lower Upper
107 100
121 52.6 41.6 62.4
122 88.4 68.7 103.1





#27
Bis(2-Chloroethoxy)methane
Concen: 31.330 ng/ul
RT: 10.116 min Scan# 1212
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion: 93 Resp: 1481409
Ion Ratio Lower Upper
93 100
95 32.6 26.4 39.6
123 12.1 9.8 14.8

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#29
2,4-Dichlorophenol
Concen: 33.401 ng/ul
RT: 10.351 min Scan# 1252
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:162 Resp: 1178883
Ion Ratio Lower Upper
162 100
164 65.4 52.5 78.7
98 32.4 27.3 40.9





#30
Naphthalene
Concen: 31.174 ng/ul
RT: 10.757 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:128 Resp: 3645608
Ion Ratio Lower Upper
128 100
129 11.1 8.8 13.2
127 13.1 10.0 15.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#32
4-Chloroaniline
Concen: 31.137 ng/ul
RT: 10.863 min Scan# 1339
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:127 Resp: 1533579
Ion Ratio Lower Upper
127 100
129 32.2 26.6 39.8





#33
Hexachlorobutadiene
Concen: 31.681 ng/ul
RT: 11.057 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:225 Resp: 775289
Ion Ratio Lower Upper
225 100
223 64.2 51.0 76.4
227 64.7 51.4 77.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#34
Caprolactam
Concen: 37.411 ng/ul m
RT: 11.616 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:113 Resp: 376221
Ion Ratio Lower Upper
113 100
55 131.2 111.3 166.9
56 99.4 84.6 127.0





#35
4-Chloro-3-methylphenol
Concen: 34.477 ng/ul
RT: 11.987 min Scan# 119298
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:107 Resp: 119298
Ion Ratio Lower Upper
107 100
144 28.8 21.9 32.9
142 87.8 68.7 103.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#36
2-Methylnaphthalene
Concen: 31.903 ng/ul
RT: 12.369 min Scan# 1595
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:142 Resp: 2640009
Ion Ratio Lower Upper
142 100
141 87.8 71.8 107.8





#37
1-Methylnaphthalene
Concen: 32.025 ng/ul
RT: 12.587 min Scan# 1632
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:142 Resp: 2697948
Ion Ratio Lower Upper
142 100
141 91.6 73.6 110.4
116 3.3 3.0 4.4




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#39
1,2,4,5-Tetrachlorobenzene
Concen: 32.034 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:216 Resp: 1528430
Ion Ratio Lower Upper
216 100
214 79.0 63.3 94.9
179 20.0 16.6 24.8
108 14.6 12.3 18.5





#40
Hexachlorocyclopentadiene
Concen: 28.193 ng/ul
RT: 12.722 min Scan# 1655
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:237 Resp: 900258
Ion Ratio Lower Upper
237 100
235 63.0 50.6 76.0
272 13.7 10.8 16.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#41
2,4,6-Trichlorophenol
Concen: 35.869 ng/ul
RT: 12.969 min Scan# 1697
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:196 Resp: 971820
Ion Ratio Lower Upper
196 100
198 97.2 76.4 114.6
200 31.2 24.4 36.6





#42
2,4,5-Trichlorophenol
Concen: 36.206 ng/ul
RT: 13.040 min Scan# 1708
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:196 Resp: 106752
Ion Ratio Lower Upper
196 100
198 96.4 77.4 116.0
200 31.4 24.9 37.3

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#43
1,1'-Biphenyl
Concen: 31.544 ng/ul
RT: 13.375 min Scan# 1766
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:154 Resp: 3614576
Ion Ratio Lower Upper
154 100
153 40.2 32.1 48.1
76 11.6 9.7 14.5





#44
2-Chloronaphthalene
Concen: 31.818 ng/ul
RT: 13.410 min Scan# 1772
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:162 Resp: 281112
Ion Ratio Lower Upper
162 100
164 32.8 26.2 39.4
127 35.7 29.3 43.9




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#45
2-Nitroaniline
Concen: 36.548 ng/ul
RT: 13.604 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion: 65 Resp: 649008
Ion Ratio Lower Upper
65 100
92 78.5 60.1 90.1
138 128.9 98.5 147.7





#47
Dimethylphthalate
Concen: 32.910 ng/ul
RT: 13.992 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:163 Resp: 3586004
Ion Ratio Lower Upper
163 100
194 4.5 3.5 5.3
164 10.4 8.1 12.1




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#48
2,6-Dinitrotoluene
Concen: 37.610 ng/ul
RT: 14.104 min Scan# 1890
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:165 Resp: 728480
Ion Ratio Lower Upper
165 100
89 46.5 42.1 63.1
121 19.0 17.4 26.0





#50
Acenaphthylene
Concen: 32.221 ng/ul
RT: 14.257 min Scan# 1916
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:152 Resp: 4502411
Ion Ratio Lower Upper
152 100
151 20.0 16.0 24.0
153 13.3 10.8 16.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#51
3-Nitroaniline
Concen: 36.263 ng/ul
RT: 14.428 min Scan# 1945
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:138 Resp: 723813
Ion Ratio Lower Upper
138 100
108 11.3 12.4 18.6#
92 106.2 85.0 127.4





#52
Acenaphthene
Concen: 31.817 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:153 Resp: 290313
Ion Ratio Lower Upper
153 100
152 47.3 38.3 57.5
154 89.1 71.3 106.9

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#53
2,4-Dinitrophenol
Concen: 32.317 ng/ul
RT: 14.634 min Scan# 1980
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:184 Resp: 264212
Ion Ratio Lower Upper
184 100
63 46.4 41.8 62.8
154 55.8 52.0 78.0





#55
4-Nitrophenol
Concen: 33.116 ng/ul
RT: 14.734 min Scan# 1996
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:109 Resp: 398624
Ion Ratio Lower Upper
109 100
139 160.1 123.5 185.3
65 126.2 101.5 152.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#56
Dibenzofuran
Concen: 31.843 ng/ul
RT: 14.934 min Scan# 2031
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:168 Resp: 4248718
Ion Ratio Lower Upper
168 100
139 34.4 28.0 42.0





#57
2,4-Dinitrotoluene
Concen: 38.013 ng/ul
RT: 14.886 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:165 Resp: 101874
Ion Ratio Lower Upper
165 100
63 32.9 28.3 42.5
182 3.0 2.4 3.6




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#58
2,3,4,6-Tetrachlorophenol
Concen: 37.056 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:232 Resp: 848715
Ion Ratio Lower Upper
232 100
131 37.6 32.3 48.5
130 1.9 1.7 2.5
166 28.8 23.8 35.6





#59
Diethylphthalate
Concen: 32.945 ng/ul
RT: 15.363 min Scan# 2104
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:149 Resp: 3506449
Ion Ratio Lower Upper
149 100
177 24.1 18.8 28.2
150 12.7 9.8 14.8




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#61
Fluorene
Concen: 32.288 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:166 Resp: 3432242
Ion Ratio Lower Upper
166 100
165 97.1 77.4 116.0
167 13.5 11.0 16.4





#62
4-Chlorophenyl-phenylether
Concen: 32.590 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:204 Resp: 1837610
Ion Ratio Lower Upper
204 100
206 33.0 26.6 39.8
141 50.7 42.5 63.7

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#63
4-Nitroaniline
Concen: 39.598 ng/ul
RT: 15.592 min Scan# 2143
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:138 Resp: 724434
Ion Ratio Lower Upper
138 100
92 45.1 39.0 58.4
108 64.0 54.9 82.3





#66
4,6-Dinitro-2-methylphenol
Concen: 33.087 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:198 Resp: 45717
Ion Ratio Lower Upper
198 100
182 3.7 3.4 5.0
77 19.8 16.5 24.7




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#67
N-Nitrosodiphenylamine
Concen: 33.616 ng/ul
RT: 15.792 min Scan# 2177
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:169 Resp: 3049650
Ion Ratio Lower Upper
169 100
168 66.0 53.3 79.9
167 36.0 28.7 43.1





#68
4-Bromophenyl-phenylether
Concen: 34.202 ng/ul
RT: 16.475 min Scan# 2293
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:248 Resp: 116364
Ion Ratio Lower Upper
248 100
250 97.6 78.6 118.0
141 61.7 52.9 79.3




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#69
Hexachlorobenzene
Concen: 33.863 ng/ul
RT: 16.592 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:284 Resp: 1341381
Ion Ratio Lower Upper
284 100
142 28.4 24.6 37.0
249 29.5 24.1 36.1





#70
Atrazine
Concen: 30.430 ng/ul
RT: 16.745 min Scan# 2339
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:200 Resp: 104754
Ion Ratio Lower Upper
200 100
173 26.2 20.9 31.3
215 49.8 38.8 58.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#71
Pentachlorophenol
Concen: 33.816 ng/ul
RT: 16.933 min Scan# 2371
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:266 Resp: 684675
Ion Ratio Lower Upper
266 100
264 64.3 49.1 73.7
268 64.3 49.8 74.6





#72
Phenanthrene
Concen: 32.953 ng/ul
RT: 17.328 min Scan# 2438
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:178 Resp: 5384162
Ion Ratio Lower Upper
178 100
179 15.5 12.3 18.5
176 19.3 15.4 23.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#74
Anthracene
Concen: 32.353 ng/ul
RT: 17.416 min Scan# 2453
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:178 Resp: 5300578
Ion Ratio Lower Upper
178 100
179 15.8 12.4 18.6
176 19.0 14.8 22.2





#75
1,2,3,4-Tetrachlorobenzene
Concen: 31.074 ng/uL
RT: 13.340 min Scan# 1760
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:216 Resp: 156374
Ion Ratio Lower Upper
216 100
214 76.9 63.4 95.0
218 48.7 38.2 57.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#76
Pentachlorobenzene
Concen: 33.033 ng/uL
RT: 14.857 min Scan# 1818
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:250 Resp: 1525996
Ion Ratio Lower Upper
250 100
248 63.6 49.5 74.3
252 63.9 50.5 75.7





#77
 Carbazole
 Concen: 33.470 ng/ul
 RT: 17.680 min Scan# 2498
 Delta R.T. -0.000 min
 Lab File: BP008529.D
 Acq: 23 Dec 2021 17:42




Instrument :
 BNA_P
 ClientSampleId :
 SLCS621




Tgt Ion:167 Resp: 4672814
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 12.2 10.0 15.0

Manual Integrations
 APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
 Supervised By :mohammad ahmed 12/31/2021




#78
 Di-n-butylphthalate
 Concen: 34.903 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. -0.000 min
 Lab File: BP008529.D
 Acq: 23 Dec 2021 17:42




Tgt Ion:149 Resp: 5689570
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.2
 104 4.6 3.6 5.4





#80
Fluoranthene
Concen: 39.167 ng/ul
RT: 19.286 min Scan# 2771
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:202 Resp: 5807649
Ion Ratio Lower Upper
202 100
101 10.5 9.0 13.4
100 8.1 6.6 10.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#82
Pyrene
Concen: 38.038 ng/ul
RT: 19.639 min Scan# 2831
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:202 Resp: 5683500
Ion Ratio Lower Upper
202 100
101 12.1 10.3 15.5
100 9.6 8.6 13.0





#83
Butylbenzylphthalate
Concen: 40.095 ng/ul
RT: 20.516 min Scan# 2981
Delta R.T. -0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:149 Resp: 203050
Ion Ratio Lower Upper
149 100
91 65.8 53.4 80.2
206 23.2 18.2 27.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#84
3,3'-Dichlorobenzidine
Concen: 32.449 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:252 Resp: 1430432
Ion Ratio Lower Upper
252 100
254 64.6 51.3 76.9
126 9.3 8.1 12.1





#85
Benzo(a)anthracene
Concen: 33.762 ng/ul
RT: 21.351 min Scan# 3112
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:228 Resp: 453638
Ion Ratio Lower Upper
228 100
229 19.9 16.2 24.2
226 27.7 22.2 33.2

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#86
Bis(2-ethylhexyl)phthalate
Concen: 37.177 ng/ul
RT: 21.286 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:149 Resp: 2857642
Ion Ratio Lower Upper
149 100
167 29.4 23.0 34.4
279 5.1 3.7 5.5





#87
Chrysene
Concen: 33.430 ng/ul
RT: 21.398 min Scan# 3130
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:228 Resp: 4336835
Ion Ratio Lower Upper
228 100
226 30.6 24.2 36.4
229 20.2 16.0 24.0




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#89
Di-n-octyl phthalate
Concen: 35.002 ng/ul
RT: 22.221 min Scan# 3270
Delta R.T. -0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:149 Resp: 4393081





#90
Benzo(b)fluoranthene
Concen: 32.348 ng/ul
RT: 23.051 min Scan# 3412
Delta R.T. -0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621



Tgt Ion:252 Resp: 4224840
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 9.1 7.4 11.2




Manual Integrations
APPROVED


Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#91
Benzo(k)fluoranthene
Concen: 32.732 ng/ul
RT: 23.104 min Scan# 3420
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42


Tgt Ion:252 Resp: 4125884
Ion Ratio Lower Upper
252 100
253 21.8 17.7 26.5
125 8.8 7.4 11.0





#93
Benzo(a)pyrene
Concen: 33.407 ng/ul
RT: 23.686 min Scan# 3519
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:252 Resp: 417648
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 10.1 8.0 12.0

Manual Integrations
APPROVED



Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021




#94
Indeno(1,2,3-cd)pyrene
Concen: 37.461 ng/ul
RT: 26.333 min Scan# 3969
Delta R.T. 0.006 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Tgt Ion:276 Resp: 5164585
Ion Ratio Lower Upper
276 100
138 22.7 19.1 28.7
277 25.3 20.3 30.5





#95
Dibenzo(a,h)anthracene
Concen: 36.984 ng/ul
RT: 26.351 min Scan# 3972
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42




Instrument :
BNA_P
ClientSampleId :
SLCS621




Tgt Ion:278 Resp: 442424
Ion Ratio Lower Upper
278 100
139 14.9 12.0 18.0
279 23.8 19.4 29.0

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 12/27/2021
Supervised By :mohammad ahmed 12/31/2021



#96
Benzo(g,h,i)perylene
Concen: 38.668 ng/ul
RT: 27.103 min Scan# 4100
Delta R.T. -0.000 min
Lab File: BP008529.D
Acq: 23 Dec 2021 17:42



Tgt Ion:276 Resp: 4406093
Ion Ratio Lower Upper
276 100
138 19.4 16.2 24.4
277 23.4 19.1 28.7

