

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
 Data File : BP008541.D
 Acq On : 24 Dec 2021 00:28
 Operator : CG/JU
 Sample : M5126-16
 Misc :
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 C0J10

Quant Time: Dec 24 04:15:30 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.905	152	583776	20.000	ng/u1	0.00
20) Naphthalene-d8	10.705	136	2481186	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	1767256	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.281	188	3852769	20.000	ng/u1	0.00
79) Chrysene-d12	21.363	240	3827343	20.000	ng/u1	0.00
88) Perylene-d12	23.792	264	3634081	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	58616	4.526	ng/uL	0.00
4) Pyridine-d5	3.758	84	302230	8.339	ng/u1	0.00
7) Phenol-d5	7.058	99	239323	5.139	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	836316	31.617	ng/u1	0.00
11) 2-Chlorophenol-d4	7.434	132	859483	23.200	ng/u1	0.00
15) 4-Methylphenol-d8	8.605	113	515842	13.807	ng/u1	0.00
21) Nitrobenzene-d5	9.058	128	601380	34.253	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	566002	35.159	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	1126829	28.515	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	1333577	24.448	ng/u1	0.00
46) Dimethylphthalate-d6	13.946	166	4675061	35.804	ng/u1	0.00
49) Acenaphthylene-d8	14.228	160	5405349	33.107	ng/u1	0.00
54) 4-Nitrophenol-d4	14.710	143	106860	5.163	ng/u1	0.00
60) Fluorene-d10	15.528	176	3971464	35.032	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.634	200	505521	29.807	ng/u1	0.00
73) Anthracene-d10	17.381	188	6829801	37.408	ng/u1	0.00
81) Pyrene-d10	19.610	212	7978733	34.790	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.633	264	7283687	38.942	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122321\
Data File : BP008541.D
Acq On : 24 Dec 2021 00:28
Operator : CG/JU
Sample : M5126-16
Misc :
ALS Vial : 27 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
C0J10

Quant Time: Dec 24 04:15:30 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Dec 23 14:48:11 2021
Response via : Initial Calibration

