

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122721\
 Data File : BP008560.D
 Acq On : 27 Dec 2021 17:37
 Operator : CG/JU
 Sample : PB141663BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS663

Quant Time: Dec 28 01:30:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.904	152	465940	20.000	ng/u1	0.00
20) Naphthalene-d8	10.704	136	1783028	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.534	164	1038365	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.280	188	1920088	20.000	ng/u1	0.00
79) Chrysene-d12	21.363	240	1869146	20.000	ng/u1	0.00
88) Perylene-d12	23.792	264	2043298	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.352	96	71993	6.964	ng/uL	0.00
4) Pyridine-d5	3.758	84	945069	32.669	ng/u1	0.00
7) Phenol-d5	7.063	99	1220677	32.840	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.228	67	681806	32.294	ng/u1	0.00
11) 2-Chlorophenol-d4	7.434	132	1008066	34.092	ng/u1	0.00
15) 4-Methylphenol-d8	8.610	113	941196	31.564	ng/u1	0.00
21) Nitrobenzene-d5	9.057	128	457284	36.244	ng/u1	0.00
24) 2-Nitrophenol-d4	9.781	143	421808	36.462	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.322	165	984044	34.653	ng/u1	0.00
31) 4-Chloroaniline-d4	10.834	131	1145884	29.232	ng/u1	0.00
46) Dimethylphthalate-d6	13.945	166	2549029	33.225	ng/u1	0.00
49) Acenaphthylene-d8	14.222	160	3409995	35.546	ng/u1	0.00
54) 4-Nitrophenol-d4	14.716	143	406235	33.404	ng/u1	0.00
60) Fluorene-d10	15.522	176	2248793	33.760	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.639	200	280140	33.144	ng/u1	0.00
73) Anthracene-d10	17.380	188	3330791	36.607	ng/u1	0.00
81) Pyrene-d10	19.610	212	3618119	32.304	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.639	264	3824088	36.363	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.381	88	147798	13.225	ng/uL	95
5) Pyridine	3.775	79	987612	33.273	ng/u1	96
6) Benzaldehyde	7.034	77	805690	38.028	ng/u1	97
8) Phenol	7.087	94	1269972	33.251	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.322	93	985471	32.239	ng/u1	99
12) 2-Chlorophenol	7.463	128	1036332	33.581	ng/u1	98
13) 2-Methylphenol	8.340	108	943212	31.874	ng/u1	100
14) 2,2'-oxybis(1-Chloropr...	8.440	45	1091578	31.002	ng/u1	100
16) Acetophenone	8.722	105	1463916	31.736	ng/u1	97
17) N-Nitroso-di-n-propyla...	8.716	70	694059	30.231	ng/u1	99
18) 4-Methylphenol	8.669	108	1013856	32.002	ng/u1	100
19) Hexachloroethane	8.987	117	411035	32.718	ng/u1	98
22) Nitrobenzene	9.099	77	1021411	35.085	ng/u1	97
23) Isophorone	9.628	82	1907315	31.240	ng/u1	99
25) 2-Nitrophenol	9.810	139	491976	36.953	ng/u1	99
26) 2,4-Dimethylphenol	9.875	107	1054232	33.507	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.116	93	1242979	32.572	ng/u1	99
29) 2,4-Dichlorophenol	10.345	162	975319	34.241	ng/u1	99
30) Naphthalene	10.757	128	3198288	33.887	ng/u1	99
32) 4-Chloroaniline	10.857	127	1169553	29.423	ng/u1	97
33) Hexachlorobutadiene	11.057	225	677216	34.289	ng/u1	98
34) Caprolactam	11.616	113	241920	29.808	ng/u1	97
35) 4-Chloro-3-methylphenol	11.981	107	885895	31.723	ng/u1	98
36) 2-Methylnaphthalene	12.363	142	2164522	32.411	ng/u1	99


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122721\
 Data File : BP008560.D
 Acq On : 27 Dec 2021 17:37
 Operator : CG/JU
 Sample : PB141663BS
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS663

Quant Time: Dec 28 01:30:07 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SFAM-EPA-BP122321.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Dec 23 14:48:11 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.581	142	2213239	32.553	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.734	216	1235024	36.849	ng/ul	100
40) Hexachlorocyclopentadiene	12.722	237	747155	33.310	ng/ul	100
41) 2,4,6-Trichlorophenol	12.963	196	702451	36.909	ng/ul	99
42) 2,4,5-Trichlorophenol	13.034	196	771168	37.234	ng/ul	99
43) 1,1'-Biphenyl	13.369	154	2888529	35.885	ng/ul	100
44) 2-Chloronaphthalene	13.410	162	2240324	36.099	ng/ul	98
45) 2-Nitroaniline	13.604	65	452951	36.312	ng/ul	95
47) Dimethylphthalate	13.992	163	2553221	33.357	ng/ul	100
48) 2,6-Dinitrotoluene	14.098	165	489477	35.975	ng/ul	96
50) Acenaphthylene	14.251	152	3457526	35.224	ng/ul	99
51) 3-Nitroaniline	14.422	138	459535	32.775	ng/ul#	98
52) Acenaphthene	14.598	153	2234715	34.866	ng/ul	99
53) 2,4-Dinitrophenol	14.628	184	178430	31.069	ng/ul	95
55) 4-Nitrophenol	14.728	109	284759	33.678	ng/ul	97
56) Dibenzofuran	14.928	168	3234228	34.508	ng/ul	99
57) 2,4-Dinitrotoluene	14.886	165	676910	35.957	ng/ul	94
58) 2,3,4,6-Tetrachlorophenol	15.157	232	554429	34.461	ng/ul	98
59) Diethylphthalate	15.357	149	2375317	31.771	ng/ul	99
61) Fluorene	15.581	166	2536843	33.974	ng/ul	99
62) 4-Chlorophenyl-phenyle...	15.575	204	1331920	33.628	ng/ul	99
63) 4-Nitroaniline	15.586	138	503119	39.150	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.651	198	302314	33.803	ng/ul	96
67) N-Nitrosodiphenylamine	15.786	169	2161802	36.816	ng/ul	100
68) 4-Bromophenyl-phenylether	16.469	248	770348	34.982	ng/ul	99
69) Hexachlorobenzene	16.592	284	914083	35.652	ng/ul	97
70) Atrazine	16.739	200	744696	33.423	ng/ul	99
71) Pentachlorophenol	16.927	266	435057	33.198	ng/ul	98
72) Phenanthrene	17.322	178	3876620	36.658	ng/ul	100
74) Anthracene	17.416	178	3888210	36.667	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.334	216	1210190	37.155	ng/ul	99
76) Pentachlorobenzene	14.857	250	1136002	37.992	ng/ul	99
77) Carbazole	17.680	167	3364902	37.238	ng/ul	100
78) Di-n-butylphthalate	18.245	149	3657508	34.666	ng/ul	99
80) Fluoranthene	19.286	202	4346386	32.638	ng/ul	98
82) Pyrene	19.639	202	4439749	33.087	ng/ul	97
83) Butylbenzylphthalate	20.516	149	1429464	31.430	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.280	252	1425316	36.003	ng/ul	99
85) Benzo(a)anthracene	21.351	228	4346318	36.019	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.286	149	2285591	33.110	ng/ul	99
87) Chrysene	21.398	228	4231532	36.320	ng/ul	100
89) Di-n-octyl phthalate	22.221	149	3709327	28.508	ng/ul	100
90) Benzo(b)fluoranthene	23.051	252	4651446	34.354	ng/ul	100
91) Benzo(k)fluoranthene	23.104	252	4607342	35.258	ng/ul	100
93) Benzo(a)pyrene	23.692	252	4721932	36.433	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.327	276	5806674	40.628	ng/ul	97
95) Dibenzo(a,h)anthracene	26.351	278	4998915	40.309	ng/ul	100
96) Benzo(g,h,i)perylene	27.103	276	4872116	41.244	ng/ul	99





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 13.225 ng/uL
RT: 3.381 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion: 88 Resp: 147798
Ion Ratio Lower Upper
88 100
43 27.6 20.2 30.2
58 66.1 49.6 74.4



#5
Pyridine
Concen: 33.273 ng/uL
RT: 3.775 min Scan# 134
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion: 79 Resp: 987612
Ion Ratio Lower Upper
79 100
52 61.3 51.6 77.4
51 29.5 25.2 37.8









#6
Benzaldehyde
Concen: 38.028 ng/ul
RT: 7.034 min Scan# 61
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion: 77 Resp: 805690
Ion Ratio Lower Upper
77 100
105 101.1 78.8 118.2
106 99.7 76.4 114.6



#8
Phenol
Concen: 33.251 ng/ul
RT: 7.087 min Scan# 697
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion: 94 Resp: 1269972
Ion Ratio Lower Upper
94 100
65 25.5 20.6 31.0
66 33.2 26.3 39.5









#10
Bis(2-Chloroethyl)ether
Concen: 32.239 ng/ul
RT: 7.322 min Scan# 737
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion: 93 Resp: 985471
Ion Ratio Lower Upper
93 100
63 69.3 56.3 84.5
95 32.9 26.5 39.7



#12
2-Chlorophenol
Concen: 33.581 ng/ul
RT: 7.463 min Scan# 761
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:128 Resp: 1036332
Ion Ratio Lower Upper
128 100
64 40.8 34.2 51.4
130 32.1 26.2 39.2









#13
2-Methylphenol
Concen: 31.874 ng/ul
RT: 8.340 min Scan# 910
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:108 Resp: 943212
Ion Ratio Lower Upper
108 100
107 90.6 72.3 108.5



#14
2,2'-oxybis(1-Chloropropane)
Concen: 31.002 ng/ul
RT: 8.440 min Scan# 927
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion: 45 Resp: 1091578
Ion Ratio Lower Upper
45 100
77 17.0 13.5 20.3
79 12.9 10.3 15.5









#16
Acetophenone
Concen: 31.736 ng/ul
RT: 8.722 min Scan# 91
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:105 Resp: 1463916
Ion Ratio Lower Upper
105 100
77 72.8 60.8 91.2
51 21.4 18.2 27.4



#17
N-Nitroso-di-n-propylamine
Concen: 30.231 ng/ul
RT: 8.716 min Scan# 974
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion: 70 Resp: 694059
Ion Ratio Lower Upper
70 100
42 48.8 39.9 59.9
101 10.7 8.2 12.4
130 23.8 19.0 28.4









#18
4-Methylphenol
Concen: 32.002 ng/ul
RT: 8.669 min Scan# 90
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:108 Resp: 1013856
Ion Ratio Lower Upper
108 100
107 110.8 88.3 132.5



#19
Hexachloroethane
Concen: 32.718 ng/ul
RT: 8.987 min Scan# 1020
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:117 Resp: 411035
Ion Ratio Lower Upper
117 100
201 112.3 91.0 136.4
199 69.3 56.7 85.1









#22
Nitrobenzene
Concen: 35.085 ng/ul
RT: 9.099 min Scan# 1039
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion: 77 Resp: 1021411
Ion Ratio Lower Upper
77 100
123 49.5 37.8 56.8
65 12.5 9.8 14.6



#23
Isophorone
Concen: 31.240 ng/ul
RT: 9.628 min Scan# 1129
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion: 82 Resp: 1907315
Ion Ratio Lower Upper
82 100
95 6.8 5.5 8.3
138 19.8 15.3 22.9









#25
2-Nitrophenol
Concen: 36.953 ng/ul
RT: 9.810 min Scan# 1161
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:139 Resp: 491976
Ion Ratio Lower Upper
139 100
65 43.1 34.9 52.3
109 26.5 21.3 31.9



#26
2,4-Dimethylphenol
Concen: 33.507 ng/ul
RT: 9.875 min Scan# 1171
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:107 Resp: 1054232
Ion Ratio Lower Upper
107 100
121 52.4 41.6 62.4
122 89.1 68.7 103.1









#27
Bis(2-Chloroethoxy)methane
Concen: 32.572 ng/ul
RT: 10.116 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion: 93 Resp: 1242979
Ion Ratio Lower Upper
93 100
95 32.4 26.4 39.6
123 11.9 9.8 14.8



#29
2,4-Dichlorophenol
Concen: 34.241 ng/ul
RT: 10.345 min Scan# 1251
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:162 Resp: 975319
Ion Ratio Lower Upper
162 100
164 66.2 52.5 78.7
98 33.3 27.3 40.9









#30
Naphthalene
Concen: 33.887 ng/ul
RT: 10.757 min Scan# 1321
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:128 Resp: 3198288
Ion Ratio Lower Upper
128 100
129 11.0 8.8 13.2
127 13.0 10.0 15.0



#32
4-Chloroaniline
Concen: 29.423 ng/ul
RT: 10.857 min Scan# 1338
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:127 Resp: 1169553
Ion Ratio Lower Upper
127 100
129 31.7 26.6 39.8









#33
Hexachlorobutadiene
Concen: 34.289 ng/ul
RT: 11.057 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:225 Resp: 677216
Ion Ratio Lower Upper
225 100
223 61.3 51.0 76.4
227 64.6 51.4 77.0



#34
Caprolactam
Concen: 29.808 ng/ul
RT: 11.616 min Scan# 1467
Delta R.T. 0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:113 Resp: 241920
Ion Ratio Lower Upper
113 100
55 136.6 111.3 166.9
56 100.9 84.6 127.0









#35
4-Chloro-3-methylphenol
Concen: 31.723 ng/ul
RT: 11.981 min Scan# 1529
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:107 Resp: 885895
Ion Ratio Lower Upper
107 100
144 28.4 21.9 32.9
142 87.8 68.7 103.1



#36
2-Methylnaphthalene
Concen: 32.411 ng/ul
RT: 12.363 min Scan# 1594
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:142 Resp: 2164522
Ion Ratio Lower Upper
142 100
141 88.7 71.8 107.8









#37
1-Methylnaphthalene
Concen: 32.553 ng/ul
RT: 12.581 min Scan# 1631
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:142 Resp: 2213239
Ion Ratio Lower Upper
142 100
141 90.8 73.6 110.4
116 3.4 3.0 4.4



#39
1,2,4,5-Tetrachlorobenzene
Concen: 36.849 ng/ul
RT: 12.734 min Scan# 1657
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:216 Resp: 1235024
Ion Ratio Lower Upper
216 100
214 78.8 63.3 94.9
179 20.4 16.6 24.8
108 14.9 12.3 18.5









#40
Hexachlorocyclopentadiene
Concen: 33.310 ng/ul
RT: 12.722 min Scan# 1655
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:237	Resp:	747155
Ion Ratio	Lower	Upper
237	100	
235	63.2	50.6 76.0
272	13.4	10.8 16.2



#41
2,4,6-Trichlorophenol
Concen: 36.909 ng/ul
RT: 12.963 min Scan# 1696
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:196	Resp:	702451
Ion Ratio	Lower	Upper
196	100	
198	94.5	76.4 114.6
200	30.3	24.4 36.6









#42
2,4,5-Trichlorophenol
Concen: 37.234 ng/ul
RT: 13.034 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:196 Resp: 771168
Ion Ratio Lower Upper
196 100
198 95.5 77.4 116.0
200 30.8 24.9 37.3



#43
1,1'-Biphenyl
Concen: 35.885 ng/ul
RT: 13.369 min Scan# 1765
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:154 Resp: 2888529
Ion Ratio Lower Upper
154 100
153 39.8 32.1 48.1
76 11.9 9.7 14.5





#44

2-Chloronaphthalene

Concen: 36.099 ng/ul

RT: 13.410 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP008560.D

Acq: 27 Dec 2021 17:37

Instrument :

BNA_P

ClientSampleId :

SLCS663





Tgt Ion:162 Resp: 2240324

Ion Ratio Lower Upper

162 100

164 32.7 26.2 39.4

127 34.9 29.3 43.9



#45

2-Nitroaniline

Concen: 36.312 ng/ul

RT: 13.604 min Scan# 1805

Delta R.T. -0.000 min

Lab File: BP008560.D

Acq: 27 Dec 2021 17:37




Tgt Ion: 65 Resp: 452951


Ion Ratio Lower Upper

65 100

92 78.2 60.1 90.1

138 130.3 98.5 147.7









#47
Dimethylphthalate
Concen: 33.357 ng/ul
RT: 13.992 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:163 Resp: 2553221
Ion Ratio Lower Upper
163 100
194 4.4 3.5 5.3
164 10.0 8.1 12.1



#48
2,6-Dinitrotoluene
Concen: 35.975 ng/ul
RT: 14.098 min Scan# 1889
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:165 Resp: 489477
Ion Ratio Lower Upper
165 100
89 50.3 42.1 63.1
121 19.7 17.4 26.0









#50
Acenaphthylene
Concen: 35.224 ng/ul
RT: 14.251 min Scan# 1915
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:152 Resp: 3457526
Ion Ratio Lower Upper
152 100
151 19.8 16.0 24.0
153 13.2 10.8 16.2



#51
3-Nitroaniline
Concen: 32.775 ng/ul
RT: 14.422 min Scan# 1944
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37


Tgt Ion:138 Resp: 459535
Ion Ratio Lower Upper
138 100
108 12.1 12.4 18.6#
92 107.2 85.0 127.4








#52
Acenaphthene
Concen: 34.866 ng/ul
RT: 14.598 min Scan# 1974
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37


Instrument :
BNA_P
ClientSampleId :
SLCS663





Tgt Ion:153 Resp: 2234715
Ion Ratio Lower Upper
153 100
152 47.3 38.3 57.5
154 88.8 71.3 106.9




#53
2,4-Dinitrophenol
Concen: 31.069 ng/ul
RT: 14.628 min Scan# 1979
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37



Tgt Ion:184 Resp: 178430
Ion Ratio Lower Upper
184 100
63 49.2 41.8 62.8
154 60.7 52.0 78.0









#55
4-Nitrophenol
Concen: 33.678 ng/ul
RT: 14.728 min Scan# 1996
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:109 Resp: 284759
Ion Ratio Lower Upper
109 100
139 151.3 123.5 185.3
65 123.3 101.5 152.3



#56
Dibenzofuran
Concen: 34.508 ng/ul
RT: 14.928 min Scan# 2030
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:168 Resp: 3234228
Ion Ratio Lower Upper
168 100
139 34.7 28.0 42.0









#57
2,4-Dinitrotoluene
Concen: 35.957 ng/ul
RT: 14.886 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:165 Resp: 676910
Ion Ratio Lower Upper
165 100
63 31.8 28.3 42.5
182 3.0 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 34.461 ng/ul
RT: 15.157 min Scan# 2069
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:232 Resp: 554429
Ion Ratio Lower Upper
232 100
131 38.6 32.3 48.5
130 2.0 1.7 2.5
166 29.4 23.8 35.6









#59
Diethylphthalate
Concen: 31.771 ng/ul
RT: 15.357 min Scan# 2103
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:149 Resp: 2375317
Ion Ratio Lower Upper
149 100
177 24.2 18.8 28.2
150 12.5 9.8 14.8



#61
Fluorene
Concen: 33.974 ng/ul
RT: 15.581 min Scan# 2141
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:166 Resp: 2536843
Ion Ratio Lower Upper
166 100
165 96.0 77.4 116.0
167 13.6 11.0 16.4









#62
4-Chlorophenyl-phenylether
Concen: 33.628 ng/ul
RT: 15.575 min Scan# 2140
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:204 Resp: 1331920
Ion Ratio Lower Upper
204 100
206 33.2 26.6 39.8
141 51.5 42.5 63.7



#63
4-Nitroaniline
Concen: 39.150 ng/ul
RT: 15.586 min Scan# 2142
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:138 Resp: 503119
Ion Ratio Lower Upper
138 100
92 46.3 39.0 58.4
108 65.2 54.9 82.3









#66
4,6-Dinitro-2-methylphenol
Concen: 33.803 ng/ul
RT: 15.651 min Scan# 2153
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:198 Resp: 302314
Ion Ratio Lower Upper
198 100
182 3.8 3.4 5.0
77 18.7 16.5 24.7



#67
N-Nitrosodiphenylamine
Concen: 36.816 ng/ul
RT: 15.786 min Scan# 2176
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:169 Resp: 2161802
Ion Ratio Lower Upper
169 100
168 66.4 53.3 79.9
167 35.7 28.7 43.1









#68
4-Bromophenyl-phenylether
Concen: 34.982 ng/ul
RT: 16.469 min Scan# 2292
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:248 Resp: 770348
Ion Ratio Lower Upper
248 100
250 97.1 78.6 118.0
141 66.0 52.9 79.3



#69
Hexachlorobenzene
Concen: 35.652 ng/ul
RT: 16.592 min Scan# 2313
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:284 Resp: 914083
Ion Ratio Lower Upper
284 100
142 28.0 24.6 37.0
249 29.7 24.1 36.1









#70
Atrazine
Concen: 33.423 ng/ul
RT: 16.739 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:200 Resp: 744696
Ion Ratio Lower Upper
200 100
173 26.4 20.9 31.3
215 49.4 38.8 58.2



#71
Pentachlorophenol
Concen: 33.198 ng/ul
RT: 16.927 min Scan# 2370
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:266 Resp: 435057
Ion Ratio Lower Upper
266 100
264 62.4 49.1 73.7
268 63.7 49.8 74.6









#72
Phenanthrene
Concen: 36.658 ng/ul
RT: 17.322 min Scan# 2438
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:178 Resp: 3876620
Ion Ratio Lower Upper
178 100
179 15.7 12.3 18.5
176 19.3 15.4 23.2



#74
Anthracene
Concen: 36.667 ng/ul
RT: 17.416 min Scan# 2453
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:178 Resp: 3888210
Ion Ratio Lower Upper
178 100
179 15.5 12.4 18.6
176 18.6 14.8 22.2









#75
1,2,3,4-Tetrachlorobenzene
Concen: 37.155 ng/uL
RT: 13.334 min Scan# 1759
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:216 Resp: 1210190
Ion Ratio Lower Upper
216 100
214 77.6 63.4 95.0
218 48.1 38.2 57.2



#76
Pentachlorobenzene
Concen: 37.992 ng/uL
RT: 14.857 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:250 Resp: 1136002
Ion Ratio Lower Upper
250 100
248 61.3 49.5 74.3
252 63.8 50.5 75.7









#77
 Carbazole
 Concen: 37.238 ng/ul
 RT: 17.680 min Scan# 2498
 Delta R.T. -0.000 min
 Lab File: BP008560.D
 Acq: 27 Dec 2021 17:37




Instrument :
 BNA_P
 ClientSampleId :
 SLCS663


Tgt Ion:167 Resp: 3364902
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.4 26.0
 139 12.2 10.0 15.0



#78
 Di-n-butylphthalate
 Concen: 34.666 ng/ul
 RT: 18.245 min Scan# 2594
 Delta R.T. -0.000 min
 Lab File: BP008560.D
 Acq: 27 Dec 2021 17:37

Tgt Ion:149 Resp: 3657508
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 4.6 3.6 5.4









#80
Fluoranthene
Concen: 32.638 ng/ul
RT: 19.286 min Scan# 2771
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:202 Resp: 4346386
Ion Ratio Lower Upper
202 100
101 10.1 9.0 13.4
100 7.6 6.6 10.0



#82
Pyrene
Concen: 33.087 ng/ul
RT: 19.639 min Scan# 2831
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:202 Resp: 4439749
Ion Ratio Lower Upper
202 100
101 11.9 10.3 15.5
100 9.6 8.6 13.0









#83
Butylbenzylphthalate
Concen: 31.430 ng/ul
RT: 20.516 min Scan# 2980
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:149 Resp: 1429464
Ion Ratio Lower Upper
149 100
91 66.8 53.4 80.2
206 22.9 18.2 27.2



#84
3,3'-Dichlorobenzidine
Concen: 36.003 ng/ul
RT: 21.280 min Scan# 3110
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:252 Resp: 1425316
Ion Ratio Lower Upper
252 100
254 64.4 51.3 76.9
126 9.1 8.1 12.1









#85
Benzo(a)anthracene
Concen: 36.019 ng/ul
RT: 21.351 min Scan# 3122
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:228 Resp: 4346318
Ion Ratio Lower Upper
228 100
229 20.1 16.2 24.2
226 27.6 22.2 33.2



#86
Bis(2-ethylhexyl)phthalate
Concen: 33.110 ng/ul
RT: 21.286 min Scan# 3111
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:149 Resp: 2285591
Ion Ratio Lower Upper
149 100
167 29.2 23.0 34.4
279 5.3 3.7 5.5









#87
Chrysene
Concen: 36.320 ng/ul
RT: 21.398 min Scan# 3130
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:228 Resp: 4231532
Ion Ratio Lower Upper
228 100
226 30.6 24.2 36.4
229 20.2 16.0 24.0



#89
Di-n-octyl phthalate
Concen: 28.508 ng/ul
RT: 22.221 min Scan# 3270
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:149 Resp: 3709327









#90
Benzo(b)fluoranthene
Concen: 34.354 ng/ul
RT: 23.051 min Scan# 3411
Delta R.T. -0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:252 Resp: 4651446
Ion Ratio Lower Upper
252 100
253 22.3 17.7 26.5
125 9.3 7.4 11.2



#91
Benzo(k)fluoranthene
Concen: 35.258 ng/ul
RT: 23.104 min Scan# 3420
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:252 Resp: 4607342
Ion Ratio Lower Upper
252 100
253 22.1 17.7 26.5
125 8.8 7.4 11.0









#93
Benzo(a)pyrene
Concen: 36.433 ng/ul
RT: 23.692 min Scan# 3520
Delta R.T. 0.006 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37




Instrument :
BNA_P
ClientSampleId :
SLCS663


Tgt Ion:252 Resp: 4721932
Ion Ratio Lower Upper
252 100
253 22.0 17.4 26.0
125 9.7 8.0 12.0



#94
Indeno(1,2,3-cd)pyrene
Concen: 40.628 ng/ul
RT: 26.327 min Scan# 3968
Delta R.T. -0.000 min
Lab File: BP008560.D
Acq: 27 Dec 2021 17:37

Tgt Ion:276 Resp: 5806674
Ion Ratio Lower Upper
276 100
138 21.5 19.1 28.7
277 24.8 20.3 30.5









#95
 Dibenzo(a,h)anthracene
 Concen: 40.309 ng/ul
 RT: 26.351 min Scan# 3971
 Delta R.T. -0.000 min
 Lab File: BP008560.D
 Acq: 27 Dec 2021 17:37

Instrument :
 BNA_P
 ClientSampleId :
 SLCS663

Tgt Ion:278 Resp: 4998915
 Ion Ratio Lower Upper
 278 100
 139 14.8 12.0 18.0
 279 24.0 19.4 29.0



#96
 Benzo(g,h,i)perylene
 Concen: 41.244 ng/ul
 RT: 27.103 min Scan# 4100
 Delta R.T. -0.000 min
 Lab File: BP008560.D
 Acq: 27 Dec 2021 17:37

Tgt Ion:276 Resp: 4872116
 Ion Ratio Lower Upper
 276 100
 138 19.7 16.2 24.4
 277 24.1 19.1 28.7

