

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP122819\  
 Data File : BP001497.D  
 Acq On : 29 Dec 2019 02:02  
 Operator : JU  
 Sample : K6439-14  
 Misc :  
 ALS Vial : 20 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 LIED-BF011-01

Quant Time: Dec 29 08:35:36 2019  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP121919.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Dec 24 16:04:49 2019  
 Response via : Initial Calibration

| Internal Standards          | R.T.  | QIon | Response | Conc   | Units | Dev(Min) |
|-----------------------------|-------|------|----------|--------|-------|----------|
| 1) 1,4-Dichlorobenzene-d4   | 8.27  | 152  | 22171    | 20.00  | ng    | -0.01    |
| 21) Naphthalene-d8          | 10.31 | 136  | 74580    | 20.00  | ng    | -0.01    |
| 39) Acenaphthene-d10        | 13.17 | 164  | 45295    | 20.00  | ng    | 0.00     |
| 64) Phenanthrene-d10        | 15.58 | 188  | 94812    | 20.00  | ng    | 0.00     |
| 76) Chrysene-d12            | 19.23 | 240  | 76838    | 20.00  | ng    | 0.00     |
| 87) Perylene-d12            | 21.01 | 264  | 81368    | 20.00  | ng    | 0.02     |
| System Monitoring Compounds |       |      |          |        |       |          |
| 5) 2-Fluorophenol           | 6.19  | 112  | 148462   | 108.22 | ng    | 0.00     |
| 7) Phenol-d6                | 7.74  | 99   | 199754   | 111.59 | ng    | -0.01    |
| 23) Nitrobenzene-d5         | 9.16  | 82   | 133196   | 68.58  | ng    | -0.02    |
| 42) 2,4,6-Tribromophenol    | 14.47 | 330  | 69757    | 123.18 | ng    | 0.00     |
| 45) 2-Fluorobiphenyl        | 12.09 | 172  | 275244   | 76.84  | ng    | 0.00     |
| 79) Terphenyl-d14           | 17.87 | 244  | 380835   | 84.89  | ng    | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP122819\  
Data File : BP001497.D  
Acq On : 29 Dec 2019 02:02  
Operator : JU  
Sample : K6439-14  
Misc :  
ALS Vial : 20 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
LIED-BF011-01

Quant Time: Dec 29 08:35:36 2019  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA P\METHODS\8270-BP121919.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Dec 24 16:04:49 2019  
Response via : Initial Calibration

