

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122821\
 Data File : BP008588.D
 Acq On : 28 Dec 2021 21:08
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP122821

Quant Time: Dec 30 08:46:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 30 08:44:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	88	0.00
2	1,4-Dioxane	40.000	38.433	3.9	86	0.00
3	Pyridine	40.000	40.461	-1.2	87	0.00
4	n-Nitrosodimethylamine	40.000	39.870	0.3	87	0.00
5 S	2-Fluorophenol	80.000	81.193	-1.5	88	0.00
6	Aniline	40.000	40.675	-1.7	87	0.00
7 S	Phenol-d6	80.000	82.225	-2.8	88	0.00
8	2-Chlorophenol	40.000	40.933	-2.3	88	0.00
9	Benzaldehyde	40.000	43.063	-7.7	90	0.00
10 C	Phenol	40.000	40.802	-2.0	87	0.00
11	bis(2-Chloroethyl)ether	40.000	40.773	-1.9	88	0.00
12	1,3-Dichlorobenzene	40.000	40.253	-0.6	88	0.00
13 C	1,4-Dichlorobenzene	40.000	40.299	-0.7	89	0.00
14	1,2-Dichlorobenzene	40.000	40.093	-0.2	88	0.00
15	Benzyl Alcohol	40.000	41.583	-4.0	88	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.789	0.5	87	0.00
17	2-Methylphenol	40.000	41.387	-3.5	88	0.00
18	Hexachloroethane	40.000	40.483	-1.2	87	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.930	-2.3	87	0.00
20	3+4-Methylphenols	40.000	41.779	-4.4	88	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	87	0.00
22	Acetophenone	40.000	40.593	-1.5	87	0.00
23 S	Nitrobenzene-d5	80.000	82.762	-3.5	88	0.00
24	Nitrobenzene	40.000	41.022	-2.6	88	0.00
25	Isophorone	40.000	41.190	-3.0	87	0.00
26 C	2-Nitrophenol	40.000	44.111	-10.3	90	0.00
27	2,4-Dimethylphenol	40.000	41.252	-3.1	88	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.240	-0.6	86	0.00
29 C	2,4-Dichlorophenol	40.000	42.119	-5.3	88	0.00
30	1,2,4-Trichlorobenzene	40.000	40.562	-1.4	88	0.00
31	Naphthalene	40.000	40.477	-1.2	88	0.00
32	Benzoic acid	40.000	43.534	-8.8	89	-0.01
33	4-Chloroaniline	40.000	40.898	-2.2	87	0.00
34 C	Hexachlorobutadiene	40.000	40.735	-1.8	88	0.00
35	Caprolactam	40.000	42.939	-7.3	88	0.00
36 C	4-Chloro-3-methylphenol	40.000	41.677	-4.2	87	0.00
37	2-Methylnaphthalene	40.000	40.811	-2.0	87	0.00
38	1-Methylnaphthalene	40.000	40.433	-1.1	87	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.876	-2.2	88	0.00
41 P	Hexachlorocyclopentadiene	40.000	41.167	-2.9	86	0.00
42 S	2,4,6-Tribromophenol	80.000	87.111	-8.9	94	0.00
43 C	2,4,6-Trichlorophenol	40.000	42.480	-6.2	88	0.00
44	2,4,5-Trichlorophenol	40.000	41.985	-5.0	87	0.00
45 S	2-Fluorobiphenyl	80.000	82.011	-2.5	88	0.00
46	1,1'-Biphenyl	40.000	40.391	-1.0	88	0.00
47	2-Chloronaphthalene	40.000	40.532	-1.3	88	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122821\
 Data File : BP008588.D
 Acq On : 28 Dec 2021 21:08
 Operator : CG/JU
 Sample : SSTDICV040
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 ICVBP122821

Quant Time: Dec 30 08:46:55 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 30 08:44:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.257	-8.1	88	0.00
49	Acenaphthylene	40.000	40.950	-2.4	88	0.00
50	Dimethylphthalate	40.000	40.665	-1.7	88	0.00
51	2,6-Dinitrotoluene	40.000	42.378	-5.9	88	0.00
52 C	Acenaphthene	40.000	40.617	-1.5	87	0.00
53	3-Nitroaniline	40.000	42.804	-7.0	89	0.00
54 P	2,4-Dinitrophenol	40.000	43.641	-9.1	87	0.00
55	Dibenzofuran	40.000	40.325	-0.8	88	0.00
56 P	4-Nitrophenol	40.000	43.977	-9.9	90	0.00
57	2,4-Dinitrotoluene	40.000	43.347	-8.4	88	0.00
58	Fluorene	40.000	40.680	-1.7	88	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	42.398	-6.0	89	0.00
60	Diethylphthalate	40.000	41.001	-2.5	88	0.00
61	4-Chlorophenyl-phenylether	40.000	40.842	-2.1	88	0.00
62	4-Nitroaniline	40.000	42.690	-6.7	89	0.00
63	Azobenzene	40.000	40.394	-1.0	87	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	88	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.377	-10.9	89	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.887	-2.2	89	0.00
67	4-Bromophenyl-phenylether	40.000	41.828	-4.6	98	0.00
68	Hexachlorobenzene	40.000	40.527	-1.3	89	0.00
69	Atrazine	40.000	42.210	-5.5	90	0.00
70 C	Pentachlorophenol	40.000	43.751	-9.4	89	0.00
71	Phenanthrene	40.000	41.206	-3.0	89	0.00
72	Anthracene	40.000	41.309	-3.3	88	0.00
73	Carbazole	40.000	41.551	-3.9	89	0.00
74	Di-n-butylphthalate	40.000	43.692	-9.2	89	0.00
75 C	Fluoranthene	40.000	43.096	-7.7	90	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	92	0.00
77	Benzydine	40.000	40.956	-2.4	86	0.00
78	Pyrene	40.000	40.956	-2.4	91	0.00
79 S	Terphenyl-d14	80.000	78.247	2.2	98	0.00
80	Butylbenzylphthalate	40.000	42.510	-6.3	91	0.00
81	Benzo(a)anthracene	40.000	41.200	-3.0	92	0.00
82	3,3'-Dichlorobenzidine	40.000	42.958	-7.4	92	0.00
83	Chrysene	40.000	41.070	-2.7	92	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.951	-4.9	90	0.00
85 c	Di-n-octyl phthalate	40.000	45.323	-13.3	92	0.00
86 I	Perylene-d12	20.000	20.000	0.0	96	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.105	-0.3	98	0.00
88	Benzo(b)fluoranthene	40.000	40.662	-1.7	92	0.00
89	Benzo(k)fluoranthene	40.000	41.719	-4.3	97	0.00
90 C	Benzo(a)pyrene	40.000	41.640	-4.1	96	0.00
91	Dibenzo(a,h)anthracene	40.000	40.354	-0.9	99	0.00
92	Benzo(g,h,i)perylene	40.000	39.373	1.6	99	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122821\
Data File : BP008588.D
Acq On : 28 Dec 2021 21:08
Operator : CG/JU
Sample : SSTDICV040
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
ICVBP122821

Quant Time: Dec 30 08:46:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 30 08:44:33 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0