



Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122821\
 Data File : BP008580.D
 Acq On : 28 Dec 2021 16:24
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Dec 29 09:31:54 2021






AutoFind: Scans 2506, 2507, 2508; Background Corrected with Scan 2498

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	21.4	381439	PASS
68	69	0.00	2	1.7	7735	PASS
69	198	0.00	100	25.8	458922	PASS
70	69	0.00	2	0.5	2147	PASS
127	198	10	80	37.7	671106	PASS
197	198	0.00	2	0.1	1947	PASS
198	198	100	100	100.0	1779476	PASS
199	198	5	9	6.9	122125	PASS
275	198	10	60	28.4	505728	PASS
365	198	1	100	3.4	60221	PASS
441	198	0.01	100	11.9	211405	PASS
442	442	100	100	100.0	1381547	PASS
443	442	15	24	19.2	265334	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122821\
 Data File : BP008580.D
 Acq On : 28 Dec 2021 16:24
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Dec 29 10:47:27 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 29 09:31:54 2021
 Response via : Initial Calibration



TIC: BP008580.D\data.ms

(70) Pentachlorophenol (C)

16.928min (+ 0.000) 9808541.03 ng




response 2080953

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	63.60	63.56
264.00	62.90	63.04
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP122821\
 Data File : BP008580.D
 Acq On : 28 Dec 2021 16:24
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Quant Time: Dec 29 10:47:27 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Dec 29 09:31:54 2021
 Response via : Initial Calibration



TIC: BP008580.D\data.ms

(77) Benzidine**19.463min (+ 0.006) 2410564.94 ng****response 12285090**

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.20	16.25
183.00	11.90	13.37
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
12/28/2021	BNA_P	BP008580.D
Compound Name	Response	Retention Time
DDT	6413854	20.669
DDD	28478	20.292
DDE	768	19.745
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
29246	6443100	0.45