

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123021\
 Data File : BP008624.D
 Acq On : 31 Dec 2021 00:24
 Operator : CG/JU
 Sample : M5244-28
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 CONC-11C

Quant Time: Dec 31 08:16:25 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Dec 30 08:44:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.893	152	1030198	20.000	ng	0.00
21) Naphthalene-d8	10.693	136	4113626	20.000	ng	0.00
39) Acenaphthene-d10	14.522	164	2769663	20.000	ng	0.00
64) Phenanthrene-d10	17.275	188	5727537	20.000	ng	0.00
76) Chrysene-d12	21.351	240	4781041	20.000	ng	-0.01
86) Perylene-d12	23.780	264	4524372	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	2047637	36.279	ng	0.00
7) Phenol-d6	7.057	99	3768581	47.970	ng	0.00
23) Nitrobenzene-d5	9.046	82	2796114	40.436	ng	-0.01
42) 2,4,6-Tribromophenol	16.010	330	671258	17.913	ng	0.00
45) 2-Fluorobiphenyl	13.151	172	7943367	41.552	ng	0.00
79) Terphenyl-d14	19.827	244	10814344	46.184	ng	0.00
Target Compounds						
80) Butylbenzylphthalate	20.504	149	1144213	9.223	ng	97
84) Bis(2-ethylhexyl)phtha...	21.274	149	2920063	15.582	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP123021\
Data File : BP008624.D
Acq On : 31 Dec 2021 00:24
Operator : CG/JU
Sample : M5244-28
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
CONC-11C

Quant Time: Dec 31 08:16:25 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP122821.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Dec 30 08:44:33 2021
Response via : Initial Calibration

