

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011922\
 Data File : BP008846.D
 Acq On : 19 Jan 2022 10:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 13:06:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 18:05:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	91	0.00
2	1,4-Dioxane	40.000	39.871	0.3	108	0.00
3	Pyridine	40.000	41.829	-4.6	106	0.00
4	n-Nitrosodimethylamine	40.000	42.197	-5.5	107	0.00
5 S	2-Fluorophenol	80.000	81.301	-1.6	105	0.00
6	Aniline	40.000	40.485	-1.2	102	0.00
7 S	Phenol-d6	80.000	81.770	-2.2	103	0.00
8	2-Chlorophenol	40.000	40.769	-1.9	104	0.00
9	Benzaldehyde	40.000	38.269	4.3	98	0.00
10 C	Phenol	40.000	40.657	-1.6	103	0.00
11	bis(2-Chloroethyl)ether	40.000	40.765	-1.9	104	0.00
12	1,3-Dichlorobenzene	40.000	40.172	-0.4	103	0.00
13 C	1,4-Dichlorobenzene	40.000	40.169	-0.4	104	0.00
14	1,2-Dichlorobenzene	40.000	39.986	0.0	103	0.00
15	Benzyl Alcohol	40.000	40.603	-1.5	101	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	41.047	-2.6	103	0.00
17	2-Methylphenol	40.000	40.287	-0.7	101	0.00
18	Hexachloroethane	40.000	40.822	-2.1	104	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	40.948	-2.4	100	0.00
20	3+4-Methylphenols	40.000	40.366	-0.9	100	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	89	0.00
22	Acetophenone	40.000	40.779	-1.9	100	0.00
23 S	Nitrobenzene-d5	80.000	83.406	-4.3	102	0.00
24	Nitrobenzene	40.000	41.824	-4.6	102	0.00
25	Isophorone	40.000	41.308	-3.3	100	0.00
26 C	2-Nitrophenol	40.000	41.476	-3.7	100	0.00
27	2,4-Dimethylphenol	40.000	40.947	-2.4	99	0.00
28	bis(2-Chloroethoxy)methane	40.000	40.859	-2.1	100	0.00
29 C	2,4-Dichlorophenol	40.000	41.060	-2.7	100	0.00
30	1,2,4-Trichlorobenzene	40.000	40.424	-1.1	101	0.00
31	Naphthalene	40.000	40.136	-0.3	100	0.00
32	Benzoic acid	40.000	41.110	-2.8	94	-0.01
33	4-Chloroaniline	40.000	40.506	-1.3	98	0.00
34 C	Hexachlorobutadiene	40.000	40.243	-0.6	100	0.00
35	Caprolactam	40.000	40.840	-2.1	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	40.788	-2.0	98	0.00
37	2-Methylnaphthalene	40.000	40.798	-2.0	100	0.00
38	1-Methylnaphthalene	40.000	40.633	-1.6	99	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	88	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	40.242	-0.6	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	40.870	-2.2	93	0.00
42 S	2,4,6-Tribromophenol	80.000	81.857	-2.3	100	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.636	-1.6	98	0.00
44	2,4,5-Trichlorophenol	40.000	40.796	-2.0	99	0.00
45 S	2-Fluorobiphenyl	80.000	80.084	-0.1	100	0.00
46	1,1'-Biphenyl	40.000	40.186	-0.5	100	0.00
47	2-Chloronaphthalene	40.000	40.474	-1.2	100	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011922\
 Data File : BP008846.D
 Acq On : 19 Jan 2022 10:44
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Jan 19 13:06:01 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 18:05:00 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	41.567	-3.9	100	0.00
49	Acenaphthylene	40.000	40.563	-1.4	100	0.00
50	Dimethylphthalate	40.000	40.398	-1.0	100	0.00
51	2,6-Dinitrotoluene	40.000	41.781	-4.5	101	0.00
52 C	Acenaphthene	40.000	40.750	-1.9	101	0.00
53	3-Nitroaniline	40.000	41.383	-3.5	100	0.00
54 P	2,4-Dinitrophenol	40.000	44.417	-11.0	104	0.00
55	Dibenzofuran	40.000	40.530	-1.3	100	0.00
56 P	4-Nitrophenol	40.000	42.371	-5.9	100	0.00
57	2,4-Dinitrotoluene	40.000	42.537	-6.3	101	0.00
58	Fluorene	40.000	40.907	-2.3	101	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	41.262	-3.2	100	0.00
60	Diethylphthalate	40.000	41.488	-3.7	102	0.00
61	4-Chlorophenyl-phenylether	40.000	41.150	-2.9	101	0.00
62	4-Nitroaniline	40.000	42.208	-5.5	102	0.00
63	Azobenzene	40.000	42.317	-5.8	102	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	92	0.00
65	4,6-Dinitro-2-methylphenol	40.000	44.716	-11.8	107	0.00
66 c	n-Nitrosodiphenylamine	40.000	40.295	-0.7	102	0.00
67	4-Bromophenyl-phenylether	40.000	40.353	-0.9	101	0.00
68	Hexachlorobenzene	40.000	39.473	1.3	101	0.00
69	Atrazine	40.000	41.378	-3.4	105	0.00
70 C	Pentachlorophenol	40.000	41.589	-4.0	104	0.00
71	Phenanthrene	40.000	40.092	-0.2	103	0.00
72	Anthracene	40.000	41.215	-3.0	105	0.00
73	Carbazole	40.000	41.346	-3.4	107	0.00
74	Di-n-butylphthalate	40.000	42.450	-6.1	107	0.00
75 C	Fluoranthene	40.000	41.749	-4.4	110	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	104	0.00
77	Benzydine	40.000	43.418	-8.5	112	0.00
78	Pyrene	40.000	39.884	0.3	111	0.00
79 S	Terphenyl-d14	80.000	78.091	2.4	109	0.00
80	Butylbenzylphthalate	40.000	39.716	0.7	112	0.00
81	Benzo(a)anthracene	40.000	40.273	-0.7	117	0.00
82	3,3'-Dichlorobenzidine	40.000	41.479	-3.7	119	0.00
83	Chrysene	40.000	40.134	-0.3	117	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	40.464	-1.2	113	0.00
85 c	Di-n-octyl phthalate	40.000	40.311	-0.8	124	0.00
86 I	Perylene-d12	20.000	20.000	0.0	108	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	41.404	-3.5	126	-0.01
88	Benzo(b)fluoranthene	40.000	41.477	-3.7	127	0.00
89	Benzo(k)fluoranthene	40.000	40.905	-2.3	128	0.00
90 C	Benzo(a)pyrene	40.000	41.792	-4.5	128	0.00
91	Dibenzo(a,h)anthracene	40.000	41.250	-3.1	125	-0.01
92	Benzo(g,h,i)perylene	40.000	40.711	-1.8	123	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP011922\
Data File : BP008846.D
Acq On : 19 Jan 2022 10:44
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Jan 19 13:06:01 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 14 18:05:00 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0