

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012422\
 Data File : BP008886.D
 Acq On : 24 Jan 2022 14:47
 Operator : CG/JU
 Sample : N1210-01
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 20220027-KEANSBURG-1-COMPOSITE

Quant Time: Jan 25 04:10:23 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jan 14 18:05:00 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.846	152	287870	20.000	ng	-0.01
21) Naphthalene-d8	10.645	136	1023265	20.000	ng	-0.01
39) Acenaphthene-d10	14.486	164	586193	20.000	ng	-0.01
64) Phenanthrene-d10	17.245	188	1095564	20.000	ng	0.00
76) Chrysene-d12	21.333	240	1162236	20.000	ng	-0.01
86) Perylene-d12	23.751	264	1337860	20.000	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.434	112	1716798	92.225	ng	0.00
7) Phenol-d6	7.022	99	2015213	89.398	ng	-0.01
23) Nitrobenzene-d5	9.004	82	1287904	71.456	ng	-0.01
42) 2,4,6-Tribromophenol	15.980	330	795223	93.198	ng	-0.01
45) 2-Fluorobiphenyl	13.110	172	3078374	69.253	ng	-0.01
79) Terphenyl-d14	19.804	244	3959975	57.512	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP012422\
Data File : BP008886.D
Acq On : 24 Jan 2022 14:47
Operator : CG/JU
Sample : N1210-01
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
20220027-KEANSBURG-1-COMPOSITE

Quant Time: Jan 25 04:10:23 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jan 18:05:00 2022
Response via : Initial Calibration

