

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020122\
 Data File : BP008983.D
 Acq On : 01 Feb 2022 13:11
 Operator : CG/JU
 Sample : N1335-01
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 MH-3S

Quant Time: Feb 01 14:06:30 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 01 04:16:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.828	152	227456	20.000	ng	0.00
21) Naphthalene-d8	10.622	136	885068	20.000	ng	-0.01
39) Acenaphthene-d10	14.463	164	581406	20.000	ng	-0.01
64) Phenanthrene-d10	17.221	188	1196025	20.000	ng	-0.01
76) Chrysene-d12	21.315	240	1270515	20.000	ng	-0.01
86) Perylene-d12	23.721	264	1256504	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.416	112	1140691	77.553	ng	0.00
7) Phenol-d6	7.010	99	1423495	79.921	ng	0.00
23) Nitrobenzene-d5	8.987	82	903479	57.954	ng	-0.01
42) 2,4,6-Tribromophenol	15.963	330	691797	81.744	ng	-0.01
45) 2-Fluorobiphenyl	13.086	172	2334681	52.955	ng	-0.01
79) Terphenyl-d14	19.786	244	3639677	48.355	ng	-0.01

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020122\
Data File : BP008983.D
Acq On : 01 Feb 2022 13:11
Operator : CG/JU
Sample : N1335-01
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
MH-3S

Quant Time: Feb 01 14:06:30 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 01 04:16:03 2022
Response via : Initial Calibration

