

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020122\
 Data File : BP008996.D
 Acq On : 01 Feb 2022 20:23
 Operator : CG/JU
 Sample : N1337-01DL 5X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SS-1DL

Quant Time: Feb 02 04:06:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 01 04:16:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	234358	20.000	ng	-0.01
21) Naphthalene-d8	10.622	136	899374	20.000	ng	#-0.01
39) Acenaphthene-d10	14.463	164	581244	20.000	ng	-0.01
64) Phenanthrene-d10	17.222	188	1255428	20.000	ng	-0.01
76) Chrysene-d12	21.316	240	1274045	20.000	ng	-0.01
86) Perylene-d12	23.721	264	1354773	20.000	ng	-0.02
System Monitoring Compounds						
5) 2-Fluorophenol	5.417	112	242793	16.021	ng	0.00
7) Phenol-d6	7.011	99	294275	16.035	ng	0.00
23) Nitrobenzene-d5	8.987	82	209293	13.212	ng	-0.01
42) 2,4,6-Tribromophenol	15.963	330	136400	16.122	ng	-0.01
45) 2-Fluorobiphenyl	13.087	172	470819	10.682	ng	-0.01
79) Terphenyl-d14	19.781	244	755455	10.009	ng	-0.02
Target Compounds						Qvalue
31) Naphthalene	10.675	128	1498134	30.669	ng	97
37) 2-Methylnaphthalene	12.287	142	1484211	41.563	ng	98
38) 1-Methylnaphthalene	12.510	142	973463	29.699	ng	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020122\
 Data File : BP008996.D
 Acq On : 01 Feb 2022 20:23
 Operator : CG/JU
 Sample : N1337-01DL 5X
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SS-1DL

Quant Time: Feb 02 04:06:51 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP011422.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 01 04:16:03 2022
 Response via : Initial Calibration

