

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020123\
 Data File : BP013690.D
 Acq On : 01 Feb 2023 11:50
 Operator : CG/JU
 Sample : SST04026
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040606

Quant Time: Feb 02 00:17:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 00:14:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.140	152	301482	20.000	ng/u1	0.00
20) Naphthalene-d8	10.969	136	1256398	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.775	164	884876	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.528	188	2085107	20.000	ng/u1	0.00
79) Chrysene-d12	21.604	240	2155934	20.000	ng/u1	0.00
88) Perylene-d12	24.180	264	2438624	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.475	96	109718	15.430	ng/uL	0.00
4) Pyridine-d5	3.911	84	822625	40.319	ng/u1	0.00
7) Phenol-d5	7.287	99	1064707	40.182	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.458	67	626478	40.313	ng/u1	0.00
11) 2-Chlorophenol-d4	7.664	132	793116	40.833	ng/u1	0.00
15) 4-Methylphenol-d8	8.846	113	847588	40.126	ng/u1	0.00
21) Nitrobenzene-d5	9.316	128	361237	45.071	ng/u1	0.00
24) 2-Nitrophenol-d4	10.040	143	430711	46.112	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.581	165	906349	41.494	ng/u1	0.00
31) 4-Chloroaniline-d4	11.099	131	1158546	40.160	ng/u1	0.00
46) Dimethylphthalate-d6	14.181	166	2794296	40.061	ng/u1	0.00
49) Acenaphthylene-d8	14.469	160	3163840	41.054	ng/u1	0.00
54) 4-Nitrophenol-d4	14.963	143	497728	42.335	ng/u1	0.01
60) Fluorene-d10	15.769	176	2421404	40.275	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.881	200	499213	42.723	ng/u1	0.00
73) Anthracene-d10	17.628	188	3897951	40.105	ng/u1	0.00
81) Pyrene-d10	19.845	212	4820107	40.086	ng/u1	0.00
92) Benzo(a)pyrene-d12	24.016	264	5054369	40.601	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.517	88	117053	15.796	ng/uL	96
5) Pyridine	3.928	79	821205	39.972	ng/u1	99
6) Benzaldehyde	7.269	77	514570	48.348	ng/u1	98
8) Phenol	7.311	94	1086218	40.058	ng/u1	100
10) Bis(2-Chloroethyl)ether	7.552	93	876955	40.245	ng/u1	97
12) 2-Chlorophenol	7.699	128	804603	40.888	ng/u1	99
13) 2-Methylphenol	8.575	108	829373	40.059	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.669	45	1033024	40.041	ng/u1	99
16) Acetophenone	8.969	105	1324786	40.231	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.958	70	667719	40.479	ng/u1	99
18) 4-Methylphenol	8.911	108	900042	40.028	ng/u1	99
19) Hexachloroethane	9.234	117	315304	40.935	ng/u1	99
22) Nitrobenzene	9.358	77	953122	44.029	ng/u1	97
23) Isophorone	9.887	82	2015439	40.666	ng/u1	100
25) 2-Nitrophenol	10.075	139	463416	44.840	ng/u1	96
26) 2,4-Dimethylphenol	10.122	107	960260	41.123	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.363	93	1208547	40.237	ng/u1	99
29) 2,4-Dichlorophenol	10.610	162	875235	41.404	ng/u1	98
30) Naphthalene	11.022	128	2665608	40.143	ng/u1	99
32) 4-Chloroaniline	11.128	127	1178256	40.583	ng/u1	99
33) Hexachlorobutadiene	11.304	225	617323	41.407	ng/u1	98
34) Caprolactam	11.910	113	165308	23.148	ng/u1	97
35) 4-Chloro-3-methylphenol	12.234	107	922474	41.173	ng/u1	99
36) 2-Methylnaphthalene	12.616	142	1842792	38.987	ng/u1	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020123\
 Data File : BP013690.D
 Acq On : 01 Feb 2023 11:50
 Operator : CG/JU
 Sample : SST04026
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SST040606

Quant Time: Feb 02 00:17:36 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 00:14:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.834	142	1866574	39.100	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.975	216	1197844	40.980	ng/ul	98
40) Hexachlorocyclopentadiene	12.957	237	674443	41.039	ng/ul	97
41) 2,4,6-Trichlorophenol	13.210	196	806251	41.639	ng/ul	99
42) 2,4,5-Trichlorophenol	13.281	196	887642	42.064	ng/ul	98
43) 1,1'-Biphenyl	13.610	154	2692938	40.195	ng/ul	100
44) 2-Chloronaphthalene	13.657	162	2144605	40.461	ng/ul	100
45) 2-Nitroaniline	13.857	65	483028	47.873	ng/ul	99
47) Dimethylphthalate	14.228	163	2762346	40.198	ng/ul	100
48) 2,6-Dinitrotoluene	14.345	165	490604	50.097	ng/ul	97
50) Acenaphthylene	14.498	152	3293741	40.532	ng/ul	100
51) 3-Nitroaniline	14.675	138	459564	44.397	ng/ul	99
52) Acenaphthene	14.840	153	2249909	40.289	ng/ul	99
53) 2,4-Dinitrophenol	14.881	184	309380	40.674	ng/ul	98
55) 4-Nitrophenol	14.975	109	370077	41.293	ng/ul	99
56) Dibenzofuran	15.175	168	3247788	40.038	ng/ul	99
57) 2,4-Dinitrotoluene	15.128	165	738922	48.039	ng/ul	97
58) 2,3,4,6-Tetrachlorophenol	15.393	232	825649	41.631	ng/ul	98
59) Diethylphthalate	15.587	149	2653366	40.220	ng/ul	100
61) Fluorene	15.822	166	2610139	39.539	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.810	204	1443525	40.137	ng/ul	98
63) 4-Nitroaniline	15.840	138	445854	46.001	ng/ul	100
66) 4,6-Dinitro-2-methylph...	15.892	198	497898	42.816	ng/ul	98
67) N-Nitrosodiphenylamine	16.022	169	2290981	40.707	ng/ul	99
68) 4-Bromophenyl-phenylether	16.710	248	977178	41.192	ng/ul	99
69) Hexachlorobenzene	16.828	284	1139621	40.947	ng/ul	100
70) Atrazine	16.975	200	934229	41.244	ng/ul	100
71) Pentachlorophenol	17.169	266	734983	41.374	ng/ul	100
72) Phenanthrene	17.575	178	4394945	39.903	ng/ul	100
74) Anthracene	17.663	178	4433486	39.743	ng/ul	100
75) 1,2,3,4-Tetrachloroben...	13.581	216	1238350	41.688	ng/uL	98
76) Pentachlorobenzene	15.093	250	1265623	41.559	ng/uL	100
77) Carbazole	17.928	167	3954744	40.864	ng/ul	99
78) Di-n-butylphthalate	18.457	149	4603422	40.258	ng/ul	100
80) Fluoranthene	19.522	202	5550760	39.605	ng/ul	100
82) Pyrene	19.875	202	5697760	39.763	ng/ul	100
83) Butylbenzylphthalate	20.728	149	1724294	41.883	ng/ul	99
84) 3,3'-Dichlorobenzidine	21.510	252	2070320	42.280	ng/ul	98
85) Benzo(a)anthracene	21.586	228	5955246	40.179	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.486	149	2844226	40.296	ng/ul	98
87) Chrysene	21.645	228	5610168	40.204	ng/ul	99
89) Di-n-octyl phthalate	22.469	149	5108912	39.771	ng/ul	100
90) Benzo(b)fluoranthene	23.386	252	6478759	40.698	ng/ul	100
91) Benzo(k)fluoranthene	23.439	252	6022275	39.582	ng/ul	99
93) Benzo(a)pyrene	24.069	252	5553427	40.408	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.904	276	7088151	41.628	ng/ul	99
95) Dibenzo(a,h)anthracene	26.921	278	5886650	41.701	ng/ul	99
96) Benzo(g,h,i)perylene	27.739	276	5621750	41.733	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed







#2
 1,4-Dioxane
 Concen: 15.796 ng/uL
 RT: 3.517 min Scan# 51
 Delta R.T. 0.000 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50

Instrument :
 BNA_P
 ClientSampleId :
 SSTD040606

Tgt Ion: 88 Resp: 117053




Ion	Ratio	Lower	Upper
88	100		
43	23.4	22.2	33.4
58	66.6	54.9	82.3




#5
 Pyridine
 Concen: 39.972 ng/uL
 RT: 3.928 min Scan# 126
 Delta R.T. -0.006 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50

Tgt Ion: 79 Resp: 821205

Ion	Ratio	Lower	Upper
79	100		
52	68.1	53.8	80.6
51	30.6	25.1	37.7









#6
Benzaldehyde
Concen: 48.348 ng/ul
RT: 7.269 min Scan# 694
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion: 77 Resp: 514570
Ion Ratio Lower Upper
77 100
105 99.2 80.6 121.0
106 95.2 75.0 112.4



#8
Phenol
Concen: 40.058 ng/ul
RT: 7.311 min Scan# 701
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion: 94 Resp: 1086218
Ion Ratio Lower Upper
94 100
65 26.2 21.0 31.4
66 34.7 27.4 41.2









#10
 Bis(2-Chloroethyl)ether
 Concen: 40.245 ng/ul
 RT: 7.552 min Scan# 742
 Delta R.T. 0.000 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SSTD040606


Tgt Ion:	93	Resp:	876955
Ion	Ratio	Lower	Upper
93	100		
63	72.4	55.8	83.6
95	32.0	25.1	37.7



#12
 2-Chlorophenol
 Concen: 40.888 ng/ul
 RT: 7.699 min Scan# 767
 Delta R.T. 0.006 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50

Tgt Ion:	128	Resp:	804603
Ion	Ratio	Lower	Upper
128	100		
64	46.0	37.9	56.9
130	31.2	25.2	37.8





#13

2-Methylphenol

Concen: 40.059 ng/ul

RT: 8.575 min Scan# 916

Delta R.T. -0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :





SSTD040606

Tgt Ion:108 Resp: 829373

Ion Ratio Lower Upper

108 100

107 89.4 70.9 106.3



#14

2,2'-oxybis(1-Chloropropane)

Concen: 40.041 ng/ul

RT: 8.669 min Scan# 932

Delta R.T. 0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion: 45 Resp: 1033024


Ion Ratio Lower Upper

45 100

77 16.4 12.6 19.0

79 12.5 10.0 15.0









#16
Acetophenone
Concen: 40.231 ng/ul
RT: 8.969 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:105 Resp: 1324786
Ion Ratio Lower Upper
105 100
77 77.1 61.2 91.8
51 23.3 18.5 27.7



#17
N-Nitroso-di-n-propylamine
Concen: 40.479 ng/ul
RT: 8.958 min Scan# 981
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion: 70 Resp: 667719
Ion Ratio Lower Upper
70 100
42 48.6 38.2 57.4
101 9.5 7.4 11.2
130 18.6 15.8 23.6









#18
4-Methylphenol
Concen: 40.028 ng/ul
RT: 8.911 min Scan# 91
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:108 Resp: 900042
Ion Ratio Lower Upper
108 100
107 110.9 89.9 134.9



#19
Hexachloroethane
Concen: 40.935 ng/ul
RT: 9.234 min Scan# 1028
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:117 Resp: 315304
Ion Ratio Lower Upper
117 100
201 117.2 94.8 142.2
199 73.8 60.1 90.1









#22
Nitrobenzene
Concen: 44.029 ng/ul
RT: 9.358 min Scan# 1048
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion: 77 Resp: 953122
Ion Ratio Lower Upper
77 100
123 44.8 34.1 51.1
65 12.7 9.8 14.8



#23
Isophorone
Concen: 40.666 ng/ul
RT: 9.887 min Scan# 1139
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion: 82 Resp: 2015439
Ion Ratio Lower Upper
82 100
95 7.0 5.4 8.0
138 17.0 13.6 20.4









#25
2-Nitrophenol
Concen: 44.840 ng/ul
RT: 10.075 min Scan# 1170
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:139 Resp: 463416
Ion Ratio Lower Upper
139 100
65 47.7 40.9 61.3
109 28.6 24.3 36.5



#26
2,4-Dimethylphenol
Concen: 41.123 ng/ul
RT: 10.122 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:107 Resp: 960260
Ion Ratio Lower Upper
107 100
121 49.5 39.8 59.8
122 85.5 67.6 101.4









#27
 Bis(2-Chloroethoxy)methane
 Concen: 40.237 ng/ul
 RT: 10.363 min Scan# 1220
 Delta R.T. 0.000 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SST040606


Tgt Ion: 93 Resp: 1208547
 Ion Ratio Lower Upper
 93 100
 95 32.0 25.3 37.9
 123 10.6 8.6 12.8



#29
 2,4-Dichlorophenol
 Concen: 41.404 ng/ul
 RT: 10.610 min Scan# 1262
 Delta R.T. 0.006 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50

Tgt Ion: 162 Resp: 875235
 Ion Ratio Lower Upper
 162 100
 164 67.1 52.1 78.1
 98 33.8 27.7 41.5









#30
Naphthalene
Concen: 40.143 ng/ul
RT: 11.022 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:128 Resp: 2665608
Ion Ratio Lower Upper
128 100
129 10.9 9.0 13.6
127 13.1 10.2 15.4



#32
4-Chloroaniline
Concen: 40.583 ng/ul
RT: 11.128 min Scan# 1350
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:127 Resp: 1178256
Ion Ratio Lower Upper
127 100
129 32.0 26.1 39.1





#33

Hexachlorobutadiene

Concen: 41.407 ng/ul

RT: 11.304 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:225 Resp: 617323

Ion Ratio Lower Upper

225 100

223 61.8 51.4 77.2

227 63.4 50.9 76.3



#34

Caprolactam

Concen: 23.148 ng/ul

RT: 11.910 min Scan# 1483

Delta R.T. 0.012 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion:113 Resp: 165308


Ion Ratio Lower Upper

113 100

55 155.8 128.4 192.6

56 118.2 92.2 138.2









#35
4-Chloro-3-methylphenol
Concen: 41.173 ng/ul
RT: 12.234 min Scan# 1538
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:107 Resp: 922474
Ion Ratio Lower Upper
107 100
144 25.9 21.4 32.0
142 81.3 65.6 98.4



#36
2-Methylnaphthalene
Concen: 38.987 ng/ul
RT: 12.616 min Scan# 1603
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:142 Resp: 1842792
Ion Ratio Lower Upper
142 100
141 87.7 71.4 107.2









#37
1-Methylnaphthalene
Concen: 39.100 ng/ul
RT: 12.834 min Scan# 1640
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:142 Resp: 1866574
Ion Ratio Lower Upper
142 100
141 91.9 74.4 111.6
116 3.8 2.8 4.2



#39
1,2,4,5-Tetrachlorobenzene
Concen: 40.980 ng/ul
RT: 12.975 min Scan# 1664
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:216 Resp: 1197844
Ion Ratio Lower Upper
216 100
214 78.6 61.5 92.3
179 19.4 15.0 22.4
108 14.2 11.0 16.6









#40
Hexachlorocyclopentadiene
Concen: 41.039 ng/ul
RT: 12.957 min Scan# 1661
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:237 Resp: 674443
Ion Ratio Lower Upper
237 100
235 60.5 50.7 76.1
272 13.2 10.6 16.0



#41
2,4,6-Trichlorophenol
Concen: 41.639 ng/ul
RT: 13.210 min Scan# 1704
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:196 Resp: 806251
Ion Ratio Lower Upper
196 100
198 97.0 79.0 118.6
200 31.3 25.2 37.8









#42
2,4,5-Trichlorophenol
Concen: 42.064 ng/ul
RT: 13.281 min Scan# 1716
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:196 Resp: 887642
Ion Ratio Lower Upper
196 100
198 95.5 77.8 116.8
200 31.1 26.2 39.2



#43
1,1'-Biphenyl
Concen: 40.195 ng/ul
RT: 13.610 min Scan# 1772
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:154 Resp: 2692938
Ion Ratio Lower Upper
154 100
153 40.1 32.2 48.2
76 13.0 10.0 15.0





#44

2-Chloronaphthalene

Concen: 40.461 ng/ul

RT: 13.657 min Scan# 1780

Delta R.T. -0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:162 Resp: 2144605

Ion Ratio Lower Upper

162 100

164 33.2 26.6 39.8

127 34.9 27.5 41.3



#45

2-Nitroaniline

Concen: 47.873 ng/ul

RT: 13.857 min Scan# 1814

Delta R.T. 0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion: 65 Resp: 483028


Ion Ratio Lower Upper

65 100

92 73.8 60.6 91.0

138 111.6 89.9 134.9









#47
Dimethylphthalate
Concen: 40.198 ng/ul
RT: 14.228 min Scan# 1876
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:163 Resp: 2762346
Ion Ratio Lower Upper
163 100
194 4.4 3.6 5.4
164 10.1 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 50.097 ng/ul
RT: 14.345 min Scan# 1897
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:165 Resp: 490604
Ion Ratio Lower Upper
165 100
89 53.3 44.6 67.0
121 19.5 15.8 23.6









#50
Acenaphthylene
Concen: 40.532 ng/ul
RT: 14.498 min Scan# 1923
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:152 Resp: 3293741
Ion Ratio Lower Upper
152 100
151 19.8 15.9 23.9
153 13.4 10.7 16.1



#51
3-Nitroaniline
Concen: 44.397 ng/ul
RT: 14.675 min Scan# 1953
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:138 Resp: 459564
Ion Ratio Lower Upper
138 100
108 11.6 10.0 15.0
92 119.9 94.8 142.2





#52

Acenaphthene

Concen: 40.289 ng/ul

RT: 14.840 min Scan# 1981

Delta R.T. 0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:153 Resp: 2249909

Ion Ratio Lower Upper

153 100

152 47.4 38.1 57.1

154 89.0 70.6 105.8



#53

2,4-Dinitrophenol

Concen: 40.674 ng/ul

RT: 14.881 min Scan# 1988

Delta R.T. 0.006 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion:184 Resp: 309380


Ion Ratio Lower Upper

184 100

63 50.6 42.5 63.7

154 57.0 46.0 69.0









#55
4-Nitrophenol
Concen: 41.293 ng/ul
RT: 14.975 min Scan# 2004
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:109 Resp: 370077
Ion Ratio Lower Upper
109 100
139 137.4 109.1 163.7
65 128.3 103.5 155.3



#56
Dibenzofuran
Concen: 40.038 ng/ul
RT: 15.175 min Scan# 2038
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:168 Resp: 3247788
Ion Ratio Lower Upper
168 100
139 33.5 27.2 40.8









#57
2,4-Dinitrotoluene
Concen: 48.039 ng/ul
RT: 15.128 min Scan# 2030
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:165 Resp: 738922
Ion Ratio Lower Upper
165 100
63 38.8 29.4 44.2
182 2.7 2.0 3.0



#58
2,3,4,6-Tetrachlorophenol
Concen: 41.631 ng/ul
RT: 15.393 min Scan# 2075
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:232 Resp: 825649
Ion Ratio Lower Upper
232 100
131 34.1 28.3 42.5
130 1.8 1.5 2.3
166 27.5 21.4 32.2









#59
Diethylphthalate
Concen: 40.220 ng/ul
RT: 15.587 min Scan# 2107
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:149 Resp: 2653366
Ion Ratio Lower Upper
149 100
177 24.2 19.2 28.8
150 12.5 10.0 15.0



#61
Fluorene
Concen: 39.539 ng/ul
RT: 15.822 min Scan# 2148
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:166 Resp: 2610139
Ion Ratio Lower Upper
166 100
165 97.7 78.6 117.8
167 13.7 10.9 16.3









#62
4-Chlorophenyl-phenylether
Concen: 40.137 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:204 Resp: 1443525
Ion Ratio Lower Upper
204 100
206 33.8 26.2 39.2
141 47.9 39.7 59.5



#63
4-Nitroaniline
Concen: 46.001 ng/ul
RT: 15.840 min Scan# 2151
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:138 Resp: 445854
Ion Ratio Lower Upper
138 100
92 52.6 41.7 62.5
108 77.7 62.0 93.0









#66
4,6-Dinitro-2-methylphenol
Concen: 42.816 ng/ul
RT: 15.892 min Scan# 2160
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:198 Resp: 497898
Ion Ratio Lower Upper
198 100
182 4.4 3.5 5.3
77 22.0 16.7 25.1



#67
N-Nitrosodiphenylamine
Concen: 40.707 ng/ul
RT: 16.022 min Scan# 2182
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:169 Resp: 2290981
Ion Ratio Lower Upper
169 100
168 67.1 53.0 79.4
167 36.1 28.9 43.3









#68
4-Bromophenyl-phenylether
Concen: 41.192 ng/ul
RT: 16.710 min Scan# 2119
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:248 Resp: 977178
Ion Ratio Lower Upper
248 100
250 96.7 78.4 117.6
141 54.2 44.2 66.2



#69
Hexachlorobenzene
Concen: 40.947 ng/ul
RT: 16.828 min Scan# 2319
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:284 Resp: 1139621
Ion Ratio Lower Upper
284 100
142 26.6 21.4 32.2
249 30.8 24.5 36.7





#70

Atrazine

Concen: 41.244 ng/ul

RT: 16.975 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:200 Resp: 934229

Ion Ratio Lower Upper

200 100

173 24.9 19.4 29.2

215 50.3 40.2 60.4



#71

Pentachlorophenol

Concen: 41.374 ng/ul

RT: 17.169 min Scan# 2377

Delta R.T. 0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion:266 Resp: 734983


Ion Ratio Lower Upper

266 100

264 63.7 51.2 76.8

268 64.3 51.8 77.8









#72
Phenanthrene
Concen: 39.903 ng/ul
RT: 17.575 min Scan# 2445
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:178 Resp: 4394945
Ion Ratio Lower Upper
178 100
179 15.6 12.6 18.8
176 19.3 15.4 23.2



#74
Anthracene
Concen: 39.743 ng/ul
RT: 17.663 min Scan# 2461
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:178 Resp: 4433486
Ion Ratio Lower Upper
178 100
179 15.7 12.2 18.4
176 18.7 15.0 22.4









#75
1,2,3,4-Tetrachlorobenzene
Concen: 41.688 ng/uL
RT: 13.581 min Scan# 1767
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:216 Resp: 1238350
Ion Ratio Lower Upper
216 100
214 78.4 61.4 92.0
218 49.6 38.4 57.6



#76
Pentachlorobenzene
Concen: 41.559 ng/uL
RT: 15.093 min Scan# 2024
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:250 Resp: 1265623
Ion Ratio Lower Upper
250 100
248 63.1 50.3 75.5
252 64.3 51.8 77.8









#77
 Carbazole
 Concen: 40.864 ng/ul
 RT: 17.928 min Scan# 2506
 Delta R.T. 0.006 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50




Instrument :
 BNA_P
 ClientSampleId :
 SST040606


Tgt Ion:167 Resp: 3954744
 Ion Ratio Lower Upper
 167 100
 166 21.7 17.1 25.7
 139 11.5 9.4 14.2



#78
 Di-n-butylphthalate
 Concen: 40.258 ng/ul
 RT: 18.457 min Scan# 2596
 Delta R.T. -0.000 min
 Lab File: BP013690.D
 Acq: 01 Feb 2023 11:50

Tgt Ion:149 Resp: 4603422
 Ion Ratio Lower Upper
 149 100
 150 9.4 7.4 11.0
 104 4.7 3.8 5.8









#80
Fluoranthene
Concen: 39.605 ng/ul
RT: 19.522 min Scan# 2777
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:202 Resp: 5550760
Ion Ratio Lower Upper
202 100
101 9.4 7.6 11.4
100 7.1 5.7 8.5



#82
Pyrene
Concen: 39.763 ng/ul
RT: 19.875 min Scan# 2837
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:202 Resp: 5697760
Ion Ratio Lower Upper
202 100
101 11.1 8.8 13.2
100 9.0 7.3 10.9





#83

Butylbenzylphthalate

Concen: 41.883 ng/ul

RT: 20.728 min Scan# 2982

Delta R.T. -0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:149 Resp: 1724294

Ion Ratio Lower Upper

149 100

91 67.8 53.9 80.9

206 25.7 19.4 29.2



#84

3,3'-Dichlorobenzidine

Concen: 42.280 ng/ul

RT: 21.510 min Scan# 3115

Delta R.T. 0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion:252 Resp: 2070320


Ion Ratio Lower Upper

252 100

254 64.3 52.6 79.0

126 8.4 7.0 10.6









#85
Benzo(a)anthracene
Concen: 40.179 ng/ul
RT: 21.586 min Scan# 3128
Delta R.T. 0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:228 Resp: 5955246
Ion Ratio Lower Upper
228 100
229 20.1 15.8 23.8
226 27.5 22.0 33.0



#86
Bis(2-ethylhexyl)phthalate
Concen: 40.296 ng/ul
RT: 21.486 min Scan# 3111
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:149 Resp: 2844226
Ion Ratio Lower Upper
149 100
167 29.3 22.7 34.1
279 5.3 3.9 5.9





#87

Chrysene

Concen: 40.204 ng/ul

RT: 21.645 min Scan# 3138

Delta R.T. -0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:228 Resp: 5610168

Ion Ratio Lower Upper

228 100

226 31.1 24.5 36.7

229 20.3 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 39.771 ng/ul




RT: 22.469 min Scan# 3278


Delta R.T. -0.000 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Tgt Ion:149 Resp: 5108912









#90
Benzo(b)fluoranthene
Concen: 40.698 ng/ul
RT: 23.386 min Scan# 3434
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50




Instrument :
BNA_P
ClientSampleId :
SSTD040606


Tgt Ion:252 Resp: 6478759
Ion Ratio Lower Upper
252 100
253 22.3 17.8 26.8
125 7.8 6.5 9.7



#91
Benzo(k)fluoranthene
Concen: 39.582 ng/ul
RT: 23.439 min Scan# 3443
Delta R.T. -0.000 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:252 Resp: 6022275
Ion Ratio Lower Upper
252 100
253 22.3 17.5 26.3
125 8.0 6.7 10.1





#93

Benzo(a)pyrene

Concen: 40.408 ng/ul

RT: 24.069 min Scan# 3

Delta R.T. 0.006 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50

Instrument :

BNA_P

ClientSampleId :

SSTD040606





Tgt Ion:252 Resp: 5553427

Ion Ratio Lower Upper

252 100

253 22.2 17.5 26.3

125 8.7 7.3 10.9



#94

Indeno(1,2,3-cd)pyrene

Concen: 41.628 ng/ul

RT: 26.904 min Scan# 4032

Delta R.T. 0.012 min

Lab File: BP013690.D

Acq: 01 Feb 2023 11:50




Tgt Ion:276 Resp: 7088151


Ion Ratio Lower Upper

276 100

138 19.2 15.0 22.6

277 25.6 20.2 30.2









#95
Dibenzo(a,h)anthracene
Concen: 41.701 ng/ul
RT: 26.921 min Scan# 4034
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Instrument :
BNA_P
ClientSampleId :
SSTD040606

Tgt Ion:278 Resp: 5886650
Ion Ratio Lower Upper
278 100
139 12.9 10.2 15.4
279 24.2 19.0 28.6



#96
Benzo(g,h,i)perylene
Concen: 41.733 ng/ul
RT: 27.739 min Scan# 4174
Delta R.T. 0.006 min
Lab File: BP013690.D
Acq: 01 Feb 2023 11:50

Tgt Ion:276 Resp: 5621750
Ion Ratio Lower Upper
276 100
138 17.7 14.3 21.5
277 24.0 19.2 28.8

