

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP020223\  
 Data File : BP013716.D  
 Acq On : 02 Feb 2023 18:55  
 Operator : CG/JU  
 Sample : PB150547BL  
 Misc :  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 BNA\_P  
 ClientSampleId :  
 SBLK547

Quant Time: Feb 02 23:46:19 2023  
 Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP020123.M  
 Quant Title : SVOA CALIBRATION  
 QLast Update : Thu Feb 02 00:55:26 2023  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|-------------------------------|--------|------|----------|--------|-------|----------|
| -----                         |        |      |          |        |       |          |
| Internal Standards            |        |      |          |        |       |          |
| 1) 1,4-Dichlorobenzene-d4     | 8.134  | 152  | 248817   | 20.000 | ng/ul | 0.00     |
| 20) Naphthalene-d8            | 10.963 | 136  | 1096039  | 20.000 | ng/ul | 0.00     |
| 38) Acenaphthene-d10          | 14.769 | 164  | 827053   | 20.000 | ng/ul | 0.00     |
| 64) Phenanthrene-d10          | 17.528 | 188  | 2029056  | 20.000 | ng/ul | 0.00     |
| 79) Chrysene-d12              | 21.604 | 240  | 2142947  | 20.000 | ng/ul | 0.00     |
| 88) Perylene-d12              | 24.180 | 264  | 2243746  | 20.000 | ng/ul | 0.00     |
| System Monitoring Compounds   |        |      |          |        |       |          |
| 3) 1,4-Dioxane-d8             | 3.475  | 96   | 42329    | 7.213  | ng/uL | 0.00     |
| 4) Pyridine-d5                | 3.911  | 84   | 618585   | 36.736 | ng/ul | 0.00     |
| 7) Phenol-d5                  | 7.281  | 99   | 753921   | 34.476 | ng/ul | 0.00     |
| 9) Bis-(2-Chloroethyl)eth...  | 7.452  | 67   | 464797   | 36.240 | ng/ul | 0.00     |
| 11) 2-Chlorophenol-d4         | 7.658  | 132  | 561496   | 35.027 | ng/ul | 0.00     |
| 15) 4-Methylphenol-d8         | 8.840  | 113  | 613788   | 35.208 | ng/ul | 0.00     |
| 21) Nitrobenzene-d5           | 9.310  | 128  | 257456   | 36.822 | ng/ul | 0.00     |
| 24) 2-Nitrophenol-d4          | 10.034 | 143  | 312723   | 38.379 | ng/ul | 0.00     |
| 28) 2,4-Dichlorophenol-d3     | 10.575 | 165  | 636551   | 33.406 | ng/ul | 0.00     |
| 31) 4-Chloroaniline-d4        | 11.099 | 131  | 879422   | 34.944 | ng/ul | 0.00     |
| 46) Dimethylphthalate-d6      | 14.175 | 166  | 2322720  | 35.628 | ng/ul | 0.00     |
| 49) Acenaphthylene-d8         | 14.469 | 160  | 2525395  | 35.061 | ng/ul | 0.00     |
| 54) 4-Nitrophenol-d4          | 14.951 | 143  | 328424   | 29.888 | ng/ul | 0.00     |
| 60) Fluorene-d10              | 15.763 | 176  | 2050348  | 36.487 | ng/ul | 0.00     |
| 65) 4,6-Dinitro-2-methylph... | 15.875 | 200  | 298462   | 26.248 | ng/ul | 0.00     |
| 73) Anthracene-d10            | 17.628 | 188  | 3457198  | 36.553 | ng/ul | 0.00     |
| 81) Pyrene-d10                | 19.845 | 212  | 4357251  | 36.456 | ng/ul | 0.00     |
| 92) Benzo(a)pyrene-d12        | 24.016 | 264  | 4282881  | 37.392 | ng/ul | 0.00     |

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA\_P\Data\BP020223\  
Data File : BP013716.D  
Acq On : 02 Feb 2023 18:55  
Operator : CG/JU  
Sample : PB150547BL  
Misc :  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
BNA\_P  
ClientSampleId :  
SBLK547

Quant Time: Feb 02 23:46:19 2023  
Quant Method : Z:\svoasrv\HPCHEM1\BNA\_P\Methods\SFAM-EPA-BP020123.M  
Quant Title : SVOA CALIBRATION  
QLast Update : Thu Feb 02 00:55:26 2023  
Response via : Initial Calibration

