

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020223\
 Data File : BP013719.D
 Acq On : 02 Feb 2023 20:38
 Operator : CG/JU
 Sample : 01314-14
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Feb 02 23:46:56 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020123.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Feb 02 00:55:26 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.134	152	303033	20.000	ng/ul	0.00
20) Naphthalene-d8	10.963	136	1339857	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.769	164	995308	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.522	188	2433314	20.000	ng/ul	0.00
79) Chrysene-d12	21.604	240	2360556	20.000	ng/ul	0.00
88) Perylene-d12	24.174	264	2366233	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.481	96	32945	4.609	ng/uL	0.00
4) Pyridine-d5	3.916	84	226778	11.058	ng/ul	0.00
7) Phenol-d5	7.275	99	159283	5.981	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.452	67	488286	31.260	ng/ul	0.00
11) 2-Chlorophenol-d4	7.657	132	448860	22.991	ng/ul	0.00
15) 4-Methylphenol-d8	8.834	113	316409	14.903	ng/ul	-0.01
21) Nitrobenzene-d5	9.310	128	277894	32.512	ng/ul	0.00
24) 2-Nitrophenol-d4	10.034	143	315836	31.707	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.575	165	646266	27.744	ng/ul	0.00
31) 4-Chloroaniline-d4	11.093	131	798715	25.962	ng/ul	0.00
46) Dimethylphthalate-d6	14.169	166	2559297	32.621	ng/ul	0.00
49) Acenaphthylene-d8	14.463	160	2686021	30.987	ng/ul	0.00
54) 4-Nitrophenol-d4	14.951	143	42783	3.235	ng/ul	0.00
60) Fluorene-d10	15.763	176	2238330	33.099	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.869	200	284205	20.842	ng/ul	0.00
73) Anthracene-d10	17.622	188	3928306	34.634	ng/ul	0.00
81) Pyrene-d10	19.839	212	4924806	37.406	ng/ul	0.00
92) Benzo(a)pyrene-d12	24.010	264	4270923	35.357	ng/ul	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020223\
Data File : BP013719.D
Acq On : 02 Feb 2023 20:38
Operator : CG/JU
Sample : 01314-14
Misc :
ALS Vial : 19 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Feb 02 23:46:56 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020123.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Feb 02 00:55:26 2023
Response via : Initial Calibration

