


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP020720\
 Data File : BP001722.D
 Acq On : 07 Feb 2020 15:37
 Operator : CG/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\8270-BP020320.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Mon Feb 03 15:16:18 2020



AutoFind: Scans 2514, 2515, 2516; Background Corrected with Scan 2506

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	33.0	76469	PASS
68	69	0.00	2	1.6	1409	PASS
69	198	0.00	100	37.6	87120	PASS
70	69	0.00	2	0.7	585	PASS
127	198	10	80	48.4	112157	PASS
197	198	0.00	2	0.3	747	PASS
198	198	100	100	100.0	231893	PASS
199	198	5	9	6.9	15927	PASS
275	198	10	60	25.3	58733	PASS
365	198	1	100	3.0	7047	PASS
441	198	0.01	100	14.6	33914	PASS
442	198	50	100	94.4	218906	PASS
443	442	15	24	18.9	41402	PASS