

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021121\
 Data File : BP004725.D
 Acq On : 11 Feb 2021 20:13
 Operator : CG/JU
 Sample : M1368-04
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 EXTERIOR-SLAB

Manual Integrations
 APPROVED

mohammad
 2/12/2021 11:24:55 AM

Quant Time: Feb 11 23:40:38 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.769	152	53670	20.00	ng	# 0.00
21) Naphthalene-d8	10.563	136	202249	20.00	ng	# 0.00
39) Acenaphthene-d10	14.416	164	116932	20.00	ng	0.00
64) Phenanthrene-d10	17.169	188	255442	20.00	ng	0.00
76) Chrysene-d12	21.251	240	274357	20.00	ng	# 0.00
86) Perylene-d12	23.557	264	271691	20.00	ng	-0.01
System Monitoring Compounds						
5) 2-Fluorophenol	5.358	112	178707	51.41	ng	0.00
7) Phenol-d6	6.940	99	360753	75.47	ng	0.00
23) Nitrobenzene-d5	8.928	82	272162	58.63	ng	0.00
42) 2,4,6-Tribromophenol	15.904	330	45735	29.44	ng	-0.01
45) 2-Fluorobiphenyl	13.034	172	555535	58.07	ng	0.00
79) Terphenyl-d14	19.733	244	851140	52.11	ng	0.00
Target Compounds						
10) Phenol	6.964	94	14024	3.03	ng	Qvalue 73
32) Benzoic acid	9.781	122	2173m	7.63	ng	
50) Dimethylphthalate	13.869	163	24205	2.35	ng	# 83
84) Bis(2-ethylhexyl)phtha...	21.169	149	66358	5.78	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021121\
Data File : BP004725.D
Acq On : 11 Feb 2021 20:13
Operator : CG/JU
Sample : M1368-04
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
EXTERIOR-SLAB

Manual Integrations
APPROVED

mohammad
2/12/2021 11:24:55 AM

Quant Time: Feb 11 23:40:38 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 18:04:48 2021
Response via : Initial Calibration

