

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021121\
 Data File : BP004709.D
 Acq On : 11 Feb 2021 09:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 11:23:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	96	0.00
2	1,4-Dioxane	40.000	37.260	6.9	96	0.00
3	Pyridine	40.000	38.639	3.4	95	0.00
4	n-Nitrosodimethylamine	40.000	38.643	3.4	99	0.00
5 S	2-Fluorophenol	80.000	73.538	8.1	98	0.00
6	Aniline	40.000	38.064	4.8	100	0.00
7 S	Phenol-d6	80.000	76.704	4.1	100	0.00
8	2-Chlorophenol	40.000	37.492	6.3	101	0.00
9	Benzaldehyde	40.000	46.722	-16.8	108	0.00
10 C	Phenol	40.000	37.401	6.5	98	0.00
11	bis(2-Chloroethyl)ether	40.000	37.120	7.2	98	0.00
12	1,3-Dichlorobenzene	40.000	36.665	8.3	99	0.00
13 C	1,4-Dichlorobenzene	40.000	37.042	7.4	99	0.00
14	1,2-Dichlorobenzene	40.000	36.775	8.1	99	0.00
15	Benzyl Alcohol	40.000	38.812	3.0	98	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	39.125	2.2	103	0.00
17	2-Methylphenol	40.000	38.763	3.1	101	0.00
18	Hexachloroethane	40.000	37.557	6.1	97	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.332	4.2	97	0.00
20	3+4-Methylphenols	40.000	39.574	1.1	102	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	102	0.00
22	Acetophenone	40.000	36.402	9.0	99	0.00
23 S	Nitrobenzene-d5	80.000	74.039	7.5	100	0.00
24	Nitrobenzene	40.000	36.537	8.7	99	0.00
25	Isophorone	40.000	36.784	8.0	98	0.00
26 C	2-Nitrophenol	40.000	38.369	4.1	102	0.00
27	2,4-Dimethylphenol	40.000	37.136	7.2	101	0.00
28	bis(2-Chloroethoxy)methane	40.000	36.384	9.0	100	0.00
29 C	2,4-Dichlorophenol	40.000	37.012	7.5	100	0.00
30	1,2,4-Trichlorobenzene	40.000	36.419	9.0	101	0.00
31	Naphthalene	40.000	36.039	9.9	101	0.00
32	Benzoic acid	40.000	37.449	6.4	104	0.00
33	4-Chloroaniline	40.000	37.657	5.9	101	0.00
34 C	Hexachlorobutadiene	40.000	37.045	7.4	102	0.00
35	Caprolactam	40.000	39.487	1.3	98	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.468	6.3	101	0.00
37	2-Methylnaphthalene	40.000	36.509	8.7	100	0.00
38	1-Methylnaphthalene	40.000	36.742	8.1	101	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	102	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	36.605	8.5	100	0.00
41 P	Hexachlorocyclopentadiene	40.000	36.990	7.5	101	0.00
42 S	2,4,6-Tribromophenol	80.000	77.170	3.5	105	0.00
43 C	2,4,6-Trichlorophenol	40.000	38.175	4.6	103	0.00
44	2,4,5-Trichlorophenol	40.000	37.345	6.6	102	0.00
45 S	2-Fluorobiphenyl	80.000	73.198	8.5	101	0.00
46	1,1'-Biphenyl	40.000	36.447	8.9	101	0.00
47	2-Chloronaphthalene	40.000	36.584	8.5	102	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021121\
 Data File : BP004709.D
 Acq On : 11 Feb 2021 09:40
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Feb 11 11:23:14 2021
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Feb 05 18:04:48 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	38.917	2.7	101	0.00
49	Acenaphthylene	40.000	36.911	7.7	100	0.00
50	Dimethylphthalate	40.000	36.867	7.8	101	0.00
51	2,6-Dinitrotoluene	40.000	37.536	6.2	99	0.00
52 C	Acenaphthene	40.000	36.698	8.3	100	0.00
53	3-Nitroaniline	40.000	38.562	3.6	101	0.00
54 P	2,4-Dinitrophenol	40.000	38.250	4.4	98	0.00
55	Dibenzofuran	40.000	36.563	8.6	101	0.00
56 P	4-Nitrophenol	40.000	37.959	5.1	100	0.00
57	2,4-Dinitrotoluene	40.000	38.529	3.7	102	0.00
58	Fluorene	40.000	37.024	7.4	102	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	36.419	9.0	98	0.00
60	Diethylphthalate	40.000	36.985	7.5	101	0.00
61	4-Chlorophenyl-phenylether	40.000	36.486	8.8	103	0.00
62	4-Nitroaniline	40.000	39.082	2.3	104	0.00
63	Azobenzene	40.000	37.615	6.0	101	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	101	0.00
65	4,6-Dinitro-2-methylphenol	40.000	39.403	1.5	103	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.247	6.9	101	0.00
67	4-Bromophenyl-phenylether	40.000	36.877	7.8	101	0.00
68	Hexachlorobenzene	40.000	36.905	7.7	100	0.00
69	Atrazine	40.000	38.511	3.7	101	0.00
70 C	Pentachlorophenol	40.000	38.858	2.9	102	0.00
71	Phenanthrene	40.000	36.610	8.5	101	0.00
72	Anthracene	40.000	36.961	7.6	101	0.00
73	Carbazole	40.000	37.234	6.9	102	0.00
74	Di-n-butylphthalate	40.000	38.035	4.9	103	0.00
75 C	Fluoranthene	40.000	37.195	7.0	102	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	103	0.00
77	Benzydine	40.000	46.739	-16.8	115	0.00
78	Pyrene	40.000	36.828	7.9	102	0.00
79 S	Terphenyl-d14	80.000	73.970	7.5	102	0.00
80	Butylbenzylphthalate	40.000	37.623	5.9	100	0.00
81	Benzo(a)anthracene	40.000	36.851	7.9	102	0.00
82	3,3'-Dichlorobenzidine	40.000	38.351	4.1	103	0.00
83	Chrysene	40.000	36.681	8.3	103	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	37.249	6.9	101	0.00
85 c	Di-n-octyl phthalate	40.000	37.360	6.6	102	0.00
86 I	Perylene-d12	20.000	20.000	0.0	103	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	37.574	6.1	104	0.00
88	Benzo(b)fluoranthene	40.000	37.610	6.0	104	0.00
89	Benzo(k)fluoranthene	40.000	36.767	8.1	103	0.00
90 C	Benzo(a)pyrene	40.000	37.532	6.2	103	-0.01
91	Dibenzo(a,h)anthracene	40.000	37.647	5.9	105	-0.01
92	Benzo(g,h,i)perylene	40.000	37.192	7.0	103	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP021121\
Data File : BP004709.D
Acq On : 11 Feb 2021 09:40
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Feb 11 11:23:14 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP020521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Feb 05 18:04:48 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0