

Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP021320\
 Data File : BP001744.D
 Acq On : 13 Feb 2020 19:02
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Feb 14 12:53:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP021320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 14 12:45:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.71	152	301195	20.00	ng/ul	0.00
18) Naphthalene-d8	10.48	136	1403231	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.34	164	931339	20.00	ng/ul	0.00
62) Phenanthrene-d10	17.09	188	2053307	20.00	ng/ul	0.00
78) Chrysene-d12	21.18	240	2205232	20.00	ng/ul	0.00
86) Perylene-d12	23.42	264	2628402	20.00	ng/ul	0.00

System Monitoring Compounds

3) 1,4-Dioxane-d8	3.24	96	46599	7.39	ng/uL	0.00
5) Phenol-d5	6.88	99	485418	19.04	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.05	67	330448	21.91	ng/ul	0.00
9) 2-Chlorophenol-d4	7.25	132	388212	20.82	ng/ul	0.00
13) 4-Methylphenol-d8	8.41	113	402015	19.22	ng/ul	0.00
19) Nitrobenzene-d5	8.85	128	226427	20.98	ng/ul	0.00
22) 2-Nitrophenol-d4	9.57	143	193616	22.86	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.10	165	433972	21.85	ng/ul	0.00
29) 4-Chloroaniline-d4	10.61	131	630061	25.19	ng/ul	0.00
44) Dimethylphthalate-d6	13.75	166	1417119	21.15	ng/ul	0.00
47) Acenaphthylene-d8	14.03	160	1711423	20.23	ng/ul	0.00
52) 4-Nitrophenol-d4	14.53	143	245340	19.73	ng/ul	0.00
58) Fluorene-d10	15.33	176	1215696	20.02	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.45	200	150161	18.15	ng/ul	0.00
71) Anthracene-d10	17.19	188	1919821	20.57	ng/ul	0.00
79) Pyrene-d10	19.43	212	2268451	21.69	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.28	264	2655861	21.35	ng/ul	0.00

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.28	88	50617	7.303	ng/uL 94
4) Benzaldehyde	6.85	77	362124	21.810	ng/ul 99
6) Phenol	6.91	94	511452	18.732	ng/ul 98
8) Bis(2-Chloroethyl)ether	7.14	93	420045	19.417	ng/ul 99
10) 2-Chlorophenol	7.28	128	403770	20.333	ng/ul 99
11) 2-Methylphenol	8.15	108	403276	19.116	ng/ul 97
12) 2,2'-oxybis(1-Chloropropan	8.25	45	540008	19.460	ng/ul 99
14) Acetophenone	8.52	105	702720	20.350	ng/ul 99
15) N-Nitroso-di-n-propylamine	8.51	70	356331	20.760	ng/ul 99
16) 4-Methylphenol	8.48	108	442677	19.030	ng/ul 99
17) Hexachloroethane	8.79	117	171372	19.818	ng/ul 99
20) Nitrobenzene	8.89	77	515793	19.655	ng/ul 99
21) Isophorone	9.42	82	998513	20.047	ng/ul 99
23) 2-Nitrophenol	9.60	139	223350	22.066	ng/ul 99
24) 2,4-Dimethylphenol	9.67	107	477919	20.354	ng/ul 99
25) Bis(2-Chloroethoxy)methane	9.90	93	613295	19.627	ng/ul 100
27) 2,4-Dichlorophenol	10.13	162	429693	21.028	ng/ul 99
28) Naphthalene	10.53	128	1481420	19.541	ng/ul 99
30) 4-Chloroaniline	10.63	127	656849	25.807	ng/ul 100
31) Hexachlorobutadiene	10.83	225	278911	19.652	ng/ul 99
32) Caprolactam	11.38	113	150157	19.988	ng/ul 95
33) 4-Chloro-3-methylphenol	11.77	107	442222	21.018	ng/ul 98
34) 2-Methylnaphthalene	12.15	142	1113375	20.440	ng/ul 98

Data Path : Z:\SVOASRV\HPCHEM1\BNA P\DATA\BP021320\
 Data File : BP001744.D
 Acq On : 13 Feb 2020 19:02
 Operator : CG/JU
 Sample : SSTDICV020
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :

Quant Time: Feb 14 12:53:32 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP021320MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Fri Feb 14 12:45:21 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 1-Methylnaphthalene	12.37	142	1036382	19.161	ng/ul	99
37) 1,2,4,5-Tetrachlorobenzene	12.52	216	588428	20.188	ng/ul	99
38) Hexachlorocyclopentadiene	12.52	237	339174	19.538	ng/ul	99
39) 2,4,6-Trichlorophenol	12.76	196	324208	21.867	ng/ul	99
40) 2,4,5-Trichlorophenol	12.83	196	367107	21.968	ng/ul	98
41) 1,1'-Biphenyl	13.17	154	1438898	20.153	ng/ul	100
42) 2-Chloronaphthalene	13.20	162	1124099	19.895	ng/ul	99
43) 2-Nitroaniline	13.40	65	304523	21.778	ng/ul	99
45) Dimethylphthalate	13.80	163	1414045	19.983	ng/ul	100
46) 2,6-Dinitrotoluene	13.90	165	307169	22.417	ng/ul	97
48) Acenaphthylene	14.06	152	1846493	20.225	ng/ul	100
49) 3-Nitroaniline	14.23	138	326460	24.573	ng/ul	99
50) Acenaphthene	14.40	153	1209272	19.744	ng/ul	99
51) 2,4-Dinitrophenol	14.44	184	81633	16.462	ng/ul	97
53) 4-Nitrophenol	14.55	109	190313	19.610	ng/ul	96
54) Dibenzofuran	14.74	168	1736451	20.117	ng/ul	99
55) 2,4-Dinitrotoluene	14.70	165	423658	21.001	ng/ul	99
56) 2,3,4,6-Tetrachlorophenol	14.97	232	341690	24.036	ng/ul	98
57) Diethylphthalate	15.18	149	1382105	20.053	ng/ul	100
59) Fluorene	15.39	166	1417988	19.766	ng/ul	99
60) 4-Chlorophenyl-phenylether	15.39	204	713276	19.658	ng/ul	99
61) 4-Nitroaniline	15.40	138	380442	23.604	ng/ul	99
64) 4,6-Dinitro-2-methylphenol	15.46	198	172616	18.476	ng/ul	95
65) N-Nitrosodiphenylamine	15.60	169	1233568	21.233	ng/ul	99
66) 4-Bromophenyl-phenylether	16.29	248	424612	19.431	ng/ul	97
67) Hexachlorobenzene	16.40	284	494444	19.143	ng/ul	100
68) Atrazine	16.56	200	383706	17.659	ng/ul	97
69) Pentachlorophenol	16.75	266	232288	18.543	ng/ul	98
70) Phenanthrene	17.13	178	2291739	19.819	ng/ul	99
72) Anthracene	17.22	178	2347382	20.212	ng/ul	100
73) 1,2,3,4-Tetrachlorobenzene	13.13	216	578657	18.735	ng/uL	99
74) Pentachlorobenzene	14.66	250	573106	18.238	ng/uL	98
75) Carbazole	17.49	167	2149864	21.905	ng/ul	100
76) Di-n-butylphthalate	18.07	149	2228759	21.553	ng/ul	100
77) Fluoranthene	19.10	202	2768445	21.726	ng/ul	100
80) Pvrene	19.46	202	2923557	20.581	ng/ul	99
81) Butylbenzylphthalate	20.35	149	942199	22.364	ng/ul	96
82) 3,3'-Dichlorobenzidine	21.10	252	940657	21.861	ng/ul#	97
83) Benzo(a)anthracene	21.16	228	2992598	20.721	ng/ul	97
84) Bis(2-ethylhexyl)phthalate	21.12	149	1574750	22.884	ng/ul	100
85) Chrysene	21.22	228	2838909	20.343	ng/ul	99
87) Di-n-octyl phthalate	22.00	149	2479821	21.124	ng/ul	100
88) Benzo(b)fluoranthene	22.75	252	3056712	20.091	ng/ul	100
89) Benzo(k)fluoranthene	22.79	252	3226506	20.992	ng/ul	98
91) Benzo(a)pyrene	23.32	252	3144059	21.780	ng/ul	100
92) Indeno(1,2,3-cd)pyrene	25.69	276	3868668	20.458	ng/ul	99
93) Dibenzo(a,h)anthracene	25.71	278	3262520	20.679	ng/ul	99
94) Benzo(a,h,i)perylene	26.38	276	3221888	20.014	ng/ul	98


Data Path : Z:\SVOASRV\HPCHEM1\BNA_P\DATA\BP021320\
Data File : BP001744.D
Acq On : 13 Feb 2020 19:02
Operator : CG/JU
Sample : SSTDICV020
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :

Quant Time: Feb 14 12:53:32 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_P\METHODS\SOM-EPA-BP021320MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Fri Feb 14 12:45:21 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)

(#)=qualifier out of range (m)=manual integration (+)=signals summed						



#2

1,4-Dioxane

Concen: 7.303 ng/uL

RT: 3.28 min Scan# 15

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampleId :





Tot Ion: 88 Resp: 50617

Ion	Ratio	Lower	Upper
88	100		
43	25.3	23.0	34.6
58	61.4	53.1	79.7

88 100

43 25.3 23.0 34.6

58 61.4 53.1 79.7



#4

Benzaldehyde

Concen: 21.810 ng/uL

RT: 6.85 min Scan# 623

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02




Tgt Ion: 77 Resp: 362124


Ion	Ratio	Lower	Upper
77	100		
105	98.2	79.3	118.9
106	97.5	77.0	115.4

77 100

105 98.2 79.3 118.9

106 97.5 77.0 115.4





#6

Phenol

Concen: 18.732 ng/ul

RT: 6.91 min Scan# 633

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampleId :





Tot Ion: 94 Resp: 511452

Ion Ratio Lower Upper

94 100

65 26.9 20.7 31.1

66 35.1 26.9 40.3



#8

Bis(2-Chloroethyl)ether

Concen: 19.417 ng/ul

RT: 7.14 min Scan# 672

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02




Tgt Ion: 93 Resp: 420045


Ion Ratio Lower Upper

93 100

63 72.0 58.1 87.1

95 32.6 26.0 39.0





#10

2-Chlorophenol

Concen: 20.333 ng/ul

RT: 7.28 min Scan# 695

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampleId :


Tot Ion:128 Resp: 403770

Ion Ratio Lower Upper

128 100

64 42.8 34.7 52.1




130 32.9 25.4 38.0



Abundance Ion 128.00 (127.70 to 128.70): E

Ion 64.00 (63.70 to 64.70): BP0

Ion 130.00 (129.70 to 130.70): E



#11

2-Methylphenol

Concen: 19.116 ng/ul

RT: 8.15 min Scan# 843

Delta R.T. -0.01 min

Lab File: BP001744.D


Acq: 13 Feb 2020 19:02

Tgt Ion:108 Resp: 403276

Ion Ratio Lower Upper



108 100


107 90.5 70.2 105.4



Abundance Ion 108.00 (107.70 to 108.70): E

Ion 107.00 (106.70 to 107.70): E









#12
2,2'-oxybis(1-Chloropropane)
Concen: 19.460 ng/ul
RT: 8.25 min Scan# 860
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampled :


Tot Ion	Ratio	Lower	Upper
45	100		
77	15.3	12.6	19.0
79	12.5	9.7	14.5



#14
Acetophenone
Concen: 20.350 ng/ul
RT: 8.52 min Scan# 907
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
105	100		
77	77.1	60.6	91.0
51	23.9	18.8	28.2









#15
N-Nitroso-di-n-propylamine
Concen: 20.760 ng/ul
RT: 8.51 min Scan# 905
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion:	70	Resp:	356331
Ion	Ratio	Lower	Upper
70	100		
42	55.1	45.0	67.4
101	9.5	7.9	11.9
130	22.4	18.2	27.2



#16
4-Methylphenol
Concen: 19.030 ng/ul
RT: 8.48 min Scan# 899
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:	108	Resp:	442677
Ion	Ratio	Lower	Upper
108	100		
107	107.8	87.3	130.9









#17
Hexachloroethane
Concen: 19.818 ng/ul
RT: 8.79 min Scan# 952
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampled :


Tot Ion:117 Resp: 171372
Ion Ratio Lower Upper
117 100
201 108.5 85.5 128.3
199 68.5 55.3 82.9



#20
Nitrobenzene
Concen: 19.655 ng/ul
RT: 8.89 min Scan# 970
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion: 77 Resp: 515793
Ion Ratio Lower Upper
77 100
123 46.4 36.6 54.8
65 12.9 9.9 14.9





#21

Isophorone

Concen: 20.047 ng/ul

RT: 9.42 min Scan# 1059

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampleId :





Tot Ion: 82 Resp: 998513

Ion Ratio Lower Upper

82 100

95 6.7 5.5 8.3

138 19.0 15.0 22.4



#23

2-Nitrophenol

Concen: 22.066 ng/ul

RT: 9.60 min Scan# 1090

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02




Tgt Ion: 139 Resp: 223350


Ion Ratio Lower Upper

139 100

65 49.9 39.8 59.6

109 28.3 22.2 33.2





#24

2,4-Dimethylphenol

Concen: 20.354 ng/ul

RT: 9.67 min Scan# 1102

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampleId :





Tot Ion: 107 Resp: 477919

Ion Ratio Lower Upper

107 100

121 50.7 40.7 61.1

122 85.7 69.6 104.4



#25

Bis(2-Chloroethoxy)methane

Concen: 19.627 ng/ul

RT: 9.90 min Scan# 1142

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02




Tgt Ion: 93 Resp: 613295


Ion Ratio Lower Upper

93 100

95 32.3 25.8 38.8

123 11.9 9.0 13.6





#27

2,4-Dichlorophenol

Concen: 21.028 ng/ul

RT: 10.13 min Scan# 1181





Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :
BNA_P
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
162	100		
164	65.8	52.2	78.2
98	34.1	26.6	40.0



#28

Naphthalene

Concen: 19.541 ng/ul




RT: 10.53 min Scan# 1249


Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.1	8.6	12.8
127	12.9	10.2	15.2









#30
4-Chloroaniline
Concen: 25.807 ng/ul
RT: 10.63 min Scan# 1266
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampled :


Tot Ion:127 Resp: 656849
Ion Ratio Lower Upper
127 100
129 31.6 25.3 37.9



#31
Hexachlorobutadiene
Concen: 19.652 ng/ul
RT: 10.83 min Scan# 1300
Delta R.T. -0.01 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:225 Resp: 278911
Ion Ratio Lower Upper
225 100
223 63.5 50.3 75.5
227 63.5 51.4 77.0









#32
 Caprolactam
 Concen: 19.988 ng/ul
 RT: 11.38 min Scan# 1392
 Delta R.T. -0.01 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02




Instrument :
 BNA_P
 ClientSampleId :


Tot Ion:113 Resp: 150157
 Ion Ratio Lower Upper
 113 100
 55 154.8 117.3 175.9
 56 117.7 90.7 136.1



#33
 4-Chloro-3-methylphenol
 Concen: 21.018 ng/ul
 RT: 11.77 min Scan# 1459
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion:107 Resp: 442222
 Ion Ratio Lower Upper
 107 100
 144 28.2 22.0 33.0
 142 84.7 66.3 99.5









#34
 2-Methylnaphthalene
 Concen: 20.440 ng/ul
 RT: 12.15 min Scan# 1523
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02




Instrument :
 BNA_P
 ClientSampleId :


Tot Ion:142 Resp: 1113375
 Ion Ratio Lower Upper
 142 100
 141 86.3 70.7 106.1



#35
 1-Methylnaphthalene
 Concen: 19.161 ng/ul
 RT: 12.37 min Scan# 1561
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion:142 Resp: 1036382
 Ion Ratio Lower Upper
 142 100
 141 91.7 74.4 111.6
 116 3.4 3.0 4.4









#37
 1,2,4,5-Tetrachlorobenzene
 Concen: 20.188 ng/ul
 RT: 12.52 min Scan# 1587
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02




Instrument :
 BNA_P
 ClientSampleId :


Tot Ion:	216	Resp:	588428
Ion	Ratio	Lower	Upper
216	100		
214	78.3	63.2	94.8
179	19.7	16.2	24.4
108	16.8	13.8	20.6



#38
 Hexachlorocyclopentadiene
 Concen: 19.538 ng/ul
 RT: 12.52 min Scan# 1586
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion:	237	Resp:	339174
Ion	Ratio	Lower	Upper
237	100		
235	62.5	50.3	75.5
272	13.4	10.5	15.7









#39
 2,4,6-Trichlorophenol
 Concen: 21.867 ng/ul
 RT: 12.76 min Scan# 1628
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02




Instrument :
 BNA_P
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
196	100		
198	97.4	76.6	114.8
200	31.1	24.6	37.0



#40
 2,4,5-Trichlorophenol
 Concen: 21.968 ng/ul
 RT: 12.83 min Scan# 1639
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
196	100		
198	94.7	77.3	115.9
200	30.6	24.4	36.6





#41

1,1'-Biphenyl

Concen: 20.153 ng/ul

RT: 13.17 min Scan# 1697

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02





Instrument :

BNA_P

ClientSampleId :

Tot Ion:154 Resp: 1438898

Ion	Ratio	Lower	Upper
154	100		
153	39.6	31.6	47.4
76	12.7	10.4	15.6



#42

2-Chloronaphthalene

Concen: 19.895 ng/ul

RT: 13.20 min Scan# 1703




Delta R.T. 0.00 min


Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Tgt Ion:162 Resp: 1124099

Ion	Ratio	Lower	Upper
162	100		
164	33.1	26.7	40.1
127	36.4	29.6	44.4









#43
2-Nitroaniline
Concen: 21.778 ng/ul
RT: 13.40 min Scan# 1737
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion: 65 Resp: 304523
Ion Ratio Lower Upper
65 100
92 72.4 57.7 86.5
138 117.3 92.7 139.1



#45
Dimethylphthalate
Concen: 19.983 ng/ul
RT: 13.80 min Scan# 1804
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion: 163 Resp: 1414045
Ion Ratio Lower Upper
163 100
194 4.2 3.5 5.3
164 10.0 8.0 12.0





#46

2,6-Dinitrotoluene

Concen: 22.417 ng/ul

RT: 13.90 min Scan# 1822





Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :
BNA_P
ClientSampled :

Tot Ion	Ratio	Lower	Upper
165	100		
89	56.1	46.6	70.0
121	20.4	17.4	26.0



#48

Acenaphthylene

Concen: 20.225 ng/ul




RT: 14.06 min Scan# 1848


Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
152	100		
151	19.9	15.9	23.9
153	13.0	10.6	16.0









#49
3-Nitroaniline
Concen: 24.573 ng/ul
RT: 14.23 min Scan# 1877
Delta R.T. -0.01 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion:138 Resp: 326460
Ion Ratio Lower Upper
138 100
108 9.6 7.8 11.6
92 111.1 87.6 131.4



#50
Acenaphthene
Concen: 19.744 ng/ul
RT: 14.40 min Scan# 1907
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:153 Resp: 1209272
Ion Ratio Lower Upper
153 100
152 48.0 39.4 59.2
154 87.7 70.8 106.2









#51
2,4-Dinitrophenol
Concen: 16.462 ng/ul
RT: 14.44 min Scan# 1913
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampled :


Tot Ion	Ratio	Lower	Upper
184	100		
63	59.1	51.1	76.7
154	58.0	46.6	70.0



#53
4-Nitrophenol
Concen: 19.610 ng/ul
RT: 14.55 min Scan# 1931
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
109	100		
139	145.9	112.7	169.1
65	126.2	104.1	156.1





#54

Dibenzofuran

Concen: 20.117 ng/ul

RT: 14.74 min Scan# 1964

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P





ClientSampleId :

Tot Ion:168 Resp: 1736451

Ion Ratio Lower Upper

168 100

139 35.8 29.0 43.4



#55

2,4-Dinitrotoluene

Concen: 21.001 ng/ul

RT: 14.70 min Scan# 1957

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02




Tgt Ion:165 Resp: 423658


Ion Ratio Lower Upper

165 100

63 38.5 31.4 47.2

182 2.8 2.4 3.6









#56
2,3,4,6-Tetrachlorophenol
Concen: 24.036 ng/ul
RT: 14.97 min Scan# 2003
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion:	232	Resp:	341690
Ion	Ratio	Lower	Upper
232	100		
131	41.0	34.3	51.5
130	2.3	1.8	2.8
166	29.8	24.1	36.1



#57
Diethylphthalate
Concen: 20.053 ng/ul
RT: 15.18 min Scan# 2039
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:	149	Resp:	1382105
Ion	Ratio	Lower	Upper
149	100		
177	23.0	18.6	28.0
150	12.2	9.8	14.6





#59

Fluorene

Concen: 19.766 ng/ul

RT: 15.39 min Scan# 2075

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02


Instrument :

BNA_P

ClientSampleId :

Tot Ion:166 Resp: 1417988

Ion	Ratio	Lower	Upper
166	100		
165	96.5	77.9	116.9
167	13.4	11.0	16.4






Abundance

Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E



#60

4-Chlorophenyl-phenylether

Concen: 19.658 ng/ul

RT: 15.39 min Scan# 2075


Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Tgt Ion:204 Resp: 713276

Ion	Ratio	Lower	Upper
204	100		
206	32.7	26.5	39.7
141	53.1	43.4	65.0






Abundance

Ion 204.00 (203.70 to 204.70): E

Ion 206.00 (205.70 to 206.70): E

Ion 141.00 (140.70 to 141.70): E









#61
4-Nitroaniline
Concen: 23.604 ng/ul
RT: 15.40 min Scan# 2076
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion:138 Resp: 380442
Ion Ratio Lower Upper
138 100
92 47.9 38.8 58.2
108 73.0 57.8 86.6



#64
4,6-Dinitro-2-methylphenol
Concen: 18.476 ng/ul
RT: 15.46 min Scan# 2087
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:198 Resp: 172616
Ion Ratio Lower Upper
198 100
182 3.9 3.2 4.8
77 24.9 17.7 26.5









#65
N-Nitrosodiphenylamine
Concen: 21.233 ng/ul
RT: 15.60 min Scan# 2111
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion:169 Resp: 1233568
Ion Ratio Lower Upper
169 100
168 66.6 54.1 81.1
167 36.2 28.9 43.3



#66
4-Bromophenyl-phenylether
Concen: 19.431 ng/ul
RT: 16.29 min Scan# 2227
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:248 Resp: 424612
Ion Ratio Lower Upper
248 100
250 99.6 77.4 116.0
141 67.9 56.1 84.1









#67
Hexachlorobenzene
Concen: 19.143 ng/ul
RT: 16.40 min Scan# 2246
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion	Ratio	Lower	Upper
284	100		
142	35.4	28.3	42.5
249	31.0	24.5	36.7



#68
Atrazine
Concen: 17.659 ng/ul
RT: 16.56 min Scan# 2274
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
200	100		
173	25.8	19.9	29.9
215	51.8	39.5	59.3









#69
Pentachlorophenol
Concen: 18.543 ng/ul
RT: 16.75 min Scan# 2305
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Instrument :
BNA_P
ClientSampleId :

Tot Ion	Ratio	Lower	Upper
266	100		
264	64.9	51.0	76.4
268	61.9	50.8	76.2




Abundance Ion 265.70 (265.40 to 266.40): E
Ion 264.00 (263.70 to 264.70): E
Ion 268.00 (267.70 to 268.70): E






#70
Phenanthrene
Concen: 19.819 ng/ul
RT: 17.13 min Scan# 2370
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
178	100		
179	15.6	12.3	18.5
176	19.2	15.1	22.7



Abundance Ion 178.00 (177.70 to 178.70): E
Ion 179.00 (178.70 to 179.70): E
Ion 176.00 (175.70 to 176.70): E





#72

Anthracene

Concen: 20.212 ng/uL

RT: 17.22 min Scan# 2386

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02


Instrument :

BNA_P

ClientSampled :

Tot Ion:178 Resp: 2347382

Ion	Ratio	Lower	Upper
178	100		
179	15.3	12.2	18.2
176	18.6	15.0	22.4






Abundance

Ion 178.00 (177.70 to 178.70): E

Ion 179.00 (178.70 to 179.70): E

Ion 176.00 (175.70 to 176.70): E



#73

1,2,3,4-Tetrachlorobenzene

Concen: 18.735 ng/uL

RT: 13.13 min Scan# 1690


Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Tgt Ion:216 Resp: 578657

Ion	Ratio	Lower	Upper
216	100		
214	78.5	62.2	93.2
218	48.6	37.8	56.6






Abundance

Ion 216.00 (215.70 to 216.70): E

Ion 214.00 (213.70 to 214.70): E

Ion 218.00 (217.70 to 218.70): E









#74
 Pentachlorobenzene
 Concen: 18.238 ng/uL
 RT: 14.66 min Scan# 1951
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02




Instrument :
 BNA_P
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
250	100		
248	62.6	51.1	76.7
252	62.7	51.4	77.2



#75
 Carbazole
 Concen: 21.905 ng/uL
 RT: 17.49 min Scan# 2431
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.1	16.8	25.2
139	12.5	10.2	15.2





#76

Di-n-butylphthalate

Concen: 21.553 ng/ul

RT: 18.07 min Scan# 2531

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampled :


Tot Ion:149 Resp: 2228759

Ion Ratio Lower Upper

149 100

150 9.1 7.4 11.2




104 5.0 4.2 6.2



Abundance Ion 149.00 (148.70 to 149.70): E

Ion 150.00 (149.70 to 150.70): E

Ion 104.00 (103.70 to 104.70): E



#77

Fluoranthene

Concen: 21.726 ng/ul

RT: 19.10 min Scan# 2706

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02


Tgt Ion:202 Resp: 2768445

Ion Ratio Lower Upper

202 100

101 10.9 8.9 13.3



100 8.4 6.9 10.3




Abundance Ion 202.00 (201.70 to 202.70): E

Ion 101.00 (100.70 to 101.70): E

Ion 100.00 (99.70 to 100.70): E









#80
Pvrene
Concen: 20.581 ng/ul
RT: 19.46 min Scan# 2766
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02




Instrument :
BNA_P
ClientSampleId :


Tot Ion:202 Resp: 2923557
Ion Ratio Lower Upper
202 100
101 13.0 10.2 15.2
100 10.4 8.4 12.6



#81
Butylbenzylphthalate
Concen: 22.364 ng/ul
RT: 20.35 min Scan# 2918
Delta R.T. 0.00 min
Lab File: BP001744.D
Acq: 13 Feb 2020 19:02

Tgt Ion:149 Resp: 942199
Ion Ratio Lower Upper
149 100
91 66.2 55.9 83.9
206 22.1 17.9 26.9





#82

3,3'-Dichlorobenzidine

Concen: 21.861 ng/ul

RT: 21.10 min Scan# 3045

Delta R.T. -0.01 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampled :


Tot Ion:252 Resp: 940657

Ion Ratio Lower Upper

252 100

254 65.9 51.4 77.2




126 11.7 7.5 11.3#



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 254.00 (253.70 to 254.70): E

Ion 126.00 (125.70 to 126.70): E



#83

Benzo(a)anthracene

Concen: 20.721 ng/ul

RT: 21.16 min Scan# 3056

Delta R.T. -0.01 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02


Tgt Ion:228 Resp: 2992598

Ion Ratio Lower Upper

228 100

229 19.9 14.9 22.3



226 27.7 20.8 31.2




Abundance Ion 228.00 (227.70 to 228.70): E

Ion 229.00 (228.70 to 229.70): E

Ion 226.00 (225.70 to 226.70): E





#84

Bis(2-ethylhexyl)phthalate

Concen: 22.884 ng/ul

RT: 21.12 min Scan# 3049

Delta R.T. 0.00 min





Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :
BNA_P
ClientSampleId :

Tot Ion:149 Resp: 1574750

Ion	Ratio	Lower	Upper
149	100		
167	27.8	22.1	33.1
279	4.3	3.3	4.9



#85

Chrysene

Concen: 20.343 ng/ul

RT: 21.22 min Scan# 3065




Delta R.T. 0.00 min


Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Tgt Ion:228 Resp: 2838909

Ion	Ratio	Lower	Upper
228	100		
226	30.6	24.3	36.5
229	20.0	16.2	24.4









#87
 Di-n-octyl phthalate
 Concen: 21.124 ng/ul
 RT: 22.00 min Scan# 3199
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion:149 Resp: 2479821




Instrument :
 BNA_P
 ClientSampleId :




#88
 Benzo(b)fluoranthene
 Concen: 20.091 ng/ul
 RT: 22.75 min Scan# 3326
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion:252 Resp: 3056712

Ion	Ratio	Lower	Upper
252	100		
253	22.1	17.8	26.8
125	9.8	8.0	12.0





#89

Benzo(k)fluoranthene

Concen: 20.992 ng/ul

RT: 22.79 min Scan# 3333

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampled :


Tot Ion:252 Resp: 3226506

Ion Ratio Lower Upper

252 100

253 21.9 18.4 27.6




125 10.2 8.3 12.5



Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E



#91

Benzo(a)pyrene

Concen: 21.780 ng/ul

RT: 23.32 min Scan# 3423

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02


Tgt Ion:252 Resp: 3144059

Ion Ratio Lower Upper

252 100

253 22.0 17.6 26.4



125 11.3 8.9 13.3




Abundance Ion 252.00 (251.70 to 252.70): E

Ion 253.00 (252.70 to 253.70): E

Ion 125.00 (124.70 to 125.70): E









#92
 Indeno(1,2,3-cd)pyrene
 Concen: 20.458 ng/ul
 RT: 25.69 min Scan# 3825
 Delta R.T. -0.01 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02




Instrument :
 BNA_P
 ClientSampleId :


Tot Ion	Ratio	Lower	Upper
276	100		
138	24.1	19.4	29.2
277	24.0	19.9	29.9



#93
 Dibenzo(a,h)anthracene
 Concen: 20.679 ng/ul
 RT: 25.71 min Scan# 3829
 Delta R.T. 0.00 min
 Lab File: BP001744.D
 Acq: 13 Feb 2020 19:02

Tgt Ion	Ratio	Lower	Upper
278	100		
139	16.3	12.9	19.3
279	23.0	18.7	28.1





#94

Benzo(a,h,i)perylene

Concen: 20.014 ng/ul

RT: 26.38 min Scan# 3943

Delta R.T. 0.00 min

Lab File: BP001744.D

Acq: 13 Feb 2020 19:02

Instrument :

BNA_P

ClientSampleId :


Tot Ion:276 Resp: 3221888

Ion Ratio Lower Upper

276 100

138 20.6 17.0 25.4

277 22.3 19.0 28.4



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

