

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022123\
 Data File : BP013864.D
 Acq On : 21 Feb 2023 13:23
 Operator : CG/JU
 Sample : PB150959BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK959

Quant Time: Feb 22 01:52:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.110	152	172386	20.000	ng/ul	0.00
20) Naphthalene-d8	10.940	136	737269	20.000	ng/ul	0.00
38) Acenaphthene-d10	14.751	164	531591	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.504	188	1245962	20.000	ng/ul	0.00
79) Chrysene-d12	21.580	240	1201693	20.000	ng/ul	0.00
88) Perylene-d12	24.139	264	1262148	20.000	ng/ul	0.01

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.458	96	17455	4.197	ng/uL	0.00
4) Pyridine-d5	3.887	84	235506	19.099	ng/ul	0.00
7) Phenol-d5	7.257	99	342144	22.276	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.428	67	203030	22.558	ng/ul	0.00
11) 2-Chlorophenol-d4	7.634	132	269876	23.869	ng/ul	0.00
15) 4-Methylphenol-d8	8.816	113	283577	22.588	ng/ul	0.00
21) Nitrobenzene-d5	9.287	128	136846	26.677	ng/ul	0.01
24) 2-Nitrophenol-d4	10.010	143	163707	27.640	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.551	165	288539	23.906	ng/ul	0.00
31) 4-Chloroaniline-d4	11.069	131	403074	23.560	ng/ul	0.00
46) Dimethylphthalate-d6	14.151	166	1078038	25.748	ng/ul	0.00
49) Acenaphthylene-d8	14.445	160	1104410	24.368	ng/ul	0.00
54) 4-Nitrophenol-d4	14.934	143	163781	21.957	ng/ul	0.00
60) Fluorene-d10	15.739	176	895259	25.103	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.857	200	181675	23.767	ng/ul	0.00
73) Anthracene-d10	17.604	188	1506075	26.435	ng/ul	0.00
81) Pyrene-d10	19.822	212	1839809	26.998	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.974	264	1701704	26.908	ng/ul	0.01

Target Compounds					Qvalue
2) 1,4-Dioxane	3.487	88	117	0.026 ng/ul#	11
5) Pyridine	3.911	79	173	0.014 ng/ul#	1
6) Benzaldehyde	7.252	77	1102	0.159 ng/ul#	4
8) Phenol	7.287	94	404	0.026 ng/ul#	1
10) Bis(2-Chloroethyl)ether	7.516	93	109	0.009 ng/ul#	69
12) 2-Chlorophenol	7.669	128	143	0.012 ng/ul#	1
13) 2-Methylphenol	8.416	108	22	0.002 ng/ul	96
14) 2,2'-oxybis(1-Chloropr...	8.640	45	22	0.002 ng/ul#	1
16) Acetophenone	8.934	105	137	0.007 ng/ul#	1
17) N-Nitroso-di-n-propyla...	8.934	70	62	0.006 ng/ul#	60
18) 4-Methylphenol	8.904	108	28	0.002 ng/ul	97
19) Hexachloroethane	9.204	117	72	0.015 ng/ul#	7
22) Nitrobenzene	9.334	77	261	0.019 ng/ul#	61
23) Isophorone	9.863	82	114	0.004 ng/ul#	84
25) 2-Nitrophenol	10.051	139	38	0.006 ng/ul#	1
26) 2,4-Dimethylphenol	10.081	107	60	0.004 ng/ul#	61
27) Bis(2-Chloroethoxy)met...	10.334	93	75	0.004 ng/ul#	34
29) 2,4-Dichlorophenol	10.581	162	12	0.001 ng/ul#	1
30) Naphthalene	10.981	128	125	0.003 ng/ul#	92
32) 4-Chloroaniline	11.093	127	77	0.004 ng/ul#	1
33) Hexachlorobutadiene	11.263	225	44	0.005 ng/ul#	53
34) Caprolactam	11.863	113	9	0.002 ng/ul#	35
35) 4-Chloro-3-methylphenol	12.204	107	9	0.001 ng/ul#	31
36) 2-Methylnaphthalene	12.569	142	49	0.002 ng/ul	91


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP022123\
 Data File : BP013864.D
 Acq On : 21 Feb 2023 13:23
 Operator : CG/JU
 Sample : PB150959BL
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK959

Quant Time: Feb 22 01:52:11 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP020723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Feb 22 01:41:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.810	142	198	0.007	ng/ul#	39
39) 1,2,4,5-Tetrachloroben...	12.951	216	64	0.004	ng/ul#	50
40) Hexachlorocyclopentadiene	12.928	237	11	0.001	ng/ul#	47
41) 2,4,6-Trichlorophenol	13.169	196	11	0.001	ng/ul#	40
42) 2,4,5-Trichlorophenol	13.251	196	43	0.004	ng/ul#	50
43) 1,1'-Biphenyl	13.575	154	28	0.001	ng/ul#	53
44) 2-Chloronaphthalene	13.634	162	17	0.001	ng/ul#	1
45) 2-Nitroaniline	13.828	65	14	0.002	ng/ul#	50
47) Dimethylphthalate	14.192	163	60	0.001	ng/ul#	57
48) 2,6-Dinitrotoluene	14.328	165	48	0.007	ng/ul#	58
50) Acenaphthylene	14.481	152	132	0.003	ng/ul#	76
51) 3-Nitroaniline	14.645	138	2	0.000	ng/ul#	1
52) Acenaphthene	14.816	153	437	0.013	ng/ul#	70
53) 2,4-Dinitrophenol	14.857	184	16	0.003	ng/ul#	1
55) 4-Nitrophenol	14.957	109	47	0.007	ng/ul#	1
56) Dibenzofuran	15.139	168	101	0.002	ng/ul#	87
57) 2,4-Dinitrotoluene	15.116	165	203	0.019	ng/ul#	40
58) 2,3,4,6-Tetrachlorophenol	15.369	232	50	0.004	ng/ul#	90
59) Diethylphthalate	15.551	149	459	0.011	ng/ul#	95
61) Fluorene	15.798	166	177	0.004	ng/ul#	80
62) 4-Chlorophenyl-phenyle...	15.792	204	31	0.001	ng/ul#	1
63) 4-Nitroaniline	15.810	138	20	0.003	ng/ul#	56
66) 4,6-Dinitro-2-methylph...	15.869	198	128	0.017	ng/ul#	1
67) N-Nitrosodiphenylamine	16.075	169	32	0.001	ng/ul	99
68) 4-Bromophenyl-phenylether	16.675	248	40	0.003	ng/ul#	18
69) Hexachlorobenzene	16.804	284	42	0.003	ng/ul#	76
70) Atrazine	16.939	200	6	0.000	ng/ul#	1
71) Pentachlorophenol	17.133	266	10	0.001	ng/ul#	1
72) Phenanthrene	17.545	178	403	0.006	ng/ul#	70
74) Anthracene	17.633	178	354	0.005	ng/ul#	38
75) 1,2,3,4-Tetrachloroben...	13.610	216	12	0.001	ng/uL	98
76) Pentachlorobenzene	15.210	250	17	0.001	ng/uL	89
77) Carbazole	17.904	167	175	0.003	ng/ul#	72
78) Di-n-butylphthalate	18.433	149	589	0.008	ng/ul#	80
80) Fluoranthene	19.492	202	1679	0.022	ng/ul#	95
82) Pyrene	19.851	202	4324	0.052	ng/ul#	96
83) Butylbenzylphthalate	20.704	149	324	0.011	ng/ul#	64
84) 3,3'-Dichlorobenzidine	21.492	252	361	0.013	ng/ul#	56
85) Benzo(a)anthracene	21.621	228	3530	0.043	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.463	149	5713	0.131	ng/ul	99
87) Chrysene	21.621	228	3491	0.045	ng/ul	95
89) Di-n-octyl phthalate	22.439	149	1251	0.014	ng/ul	100
90) Benzo(b)fluoranthene	23.351	252	2391	0.028	ng/ul#	14
91) Benzo(k)fluoranthene	23.404	252	1729	0.021	ng/ul#	26
93) Benzo(a)pyrene	24.027	252	668	0.009	ng/ul#	1
94) Indeno(1,2,3-cd)pyrene	26.833	276	216	0.003	ng/ul#	43
95) Dibenzo(a,h)anthracene	26.862	278	143	0.002	ng/ul#	1
96) Benzo(g,h,i)perylene	27.692	276	251	0.004	ng/ul	96





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
 1,4-Dioxane
 Concen: 0.026 ng/uL
 RT: 3.487 min Scan# 51
 Delta R.T. -0.000 min
 Lab File: BP013864.D
 Acq: 21 Feb 2023 13:23




Instrument :
 BNA_P
 ClientSampleId :
 SBLK959


Tgt Ion: 88 Resp: 117
 Ion Ratio Lower Upper
 88 100
 43 77.3 20.1 30.1#
 58 0.0 53.0 79.6#



#5
 Pyridine
 Concen: 0.014 ng/uL
 RT: 3.911 min Scan# 123
 Delta R.T. 0.006 min
 Lab File: BP013864.D
 Acq: 21 Feb 2023 13:23

Tgt Ion: 79 Resp: 173
 Ion Ratio Lower Upper
 79 100
 52 3756.4 55.1 82.7#
 51 303.2 25.7 38.5#









#6
Benzaldehyde
Concen: 0.159 ng/ul
RT: 7.252 min Scan# 691
Delta R.T. 0.018 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion: 77 Resp: 1102
Ion Ratio Lower Upper
77 100
105 2.7 79.5 119.3#
106 3.0 76.5 114.7#



#8
Phenol
Concen: 0.026 ng/ul
RT: 7.287 min Scan# 697
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion: 94 Resp: 404
Ion Ratio Lower Upper
94 100
65 52.8 22.2 33.4#
66 146.1 29.6 44.4#









#10
 Bis(2-Chloroethyl)ether
 Concen: 0.009 ng/ul
 RT: 7.516 min Scan# 710
 Delta R.T. -0.000 min
 Lab File: BP013864.D
 Acq: 21 Feb 2023 13:23




Instrument :
 BNA_P
 ClientSampleId :
 SBLK959


Tgt Ion: 93 Resp: 109
 Ion Ratio Lower Upper
 93 100
 63 46.7 59.4 89.0#
 95 48.3 25.8 38.8#



#12
 2-Chlorophenol
 Concen: 0.012 ng/ul
 RT: 7.669 min Scan# 762
 Delta R.T. 0.006 min
 Lab File: BP013864.D
 Acq: 21 Feb 2023 13:23

Tgt Ion:128 Resp: 143
 Ion Ratio Lower Upper
 128 100
 64 164.2 37.3 55.9#
 130 34.7 26.3 39.5





#13

2-Methylphenol

Concen: 0.002 ng/ul

RT: 8.416 min Scan# 81

Delta R.T. -0.130 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :





SBLK959

Tgt Ion:108 Resp: 22

Ion Ratio Lower Upper

108 100

107 85.7 71.9 107.9



#14

2,2'-oxybis(1-Chloropropane)

Concen: 0.002 ng/ul

RT: 8.640 min Scan# 927

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion: 45 Resp: 22


Ion Ratio Lower Upper

45 100

77 202.5 13.4 20.2#

79 113.6 11.0 16.4#









#16
Acetophenone
Concen: 0.007 ng/ul
RT: 8.934 min Scan# 911
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:105 Resp: 137
Ion Ratio Lower Upper
105 100
77 180.0 61.0 91.4#
51 118.6 19.6 29.4#



#17
N-Nitroso-di-n-propylamine
Concen: 0.006 ng/ul
RT: 8.934 min Scan# 977
Delta R.T. 0.018 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion: 70 Resp: 62
Ion Ratio Lower Upper
70 100
42 86.3 39.8 59.8#
101 8.4 7.4 11.2
130 7.4 14.6 22.0#





#18

4-Methylphenol

Concen: 0.002 ng/ul

RT: 8.904 min Scan# 91

Delta R.T. 0.029 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :





SBLK959

Tgt Ion:108 Resp: 28

Ion Ratio Lower Upper

108 100

107 116.3 90.9 136.3



#19

Hexachloroethane

Concen: 0.015 ng/ul

RT: 9.204 min Scan# 1023

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:117 Resp: 72


Ion Ratio Lower Upper

117 100

201 0.0 86.6 129.8#

199 6.7 55.2 82.8#









#22
Nitrobenzene
Concen: 0.019 ng/ul
RT: 9.334 min Scan# 1045
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion: 77 Resp: 261
Ion Ratio Lower Upper
77 100
123 11.6 35.1 52.7#
65 14.6 10.6 16.0



#23
Isophorone
Concen: 0.004 ng/ul
RT: 9.863 min Scan# 1135
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion: 82 Resp: 114
Ion Ratio Lower Upper
82 100
95 9.8 6.0 9.0#
138 8.6 13.7 20.5#









#25
2-Nitrophenol
Concen: 0.006 ng/ul
RT: 10.051 min Scan# 1167
Delta R.T. 0.018 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:139 Resp: 38
Ion Ratio Lower Upper
139 100
65 413.9 39.9 59.9#
109 191.7 27.4 41.0#



#26
2,4-Dimethylphenol
Concen: 0.004 ng/ul
RT: 10.081 min Scan# 1172
Delta R.T. -0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:107 Resp: 60
Ion Ratio Lower Upper
107 100
121 37.2 37.8 56.6#
122 32.6 63.7 95.5#









#27
Bis(2-Chloroethoxy)methane
Concen: 0.004 ng/ul
RT: 10.334 min Scan# 11
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion: 93 Resp: 75
Ion Ratio Lower Upper
93 100
95 78.5 25.0 37.6#
123 12.7 8.2 12.4#



#29
2,4-Dichlorophenol
Concen: 0.001 ng/ul
RT: 10.581 min Scan# 1257
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:162 Resp: 12
Ion Ratio Lower Upper
162 100
164 7384.2 51.4 77.2#
98 657.9 27.5 41.3#









#30
Naphthalene
Concen: 0.003 ng/ul
RT: 10.981 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:128 Resp: 125
Ion Ratio Lower Upper
128 100
129 11.2 8.6 12.8
127 18.9 10.6 15.8#



#32
4-Chloroaniline
Concen: 0.004 ng/ul
RT: 11.093 min Scan# 1344
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:127 Resp: 77
Ion Ratio Lower Upper
127 100
129 142.4 26.2 39.4#









#33
Hexachlorobutadiene
Concen: 0.005 ng/ul
RT: 11.263 min Scan# 11
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:225 Resp: 44
Ion Ratio Lower Upper
225 100
223 84.6 50.6 76.0#
227 9.2 49.0 73.4#



#34
Caprolactam
Concen: 0.002 ng/ul
RT: 11.863 min Scan# 1475
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:113 Resp: 9
Ion Ratio Lower Upper
113 100
55 233.3 120.9 181.3#
56 190.5 93.8 140.8#









#35
4-Chloro-3-methylphenol
Concen: 0.001 ng/ul
RT: 12.204 min Scan# 1533
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:107 Resp: 9
Ion Ratio Lower Upper
107 100
144 28.6 20.0 30.0
142 0.0 62.1 93.1#



#36
2-Methylnaphthalene
Concen: 0.002 ng/ul
RT: 12.569 min Scan# 1595
Delta R.T. -0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:142 Resp: 49
Ion Ratio Lower Upper
142 100
141 100.0 72.8 109.2





#37

1-Methylnaphthalene

Concen: 0.007 ng/ul

RT: 12.810 min Scan# 1

Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:142 Resp: 198

Ion Ratio Lower Upper

142 100

141 34.1 74.2 111.2#

116 25.2 3.1 4.7#



#39

1,2,4,5-Tetrachlorobenzene

Concen: 0.004 ng/ul

RT: 12.951 min Scan# 1660

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Tgt Ion:216 Resp: 64




Ion Ratio Lower Upper


216 100

214 103.9 62.6 93.8#

179 76.5 15.4 23.2#

108 41.2 12.7 19.1#





#40

Hexachlorocyclopentadiene

Concen: 0.001 ng/ul

RT: 12.928 min Scan# 1

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:237 Resp: 11

Ion Ratio Lower Upper

237 100

235 53.8 49.8 74.8

272 115.4 10.1 15.1#



#41

2,4,6-Trichlorophenol

Concen: 0.001 ng/ul

RT: 13.169 min Scan# 1697

Delta R.T. -0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:196 Resp: 11


Ion Ratio Lower Upper

196 100

198 142.1 75.1 112.7#

200 78.9 24.3 36.5#









#42
2,4,5-Trichlorophenol
Concen: 0.004 ng/ul
RT: 13.251 min Scan# 1711
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:196 Resp: 43
Ion Ratio Lower Upper
196 100
198 43.5 77.0 115.4#
200 10.9 25.8 38.6#



#43
1,1'-Biphenyl
Concen: 0.001 ng/ul
RT: 13.575 min Scan# 1766
Delta R.T. -0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:154 Resp: 28
Ion Ratio Lower Upper
154 100
153 41.7 31.9 47.9
76 86.1 9.9 14.9#





#44

2-Chloronaphthalene

Concen: 0.001 ng/ul

RT: 13.634 min Scan# 11

Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:162 Resp: 17

Ion Ratio Lower Upper

162 100

164 227.6 26.1 39.1#

127 179.3 31.1 46.7#



#45

2-Nitroaniline

Concen: 0.002 ng/ul

RT: 13.828 min Scan# 1809

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion: 65 Resp: 14


Ion Ratio Lower Upper

65 100

92 44.4 58.6 88.0#

138 38.9 81.2 121.8#









#47
Dimethylphthalate
Concen: 0.001 ng/ul
RT: 14.192 min Scan# 1871
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:163 Resp: 60
Ion Ratio Lower Upper
163 100
194 40.4 3.4 5.2#
164 14.9 7.8 11.8#



#48
2,6-Dinitrotoluene
Concen: 0.007 ng/ul
RT: 14.328 min Scan# 1894
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:165 Resp: 48
Ion Ratio Lower Upper
165 100
89 16.9 47.5 71.3#
121 21.1 16.6 24.8









#50
Acenaphthylene
Concen: 0.003 ng/ul
RT: 14.481 min Scan# 1918
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:152 Resp: 132
Ion Ratio Lower Upper
152 100
151 7.3 15.8 23.6#
153 21.1 10.4 15.6#



#51
3-Nitroaniline
Concen: 0.000 ng/ul
RT: 14.645 min Scan# 1948
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:138 Resp: 2
Ion Ratio Lower Upper
138 100
108 475.0 9.4 14.2#
92 250.0 97.4 146.0#









#52
Acenaphthene
Concen: 0.013 ng/ul
RT: 14.816 min Scan# 1976
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:153 Resp: 437
Ion Ratio Lower Upper
153 100
152 43.8 37.4 56.0
154 47.1 70.1 105.1#



#53
2,4-Dinitrophenol
Concen: 0.003 ng/ul
RT: 14.857 min Scan# 1984
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:184 Resp: 16
Ion Ratio Lower Upper
184 100
63 300.0 49.5 74.3#
154 0.0 54.3 81.5#









#55
4-Nitrophenol
Concen: 0.007 ng/ul
RT: 14.957 min Scan# 2001
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:109 Resp: 47
Ion Ratio Lower Upper
109 100
139 74.8 96.0 144.0#
65 445.0 95.8 143.6#



#56
Dibenzofuran
Concen: 0.002 ng/ul
RT: 15.139 min Scan# 2032
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:168 Resp: 101
Ion Ratio Lower Upper
168 100
139 29.3 29.6 44.4#





#57

2,4-Dinitrotoluene

Concen: 0.019 ng/ul

RT: 15.116 min Scan# 2025

Delta R.T. 0.018 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:165 Resp: 203

Ion Ratio Lower Upper

165 100

63 0.0 32.0 48.0#

182 3.8 2.4 3.6#



#58

2,3,4,6-Tetrachlorophenol

Concen: 0.004 ng/ul

RT: 15.369 min Scan# 2071

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Tgt Ion:232 Resp: 50




Ion Ratio Lower Upper


232 100

131 37.0 33.0 49.6

130 0.0 1.5 2.3#

166 20.4 22.6 33.8#









#59
Diethylphthalate
Concen: 0.011 ng/ul
RT: 15.551 min Scan# 2102
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:	149	Resp:	459
Ion Ratio	Lower	Upper	
149	100		
177	23.8	17.8	26.8
150	15.9	10.1	15.1#



#61
Fluorene
Concen: 0.004 ng/ul
RT: 15.798 min Scan# 2144
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:	166	Resp:	177
Ion Ratio	Lower	Upper	
166	100		
165	77.5	79.1	118.7#
167	9.4	10.8	16.2#









#62
4-Chlorophenyl-phenylether
Concen: 0.001 ng/ul
RT: 15.792 min Scan# 2143
Delta R.T. 0.018 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:204 Resp: 31
Ion Ratio Lower Upper
204 100
206 214.8 26.7 40.1#
141 48.1 44.6 66.8



#63
4-Nitroaniline
Concen: 0.003 ng/ul
RT: 15.810 min Scan# 2146
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:138 Resp: 20
Ion Ratio Lower Upper
138 100
92 129.4 43.8 65.8#
108 94.1 69.8 104.6









#66
4,6-Dinitro-2-methylphenol
Concen: 0.017 ng/ul
RT: 15.869 min Scan# 2155
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:198 Resp: 128
Ion Ratio Lower Upper
198 100
182 477.9 4.1 6.1#
77 983.4 20.3 30.5#



#67
N-Nitrosodiphenylamine
Concen: 0.001 ng/ul
RT: 16.075 min Scan# 2191
Delta R.T. 0.082 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:169 Resp: 32
Ion Ratio Lower Upper
169 100
168 68.6 53.5 80.3
167 37.1 29.8 44.6









#68
4-Bromophenyl-phenylether
Concen: 0.003 ng/ul
RT: 16.675 min Scan# 2115
Delta R.T. -0.000 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:248 Resp: 40
Ion Ratio Lower Upper
248 100
250 227.0 76.0 114.0#
141 62.2 56.0 84.0



#69
Hexachlorobenzene
Concen: 0.003 ng/ul
RT: 16.804 min Scan# 2315
Delta R.T. 0.006 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:284 Resp: 42
Ion Ratio Lower Upper
284 100
142 19.7 25.0 37.4#
249 16.4 24.9 37.3#





#70

Atrazine

Concen: 0.000 ng/ul

RT: 16.939 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:200 Resp: 6

Ion Ratio Lower Upper

200 100

173 261.5 20.4 30.6#

215 46.2 39.3 58.9



#71

Pentachlorophenol

Concen: 0.001 ng/ul

RT: 17.133 min Scan# 2371

Delta R.T. -0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:266 Resp: 10


Ion Ratio Lower Upper

266 100

264 93.8 50.5 75.7#

268 431.3 51.0 76.6#





#72

Phenanthrene

Concen: 0.006 ng/ul

RT: 17.545 min Scan# 2440

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:178 Resp: 403

Ion Ratio Lower Upper

178 100

179 41.3 12.2 18.2#

176 16.8 15.2 22.8



#74

Anthracene

Concen: 0.005 ng/ul

RT: 17.633 min Scan# 2456

Delta R.T. -0.000 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:178 Resp: 354


Ion Ratio Lower Upper

178 100

179 34.9 12.2 18.4#

176 51.8 15.0 22.4#





#75

1,2,3,4-Tetrachlorobenzene

Concen: 0.001 ng/uL

RT: 13.610 min Scan# 11

Delta R.T. 0.065 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:216 Resp: 12

Ion Ratio Lower Upper

216 100

214 76.9 62.6 94.0

218 46.2 39.0 58.4



#76

Pentachlorobenzene

Concen: 0.001 ng/uL

RT: 15.210 min Scan# 2044

Delta R.T. 0.147 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:250 Resp: 17


Ion Ratio Lower Upper

250 100

248 51.9 51.1 76.7

252 59.3 51.0 76.6









#77
 Carbazole
 Concen: 0.003 ng/ul
 RT: 17.904 min Scan# 2502
 Delta R.T. 0.006 min
 Lab File: BP013864.D
 Acq: 21 Feb 2023 13:23




Instrument :
 BNA_P
 ClientSampleId :
 SBLK959


Tgt Ion:167 Resp: 175
 Ion Ratio Lower Upper
 167 100
 166 24.0 17.3 25.9
 139 39.0 10.4 15.6#



#78
 Di-n-butylphthalate
 Concen: 0.008 ng/ul
 RT: 18.433 min Scan# 2592
 Delta R.T. 0.006 min
 Lab File: BP013864.D
 Acq: 21 Feb 2023 13:23

Tgt Ion:149 Resp: 589
 Ion Ratio Lower Upper
 149 100
 150 8.3 7.1 10.7
 104 22.2 4.1 6.1#





#80

Fluoranthene

Concen: 0.022 ng/ul

RT: 19.492 min Scan# 21

Delta R.T. -0.000 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:202 Resp: 1679

Ion Ratio Lower Upper

202 100

101 8.3 8.6 12.8#

100 9.0 6.3 9.5



#82

Pyrene

Concen: 0.052 ng/ul

RT: 19.851 min Scan# 2833

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:202 Resp: 4324


Ion Ratio Lower Upper

202 100

101 11.9 10.3 15.5

100 12.7 8.5 12.7#





#83

Butylbenzylphthalate

Concen: 0.011 ng/ul

RT: 20.704 min Scan# 2978

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:149 Resp: 324

Ion Ratio Lower Upper

149 100

91 53.1 57.7 86.5#

206 59.2 16.7 25.1#



#84

3,3'-Dichlorobenzidine

Concen: 0.013 ng/ul

RT: 21.492 min Scan# 3112

Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:252 Resp: 361


Ion Ratio Lower Upper

252 100

254 43.7 51.8 77.6#

126 68.2 7.9 11.9#





#85

Benzo(a)anthracene

Concen: 0.043 ng/ul

RT: 21.621 min Scan# 3123

Delta R.T. 0.065 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:228 Resp: 3530

Ion Ratio Lower Upper

228 100

229 20.5 15.4 23.2

226 27.0 21.8 32.6



#86

Bis(2-ethylhexyl)phthalate

Concen: 0.131 ng/ul

RT: 21.463 min Scan# 3107

Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:149 Resp: 5713


Ion Ratio Lower Upper

149 100

167 27.3 21.5 32.3

279 3.9 3.2 4.8





#87

Chrysene

Concen: 0.045 ng/ul

RT: 21.621 min Scan# 3133

Delta R.T. 0.006 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:228 Resp: 3491

Ion Ratio Lower Upper

228 100

226 27.0 25.0 37.6

229 20.5 16.4 24.6



#89

Di-n-octyl phthalate

Concen: 0.014 ng/ul




RT: 22.439 min Scan# 3273


Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Tgt Ion:149 Resp: 1251









#90
Benzo(b)fluoranthene
Concen: 0.028 ng/ul
RT: 23.351 min Scan# 3426
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23




Instrument :
BNA_P
ClientSampleId :
SBLK959


Tgt Ion:252 Resp: 2391
Ion Ratio Lower Upper
252 100
253 59.4 17.5 26.3#
125 47.8 7.8 11.6#



#91
Benzo(k)fluoranthene
Concen: 0.021 ng/ul
RT: 23.404 min Scan# 3437
Delta R.T. 0.012 min
Lab File: BP013864.D
Acq: 21 Feb 2023 13:23

Tgt Ion:252 Resp: 1729
Ion Ratio Lower Upper
252 100
253 50.4 17.1 25.7#
125 46.8 7.2 10.8#





#93

Benzo(a)pyrene

Concen: 0.009 ng/ul

RT: 24.027 min Scan# 3

Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:252 Resp: 668

Ion Ratio Lower Upper

252 100

253 82.4 17.4 26.0#

125 82.8 8.2 12.2#



#94

Indeno(1,2,3-cd)pyrene

Concen: 0.003 ng/ul

RT: 26.833 min Scan# 4020

Delta R.T. 0.012 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23




Tgt Ion:276 Resp: 216


Ion Ratio Lower Upper

276 100

138 53.7 18.4 27.6#

277 50.2 19.7 29.5#





#95

Dibenzo(a,h)anthracene

Concen: 0.002 ng/ul

RT: 26.862 min Scan# 4021

Delta R.T. 0.023 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Instrument :

BNA_P

ClientSampleId :

SBLK959





Tgt Ion:278 Resp: 143

Ion Ratio Lower Upper

278 100

139 103.3 12.0 18.0#

279 114.9 18.9 28.3#



#96

Benzo(g,h,i)perylene

Concen: 0.004 ng/ul

RT: 27.692 min Scan# 4166

Delta R.T. 0.035 min

Lab File: BP013864.D

Acq: 21 Feb 2023 13:23

Tgt Ion:276 Resp: 251

Ion Ratio Lower Upper

276 100

138 22.4 15.9 23.9

277 22.6 19.0 28.6

