

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030822\
 Data File : BP009297.D
 Acq On : 08 Mar 2022 10:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 09 02:59:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 01:55:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	125	0.00
2	1,4-Dioxane	40.000	34.751	13.1	121	0.00
3	Pyridine	40.000	36.295	9.3	122	0.00
4	n-Nitrosodimethylamine	40.000	38.590	3.5	127	0.00
5 S	2-Fluorophenol	80.000	72.733	9.1	125	0.00
6	Aniline	40.000	36.398	9.0	124	0.00
7 S	Phenol-d6	80.000	72.821	9.0	125	0.00
8	2-Chlorophenol	40.000	37.260	6.9	126	0.00
9	Benzaldehyde	40.000	36.334	9.2	122	0.00
10 C	Phenol	40.000	36.217	9.5	124	0.00
11	bis(2-Chloroethyl)ether	40.000	36.521	8.7	123	0.00
12	1,3-Dichlorobenzene	40.000	35.748	10.6	125	0.00
13 C	1,4-Dichlorobenzene	40.000	35.887	10.3	125	0.00
14	1,2-Dichlorobenzene	40.000	35.843	10.4	124	0.00
15	Benzyl Alcohol	40.000	37.331	6.7	126	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	34.035	14.9	118	0.00
17	2-Methylphenol	40.000	36.826	7.9	125	0.00
18	Hexachloroethane	40.000	36.702	8.2	127	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	37.515	6.2	126	0.00
20	3+4-Methylphenols	40.000	36.964	7.6	125	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	124	0.00
22	Acetophenone	40.000	37.391	6.5	125	0.00
23 S	Nitrobenzene-d5	80.000	78.701	1.6	127	0.00
24	Nitrobenzene	40.000	38.377	4.1	126	0.00
25	Isophorone	40.000	38.377	4.1	126	0.00
26 C	2-Nitrophenol	40.000	44.969	-12.4	141	0.00
27	2,4-Dimethylphenol	40.000	36.969	7.6	123	0.00
28	bis(2-Chloroethoxy)methane	40.000	37.596	6.0	125	0.00
29 C	2,4-Dichlorophenol	40.000	38.849	2.9	126	0.00
30	1,2,4-Trichlorobenzene	40.000	37.495	6.3	128	0.00
31	Naphthalene	40.000	36.698	8.3	124	0.00
32	Benzoic acid	40.000	48.802	-22.0	155	0.02
33	4-Chloroaniline	40.000	37.616	6.0	125	0.00
34 C	Hexachlorobutadiene	40.000	38.487	3.8	131	0.00
35	Caprolactam	40.000	41.070	-2.7	129	0.00
36 C	4-Chloro-3-methylphenol	40.000	37.895	5.3	125	0.00
37	2-Methylnaphthalene	40.000	36.934	7.7	124	0.00
38	1-Methylnaphthalene	40.000	36.824	7.9	125	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	125	0.00
40	1,2,4,5-Tetrachlorobenzene	40.000	38.631	3.4	130	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.364	1.6	131	0.00
42 S	2,4,6-Tribromophenol	80.000	87.771	-9.7	142	0.00
43 C	2,4,6-Trichlorophenol	40.000	40.534	-1.3	131	0.00
44	2,4,5-Trichlorophenol	40.000	39.575	1.1	128	0.00
45 S	2-Fluorobiphenyl	80.000	74.855	6.4	127	0.00
46	1,1'-Biphenyl	40.000	37.007	7.5	125	0.00
47	2-Chloronaphthalene	40.000	37.422	6.4	126	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030822\
 Data File : BP009297.D
 Acq On : 08 Mar 2022 10:45
 Operator : CG/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Mar 09 02:59:36 2022
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Mar 05 01:55:59 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	43.599	-9.0	135	0.00
49	Acenaphthylene	40.000	38.033	4.9	126	0.00
50	Dimethylphthalate	40.000	37.833	5.4	127	0.00
51	2,6-Dinitrotoluene	40.000	41.355	-3.4	130	0.00
52 C	Acenaphthene	40.000	37.799	5.5	126	0.00
53	3-Nitroaniline	40.000	40.801	-2.0	129	0.00
54 P	2,4-Dinitrophenol	40.000	48.845	-22.1	150	0.00
55	Dibenzofuran	40.000	36.747	8.1	125	0.00
56 P	4-Nitrophenol	40.000	40.873	-2.2	127	0.00
57	2,4-Dinitrotoluene	40.000	43.237	-8.1	132	0.00
58	Fluorene	40.000	37.100	7.2	127	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	40.158	-0.4	132	0.00
60	Diethylphthalate	40.000	38.611	3.5	128	0.00
61	4-Chlorophenyl-phenylether	40.000	37.793	5.5	129	0.00
62	4-Nitroaniline	40.000	42.090	-5.2	131	0.00
63	Azobenzene	40.000	37.891	5.3	125	0.00
64 I	Phenanthrene-d10	20.000	20.000	0.0	126	0.00
65	4,6-Dinitro-2-methylphenol	40.000	47.473	-18.7	145	0.00
66 c	n-Nitrosodiphenylamine	40.000	37.021	7.4	124	0.00
67	4-Bromophenyl-phenylether	40.000	40.569	-1.4	136	0.00
68	Hexachlorobenzene	40.000	39.120	2.2	132	0.00
69	Atrazine	40.000	40.351	-0.9	130	0.00
70 C	Pentachlorophenol	40.000	42.128	-5.3	133	0.00
71	Phenanthrene	40.000	36.550	8.6	124	0.00
72	Anthracene	40.000	37.343	6.6	125	0.00
73	Carbazole	40.000	37.019	7.5	122	0.00
74	Di-n-butylphthalate	40.000	40.464	-1.2	129	0.00
75 C	Fluoranthene	40.000	37.078	7.3	122	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	125	0.00
77	Benzydine	40.000	43.426	-8.6	128	0.00
78	Pyrene	40.000	36.692	8.3	122	0.00
79 S	Terphenyl-d14	80.000	74.874	6.4	126	0.00
80	Butylbenzylphthalate	40.000	42.595	-6.5	132	0.00
81	Benzo(a)anthracene	40.000	36.677	8.3	123	0.00
82	3,3'-Dichlorobenzidine	40.000	40.334	-0.8	127	0.00
83	Chrysene	40.000	36.702	8.2	124	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.014	-2.5	129	0.00
85 c	Di-n-octyl phthalate	40.000	43.591	-9.0	134	0.00
86 I	Perylene-d12	20.000	20.000	0.0	126	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	39.156	2.1	131	0.00
88	Benzo(b)fluoranthene	40.000	37.722	5.7	124	0.00
89	Benzo(k)fluoranthene	40.000	37.320	6.7	127	0.00
90 C	Benzo(a)pyrene	40.000	38.159	4.6	126	0.00
91	Dibenzo(a,h)anthracene	40.000	39.289	1.8	130	0.00
92	Benzo(g,h,i)perylene	40.000	39.323	1.7	131	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030822\
Data File : BP009297.D
Acq On : 08 Mar 2022 10:45
Operator : CG/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Mar 09 02:59:36 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 01:55:59 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0