

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP014009.D
 Acq On : 09 Mar 2023 14:22
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771

Quant Time: Mar 10 00:18:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	129788	20.000	ng/u1	0.00
20) Naphthalene-d8	10.905	136	540004	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.716	164	377026	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.475	188	893211	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	982081	20.000	ng/u1	0.00
88) Perylene-d12	24.092	264	1044907	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.434	96	28892	8.411	ng/uL	0.00
4) Pyridine-d5	3.870	84	197478	20.763	ng/u1	0.00
7) Phenol-d5	7.228	99	248193	21.670	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.399	67	146987	22.005	ng/u1	0.00
11) 2-Chlorophenol-d4	7.605	132	188840	21.641	ng/u1	0.00
15) 4-Methylphenol-d8	8.787	113	205467	21.962	ng/u1	0.00
21) Nitrobenzene-d5	9.252	128	91840	21.315	ng/u1	0.00
24) 2-Nitrophenol-d4	9.975	143	108580	21.052	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	201985	20.813	ng/u1	0.00
31) 4-Chloroaniline-d4	11.040	131	262273	20.422	ng/u1	0.00
46) Dimethylphthalate-d6	14.122	166	640187	20.008	ng/u1	0.00
49) Acenaphthylene-d8	14.410	160	707981	20.887	ng/u1	0.00
54) 4-Nitrophenol-d4	14.910	143	111230	19.726	ng/u1	0.00
60) Fluorene-d10	15.710	176	543434	20.040	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.828	200	129749	19.219	ng/u1	0.00
73) Anthracene-d10	17.569	188	894668	20.633	ng/u1	0.00
81) Pyrene-d10	19.792	212	1138216	21.014	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.922	264	1159283	20.857	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.470	88	29749	8.022	ng/uL	91
5) Pyridine	3.887	79	202081	20.951	ng/u1	96
6) Benzaldehyde	7.211	77	148523	26.062	ng/u1	99
8) Phenol	7.258	94	255451	21.627	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.493	93	198662	21.954	ng/u1	99
12) 2-Chlorophenol	7.640	128	190931	21.717	ng/u1	98
13) 2-Methylphenol	8.522	108	193607	22.044	ng/u1	97
14) 2,2'-oxybis(1-Chloropr...	8.611	45	223198	22.840	ng/u1	98
16) Acetophenone	8.911	105	334815	22.153	ng/u1	99
17) N-Nitroso-di-n-propyla...	8.893	70	173731	22.469	ng/u1	97
18) 4-Methylphenol	8.852	108	215799	22.338	ng/u1	99
19) Hexachloroethane	9.169	117	79840	20.734	ng/u1	94
22) Nitrobenzene	9.293	77	248481	21.559	ng/u1	98
23) Isophorone	9.822	82	466671	21.025	ng/u1	98
25) 2-Nitrophenol	10.011	139	112622	21.187	ng/u1	99
26) 2,4-Dimethylphenol	10.063	107	256901	22.080	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.305	93	272932	21.259	ng/u1	100
29) 2,4-Dichlorophenol	10.546	162	199871	20.879	ng/u1	98
30) Naphthalene	10.952	128	612749	20.664	ng/u1	99
32) 4-Chloroaniline	11.063	127	264641	20.478	ng/u1	98
33) Hexachlorobutadiene	11.234	225	152627	20.036	ng/u1	99
34) Caprolactam	11.846	113	63532	20.439	ng/u1	97
35) 4-Chloro-3-methylphenol	12.175	107	234039	21.142	ng/u1	99
36) 2-Methylnaphthalene	12.552	142	423048	20.340	ng/u1	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP014009.D
 Acq On : 09 Mar 2023 14:22
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771

Quant Time: Mar 10 00:18:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.769	142	426786	20.415	ng/ul	99
39) 1,2,4,5-Tetrachloroben...	12.916	216	280205	20.552	ng/ul	98
40) Hexachlorocyclopentadiene	12.893	237	171614	17.578	ng/ul	99
41) 2,4,6-Trichlorophenol	13.157	196	181712	21.217	ng/ul	95
42) 2,4,5-Trichlorophenol	13.228	196	198063	21.072	ng/ul	99
43) 1,1'-Biphenyl	13.552	154	602139	20.744	ng/ul	99
44) 2-Chloronaphthalene	13.599	162	483837	20.987	ng/ul	98
45) 2-Nitroaniline	13.804	65	150770	22.112	ng/ul	98
47) Dimethylphthalate	14.169	163	636790	20.106	ng/ul	100
48) 2,6-Dinitrotoluene	14.293	165	129401	20.666	ng/ul	98
50) Acenaphthylene	14.440	152	734328	20.643	ng/ul	100
51) 3-Nitroaniline	14.622	138	122156	22.033	ng/ul	99
52) Acenaphthene	14.781	153	514153	20.675	ng/ul	99
53) 2,4-Dinitrophenol	14.828	184	79096	16.915	ng/ul	95
55) 4-Nitrophenol	14.928	109	123580	20.290	ng/ul	99
56) Dibenzofuran	15.116	168	736357	20.354	ng/ul	99
57) 2,4-Dinitrotoluene	15.081	165	189999	20.309	ng/ul	96
58) 2,3,4,6-Tetrachlorophenol	15.340	232	188333	20.944	ng/ul	99
59) Diethylphthalate	15.528	149	637455	19.844	ng/ul	100
61) Fluorene	15.763	166	611094	20.378	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.751	204	331586	20.033	ng/ul	98
63) 4-Nitroaniline	15.787	138	125449	22.013	ng/ul	96
66) 4,6-Dinitro-2-methylph...	15.840	198	123206	19.129	ng/ul	97
67) N-Nitrosodiphenylamine	15.969	169	524032	20.773	ng/ul	100
68) 4-Bromophenyl-phenylether	16.651	248	221076	20.474	ng/ul	98
69) Hexachlorobenzene	16.769	284	260003	20.433	ng/ul	95
70) Atrazine	16.922	200	221222	20.332	ng/ul	98
71) Pentachlorophenol	17.116	266	165062	19.845	ng/ul	98
72) Phenanthrene	17.516	178	1011021	20.753	ng/ul	100
74) Anthracene	17.610	178	1022542	20.700	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.522	216	287255	20.955	ng/ul	99
76) Pentachlorobenzene	15.034	250	294657	20.693	ng/ul	99
77) Carbazole	17.875	167	912955	20.953	ng/ul	99
78) Di-n-butylphthalate	18.404	149	1117520	20.522	ng/ul	99
80) Fluoranthene	19.469	202	1301708	21.084	ng/ul	99
82) Pyrene	19.822	202	1351547	20.986	ng/ul	100
83) Butylbenzylphthalate	20.675	149	538590	20.740	ng/ul	100
84) 3,3'-Dichlorobenzidine	21.457	252	499806	19.880	ng/ul	99
85) Benzo(a)anthracene	21.533	228	1419983	20.557	ng/ul	100
86) Bis(2-ethylhexyl)phtha...	21.428	149	799169	20.584	ng/ul	99
87) Chrysene	21.592	228	1359517	20.652	ng/ul	100
89) Di-n-octyl phthalate	22.398	149	1369462	20.771	ng/ul	100
90) Benzo(b)fluoranthene	23.310	252	1484141	21.316	ng/ul	99
91) Benzo(k)fluoranthene	23.357	252	1422091	21.419	ng/ul	99
93) Benzo(a)pyrene	23.980	252	1260763	20.818	ng/ul	100
94) Indeno(1,2,3-cd)pyrene	26.763	276	1475879	18.409	ng/ul	99
95) Dibenzo(a,h)anthracene	26.774	278	1233012	18.505	ng/ul	98
96) Benzo(g,h,i)perylene	27.580	276	1134637	17.707	ng/ul	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP014009.D
 Acq On : 09 Mar 2023 14:22
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020771

Quant Time: Mar 10 00:18:37 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 8.022 ng/uL
RT: 3.470 min Scan# 41
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 88	Resp: 29749
Ion Ratio	Lower Upper
88 100	
43 30.6	23.9 35.9
58 72.2	50.2 75.2



#5
Pyridine
Concen: 20.951 ng/uL
RT: 3.887 min Scan# 119
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion: 79	Resp: 202081
Ion Ratio	Lower Upper
79 100	
52 72.9	55.1 82.7
51 37.4	28.9 43.3









#6
Benzaldehyde
Concen: 26.062 ng/ul
RT: 7.211 min Scan# 684
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 77 Resp: 148523
Ion Ratio Lower Upper
77 100
105 90.8 74.2 111.2
106 87.7 69.8 104.8



#8
Phenol
Concen: 21.627 ng/ul
RT: 7.258 min Scan# 692
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion: 94 Resp: 255451
Ion Ratio Lower Upper
94 100
65 32.3 25.2 37.8
66 46.5 37.2 55.8









#10
Bis(2-Chloroethyl)ether
Concen: 21.954 ng/ul
RT: 7.493 min Scan# 732
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 93 Resp: 198662
Ion Ratio Lower Upper
93 100
63 74.6 58.8 88.2
95 32.8 25.8 38.6



#12
2-Chlorophenol
Concen: 21.717 ng/ul
RT: 7.640 min Scan# 757
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:128 Resp: 190931
Ion Ratio Lower Upper
128 100
64 50.2 38.3 57.5
130 32.5 25.7 38.5





#13

2-Methylphenol

Concen: 22.044 ng/ul

RT: 8.522 min Scan# 907

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :





SSTD020771

Tgt Ion:108 Resp: 193607

Ion Ratio Lower Upper

108 100

107 91.2 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 22.840 ng/ul

RT: 8.611 min Scan# 922

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion: 45 Resp: 223198


Ion Ratio Lower Upper

45 100

77 18.2 14.2 21.2

79 14.9 10.6 15.8









#16
Acetophenone
Concen: 22.153 ng/ul
RT: 8.911 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:105 Resp: 334815
Ion Ratio Lower Upper
105 100
77 82.5 65.4 98.2
51 27.5 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.469 ng/ul
RT: 8.893 min Scan# 970
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion: 70 Resp: 173731
Ion Ratio Lower Upper
70 100
42 53.8 40.9 61.3
101 9.4 7.9 11.9
130 20.0 16.6 25.0





#18

4-Methylphenol

Concen: 22.338 ng/ul

RT: 8.852 min Scan# 9

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :





SSTD020771

Tgt Ion:108 Resp: 215799

Ion Ratio Lower Upper

108 100

107 110.6 89.8 134.6



#19

Hexachloroethane

Concen: 20.734 ng/ul

RT: 9.169 min Scan# 1017

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:117 Resp: 79840


Ion Ratio Lower Upper

117 100

201 106.9 91.7 137.5

199 71.4 59.1 88.7









#22
Nitrobenzene
Concen: 21.559 ng/ul
RT: 9.293 min Scan# 1039
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion: 77 Resp: 248481
Ion Ratio Lower Upper
77 100
123 39.2 32.6 48.8
65 14.5 11.2 16.8



#23
Isophorone
Concen: 21.025 ng/ul
RT: 9.822 min Scan# 1128
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion: 82 Resp: 466671
Ion Ratio Lower Upper
82 100
95 8.0 7.1 10.7
138 16.0 13.3 19.9









#25
2-Nitrophenol
Concen: 21.187 ng/ul
RT: 10.011 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:139 Resp: 112622
Ion Ratio Lower Upper
139 100
65 56.8 46.6 70.0
109 42.6 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 22.080 ng/ul
RT: 10.063 min Scan# 1169
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:107 Resp: 256901
Ion Ratio Lower Upper
107 100
121 45.7 37.1 55.7
122 76.1 61.2 91.8





#27

Bis(2-Chloroethoxy)methane

Concen: 21.259 ng/ul

RT: 10.305 min Scan# 1210

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion: 93 Resp: 272932

Ion Ratio Lower Upper

93 100

95 31.7 25.6 38.4

123 10.7 8.5 12.7



#29

2,4-Dichlorophenol

Concen: 20.879 ng/ul

RT: 10.546 min Scan# 1251

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:162 Resp: 199871


Ion Ratio Lower Upper

162 100

164 63.6 52.0 78.0

98 37.4 29.1 43.7









#30
Naphthalene
Concen: 20.664 ng/ul
RT: 10.952 min Scan# 11
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:128 Resp: 612749
Ion Ratio Lower Upper
128 100
129 11.3 8.5 12.7
127 13.4 11.0 16.4



#32
4-Chloroaniline
Concen: 20.478 ng/ul
RT: 11.063 min Scan# 1339
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:127 Resp: 264641
Ion Ratio Lower Upper
127 100
129 32.3 24.9 37.3









#33
Hexachlorobutadiene
Concen: 20.036 ng/ul
RT: 11.234 min Scan# 1368
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:225 Resp: 152627
Ion Ratio Lower Upper
225 100
223 61.2 48.0 72.0
227 63.3 50.3 75.5



#34
Caprolactam
Concen: 20.439 ng/ul
RT: 11.846 min Scan# 1472
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:113 Resp: 63532
Ion Ratio Lower Upper
113 100
55 158.3 124.0 186.0
56 123.1 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 21.142 ng/ul
RT: 12.175 min Scan# 1528
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:107 Resp: 234039
Ion Ratio Lower Upper
107 100
144 23.1 19.4 29.0
142 74.2 59.6 89.4



#36
2-Methylnaphthalene
Concen: 20.340 ng/ul
RT: 12.552 min Scan# 1592
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:142 Resp: 423048
Ion Ratio Lower Upper
142 100
141 91.0 71.7 107.5





#37

1-Methylnaphthalene

Concen: 20.415 ng/ul

RT: 12.769 min Scan# 1629

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:142 Resp: 426786

Ion Ratio Lower Upper

142 100

141 93.2 73.8 110.6

116 4.7 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 20.552 ng/ul

RT: 12.916 min Scan# 1654

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Tgt Ion:216 Resp: 280205




Ion Ratio Lower Upper


216 100

214 79.0 61.7 92.5

179 20.3 16.2 24.4

108 16.1 13.2 19.8









#40
Hexachlorocyclopentadiene
Concen: 17.578 ng/ul
RT: 12.893 min Scan# 1650
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:237 Resp: 171614
Ion Ratio Lower Upper
237 100
235 62.0 49.0 73.6
272 11.8 9.6 14.4



#41
2,4,6-Trichlorophenol
Concen: 21.217 ng/ul
RT: 13.157 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:196 Resp: 181712
Ion Ratio Lower Upper
196 100
198 92.5 78.5 117.7
200 29.7 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 21.072 ng/ul

RT: 13.228 min Scan# 1707

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:196 Resp: 198063

Ion Ratio Lower Upper

196 100

198 97.6 77.6 116.4

200 32.9 25.3 37.9



#43

1,1'-Biphenyl

Concen: 20.744 ng/ul

RT: 13.552 min Scan# 1762

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:154 Resp: 602139


Ion Ratio Lower Upper

154 100

153 41.2 32.6 49.0

76 11.5 9.4 14.0









#44
2-Chloronaphthalene
Concen: 20.987 ng/ul
RT: 13.599 min Scan# 1770
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:162 Resp: 483837
Ion Ratio Lower Upper
162 100
164 32.8 26.0 39.0
127 41.0 31.3 46.9



#45
2-Nitroaniline
Concen: 22.112 ng/ul
RT: 13.804 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion: 65 Resp: 150770
Ion Ratio Lower Upper
65 100
92 70.2 56.3 84.5
138 99.2 81.4 122.2









#47
Dimethylphthalate
Concen: 20.106 ng/ul
RT: 14.169 min Scan# 1867
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:163 Resp: 636790
Ion Ratio Lower Upper
163 100
194 4.1 3.6 5.4
164 10.3 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.666 ng/ul
RT: 14.293 min Scan# 1888
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:165 Resp: 129401
Ion Ratio Lower Upper
165 100
89 66.3 51.7 77.5
121 22.2 17.5 26.3









#50
Acenaphthylene
Concen: 20.643 ng/ul
RT: 14.440 min Scan# 1913
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:152 Resp: 734328
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.3 10.4 15.6



#51
3-Nitroaniline
Concen: 22.033 ng/ul
RT: 14.622 min Scan# 1944
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:138 Resp: 122156
Ion Ratio Lower Upper
138 100
108 11.9 10.5 15.7
92 124.4 98.5 147.7









#52
Acenaphthene
Concen: 20.675 ng/ul
RT: 14.781 min Scan# 1971
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:153 Resp: 514153
Ion Ratio Lower Upper
153 100
152 47.5 37.7 56.5
154 88.9 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 16.915 ng/ul
RT: 14.828 min Scan# 1979
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:184 Resp: 79096
Ion Ratio Lower Upper
184 100
63 71.3 51.9 77.9
154 75.0 58.8 88.2









#55
4-Nitrophenol
Concen: 20.290 ng/ul
RT: 14.928 min Scan# 1996
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:109 Resp: 123580
Ion Ratio Lower Upper
109 100
139 92.5 74.5 111.7
65 106.4 86.6 130.0



#56
Dibenzofuran
Concen: 20.354 ng/ul
RT: 15.116 min Scan# 2028
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:168 Resp: 736357
Ion Ratio Lower Upper
168 100
139 40.3 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 20.309 ng/ul
RT: 15.081 min Scan# 2022
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:165 Resp: 189999
Ion Ratio Lower Upper
165 100
63 45.8 34.6 52.0
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 20.944 ng/ul
RT: 15.340 min Scan# 2066
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:232 Resp: 188333
Ion Ratio Lower Upper
232 100
131 41.4 34.1 51.1
130 2.7 2.1 3.1
166 28.9 23.0 34.6









#59
Diethylphthalate
Concen: 19.844 ng/ul
RT: 15.528 min Scan# 2098
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:149 Resp: 637455
Ion Ratio Lower Upper
149 100
177 22.9 18.2 27.2
150 12.5 10.0 15.0



#61
Fluorene
Concen: 20.378 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:166 Resp: 611094
Ion Ratio Lower Upper
166 100
165 98.8 78.9 118.3
167 13.3 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 20.033 ng/ul
RT: 15.751 min Scan# 2136
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:204 Resp: 331586
Ion Ratio Lower Upper
204 100
206 33.6 26.6 39.8
141 63.3 49.2 73.8



#63
4-Nitroaniline
Concen: 22.013 ng/ul
RT: 15.787 min Scan# 2142
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:138 Resp: 125449
Ion Ratio Lower Upper
138 100
92 58.3 44.5 66.7
108 113.5 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 19.129 ng/ul
RT: 15.840 min Scan# 2152
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:198 Resp: 123206
Ion Ratio Lower Upper
198 100
182 4.1 3.0 4.6
77 27.1 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 20.773 ng/ul
RT: 15.969 min Scan# 2173
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:169 Resp: 524032
Ion Ratio Lower Upper
169 100
168 67.4 53.7 80.5
167 36.2 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 20.474 ng/ul
RT: 16.651 min Scan# 2109
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:248 Resp: 221076
Ion Ratio Lower Upper
248 100
250 95.6 78.8 118.2
141 76.5 60.8 91.2



#69
Hexachlorobenzene
Concen: 20.433 ng/ul
RT: 16.769 min Scan# 2309
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:284 Resp: 260003
Ion Ratio Lower Upper
284 100
142 34.2 23.2 34.8
249 32.4 25.7 38.5





#70

Atrazine

Concen: 20.332 ng/ul

RT: 16.922 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:200 Resp: 221222

Ion Ratio Lower Upper

200 100

173 23.1 19.0 28.4

215 46.1 38.5 57.7



#71

Pentachlorophenol

Concen: 19.845 ng/ul

RT: 17.116 min Scan# 2368

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:266 Resp: 165062


Ion Ratio Lower Upper

266 100

264 63.0 48.4 72.6

268 63.3 50.4 75.6





#72

Phenanthrene

Concen: 20.753 ng/ul

RT: 17.516 min Scan# 2436

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:178 Resp: 1011021

Ion Ratio Lower Upper

178 100

179 15.3 12.2 18.4

176 19.3 15.4 23.0



#74

Anthracene

Concen: 20.700 ng/ul

RT: 17.610 min Scan# 2452

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:178 Resp: 1022542


Ion Ratio Lower Upper

178 100

179 15.0 12.5 18.7

176 18.5 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 20.955 ng/uL

RT: 13.522 min Scan# 1757

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:216 Resp: 287255

Ion Ratio Lower Upper

216 100

214 77.0 61.4 92.2

218 48.1 39.1 58.7



#76

Pentachlorobenzene

Concen: 20.693 ng/uL

RT: 15.034 min Scan# 2014

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:250 Resp: 294657


Ion Ratio Lower Upper

250 100

248 63.8 51.0 76.4

252 64.9 50.8 76.2





#77

Carbazole

Concen: 20.953 ng/ul

RT: 17.875 min Scan# 2497

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:167 Resp: 912955

Ion Ratio Lower Upper

167 100

166 21.5 17.5 26.3

139 13.4 10.5 15.7



#78

Di-n-butylphthalate

Concen: 20.522 ng/ul

RT: 18.404 min Scan# 2587

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:149 Resp: 1117520


Ion Ratio Lower Upper

149 100

150 8.9 7.4 11.2

104 5.8 4.6 6.8





#80

Fluoranthene

Concen: 21.084 ng/ul

RT: 19.469 min Scan# 21

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:202 Resp: 1301708

Ion Ratio Lower Upper

202 100

101 7.8 6.6 10.0

100 6.1 5.0 7.4



#82

Pyrene

Concen: 20.986 ng/ul

RT: 19.822 min Scan# 2828

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:202 Resp: 1351547


Ion Ratio Lower Upper

202 100

101 9.8 7.8 11.8

100 8.2 6.5 9.7





#83

Butylbenzylphthalate

Concen: 20.740 ng/ul

RT: 20.675 min Scan# 2973

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:149 Resp: 538590

Ion Ratio Lower Upper

149 100

91 72.4 57.9 86.9

206 22.7 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 19.880 ng/ul

RT: 21.457 min Scan# 3106

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:252 Resp: 499806


Ion Ratio Lower Upper

252 100

254 65.0 51.3 76.9

126 8.8 6.9 10.3





#85

Benzo(a)anthracene

Concen: 20.557 ng/ul

RT: 21.533 min Scan# 3119

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:228 Resp: 1419983

Ion Ratio Lower Upper

228 100

229 19.8 15.5 23.3

226 27.6 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 20.584 ng/ul

RT: 21.428 min Scan# 3101

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:149 Resp: 799169


Ion Ratio Lower Upper

149 100

167 27.7 22.3 33.5

279 4.7 4.3 6.5





#87

Chrysene

Concen: 20.652 ng/ul

RT: 21.592 min Scan# 3128

Delta R.T. 0.000 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:228 Resp: 1359517

Ion Ratio Lower Upper

228 100

226 30.2 24.2 36.2

229 19.3 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 20.771 ng/ul




RT: 22.398 min Scan# 3266


Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Tgt Ion:149 Resp: 1369462





#90

Benzo(b)fluoranthene

Concen: 21.316 ng/ul

RT: 23.310 min Scan# 3420

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:252 Resp: 1484141

Ion Ratio Lower Upper

252 100

253 22.1 17.3 25.9

125 7.0 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 21.419 ng/ul

RT: 23.357 min Scan# 3429

Delta R.T. -0.006 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22




Tgt Ion:252 Resp: 1422091


Ion Ratio Lower Upper

252 100

253 22.0 17.4 26.0

125 7.0 5.5 8.3









#93
Benzo(a)pyrene
Concen: 20.818 ng/ul
RT: 23.980 min Scan# 31
Delta R.T. 0.000 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22




Instrument :
BNA_P
ClientSampleId :
SSTD020771


Tgt Ion:252 Resp: 1260763
Ion Ratio Lower Upper
252 100
253 21.7 17.4 26.2
125 7.7 6.2 9.2



#94
Indeno(1,2,3-cd)pyrene
Concen: 18.409 ng/ul
RT: 26.763 min Scan# 4008
Delta R.T. -0.006 min
Lab File: BP014009.D
Acq: 09 Mar 2023 14:22

Tgt Ion:276 Resp: 1475879
Ion Ratio Lower Upper
276 100
138 17.5 13.1 19.7
277 25.5 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 18.505 ng/ul

RT: 26.774 min Scan# 4010

Delta R.T. -0.012 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Instrument :

BNA_P

ClientSampleId :

SSTD020771





Tgt Ion:278 Resp: 1233012

Ion Ratio Lower Upper

278 100

139 12.4 8.7 13.1

279 24.2 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 17.707 ng/ul

RT: 27.580 min Scan# 4147

Delta R.T. -0.018 min

Lab File: BP014009.D

Acq: 09 Mar 2023 14:22

Tgt Ion:276 Resp: 1134637

Ion Ratio Lower Upper

276 100

138 15.1 11.9 17.9

277 23.8 18.7 28.1

