

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP013976.D
 Acq On : 08 Mar 2023 14:19
 Operator : CG/JU
 Sample : 01746-20
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAC5

Quant Time: Mar 09 01:13:08 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.081	152	110716	20.000	ng/u1	0.00
20) Naphthalene-d8	10.904	136	469427	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.722	164	353242	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.475	188	881603	20.000	ng/u1	0.00
79) Chrysene-d12	21.551	240	1014303	20.000	ng/u1	0.00
88) Perylene-d12	24.092	264	1130601	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	12224	4.172	ng/uL	0.00
4) Pyridine-d5	3.881	84	45815	5.647	ng/u1	0.01
7) Phenol-d5	7.234	99	54911	5.620	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.405	67	158376	27.794	ng/u1	0.00
11) 2-Chlorophenol-d4	7.610	132	155876	20.940	ng/u1	0.00
15) 4-Methylphenol-d8	8.793	113	112940	14.151	ng/u1	0.00
21) Nitrobenzene-d5	9.257	128	107124	28.600	ng/u1	0.00
24) 2-Nitrophenol-d4	9.981	143	124398	27.745	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.522	165	210568	24.960	ng/u1	0.00
31) 4-Chloroaniline-d4	11.046	131	222214	19.904	ng/u1	0.00
46) Dimethylphthalate-d6	14.122	166	946002	31.556	ng/u1	0.00
49) Acenaphthylene-d8	14.416	160	933322	29.389	ng/u1	0.00
54) 4-Nitrophenol-d4	14.916	143	29701	5.622	ng/u1	0.00
60) Fluorene-d10	15.710	176	784167	30.864	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.828	200	199565	29.950	ng/u1	0.00
73) Anthracene-d10	17.575	188	1511471	35.317	ng/u1	0.00
81) Pyrene-d10	19.798	212	1976926	35.339	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.927	264	2154711	35.829	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
Data File : BP013976.D
Acq On : 08 Mar 2023 14:19
Operator : CG/JU
Sample : 01746-20
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
DCAC5

Quant Time: Mar 09 01:13:08 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Thu Mar 09 01:10:12 2023
Response via : Initial Calibration

