

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP013972.D
 Acq On : 08 Mar 2023 11:53
 Operator : CG/JU
 Sample : 01746-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAG9

Quant Time: Mar 09 01:25:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.093	152	126278	20.000	ng/u1	0.01
20) Naphthalene-d8	10.916	136	542019	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.734	164	371246	20.000	ng/u1	0.01
64) Phenanthrene-d10	17.486	188	924979	20.000	ng/u1	0.01
79) Chrysene-d12	21.569	240	1155762	20.000	ng/u1	0.01
88) Perylene-d12	24.115	264	1352850	20.000	ng/u1	0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.440	96	21755	6.509	ng/uL	0.00
4) Pyridine-d5	3.869	84	286027	30.909	ng/u1	0.00
7) Phenol-d5	7.240	99	372290	33.408	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.410	67	226024	34.778	ng/u1	0.00
11) 2-Chlorophenol-d4	7.616	132	298687	35.180	ng/u1	0.00
15) 4-Methylphenol-d8	8.799	113	333637	36.653	ng/u1	0.00
21) Nitrobenzene-d5	9.269	128	154796	35.793	ng/u1	0.01
24) 2-Nitrophenol-d4	9.993	143	182235	35.201	ng/u1	0.01
28) 2,4-Dichlorophenol-d3	10.534	165	341434	35.052	ng/u1	0.01
31) 4-Chloroaniline-d4	11.057	131	70942	5.503	ng/u1	0.01
46) Dimethylphthalate-d6	14.134	166	1157890	36.751	ng/u1	0.01
49) Acenaphthylene-d8	14.428	160	1207819	36.188	ng/u1	0.01
54) 4-Nitrophenol-d4	14.928	143	193268	34.809	ng/u1	0.01
60) Fluorene-d10	15.728	176	920540	34.475	ng/u1	0.01
65) 4,6-Dinitro-2-methylph...	15.845	200	206026	29.470	ng/u1	0.01
73) Anthracene-d10	17.586	188	1530453	34.083	ng/u1	0.01
81) Pyrene-d10	19.810	212	2089007	32.772	ng/u1	0.01
92) Benzo(a)pyrene-d12	23.951	264	2543559	35.346	ng/u1	0.02
Target Compounds						
					Qvalue	
6) Benzaldehyde	7.222	77	19626	3.540	ng/u1	92
12) 2-Chlorophenol	7.652	128	361051	42.208	ng/u1	99
14) 2,2'-oxybis(1-Chloropr...	8.622	45	598314	62.927	ng/u1	98
16) Acetophenone	8.922	105	615654	41.867	ng/u1	100
19) Hexachloroethane	9.181	117	76059	20.302	ng/u1	98
22) Nitrobenzene	9.310	77	713650	61.687	ng/u1	100
25) 2-Nitrophenol	10.028	139	268092	50.246	ng/u1	98
29) 2,4-Dichlorophenol	10.557	162	245548	25.555	ng/u1	99
30) Naphthalene	10.969	128	474085	15.928	ng/u1	99
35) 4-Chloro-3-methylphenol	12.193	107	578036	52.024	ng/u1	99
37) 1-Methylnaphthalene	12.787	142	987916	47.082	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.928	216	282102	21.013	ng/u1	99
41) 2,4,6-Trichlorophenol	13.169	196	299809	35.550	ng/u1	99
43) 1,1'-Biphenyl	13.569	154	1203379	42.103	ng/u1	99
45) 2-Nitroaniline	13.816	65	318178	47.391	ng/u1	96
48) 2,6-Dinitrotoluene	14.310	165	285812	46.356	ng/u1	95
50) Acenaphthylene	14.457	152	1846072	52.705	ng/u1	99
52) Acenaphthene	14.792	153	109278	4.463	ng/u1	99
55) 4-Nitrophenol	14.945	109	388233	64.733	ng/u1	96
56) Dibenzofuran	15.128	168	724109	20.327	ng/u1	98
57) 2,4-Dinitrotoluene	15.092	165	457055	49.616	ng/u1	94
59) Diethylphthalate	15.545	149	1503584	47.535	ng/u1	100
61) Fluorene	15.781	166	245598	8.317	ng/u1	100
67) N-Nitrosodiphenylamine	15.986	169	1298356	49.700	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP013972.D
 Acq On : 08 Mar 2023 11:53
 Operator : CG/JU
 Sample : 01746-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAG9

Quant Time: Mar 09 01:25:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


69) Hexachlorobenzene	16.786	284	519550	39.429	ng/u1	96
71) Pentachlorophenol	17.133	266	352671	40.945	ng/u1	97
72) Phenanthrene	17.533	178	1607578	31.864	ng/u1	99
78) Di-n-butylphthalate	18.416	149	2312980	41.017	ng/u1	100
82) Pyrene	19.833	202	1211850	15.989	ng/u1	98
85) Benzo(a)anthracene	21.551	228	3281866	40.371	ng/u1	100
89) Di-n-octyl phthalate	22.416	149	2145598	25.135	ng/u1	100
90) Benzo(b)fluoranthene	23.333	252	2860713	31.734	ng/u1	99
93) Benzo(a)pyrene	23.998	252	1665200	21.237	ng/u1	98
95) Dibenzo(a,h)anthracene	26.827	278	4005858	46.434	ng/u1	99
96) Benzo(g,h,i)perylene	27.627	276	2065045	24.891	ng/u1	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP030823\
 Data File : BP013972.D
 Acq On : 08 Mar 2023 11:53
 Operator : CG/JU
 Sample : 01746-18
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAG9

Quant Time: Mar 09 01:25:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Thu Mar 09 01:10:12 2023
 Response via : Initial Calibration









#6
Benzaldehyde
Concen: 3.540 ng/ul
RT: 7.222 min Scan# 685
Delta R.T. 0.006 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion: 77 Resp: 19626
Ion Ratio Lower Upper
77 100
105 82.7 74.2 111.2
106 81.6 69.8 104.8



#12
2-Chlorophenol
Concen: 42.208 ng/ul
RT: 7.652 min Scan# 759
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion: 128 Resp: 361051
Ion Ratio Lower Upper
128 100
64 48.0 38.3 57.5
130 31.5 25.7 38.5









#14
2,2'-oxybis(1-Chloropropane)
Concen: 62.927 ng/ul
RT: 8.622 min Scan# 91
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion: 45 Resp: 598314
Ion Ratio Lower Upper
45 100
77 18.4 14.2 21.2
79 14.3 10.6 15.8



#16
Acetophenone
Concen: 41.867 ng/ul
RT: 8.922 min Scan# 975
Delta R.T. 0.006 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion:105 Resp: 615654
Ion Ratio Lower Upper
105 100
77 82.0 65.4 98.2
51 27.2 21.8 32.8









#19
Hexachloroethane
Concen: 20.302 ng/ul
RT: 9.181 min Scan# 1018
Delta R.T. 0.006 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion:117 Resp: 76059
Ion Ratio Lower Upper
117 100
201 118.1 91.7 137.5
199 73.4 59.1 88.7



#22
Nitrobenzene
Concen: 61.687 ng/ul
RT: 9.310 min Scan# 1041
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion: 77 Resp: 713650
Ion Ratio Lower Upper
77 100
123 40.8 32.6 48.8
65 14.0 11.2 16.8









#25
2-Nitrophenol
Concen: 50.246 ng/ul
RT: 10.028 min Scan# 1161
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion:139 Resp: 268092
Ion Ratio Lower Upper
139 100
65 57.2 46.6 70.0
109 41.7 34.5 51.7



#29
2,4-Dichlorophenol
Concen: 25.555 ng/ul
RT: 10.557 min Scan# 1253
Delta R.T. 0.006 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion:162 Resp: 245548
Ion Ratio Lower Upper
162 100
164 65.2 52.0 78.0
98 37.3 29.1 43.7









#30
Naphthalene
Concen: 15.928 ng/ul
RT: 10.969 min Scan# 1321
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion:128 Resp: 474085
Ion Ratio Lower Upper
128 100
129 10.9 8.5 12.7
127 14.3 11.0 16.4



#35
4-Chloro-3-methylphenol
Concen: 52.024 ng/ul
RT: 12.193 min Scan# 1531
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion:107 Resp: 578036
Ion Ratio Lower Upper
107 100
144 24.3 19.4 29.0
142 73.7 59.6 89.4





#37

1-Methylnaphthalene

Concen: 47.082 ng/ul

RT: 12.787 min Scan# 1632

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:142 Resp: 987916

Ion Ratio Lower Upper

142 100

141 92.8 73.8 110.6

116 4.6 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 21.013 ng/ul

RT: 12.928 min Scan# 1656

Delta R.T. 0.006 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Tgt Ion:216 Resp: 282102




Ion Ratio Lower Upper


216 100

214 76.8 61.7 92.5

179 21.2 16.2 24.4

108 15.7 13.2 19.8





#41

2,4,6-Trichlorophenol

Concen: 35.550 ng/ul

RT: 13.169 min Scan# 1695

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:196 Resp: 299809

Ion Ratio Lower Upper

196 100

198 98.9 78.5 117.7

200 32.1 25.4 38.0



#43

1,1'-Biphenyl

Concen: 42.103 ng/ul

RT: 13.569 min Scan# 1765

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:154 Resp: 1203379


Ion Ratio Lower Upper

154 100

153 40.4 32.6 49.0

76 11.2 9.4 14.0









#45
2-Nitroaniline
Concen: 47.391 ng/ul
RT: 13.816 min Scan# 1807
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion: 65 Resp: 318178
Ion Ratio Lower Upper
65 100
92 69.1 56.3 84.5
138 96.8 81.4 122.2



#48
2,6-Dinitrotoluene
Concen: 46.356 ng/ul
RT: 14.310 min Scan# 1891
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion: 165 Resp: 285812
Ion Ratio Lower Upper
165 100
89 60.6 51.7 77.5
121 20.4 17.5 26.3





#50

Acenaphthylene

Concen: 52.705 ng/ul

RT: 14.457 min Scan# 1914

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:152 Resp: 1846072

Ion Ratio Lower Upper

152 100

151 20.6 16.2 24.2

153 13.3 10.4 15.6



#52

Acenaphthene

Concen: 4.463 ng/ul

RT: 14.792 min Scan# 1973

Delta R.T. 0.006 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:153 Resp: 109278


Ion Ratio Lower Upper

153 100

152 47.5 37.7 56.5

154 87.5 70.4 105.6









#55
4-Nitrophenol
Concen: 64.733 ng/ul
RT: 14.945 min Scan# 1999
Delta R.T. 0.017 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion:109 Resp: 388233
Ion Ratio Lower Upper
109 100
139 88.9 74.5 111.7
65 104.0 86.6 130.0



#56
Dibenzofuran
Concen: 20.327 ng/ul
RT: 15.128 min Scan# 2030
Delta R.T. 0.006 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion:168 Resp: 724109
Ion Ratio Lower Upper
168 100
139 40.9 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 49.616 ng/ul
RT: 15.092 min Scan# 20
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion:165 Resp: 457055
Ion Ratio Lower Upper
165 100
63 47.4 34.6 52.0
182 2.9 2.2 3.2



#59
Diethylphthalate
Concen: 47.535 ng/ul
RT: 15.545 min Scan# 2101
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion:149 Resp: 1503584
Ion Ratio Lower Upper
149 100
177 22.4 18.2 27.2
150 12.5 10.0 15.0





#61

Fluorene

Concen: 8.317 ng/ul

RT: 15.781 min Scan# 2139

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:166 Resp: 245598

Ion Ratio Lower Upper

166 100

165 98.9 78.9 118.3

167 13.5 11.0 16.6



#67

N-Nitrosodiphenylamine

Concen: 49.700 ng/ul

RT: 15.986 min Scan# 2176

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:169 Resp: 1298356


Ion Ratio Lower Upper

169 100

168 67.3 53.7 80.5

167 36.8 28.6 43.0





#69

Hexachlorobenzene

Concen: 39.429 ng/ul

RT: 16.786 min Scan# 2310

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:284 Resp: 519550

Ion Ratio Lower Upper

284 100

142 32.0 23.2 34.8

249 33.6 25.7 38.5



#71

Pentachlorophenol

Concen: 40.945 ng/ul

RT: 17.133 min Scan# 2371

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:266 Resp: 352671


Ion Ratio Lower Upper

266 100

264 64.0 48.4 72.6

268 64.1 50.4 75.6









#72
Phenanthrene
Concen: 31.864 ng/ul
RT: 17.533 min Scan# 2439
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53




Instrument :
BNA_P
ClientSampleId :
DCAG9


Tgt Ion:178 Resp: 1607578
Ion Ratio Lower Upper
178 100
179 15.2 12.2 18.4
176 19.8 15.4 23.0



#78
Di-n-butylphthalate
Concen: 41.017 ng/ul
RT: 18.416 min Scan# 2589
Delta R.T. 0.012 min
Lab File: BP013972.D
Acq: 08 Mar 2023 11:53

Tgt Ion:149 Resp: 2312980
Ion Ratio Lower Upper
149 100
150 9.4 7.4 11.2
104 5.9 4.6 6.8





#82

Pyrene

Concen: 15.989 ng/ul

RT: 19.833 min Scan# 2829

Delta R.T. 0.006 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:202 Resp: 1211850

Ion Ratio Lower Upper

202 100

101 10.5 7.8 11.8

100 8.6 6.5 9.7



#85

Benzo(a)anthracene

Concen: 40.371 ng/ul

RT: 21.551 min Scan# 3122

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:228 Resp: 3281866


Ion Ratio Lower Upper

228 100

229 19.6 15.5 23.3

226 27.7 22.1 33.1





#89

Di-n-octyl phthalate

Concen: 25.135 ng/ul

RT: 22.416 min Scan# 31

Delta R.T. 0.012 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53





Instrument :

BNA_P

ClientSampleId :

DCAG9

Tgt Ion:149 Resp: 2145598



#90

Benzo(b)fluoranthene

Concen: 31.734 ng/ul

RT: 23.333 min Scan# 3425

Delta R.T. 0.017 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:252 Resp: 2860713


Ion Ratio Lower Upper

252 100

253 21.7 17.3 25.9

125 7.3 5.4 8.2





#93

Benzo(a)pyrene

Concen: 21.237 ng/ul

RT: 23.998 min Scan# 3538

Delta R.T. 0.017 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9





Tgt Ion:252 Resp: 1665200

Ion Ratio Lower Upper

252 100

253 21.2 17.4 26.2

125 8.7 6.2 9.2



#95

Dibenzo(a,h)anthracene

Concen: 46.434 ng/ul

RT: 26.827 min Scan# 4019

Delta R.T. 0.041 min

Lab File: BP013972.D

Acq: 08 Mar 2023 11:53




Tgt Ion:278 Resp: 4005858


Ion Ratio Lower Upper

278 100

139 11.8 8.7 13.1

279 23.8 19.0 28.4





#96

Benzo(g,h,i)perylene

Concen: 24.891 ng/ul

RT: 27.627 min Scan# 41

Delta R.T. 0.029 min

Lab File: BP013972.D


Acq: 08 Mar 2023 11:53

Instrument :

BNA_P

ClientSampleId :

DCAG9



Tgt Ion:276 Resp: 2065045

Ion Ratio Lower Upper

276 100

138 15.8 11.9 17.9

277 23.6 18.7 28.1

