




Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP031622\
Data File : BP009419.D
Acq On : 16 Mar 2022 12:05
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Mar 16 15:25:29 2022
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP030522.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Mar 05 01:55:59 2022
Response via : Initial Calibration



TIC: BP009419.D\data.ms

(77) Benzidine

19.422min (+ 0.006) 52624.66 ng

response 2479687

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	14.40	13.79
183.00	11.70	11.81
0.00	0.00	0.00