

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014180.D
 Acq On : 21 Mar 2023 14:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020782

Quant Time: Mar 22 01:19:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 21 04:47:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	198746	20.000	ng/ul	0.00
20) Naphthalene-d8	10.869	136	873122	20.000	ng/ul	# 0.00
38) Acenaphthene-d10	14.687	164	597557	20.000	ng/ul	0.00
64) Phenanthrene-d10	17.445	188	1388382	20.000	ng/ul	# 0.00
79) Chrysene-d12	21.522	240	1272911	20.000	ng/ul	# 0.00
88) Perylene-d12	24.039	264	1099726	20.000	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	38897	7.395	ng/uL	0.00
4) Pyridine-d5	3.846	84	277390	19.046	ng/ul	0.00
7) Phenol-d5	7.205	99	368043	20.985	ng/ul	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	229238	22.411	ng/ul	0.00
11) 2-Chlorophenol-d4	7.575	132	276958	20.727	ng/ul	0.00
15) 4-Methylphenol-d8	8.763	113	302655	21.126	ng/ul	0.00
21) Nitrobenzene-d5	9.222	128	142424	20.444	ng/ul	0.00
24) 2-Nitrophenol-d4	9.946	143	167833	20.125	ng/ul	0.00
28) 2,4-Dichlorophenol-d3	10.493	165	308477	19.659	ng/ul	0.00
31) 4-Chloroaniline-d4	11.010	131	416111	20.039	ng/ul	0.00
46) Dimethylphthalate-d6	14.093	166	973768	19.201	ng/ul	0.00
49) Acenaphthylene-d8	14.381	160	1081966	20.140	ng/ul	0.00
54) 4-Nitrophenol-d4	14.893	143	160047	17.909	ng/ul	0.00
60) Fluorene-d10	15.681	176	822621	19.140	ng/ul	0.00
65) 4,6-Dinitro-2-methylph...	15.804	200	192011	18.298	ng/ul	0.00
73) Anthracene-d10	17.545	188	1321775	19.611	ng/ul	0.00
81) Pyrene-d10	19.769	212	1576164	22.451	ng/ul	0.00
92) Benzo(a)pyrene-d12	23.874	264	1169975	20.001	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	41153	7.247	ng/uL	89
5) Pyridine	3.870	79	287962	19.497	ng/ul	98
6) Benzaldehyde	7.187	77	224915	25.773	ng/ul	100
8) Phenol	7.234	94	378543	20.929	ng/ul	98
10) Bis(2-Chloroethyl)ether	7.464	93	304306	21.961	ng/ul	98
12) 2-Chlorophenol	7.611	128	285957	21.240	ng/ul	94
13) 2-Methylphenol	8.499	108	287970	21.411	ng/ul	99
14) 2,2'-oxybis(1-Chloropr...	8.581	45	383268	25.612	ng/ul	99
16) Acetophenone	8.881	105	490490	21.193	ng/ul	98
17) N-Nitroso-di-n-propyla...	8.863	70	265208	22.399	ng/ul	98
18) 4-Methylphenol	8.828	108	320136	21.641	ng/ul	100
19) Hexachloroethane	9.134	117	126105	21.386	ng/ul	95
22) Nitrobenzene	9.263	77	395838	21.241	ng/ul	98
23) Isophorone	9.793	82	740391	20.631	ng/ul	97
25) 2-Nitrophenol	9.981	139	172272	20.044	ng/ul	95
26) 2,4-Dimethylphenol	10.034	107	382483	20.331	ng/ul	98
27) Bis(2-Chloroethoxy)met...	10.275	93	434651	20.938	ng/ul	99
29) 2,4-Dichlorophenol	10.516	162	300768	19.431	ng/ul	98
30) Naphthalene	10.922	128	967954	20.189	ng/ul	99
32) 4-Chloroaniline	11.034	127	416894	19.951	ng/ul	99
33) Hexachlorobutadiene	11.199	225	223619	18.155	ng/ul	99
34) Caprolactam	11.828	113	92591	18.423	ng/ul	94
35) 4-Chloro-3-methylphenol	12.152	107	365155	20.402	ng/ul	95
36) 2-Methylnaphthalene	12.522	142	670969	19.952	ng/ul	98


Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014180.D
 Acq On : 21 Mar 2023 14:24
 Operator : CG/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SSTD020782

Quant Time: Mar 22 01:19:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Tue Mar 21 04:47:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.740	142	663946	19.643	ng/ul	100
39) 1,2,4,5-Tetrachloroben...	12.887	216	408410	18.900	ng/ul	98
40) Hexachlorocyclopentadiene	12.857	237	240836	15.564	ng/ul	100
41) 2,4,6-Trichlorophenol	13.128	196	268993	19.816	ng/ul	99
42) 2,4,5-Trichlorophenol	13.204	196	289829	19.455	ng/ul	99
43) 1,1'-Biphenyl	13.522	154	930407	20.224	ng/ul	98
44) 2-Chloronaphthalene	13.569	162	735102	20.119	ng/ul	98
45) 2-Nitroaniline	13.775	65	243266	22.511	ng/ul	93
47) Dimethylphthalate	14.140	163	967655	19.277	ng/ul	99
48) 2,6-Dinitrotoluene	14.269	165	202252	20.380	ng/ul	99
50) Acenaphthylene	14.410	152	1144667	20.303	ng/ul	100
51) 3-Nitroaniline	14.604	138	182283	20.745	ng/ul	95
52) Acenaphthene	14.751	153	793491	20.132	ng/ul	100
53) 2,4-Dinitrophenol	14.810	184	122560	16.537	ng/ul	93
55) 4-Nitrophenol	14.910	109	155468	16.105	ng/ul	86
56) Dibenzofuran	15.087	168	1126178	19.641	ng/ul	99
57) 2,4-Dinitrotoluene	15.051	165	296180	19.975	ng/ul	90
58) 2,3,4,6-Tetrachlorophenol	15.316	232	267824	18.792	ng/ul	98
59) Diethylphthalate	15.498	149	994271	19.529	ng/ul	99
61) Fluorene	15.740	166	929385	19.554	ng/ul	100
62) 4-Chlorophenyl-phenyle...	15.722	204	492438	18.771	ng/ul	98
63) 4-Nitroaniline	15.763	138	186797	20.681	ng/ul	94
66) 4,6-Dinitro-2-methylph...	15.816	198	184802	18.460	ng/ul#	93
67) N-Nitrosodiphenylamine	15.940	169	793144	20.227	ng/ul	98
68) 4-Bromophenyl-phenylether	16.622	248	318039	18.949	ng/ul	98
69) Hexachlorobenzene	16.740	284	365741	18.492	ng/ul#	94
70) Atrazine	16.892	200	322172	19.049	ng/ul	98
71) Pentachlorophenol	17.092	266	220511	17.056	ng/ul	98
72) Phenanthrene	17.487	178	1500872	19.820	ng/ul	100
74) Anthracene	17.581	178	1542467	20.089	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.493	216	425335	19.962	ng/ul	98
76) Pentachlorobenzene	15.004	250	429885	19.422	ng/ul	99
77) Carbazole	17.845	167	1354916	20.006	ng/ul	99
78) Di-n-butylphthalate	18.375	149	1766506	20.870	ng/ul	99
80) Fluoranthene	19.439	202	1841892	23.018	ng/ul#	94
82) Pyrene	19.798	202	1905468	22.827	ng/ul	95
83) Butylbenzylphthalate	20.651	149	830771	24.682	ng/ul	98
84) 3,3'-Dichlorobenzidine	21.433	252	588355	18.055	ng/ul	99
85) Benzo(a)anthracene	21.504	228	1787117	19.961	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.404	149	1232497	24.492	ng/ul	100
87) Chrysene	21.563	228	1690894	19.817	ng/ul	100
89) Di-n-octyl phthalate	22.363	149	2064469	29.751	ng/ul	100
90) Benzo(b)fluoranthene	23.269	252	1556617	21.242	ng/ul	98
91) Benzo(k)fluoranthene	23.316	252	1485582	21.260	ng/ul	98
93) Benzo(a)pyrene	23.927	252	1273522	19.981	ng/ul	98
94) Indeno(1,2,3-cd)pyrene	26.686	276	1519243	18.006	ng/ul	98
95) Dibenzo(a,h)anthracene	26.704	278	1263330	18.015	ng/ul	98
96) Benzo(g,h,i)perylene	27.504	276	1201672	17.818	ng/ul	98





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 7.247 ng/uL
RT: 3.446 min Scan# 44
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion: 88	Resp: 41153
Ion Ratio	Lower Upper
88 100	
43 30.4	23.9 35.9
58 74.4	50.2 75.2



#5
Pyridine
Concen: 19.497 ng/uL
RT: 3.870 min Scan# 116
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion: 79	Resp: 287962
Ion Ratio	Lower Upper
79 100	
52 71.1	55.1 82.7
51 35.7	28.9 43.3









#6
Benzaldehyde
Concen: 25.773 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. -0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion: 77 Resp: 224915
Ion Ratio Lower Upper
77 100
105 92.5 74.2 111.2
106 87.5 69.8 104.8



#8
Phenol
Concen: 20.929 ng/ul
RT: 7.234 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion: 94 Resp: 378543
Ion Ratio Lower Upper
94 100
65 31.3 25.2 37.8
66 44.7 37.2 55.8









#10
 Bis(2-Chloroethyl)ether
 Concen: 21.961 ng/ul
 RT: 7.464 min Scan# 71
 Delta R.T. -0.006 min
 Lab File: BP014180.D
 Acq: 21 Mar 2023 14:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020782


Tgt Ion: 93	Resp: 304306
Ion Ratio	Lower Upper
93 100	
63 76.1	58.8 88.2
95 32.2	25.8 38.6



#12
 2-Chlorophenol
 Concen: 21.240 ng/ul
 RT: 7.611 min Scan# 752
 Delta R.T. 0.000 min
 Lab File: BP014180.D
 Acq: 21 Mar 2023 14:24

Tgt Ion:128	Resp: 285957		
Ion Ratio	Lower	Upper	
128 100			
64 54.2	38.3	57.5	
130 32.1	25.7	38.5	





#13

2-Methylphenol

Concen: 21.411 ng/ul

RT: 8.499 min Scan# 903

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :





SSTD020782

Tgt Ion:108 Resp: 287970

Ion Ratio Lower Upper

108 100

107 88.9 70.5 105.7



#14

2,2'-oxybis(1-Chloropropane)

Concen: 25.612 ng/ul

RT: 8.581 min Scan# 917

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion: 45 Resp: 383268


Ion Ratio Lower Upper

45 100

77 16.7 14.2 21.2

79 13.1 10.6 15.8









#16
Acetophenone
Concen: 21.193 ng/ul
RT: 8.881 min Scan# 968
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:105 Resp: 490490
Ion Ratio Lower Upper
105 100
77 84.1 65.4 98.2
51 28.2 21.8 32.8



#17
N-Nitroso-di-n-propylamine
Concen: 22.399 ng/ul
RT: 8.863 min Scan# 965
Delta R.T. -0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion: 70 Resp: 265208
Ion Ratio Lower Upper
70 100
42 53.3 40.9 61.3
101 9.7 7.9 11.9
130 20.3 16.6 25.0









#18
4-Methylphenol
Concen: 21.641 ng/ul
RT: 8.828 min Scan# 91
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:108 Resp: 320136
Ion Ratio Lower Upper
108 100
107 112.3 89.8 134.6



#19
Hexachloroethane
Concen: 21.386 ng/ul
RT: 9.134 min Scan# 1011
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:117 Resp: 126105
Ion Ratio Lower Upper
117 100
201 110.2 91.7 137.5
199 67.9 59.1 88.7









#22
Nitrobenzene
Concen: 21.241 ng/ul
RT: 9.263 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion: 77 Resp: 395838
Ion Ratio Lower Upper
77 100
123 39.2 32.6 48.8
65 13.5 11.2 16.8



#23
Isophorone
Concen: 20.631 ng/ul
RT: 9.793 min Scan# 1123
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion: 82 Resp: 740391
Ion Ratio Lower Upper
82 100
95 7.9 7.1 10.7
138 15.3 13.3 19.9









#25
2-Nitrophenol
Concen: 20.044 ng/ul
RT: 9.981 min Scan# 1155
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:139 Resp: 172272
Ion Ratio Lower Upper
139 100
65 60.9 46.6 70.0
109 39.2 34.5 51.7



#26
2,4-Dimethylphenol
Concen: 20.331 ng/ul
RT: 10.034 min Scan# 1164
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:107 Resp: 382483
Ion Ratio Lower Upper
107 100
121 44.4 37.1 55.7
122 74.8 61.2 91.8









#27
 Bis(2-Chloroethoxy)methane
 Concen: 20.938 ng/ul
 RT: 10.275 min Scan# 1205
 Delta R.T. 0.000 min
 Lab File: BP014180.D
 Acq: 21 Mar 2023 14:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020782


Tgt Ion: 93 Resp: 434651
 Ion Ratio Lower Upper
 93 100
 95 31.8 25.6 38.4
 123 9.7 8.5 12.7



#29
 2,4-Dichlorophenol
 Concen: 19.431 ng/ul
 RT: 10.516 min Scan# 1246
 Delta R.T. -0.006 min
 Lab File: BP014180.D
 Acq: 21 Mar 2023 14:24

Tgt Ion: 162 Resp: 300768
 Ion Ratio Lower Upper
 162 100
 164 65.2 52.0 78.0
 98 39.6 29.1 43.7









#30
Naphthalene
Concen: 20.189 ng/ul
RT: 10.922 min Scan# 1315
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:128 Resp: 967954
Ion Ratio Lower Upper
128 100
129 10.8 8.5 12.7
127 13.9 11.0 16.4



#32
4-Chloroaniline
Concen: 19.951 ng/ul
RT: 11.034 min Scan# 1334
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:127 Resp: 416894
Ion Ratio Lower Upper
127 100
129 31.6 24.9 37.3





#33

Hexachlorobutadiene

Concen: 18.155 ng/ul

RT: 11.199 min Scan# 11

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:225 Resp: 223619

Ion Ratio Lower Upper

225 100

223 59.3 48.0 72.0

227 64.2 50.3 75.5



#34

Caprolactam

Concen: 18.423 ng/ul

RT: 11.828 min Scan# 1469

Delta R.T. -0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:113 Resp: 92591


Ion Ratio Lower Upper

113 100

55 165.4 124.0 186.0

56 130.4 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 20.402 ng/ul
RT: 12.152 min Scan# 1524
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:107 Resp: 365155
Ion Ratio Lower Upper
107 100
144 23.1 19.4 29.0
142 69.0 59.6 89.4



#36
2-Methylnaphthalene
Concen: 19.952 ng/ul
RT: 12.522 min Scan# 1587
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:142 Resp: 670969
Ion Ratio Lower Upper
142 100
141 88.1 71.7 107.5





#37

1-Methylnaphthalene

Concen: 19.643 ng/ul

RT: 12.740 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:142 Resp: 663946

Ion Ratio Lower Upper

142 100

141 92.4 73.8 110.6

116 4.5 3.7 5.5



#39

1,2,4,5-Tetrachlorobenzene

Concen: 18.900 ng/ul

RT: 12.887 min Scan# 1649

Delta R.T. -0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Tgt Ion:216 Resp: 408410




Ion Ratio Lower Upper


216 100

214 78.6 61.7 92.5

179 21.0 16.2 24.4

108 16.9 13.2 19.8





#40

Hexachlorocyclopentadiene

Concen: 15.564 ng/ul

RT: 12.857 min Scan# 1

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:237 Resp: 240836

Ion Ratio Lower Upper

237 100

235 61.0 49.0 73.6

272 12.2 9.6 14.4



#41

2,4,6-Trichlorophenol

Concen: 19.816 ng/ul

RT: 13.128 min Scan# 1690

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:196 Resp: 268993


Ion Ratio Lower Upper

196 100

198 96.5 78.5 117.7

200 31.3 25.4 38.0





#42

2,4,5-Trichlorophenol

Concen: 19.455 ng/ul

RT: 13.204 min Scan# 1703

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:196 Resp: 289829

Ion Ratio Lower Upper

196 100

198 97.5 77.6 116.4

200 30.9 25.3 37.9



#43

1,1'-Biphenyl

Concen: 20.224 ng/ul

RT: 13.522 min Scan# 1757

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:154 Resp: 930407


Ion Ratio Lower Upper

154 100

153 40.0 32.6 49.0

76 13.5 9.4 14.0





#44

2-Chloronaphthalene

Concen: 20.119 ng/ul

RT: 13.569 min Scan# 11

Delta R.T. -0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:162 Resp: 735102

Ion Ratio Lower Upper

162 100

164 32.6 26.0 39.0

127 41.5 31.3 46.9



#45

2-Nitroaniline

Concen: 22.511 ng/ul

RT: 13.775 min Scan# 1800

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion: 65 Resp: 243266


Ion Ratio Lower Upper

65 100

92 67.7 56.3 84.5

138 91.7 81.4 122.2









#47
Dimethylphthalate
Concen: 19.277 ng/ul
RT: 14.140 min Scan# 1862
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:163 Resp: 967655
Ion Ratio Lower Upper
163 100
194 4.2 3.6 5.4
164 10.2 8.2 12.4



#48
2,6-Dinitrotoluene
Concen: 20.380 ng/ul
RT: 14.269 min Scan# 1884
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:165 Resp: 202252
Ion Ratio Lower Upper
165 100
89 63.8 51.7 77.5
121 21.5 17.5 26.3









#50
Acenaphthylene
Concen: 20.303 ng/ul
RT: 14.410 min Scan# 1908
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:152 Resp: 1144667
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.2
153 13.2 10.4 15.6



#51
3-Nitroaniline
Concen: 20.745 ng/ul
RT: 14.604 min Scan# 1941
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:138 Resp: 182283
Ion Ratio Lower Upper
138 100
108 12.1 10.5 15.7
92 129.2 98.5 147.7









#52
Acenaphthene
Concen: 20.132 ng/ul
RT: 14.751 min Scan# 1966
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:153 Resp: 793491
Ion Ratio Lower Upper
153 100
152 47.2 37.7 56.5
154 87.6 70.4 105.6



#53
2,4-Dinitrophenol
Concen: 16.537 ng/ul
RT: 14.810 min Scan# 1976
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:184 Resp: 122560
Ion Ratio Lower Upper
184 100
63 69.4 51.9 77.9
154 66.3 58.8 88.2









#55
4-Nitrophenol
Concen: 16.105 ng/ul
RT: 14.910 min Scan# 1993
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:109 Resp: 155468
Ion Ratio Lower Upper
109 100
139 107.1 74.5 111.7
65 123.3 86.6 130.0



#56
Dibenzofuran
Concen: 19.641 ng/ul
RT: 15.087 min Scan# 2023
Delta R.T. -0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:168 Resp: 1126178
Ion Ratio Lower Upper
168 100
139 39.2 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 19.975 ng/ul
RT: 15.051 min Scan# 2018
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:165 Resp: 296180
Ion Ratio Lower Upper
165 100
63 50.1 34.6 52.0
182 2.9 2.2 3.2



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.792 ng/ul
RT: 15.316 min Scan# 2062
Delta R.T. -0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:232 Resp: 267824
Ion Ratio Lower Upper
232 100
131 41.0 34.1 51.1
130 2.3 2.1 3.1
166 29.7 23.0 34.6









#59
Diethylphthalate
Concen: 19.529 ng/ul
RT: 15.498 min Scan# 2093
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:149 Resp: 994271
Ion Ratio Lower Upper
149 100
177 22.3 18.2 27.2
150 12.6 10.0 15.0



#61
Fluorene
Concen: 19.554 ng/ul
RT: 15.740 min Scan# 2134
Delta R.T. 0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:166 Resp: 929385
Ion Ratio Lower Upper
166 100
165 98.6 78.9 118.3
167 13.2 11.0 16.6









#62
4-Chlorophenyl-phenylether
Concen: 18.771 ng/ul
RT: 15.722 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:204 Resp: 492438
Ion Ratio Lower Upper
204 100
206 32.6 26.6 39.8
141 63.0 49.2 73.8



#63
4-Nitroaniline
Concen: 20.681 ng/ul
RT: 15.763 min Scan# 2138
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:138 Resp: 186797
Ion Ratio Lower Upper
138 100
92 60.1 44.5 66.7
108 102.3 87.2 130.8









#66
4,6-Dinitro-2-methylphenol
Concen: 18.460 ng/ul
RT: 15.816 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:198 Resp: 184802
Ion Ratio Lower Upper
198 100
182 4.6 3.0 4.6#
77 29.5 20.3 30.5



#67
N-Nitrosodiphenylamine
Concen: 20.227 ng/ul
RT: 15.940 min Scan# 2168
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:169 Resp: 793144
Ion Ratio Lower Upper
169 100
168 68.6 53.7 80.5
167 37.1 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 18.949 ng/ul
RT: 16.622 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:248 Resp: 318039
Ion Ratio Lower Upper
248 100
250 95.9 78.8 118.2
141 77.9 60.8 91.2



#69
Hexachlorobenzene
Concen: 18.492 ng/ul
RT: 16.740 min Scan# 2304
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:284 Resp: 365741
Ion Ratio Lower Upper
284 100
142 35.5 23.2 34.8#
249 32.8 25.7 38.5





#70

Atrazine

Concen: 19.049 ng/ul

RT: 16.892 min Scan# 21

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:200 Resp: 322172

Ion Ratio Lower Upper

200 100

173 24.5 19.0 28.4

215 47.0 38.5 57.7



#71

Pentachlorophenol

Concen: 17.056 ng/ul

RT: 17.092 min Scan# 2364

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:266 Resp: 220511


Ion Ratio Lower Upper

266 100

264 62.8 48.4 72.6

268 64.1 50.4 75.6





#72

Phenanthrene

Concen: 19.820 ng/ul

RT: 17.487 min Scan# 2432

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:178 Resp: 1500872

Ion Ratio Lower Upper

178 100

179 15.4 12.2 18.4

176 19.5 15.4 23.0



#74

Anthracene

Concen: 20.089 ng/ul

RT: 17.581 min Scan# 2447

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:178 Resp: 1542467


Ion Ratio Lower Upper

178 100

179 15.4 12.5 18.7

176 18.6 15.1 22.7





#75

1,2,3,4-Tetrachlorobenzene

Concen: 19.962 ng/uL

RT: 13.493 min Scan# 1

Delta R.T. -0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:216 Resp: 425335

Ion Ratio Lower Upper

216 100

214 74.8 61.4 92.2

218 48.2 39.1 58.7



#76

Pentachlorobenzene

Concen: 19.422 ng/uL

RT: 15.004 min Scan# 2009

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:250 Resp: 429885


Ion Ratio Lower Upper

250 100

248 62.8 51.0 76.4

252 64.2 50.8 76.2









#77
 Carbazole
 Concen: 20.006 ng/ul
 RT: 17.845 min Scan# 2493
 Delta R.T. -0.006 min
 Lab File: BP014180.D
 Acq: 21 Mar 2023 14:24




Instrument :
 BNA_P
 ClientSampleId :
 SSTD020782


Tgt Ion:167 Resp: 1354916
 Ion Ratio Lower Upper
 167 100
 166 21.6 17.5 26.3
 139 13.5 10.5 15.7



#78
 Di-n-butylphthalate
 Concen: 20.870 ng/ul
 RT: 18.375 min Scan# 2582
 Delta R.T. -0.006 min
 Lab File: BP014180.D
 Acq: 21 Mar 2023 14:24

Tgt Ion:149 Resp: 1766506
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.4 11.2
 104 5.6 4.6 6.8









#80
Fluoranthene
Concen: 23.018 ng/ul
RT: 19.439 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24




Instrument :
BNA_P
ClientSampleId :
SSTD020782


Tgt Ion:202 Resp: 1841892
Ion Ratio Lower Upper
202 100
101 10.7 6.6 10.0#
100 8.1 5.0 7.4#



#82
Pyrene
Concen: 22.827 ng/ul
RT: 19.798 min Scan# 2824
Delta R.T. -0.000 min
Lab File: BP014180.D
Acq: 21 Mar 2023 14:24

Tgt Ion:202 Resp: 1905468
Ion Ratio Lower Upper
202 100
101 11.7 7.8 11.8
100 9.4 6.5 9.7





#83

Butylbenzylphthalate

Concen: 24.682 ng/ul

RT: 20.651 min Scan# 2969

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:149 Resp: 830771

Ion Ratio Lower Upper

149 100

91 74.3 57.9 86.9

206 23.4 17.9 26.9



#84

3,3'-Dichlorobenzidine

Concen: 18.055 ng/ul

RT: 21.433 min Scan# 3102

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:252 Resp: 588355


Ion Ratio Lower Upper

252 100

254 64.6 51.3 76.9

126 10.2 6.9 10.3





#85

Benzo(a)anthracene

Concen: 19.961 ng/ul

RT: 21.504 min Scan# 3115

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:228 Resp: 1787117

Ion Ratio Lower Upper

228 100

229 19.9 15.5 23.3

226 27.3 22.1 33.1



#86

Bis(2-ethylhexyl)phthalate

Concen: 24.492 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:149 Resp: 1232497


Ion Ratio Lower Upper

149 100

167 28.0 22.3 33.5

279 5.3 4.3 6.5





#87

Chrysene

Concen: 19.817 ng/ul

RT: 21.563 min Scan# 3125

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:228 Resp: 1690894

Ion Ratio Lower Upper

228 100

226 30.3 24.2 36.2

229 19.6 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 29.751 ng/ul




RT: 22.363 min Scan# 3260


Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Tgt Ion:149 Resp: 2064469





#90

Benzo(b)fluoranthene

Concen: 21.242 ng/ul

RT: 23.269 min Scan# 3414

Delta R.T. 0.000 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:252 Resp: 1556617

Ion Ratio Lower Upper

252 100

253 22.4 17.3 25.9

125 8.0 5.4 8.2



#91

Benzo(k)fluoranthene

Concen: 21.260 ng/ul

RT: 23.316 min Scan# 3422

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:252 Resp: 1485582


Ion Ratio Lower Upper

252 100

253 22.3 17.4 26.0

125 8.0 5.5 8.3





#93

Benzo(a)pyrene

Concen: 19.981 ng/ul

RT: 23.927 min Scan# 3

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:252 Resp: 1273522

Ion Ratio Lower Upper

252 100

253 22.5 17.4 26.2

125 8.9 6.2 9.2



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.006 ng/ul

RT: 26.686 min Scan# 3995

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24




Tgt Ion:276 Resp: 1519243


Ion Ratio Lower Upper

276 100

138 17.9 13.1 19.7

277 24.8 20.2 30.2





#95

Dibenzo(a,h)anthracene

Concen: 18.015 ng/ul

RT: 26.704 min Scan# 3999

Delta R.T. -0.006 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Instrument :

BNA_P

ClientSampleId :

SSTD020782





Tgt Ion:278 Resp: 1263330

Ion Ratio Lower Upper

278 100

139 12.3 8.7 13.1

279 23.3 19.0 28.4



#96

Benzo(g,h,i)perylene

Concen: 17.818 ng/ul

RT: 27.504 min Scan# 4134

Delta R.T. -0.012 min

Lab File: BP014180.D

Acq: 21 Mar 2023 14:24

Tgt Ion:276 Resp: 1201672

Ion Ratio Lower Upper

276 100

138 17.1 11.9 17.9

277 23.7 18.7 28.1

