

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014198.D
 Acq On : 22 Mar 2023 01:44
 Operator : CG/JU
 Sample : PB151552BL
 Misc :
 ALS Vial : 70 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SBLK552

Quant Time: Mar 22 03:09:51 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 22 03:05:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.046	152	231121	20.000	ng/u1	0.00
20) Naphthalene-d8	10.863	136	974115	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.681	164	668628	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1538918	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.516	240	1433396	20.000	ng/u1	# 0.00
88) Perylene-d12	24.027	264	1259970	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	31949	5.223	ng/uL	0.00
4) Pyridine-d5	3.840	84	460593	27.195	ng/u1	0.00
7) Phenol-d5	7.205	99	612225	30.017	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.369	67	373265	31.380	ng/u1	0.00
11) 2-Chlorophenol-d4	7.575	132	456703	29.390	ng/u1	0.00
15) 4-Methylphenol-d8	8.757	113	498366	29.914	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	233654	30.062	ng/u1	0.00
24) 2-Nitrophenol-d4	9.946	143	275899	29.654	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	479861	27.411	ng/u1	0.00
31) 4-Chloroaniline-d4	11.004	131	686151	29.617	ng/u1	0.00
46) Dimethylphthalate-d6	14.087	166	1664434	29.332	ng/u1	0.00
49) Acenaphthylene-d8	14.381	160	1802690	29.989	ng/u1	0.00
54) 4-Nitrophenol-d4	14.892	143	245561	24.557	ng/u1	0.00
60) Fluorene-d10	15.675	176	1428676	29.708	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	277235	23.835	ng/u1	0.00
73) Anthracene-d10	17.539	188	2362072	31.618	ng/u1	0.00
81) Pyrene-d10	19.763	212	2807716	35.515	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.868	264	2110839	31.495	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
Data File : BP014198.D
Acq On : 22 Mar 2023 01:44
Operator : CG/JU
Sample : PB151552BL
Misc :
ALS Vial : 70 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
SBLK552

Quant Time: Mar 22 03:09:51 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 22 03:05:29 2023
Response via : Initial Calibration

