

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014215.D
 Acq On : 22 Mar 2023 12:05
 Operator : CG/JU
 Sample : 01956-03
 Misc :
 ALS Vial : 87 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 E0236

Quant Time: Mar 23 00:58:19 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Mar 22 03:05:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.040	152	218764	20.000	ng/u1	0.00
20) Naphthalene-d8	10.863	136	915794	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.681	164	587884	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.439	188	1264274	20.000	ng/u1	# 0.00
79) Chrysene-d12	21.516	240	927358	20.000	ng/u1	# 0.00
88) Perylene-d12	24.027	264	901030	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.411	96	20867	3.604	ng/uL	0.00
4) Pyridine-d5	3.852	84	85048	5.305	ng/u1	0.00
7) Phenol-d5	7.205	99	148685	7.702	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.363	67	347685	30.881	ng/u1	0.00
11) 2-Chlorophenol-d4	7.569	132	353907	24.062	ng/u1	0.00
15) 4-Methylphenol-d8	8.757	113	286769	18.185	ng/u1	0.00
21) Nitrobenzene-d5	9.216	128	221638	30.332	ng/u1	0.00
24) 2-Nitrophenol-d4	9.940	143	260598	29.793	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.487	165	483741	29.392	ng/u1	0.00
31) 4-Chloroaniline-d4	11.004	131	282764	12.983	ng/u1	0.00
46) Dimethylphthalate-d6	14.087	166	1739562	34.866	ng/u1	0.00
49) Acenaphthylene-d8	14.375	160	1844133	34.892	ng/u1	0.00
54) 4-Nitrophenol-d4	14.904	143	61042	6.943	ng/u1	0.01
60) Fluorene-d10	15.675	176	1454443	34.397	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.798	200	302972	31.707	ng/u1	0.00
73) Anthracene-d10	17.539	188	2273777	37.048	ng/u1	0.00
81) Pyrene-d10	19.763	212	2434329	47.595	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.868	264	1785490	37.254	ng/u1	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
Data File : BP014215.D
Acq On : 22 Mar 2023 12:05
Operator : CG/JU
Sample : 01956-03
Misc :
ALS Vial : 87 Sample Multiplier: 1

Instrument :
BNA_P
ClientSampleId :
E0236

Quant Time: Mar 23 00:58:19 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Mar 22 03:05:29 2023
Response via : Initial Calibration

