

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014138.D
 Acq On : 20 Mar 2023 13:20
 Operator : CG/JU
 Sample : 01927-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAP5

Quant Time: Mar 20 13:46:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 20 13:12:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	160355	20.000	ng/u1	0.00
20) Naphthalene-d8	10.875	136	665084	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.698	164	461962	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.451	188	1107976	20.000	ng/u1	0.00
79) Chrysene-d12	21.533	240	1141203	20.000	ng/u1	0.00
88) Perylene-d12	24.057	264	1098303	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.417	96	13983	3.295	ng/uL	0.00
4) Pyridine-d5	0.000	84	0d	0.000	ng/u1	
7) Phenol-d5	7.211	99	337329	23.838	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.375	67	196002	23.750	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	248171	23.019	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	266334	23.041	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	129804	24.460	ng/u1	0.00
24) 2-Nitrophenol-d4	9.958	143	149668	23.561	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.499	165	274350	22.953	ng/u1	0.00
31) 4-Chloroaniline-d4	11.016	131	183693	11.613	ng/u1	0.00
46) Dimethylphthalate-d6	14.099	166	1113145	28.393	ng/u1	0.00
49) Acenaphthylene-d8	14.393	160	1114343	26.831	ng/u1	0.00
54) 4-Nitrophenol-d4	14.898	143	177924	25.753	ng/u1	0.00
60) Fluorene-d10	15.687	176	946676	28.491	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.810	200	195217	23.312	ng/u1	0.00
73) Anthracene-d10	17.551	188	1719162	31.962	ng/u1	0.00
81) Pyrene-d10	19.775	212	2212669	35.155	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.898	264	1965029	33.635	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.452	88	45460	9.922	ng/uL	90
6) Benzaldehyde	7.187	77	192857	27.390	ng/u1	99
8) Phenol	7.240	94	558049	38.240	ng/u1	97
10) Bis(2-Chloroethyl)ether	7.469	93	307045	27.463	ng/u1	99
19) Hexachloroethane	9.140	117	193358	40.643	ng/u1	96
22) Nitrobenzene	9.269	77	218258	15.375	ng/u1	99
25) 2-Nitrophenol	9.987	139	196463	30.008	ng/u1	96
30) Naphthalene	10.928	128	1130897	30.965	ng/u1	99
36) 2-Methylnaphthalene	12.528	142	1164795	45.471	ng/u1	99
37) 1-Methylnaphthalene	12.746	142	589544	22.897	ng/u1	98
39) 1,2,4,5-Tetrachloroben...	12.893	216	258597	15.480	ng/u1	98
42) 2,4,5-Trichlorophenol	13.204	196	551852	47.917	ng/u1	98
43) 1,1'-Biphenyl	13.528	154	1141912	32.107	ng/u1	99
44) 2-Chloronaphthalene	13.575	162	873231	30.914	ng/u1	99
50) Acenaphthylene	14.422	152	1295556	29.724	ng/u1	100
52) Acenaphthene	14.763	153	812234	26.657	ng/u1	99
55) 4-Nitrophenol	14.916	109	158466	21.234	ng/u1	93
56) Dibenzofuran	15.093	168	616231	13.902	ng/u1	99
57) 2,4-Dinitrotoluene	15.063	165	279002	24.340	ng/u1	99
59) Diethylphthalate	15.504	149	741243	18.832	ng/u1	99
61) Fluorene	15.745	166	1174853	31.974	ng/u1	99
69) Hexachlorobenzene	16.751	284	232214	14.712	ng/u1	97
72) Phenanthrene	17.498	178	1773726	29.351	ng/u1	100
74) Anthracene	17.587	178	3081975	50.298	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014138.D
 Acq On : 20 Mar 2023 13:20
 Operator : CG/JU
 Sample : 01927-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAP5

Quant Time: Mar 20 13:46:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 20 13:12:59 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


80) Fluoranthene	19.451	202	2234890	31.152	ng/u1	98
82) Pyrene	19.804	202	2019042	26.979	ng/u1	97
85) Benzo(a)anthracene	21.516	228	1849021	23.036	ng/u1	99
87) Chrysene	21.575	228	1494588	19.538	ng/u1	100
89) Di-n-octyl phthalate	22.375	149	1806841	26.072	ng/u1	100
90) Benzo(b)fluoranthene	23.286	252	2937744	40.142	ng/u1	100
91) Benzo(k)fluoranthene	23.333	252	1483026	21.251	ng/u1	99
93) Benzo(a)pyrene	23.951	252	1571845	24.693	ng/u1	100
94) Indeno(1,2,3-cd)pyrene	26.715	276	1583263	18.789	ng/u1	100
95) Dibenzo(a,h)anthracene	26.727	278	902299	12.883	ng/u1	99
96) Benzo(g,h,i)perylene	27.539	276	2097728	31.145	ng/u1	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014138.D
 Acq On : 20 Mar 2023 13:20
 Operator : CG/JU
 Sample : 01927-01
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 DCAP5

Quant Time: Mar 20 13:46:49 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 20 13:12:59 2023
 Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 9.922 ng/uL
RT: 3.452 min Scan# 41
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion: 88 Resp: 45460
Ion Ratio Lower Upper
88 100
43 31.8 23.9 35.9
58 72.6 50.2 75.2



#6
Benzaldehyde
Concen: 27.390 ng/uL
RT: 7.187 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion: 77 Resp: 192857
Ion Ratio Lower Upper
77 100
105 92.8 74.2 111.2
106 84.9 69.8 104.8









#8
Phenol
Concen: 38.240 ng/ul
RT: 7.240 min Scan# 689
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion: 94 Resp: 558049
Ion Ratio Lower Upper
94 100
65 30.3 25.2 37.8
66 44.6 37.2 55.8



#10
Bis(2-Chloroethyl)ether
Concen: 27.463 ng/ul
RT: 7.469 min Scan# 728
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion: 93 Resp: 307045
Ion Ratio Lower Upper
93 100
63 73.0 58.8 88.2
95 31.9 25.8 38.6









#19
Hexachloroethane
Concen: 40.643 ng/ul
RT: 9.140 min Scan# 1013
Delta R.T. -0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:117 Resp: 193358
Ion Ratio Lower Upper
117 100
201 112.0 91.7 137.5
199 69.1 59.1 88.7



#22
Nitrobenzene
Concen: 15.375 ng/ul
RT: 9.269 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion: 77 Resp: 218258
Ion Ratio Lower Upper
77 100
123 41.1 32.6 48.8
65 13.6 11.2 16.8









#25
2-Nitrophenol
Concen: 30.008 ng/ul
RT: 9.987 min Scan# 1156
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:139 Resp: 196463
Ion Ratio Lower Upper
139 100
65 59.0 46.6 70.0
109 38.3 34.5 51.7



#30
Naphthalene
Concen: 30.965 ng/ul
RT: 10.928 min Scan# 1316
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:128 Resp: 1130897
Ion Ratio Lower Upper
128 100
129 11.3 8.5 12.7
127 13.8 11.0 16.4









#36
2-Methylnaphthalene
Concen: 45.471 ng/ul
RT: 12.528 min Scan# 1588
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:142 Resp: 1164795
Ion Ratio Lower Upper
142 100
141 88.5 71.7 107.5



#37
1-Methylnaphthalene
Concen: 22.897 ng/ul
RT: 12.746 min Scan# 1625
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:142 Resp: 589544
Ion Ratio Lower Upper
142 100
141 94.6 73.8 110.6
116 4.5 3.7 5.5









#39
1,2,4,5-Tetrachlorobenzene
Concen: 15.480 ng/ul
RT: 12.893 min Scan# 1650
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:	216	Resp:	258597
Ion Ratio	Lower	Upper	
216	100		
214	78.6	61.7	92.5
179	21.4	16.2	24.4
108	15.8	13.2	19.8



#42
2,4,5-Trichlorophenol
Concen: 47.917 ng/ul
RT: 13.204 min Scan# 1703
Delta R.T. -0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:	196	Resp:	551852
Ion Ratio	Lower	Upper	
196	100		
198	98.6	77.6	116.4
200	32.5	25.3	37.9









#43
1,1'-Biphenyl
Concen: 32.107 ng/ul
RT: 13.528 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:154 Resp: 1141912
Ion Ratio Lower Upper
154 100
153 40.4 32.6 49.0
76 12.7 9.4 14.0



#44
2-Chloronaphthalene
Concen: 30.914 ng/ul
RT: 13.575 min Scan# 1766
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:162 Resp: 873231
Ion Ratio Lower Upper
162 100
164 32.7 26.0 39.0
127 40.5 31.3 46.9





#50

Acenaphthylene

Concen: 29.724 ng/ul

RT: 14.422 min Scan# 1910

Delta R.T. 0.000 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20

Instrument :

BNA_P

ClientSampleId :

DCAP5





Tgt Ion:152 Resp: 1295556

Ion Ratio Lower Upper

152 100

151 20.1 16.2 24.2

153 12.9 10.4 15.6



#52

Acenaphthene

Concen: 26.657 ng/ul

RT: 14.763 min Scan# 1968

Delta R.T. 0.000 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20




Tgt Ion:153 Resp: 812234


Ion Ratio Lower Upper

153 100

152 46.9 37.7 56.5

154 86.5 70.4 105.6









#55
4-Nitrophenol
Concen: 21.234 ng/ul
RT: 14.916 min Scan# 1994
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:109 Resp: 158466
Ion Ratio Lower Upper
109 100
139 102.0 74.5 111.7
65 112.9 86.6 130.0



#56
Dibenzofuran
Concen: 13.902 ng/ul
RT: 15.093 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:168 Resp: 616231
Ion Ratio Lower Upper
168 100
139 40.5 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 24.340 ng/ul
RT: 15.063 min Scan# 2019
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:165 Resp: 279002
Ion Ratio Lower Upper
165 100
63 44.3 34.6 52.0
182 3.1 2.2 3.2



#59
Diethylphthalate
Concen: 18.832 ng/ul
RT: 15.504 min Scan# 2094
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:149 Resp: 741243
Ion Ratio Lower Upper
149 100
177 21.9 18.2 27.2
150 12.5 10.0 15.0





#61

Fluorene

Concen: 31.974 ng/ul

RT: 15.745 min Scan# 2135

Delta R.T. 0.000 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20

Instrument :

BNA_P

ClientSampleId :

DCAP5





Tgt Ion:166 Resp: 1174853

Ion Ratio Lower Upper

166 100

165 99.4 78.9 118.3

167 12.9 11.0 16.6



#69

Hexachlorobenzene

Concen: 14.712 ng/ul

RT: 16.751 min Scan# 2306

Delta R.T. 0.000 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20




Tgt Ion:284 Resp: 232214


Ion Ratio Lower Upper

284 100

142 30.9 23.2 34.8

249 31.0 25.7 38.5









#72
Phenanthrene
Concen: 29.351 ng/ul
RT: 17.498 min Scan# 2433
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:178 Resp: 1773726
Ion Ratio Lower Upper
178 100
179 15.1 12.2 18.4
176 19.1 15.4 23.0



#74
Anthracene
Concen: 50.298 ng/ul
RT: 17.587 min Scan# 2448
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:178 Resp: 3081975
Ion Ratio Lower Upper
178 100
179 15.4 12.5 18.7
176 18.7 15.1 22.7









#80
 Fluoranthene
 Concen: 31.152 ng/u1
 RT: 19.451 min Scan# 2765
 Delta R.T. 0.000 min
 Lab File: BP014138.D
 Acq: 20 Mar 2023 13:20




Instrument :
 BNA_P
 ClientSampleId :
 DCAP5


Tgt Ion:202 Resp: 2234890
 Ion Ratio Lower Upper
 202 100
 101 9.0 6.6 10.0
 100 7.0 5.0 7.4



#82
 Pyrene
 Concen: 26.979 ng/u1
 RT: 19.804 min Scan# 2825
 Delta R.T. 0.000 min
 Lab File: BP014138.D
 Acq: 20 Mar 2023 13:20

Tgt Ion:202 Resp: 2019042
 Ion Ratio Lower Upper
 202 100
 101 10.7 7.8 11.8
 100 9.0 6.5 9.7









#85
Benzo(a)anthracene
Concen: 23.036 ng/u1
RT: 21.516 min Scan# 3116
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:228 Resp: 1849021
Ion Ratio Lower Upper
228 100
229 19.6 15.5 23.3
226 26.9 22.1 33.1



#87
Chrysene
Concen: 19.538 ng/u1
RT: 21.575 min Scan# 3126
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:228 Resp: 1494588
Ion Ratio Lower Upper
228 100
226 30.1 24.2 36.2
229 19.4 15.7 23.5





#89

Di-n-octyl phthalate

Concen: 26.072 ng/ul

RT: 22.375 min Scan# 31

Delta R.T. 0.000 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20





Instrument :

BNA_P

ClientSampleId :

DCAP5

Tgt Ion:149 Resp: 1806841



#90

Benzo(b)fluoranthene

Concen: 40.142 ng/ul

RT: 23.286 min Scan# 3417

Delta R.T. 0.006 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20




Tgt Ion:252 Resp: 2937744


Ion Ratio Lower Upper

252 100

253 21.6 17.3 25.9

125 6.9 5.4 8.2









#91
Benzo(k)fluoranthene
Concen: 21.251 ng/ul
RT: 23.333 min Scan# 3425
Delta R.T. 0.000 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20




Instrument :
BNA_P
ClientSampleId :
DCAP5


Tgt Ion:252 Resp: 1483026
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 7.5 5.5 8.3



#93
Benzo(a)pyrene
Concen: 24.693 ng/ul
RT: 23.951 min Scan# 3530
Delta R.T. 0.006 min
Lab File: BP014138.D
Acq: 20 Mar 2023 13:20

Tgt Ion:252 Resp: 1571845
Ion Ratio Lower Upper
252 100
253 21.8 17.4 26.2
125 8.1 6.2 9.2





#94

Indeno(1,2,3-cd)pyrene

Concen: 18.789 ng/ul

RT: 26.715 min Scan# 4000

Delta R.T. 0.000 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20

Instrument :

BNA_P

ClientSampleId :

DCAP5





Tgt Ion:276 Resp: 1583263

Ion Ratio Lower Upper

276 100

138 16.9 13.1 19.7

277 25.2 20.2 30.2



#95

Dibenzo(a,h)anthracene

Concen: 12.883 ng/ul

RT: 26.727 min Scan# 4002

Delta R.T. -0.006 min

Lab File: BP014138.D

Acq: 20 Mar 2023 13:20




Tgt Ion:278 Resp: 902299


Ion Ratio Lower Upper

278 100

139 12.6 8.7 13.1

279 23.7 19.0 28.4





#96
 Benzo(g,h,i)perylene
 Concen: 31.145 ng/ul
 RT: 27.539 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BP014138.D
 Acq: 20 Mar 2023 13:20

Instrument :
 BNA_P
 ClientSampleId :
 DCAP5

Tgt Ion:276 Resp: 2097728

Ion	Ratio	Lower	Upper
276	100		
138	15.4	11.9	17.9
277	24.2	18.7	28.1

