

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014154.D
 Acq On : 20 Mar 2023 23:02
 Operator : CG/JU
 Sample : PB151502BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS502

Quant Time: Mar 21 00:49:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 20 13:12:59 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Christian Giraldo 03/21/2023
 Supervised By :Jagrut Upadhyay 03/21/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.052	152	169422	20.000	ng/u1	0.00
20) Naphthalene-d8	10.875	136	723822	20.000	ng/u1	0.00
38) Acenaphthene-d10	14.692	164	493505	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.445	188	1151994	20.000	ng/u1	#-0.01
79) Chrysene-d12	21.527	240	1213834	20.000	ng/u1	# 0.00
88) Perylene-d12	24.051	264	1247300	20.000	ng/u1	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.416	96	24105	5.376	ng/uL	0.00
4) Pyridine-d5	3.846	84	300384	24.194	ng/u1	0.00
7) Phenol-d5	7.210	99	482407	32.266	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.375	67	287281	32.947	ng/u1	0.00
11) 2-Chlorophenol-d4	7.581	132	359224	31.536	ng/u1	0.00
15) 4-Methylphenol-d8	8.769	113	399301	32.696	ng/u1	0.00
21) Nitrobenzene-d5	9.228	128	185248	32.076	ng/u1	0.00
24) 2-Nitrophenol-d4	9.951	143	217601	31.475	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.493	165	403551	31.023	ng/u1	0.00
31) 4-Chloroaniline-d4	11.016	131	458443	26.631	ng/u1	0.00
46) Dimethylphthalate-d6	14.098	166	1282656	30.625	ng/u1	0.00
49) Acenaphthylene-d8	14.386	160	1399170	31.536	ng/u1	0.00
54) 4-Nitrophenol-d4	14.898	143	233078	31.579	ng/u1	0.00
60) Fluorene-d10	15.686	176	1087649	30.642	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.810	200	267847	30.763	ng/u1	0.00
73) Anthracene-d10	17.551	188	1725876	30.861	ng/u1	0.00
81) Pyrene-d10	19.769	212	2202655	32.902	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.886	264	2149820	32.403	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	49253	10.175	ng/uL	94
5) Pyridine	3.864	79	320744	25.475	ng/u1	94
6) Benzaldehyde	7.187	77	246724	33.165	ng/u1	99
8) Phenol	7.234	94	484648	31.433	ng/u1	99
10) Bis(2-Chloroethyl)ether	7.469	93	374955	31.743	ng/u1	97
12) 2-Chlorophenol	7.616	128	361374	31.488	ng/u1	98
13) 2-Methylphenol	8.499	108	372790	32.516	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.581	45	467925	36.681	ng/u1	99
16) Acetophenone	8.887	105	624882	31.673	ng/u1	98
17) N-Nitroso-di-n-propyla...	8.869	70	333923	33.084	ng/u1	98
18) 4-Methylphenol	8.828	108	407963	32.351	ng/u1	97
19) Hexachloroethane	9.140	117	159890	31.810	ng/u1	98
22) Nitrobenzene	9.269	77	497891	32.228	ng/u1	100
23) Isophorone	9.798	82	950955	31.963	ng/u1	98
25) 2-Nitrophenol	9.987	139	225564	31.657	ng/u1	96
26) 2,4-Dimethylphenol	10.040	107	460356	29.518	ng/u1	99
27) Bis(2-Chloroethoxy)met...	10.275	93	547967	31.842	ng/u1	100
29) 2,4-Dichlorophenol	10.522	162	387319	30.185	ng/u1	98
30) Naphthalene	10.928	128	1220189	30.699	ng/u1	99
32) 4-Chloroaniline	11.040	127	413259	23.857	ng/u1	97
33) Hexachlorobutadiene	11.204	225	293998	28.793	ng/u1	99
34) Caprolactam	11.834	113	133576m	32.060	ng/u1	
35) 4-Chloro-3-methylphenol	12.157	107	471915	31.805	ng/u1	98
36) 2-Methylnaphthalene	12.528	142	843857	30.269	ng/u1	99

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP032023\
 Data File : BP014154.D
 Acq On : 20 Mar 2023 23:02
 Operator : CG/JU
 Sample : PB151502BS
 Misc :
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 BNA_P
 ClientSampleId :
 SLCS502


Quant Time: Mar 21 00:49:45 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\SFAM-EPA-BP030723.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Mon Mar 20 13:12:59 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : Christian Giraldo 03/21/2023
 Supervised By : Jagrut Upadhyay 03/21/2023


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.745	142	867009	30.941	ng/u1	99
39) 1,2,4,5-Tetrachloroben...	12.892	216	532900	29.861	ng/u1	99
40) Hexachlorocyclopentadiene	12.863	237	299458	23.433	ng/u1	98
41) 2,4,6-Trichlorophenol	13.128	196	346120	30.874	ng/u1	98
42) 2,4,5-Trichlorophenol	13.204	196	384207	31.228	ng/u1	99
43) 1,1'-Biphenyl	13.528	154	1188733	31.287	ng/u1	99
44) 2-Chloronaphthalene	13.575	162	954552	31.633	ng/u1	99
45) 2-Nitroaniline	13.781	65	322932	36.183	ng/u1	93
47) Dimethylphthalate	14.145	163	1278543	30.841	ng/u1	100
48) 2,6-Dinitrotoluene	14.269	165	271141	33.082	ng/u1	99
50) Acenaphthylene	14.416	152	1485256	31.899	ng/u1	100
51) 3-Nitroaniline	14.604	138	241488	33.277	ng/u1	93
52) Acenaphthene	14.757	153	1025519	31.505	ng/u1	99
53) 2,4-Dinitrophenol	14.810	184	177371	28.979	ng/u1	97
55) 4-Nitrophenol	14.910	109	227657	28.555	ng/u1	92
56) Dibenzofuran	15.092	168	1451997	30.662	ng/u1	99
57) 2,4-Dinitrotoluene	15.057	165	396874	32.409	ng/u1	96
58) 2,3,4,6-Tetrachlorophenol	15.316	232	351009	29.821	ng/u1	99
59) Diethylphthalate	15.504	149	1305689	31.053	ng/u1	100
61) Fluorene	15.739	166	1204774	30.692	ng/u1	99
62) 4-Chlorophenyl-phenyle...	15.728	204	644266	29.737	ng/u1	100
63) 4-Nitroaniline	15.769	138	276035	37.005	ng/u1	95
66) 4,6-Dinitro-2-methylph...	15.822	198	265440	31.955	ng/u1	96
67) N-Nitrosodiphenylamine	15.945	169	1032445	31.733	ng/u1	99
68) 4-Bromophenyl-phenylether	16.628	248	423957	30.442	ng/u1	99
69) Hexachlorobenzene	16.745	284	496160	30.233	ng/u1	96
70) Atrazine	16.898	200	277580	19.781	ng/u1	98
71) Pentachlorophenol	17.098	266	299018	27.875	ng/u1	98
72) Phenanthrene	17.492	178	1974206	31.420	ng/u1	100
74) Anthracene	17.580	178	1978825	31.061	ng/u1	99
75) 1,2,3,4-Tetrachloroben...	13.492	216	547958	30.994	ng/uL	99
76) Pentachlorobenzene	15.010	250	556013	30.275	ng/uL	98
77) Carbazole	17.851	167	1802539	32.077	ng/u1	100
78) Di-n-butylphthalate	18.380	149	2351369	33.481	ng/u1	100
80) Fluoranthene	19.445	202	2495526	32.704	ng/u1#	97
82) Pyrene	19.798	202	2586988	32.499	ng/u1	97
83) Butylbenzylphthalate	20.651	149	1141420	35.562	ng/u1	96
84) 3,3'-Dichlorobenzidine	21.439	252	886486	28.529	ng/u1	99
85) Benzo(a)anthracene	21.510	228	2701215	31.639	ng/u1	99
86) Bis(2-ethylhexyl)phtha...	21.404	149	1718040	35.803	ng/u1	99
87) Chrysene	21.568	228	2544958	31.278	ng/u1	100
89) Di-n-octyl phthalate	22.368	149	3032101	38.526	ng/u1	100
90) Benzo(b)fluoranthene	23.274	252	2693913	32.413	ng/u1	99
91) Benzo(k)fluoranthene	23.327	252	2642540	33.343	ng/u1	99
93) Benzo(a)pyrene	23.939	252	2293583	31.727	ng/u1	99
94) Indeno(1,2,3-cd)pyrene	26.703	276	2642005	27.608	ng/u1	98
95) Dibenzo(a,h)anthracene	26.715	278	2210169	27.787	ng/u1	98
96) Benzo(g,h,i)perylene	27.521	276	2068610	27.044	ng/u1	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2
1,4-Dioxane
Concen: 10.175 ng/uL
RT: 3.446 min Scan# 44
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion: 88 Resp: 49253
Ion Ratio Lower Upper
88 100
43 30.5 23.9 35.9
58 68.9 50.2 75.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#5
Pyridine
Concen: 25.475 ng/uL
RT: 3.864 min Scan# 115
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion: 79 Resp: 320744
Ion Ratio Lower Upper
79 100
52 76.1 55.1 82.7
51 35.1 28.9 43.3





#6
Benzaldehyde
Concen: 33.165 ng/ul
RT: 7.187 min Scan# 680
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion: 77 Resp: 246724
Ion Ratio Lower Upper
77 100
105 91.0 74.2 111.2
106 88.3 69.8 104.8

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#8
Phenol
Concen: 31.433 ng/ul
RT: 7.234 min Scan# 688
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion: 94 Resp: 484648
Ion Ratio Lower Upper
94 100
65 31.4 25.2 37.8
66 45.8 37.2 55.8





#10
Bis(2-Chloroethyl)ether
Concen: 31.743 ng/ul
RT: 7.469 min Scan# 728
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion: 93 Resp: 374955
Ion Ratio Lower Upper
93 100
63 76.5 58.8 88.2
95 32.1 25.8 38.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#12
2-Chlorophenol
Concen: 31.488 ng/ul
RT: 7.616 min Scan# 753
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:128 Resp: 361374
Ion Ratio Lower Upper
128 100
64 49.9 38.3 57.5
130 32.1 25.7 38.5





#13

2-Methylphenol

Concen: 32.516 ng/ul

RT: 8.499 min Scan# 903

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02

Instrument :

BNA_P

ClientSampleId :

SLCS502

Tgt Ion:108 Resp: 372790

Ion Ratio Lower Upper

108 100





107 90.0 70.5 105.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/21/2023

Supervised By :Jagrut Upadhyay 03/21/2023



#14

2,2'-oxybis(1-Chloropropane)

Concen: 36.681 ng/ul

RT: 8.581 min Scan# 917

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02




Tgt Ion: 45 Resp: 467925


Ion Ratio Lower Upper

45 100

77 17.1 14.2 21.2

79 12.8 10.6 15.8









#16
Acetophenone
Concen: 31.673 ng/ul
RT: 8.887 min Scan# 969
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion: 105 Resp: 624882
Ion Ratio Lower Upper
105 100
77 83.5 65.4 98.2
51 27.9 21.8 32.8




Manual Integrations
APPROVED


Reviewed By : Christian Giraldo 03/21/2023
Supervised By : Jagrut Upadhyay 03/21/2023



#17
N-Nitroso-di-n-propylamine
Concen: 33.084 ng/ul
RT: 8.869 min Scan# 966
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion: 70 Resp: 333923
Ion Ratio Lower Upper
70 100
42 52.5 40.9 61.3
101 10.1 7.9 11.9
130 19.2 16.6 25.0









#18
4-Methylphenol
Concen: 32.351 ng/ul
RT: 8.828 min Scan# 91
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:108 Resp: 407963
Ion Ratio Lower Upper
108 100
107 115.1 89.8 134.6




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#19
Hexachloroethane
Concen: 31.810 ng/ul
RT: 9.140 min Scan# 1012
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:117 Resp: 159890
Ion Ratio Lower Upper
117 100
201 111.6 91.7 137.5
199 72.9 59.1 88.7





#22
Nitrobenzene
Concen: 32.228 ng/ul
RT: 9.269 min Scan# 1034
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion: 77 Resp: 497891
Ion Ratio Lower Upper
77 100
123 40.4 32.6 48.8
65 14.1 11.2 16.8




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#23
Isophorone
Concen: 31.963 ng/ul
RT: 9.798 min Scan# 1124
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion: 82 Resp: 950955
Ion Ratio Lower Upper
82 100
95 8.3 7.1 10.7
138 15.8 13.3 19.9









#25
2-Nitrophenol
Concen: 31.657 ng/ul
RT: 9.987 min Scan# 1156
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:139 Resp: 225564
Ion Ratio Lower Upper
139 100
65 61.1 46.6 70.0
109 40.0 34.5 51.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#26
2,4-Dimethylphenol
Concen: 29.518 ng/ul
RT: 10.040 min Scan# 1165
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:107 Resp: 460356
Ion Ratio Lower Upper
107 100
121 47.2 37.1 55.7
122 76.6 61.2 91.8





#27
Bis(2-Chloroethoxy)methane
Concen: 31.842 ng/ul
RT: 10.275 min Scan# 1205
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion: 93 Resp: 547962
Ion Ratio Lower Upper
93 100
95 32.2 25.6 38.4
123 10.5 8.5 12.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#29
2,4-Dichlorophenol
Concen: 30.185 ng/ul
RT: 10.522 min Scan# 1247
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:162 Resp: 387319
Ion Ratio Lower Upper
162 100
164 65.8 52.0 78.0
98 39.2 29.1 43.7





#30
Naphthalene
Concen: 30.699 ng/ul
RT: 10.928 min Scan# 1316
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:128 Resp: 1220189
Ion Ratio Lower Upper
128 100
129 11.0 8.5 12.7
127 13.2 11.0 16.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#32
4-Chloroaniline
Concen: 23.857 ng/ul
RT: 11.040 min Scan# 1335
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion:127 Resp: 413259
Ion Ratio Lower Upper
127 100
129 32.7 24.9 37.3





#33
Hexachlorobutadiene
Concen: 28.793 ng/ul
RT: 11.204 min Scan# 1363
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion:225 Resp: 293998
Ion Ratio Lower Upper
225 100
223 60.0 48.0 72.0
227 64.4 50.3 75.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#34
Caprolactam
Concen: 32.060 ng/ul m
RT: 11.834 min Scan# 1470
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:113 Resp: 133576
Ion Ratio Lower Upper
113 100
55 163.9 124.0 186.0
56 131.9 102.0 153.0









#35
4-Chloro-3-methylphenol
Concen: 31.805 ng/ul
RT: 12.157 min Scan# 1525
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:107 Resp: 471915
Ion Ratio Lower Upper
107 100
144 23.6 19.4 29.0
142 73.1 59.6 89.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#36
2-Methylnaphthalene
Concen: 30.269 ng/ul
RT: 12.528 min Scan# 1588
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:142 Resp: 843857
Ion Ratio Lower Upper
142 100
141 88.9 71.7 107.5





#37
1-Methylnaphthalene
Concen: 30.941 ng/ul
RT: 12.745 min Scan# 1625
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion:142 Resp: 867009
Ion Ratio Lower Upper
142 100
141 93.5 73.8 110.6
116 4.7 3.7 5.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#39
1,2,4,5-Tetrachlorobenzene
Concen: 29.861 ng/ul
RT: 12.892 min Scan# 1650
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:216 Resp: 532900
Ion Ratio Lower Upper
216 100
214 76.7 61.7 92.5
179 20.6 16.2 24.4
108 15.7 13.2 19.8





#40
Hexachlorocyclopentadiene
Concen: 23.433 ng/ul
RT: 12.863 min Scan# 1645
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:237 Resp: 299458
Ion Ratio Lower Upper
237 100
235 63.1 49.0 73.6
272 12.4 9.6 14.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#41
2,4,6-Trichlorophenol
Concen: 30.874 ng/ul
RT: 13.128 min Scan# 1690
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02



Tgt Ion:196 Resp: 346120
Ion Ratio Lower Upper
196 100
198 95.4 78.5 117.7
200 31.7 25.4 38.0









#42
2,4,5-Trichlorophenol
Concen: 31.228 ng/ul
RT: 13.204 min Scan# 1704
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:196 Resp: 38420
Ion Ratio Lower Upper
196 100
198 97.3 77.6 116.4
200 30.8 25.3 37.9




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#43
1,1'-Biphenyl
Concen: 31.287 ng/ul
RT: 13.528 min Scan# 1758
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:154 Resp: 1188733
Ion Ratio Lower Upper
154 100
153 40.3 32.6 49.0
76 12.8 9.4 14.0





#44
2-Chloronaphthalene
Concen: 31.633 ng/ul
RT: 13.575 min Scan# 1766
Delta R.T. -0.000 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:162 Resp: 954552
Ion Ratio Lower Upper
162 100
164 32.1 26.0 39.0
127 40.2 31.3 46.9

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#45
2-Nitroaniline
Concen: 36.183 ng/ul
RT: 13.781 min Scan# 1801
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion: 65 Resp: 322932
Ion Ratio Lower Upper
65 100
92 67.7 56.3 84.5
138 92.8 81.4 122.2





#47
Dimethylphthalate
Concen: 30.841 ng/ul
RT: 14.145 min Scan# 1863
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:163 Resp: 1278543
Ion Ratio Lower Upper
163 100
194 4.2 3.6 5.4
164 10.3 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#48
2,6-Dinitrotoluene
Concen: 33.082 ng/ul
RT: 14.269 min Scan# 1884
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion:165 Resp: 271141
Ion Ratio Lower Upper
165 100
89 65.7 51.7 77.5
121 21.3 17.5 26.3





#50
Acenaphthylene
Concen: 31.899 ng/ul
RT: 14.416 min Scan# 1909
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:152 Resp: 1485256
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.2
153 13.2 10.4 15.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#51
3-Nitroaniline
Concen: 33.277 ng/ul
RT: 14.604 min Scan# 1941
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion:138 Resp: 241488
Ion Ratio Lower Upper
138 100
108 12.7 10.5 15.7
92 132.0 98.5 147.7





#52
Acenaphthene
Concen: 31.505 ng/ul
RT: 14.757 min Scan# 1967
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:153 Resp: 1025519
Ion Ratio Lower Upper
153 100
152 46.7 37.7 56.5
154 87.5 70.4 105.6

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#53
2,4-Dinitrophenol
Concen: 28.979 ng/ul
RT: 14.810 min Scan# 1976
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion:184 Resp: 177371
Ion Ratio Lower Upper
184 100
63 69.0 51.9 77.9
154 74.0 58.8 88.2





#55
4-Nitrophenol
Concen: 28.555 ng/ul
RT: 14.910 min Scan# 1994
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion:109 Resp: 227652
Ion Ratio Lower Upper
109 100
139 101.5 74.5 111.7
65 116.8 86.6 130.0




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#56
Dibenzofuran
Concen: 30.662 ng/ul
RT: 15.092 min Scan# 2024
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:168 Resp: 1451997
Ion Ratio Lower Upper
168 100
139 40.5 31.8 47.6









#57
2,4-Dinitrotoluene
Concen: 32.409 ng/ul
RT: 15.057 min Scan# 2019
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:165 Resp: 396874
Ion Ratio Lower Upper
165 100
63 46.3 34.6 52.0
182 2.8 2.2 3.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#58
2,3,4,6-Tetrachlorophenol
Concen: 29.821 ng/ul
RT: 15.316 min Scan# 2062
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:232 Resp: 351009
Ion Ratio Lower Upper
232 100
131 42.7 34.1 51.1
130 2.4 2.1 3.1
166 30.0 23.0 34.6





#59

Diethylphthalate

Concen: 31.053 ng/ul

RT: 15.504 min Scan# 2094

Delta R.T. -0.000 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02

Instrument :

BNA_P

ClientSampleId :

SLCS502

Tgt Ion:149 Resp: 1305689

Ion Ratio Lower Upper

149 100

177 22.9 18.2 27.2





150 12.7 10.0 15.0

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/21/2023

Supervised By :Jagrut Upadhyay 03/21/2023



#61

Fluorene

Concen: 30.692 ng/ul

RT: 15.739 min Scan# 2134

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02




Tgt Ion:166 Resp: 1204774


Ion Ratio Lower Upper

166 100

165 97.6 78.9 118.3


167 13.3 11.0 16.6





#62
4-Chlorophenyl-phenylether
Concen: 29.737 ng/ul
RT: 15.728 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion: 204 Resp: 644266
Ion Ratio Lower Upper
204 100
206 32.9 26.6 39.8
141 61.7 49.2 73.8




Manual Integrations
APPROVED


Reviewed By : Christian Giraldo 03/21/2023
Supervised By : Jagrut Upadhyay 03/21/2023



#63
4-Nitroaniline
Concen: 37.005 ng/ul
RT: 15.769 min Scan# 2139
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion: 138 Resp: 276035
Ion Ratio Lower Upper
138 100
92 59.4 44.5 66.7
108 104.4 87.2 130.8





#66
4,6-Dinitro-2-methylphenol
Concen: 31.955 ng/ul
RT: 15.822 min Scan# 2149
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion:198 Resp: 265440
Ion Ratio Lower Upper
198 100
182 3.6 3.0 4.6
77 27.6 20.3 30.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#67
N-Nitrosodiphenylamine
Concen: 31.733 ng/ul
RT: 15.945 min Scan# 2169
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:169 Resp: 1032445
Ion Ratio Lower Upper
169 100
168 68.1 53.7 80.5
167 36.6 28.6 43.0









#68
4-Bromophenyl-phenylether
Concen: 30.442 ng/ul
RT: 16.628 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:248 Resp: 423952
Ion Ratio Lower Upper
248 100
250 97.3 78.8 118.2
141 75.8 60.8 91.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#69
Hexachlorobenzene
Concen: 30.233 ng/ul
RT: 16.745 min Scan# 2305
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:284 Resp: 496160
Ion Ratio Lower Upper
284 100
142 33.5 23.2 34.8
249 32.5 25.7 38.5





#70

Atrazine

Concen: 19.781 ng/ul

RT: 16.898 min Scan# 2331

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02

Instrument :

BNA_P

ClientSampleId :

SLCS502

Tgt Ion:200 Resp: 277580

Ion Ratio Lower Upper

200 100

173 25.8 19.0 28.4





215 47.5 38.5 57.7

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/21/2023

Supervised By :Jagrut Upadhyay 03/21/2023



#71

Pentachlorophenol

Concen: 27.875 ng/ul

RT: 17.098 min Scan# 2365

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02




Tgt Ion:266 Resp: 299018


Ion Ratio Lower Upper

266 100

264 62.8 48.4 72.6


268 63.6 50.4 75.6





#72
Phenanthrene
Concen: 31.420 ng/ul
RT: 17.492 min Scan# 2432
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:178 Resp: 1974206
Ion Ratio Lower Upper
178 100
179 15.4 12.2 18.4
176 19.1 15.4 23.0

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#74
Anthracene
Concen: 31.061 ng/ul
RT: 17.580 min Scan# 2447
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02



Tgt Ion:178 Resp: 1978825
Ion Ratio Lower Upper
178 100
179 15.5 12.5 18.7
176 18.5 15.1 22.7









#75
1,2,3,4-Tetrachlorobenzene
Concen: 30.994 ng/uL
RT: 13.492 min Scan# 1753
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:216 Resp: 547958
Ion Ratio Lower Upper
216 100
214 77.6 61.4 92.2
218 49.5 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#76
Pentachlorobenzene
Concen: 30.275 ng/uL
RT: 15.010 min Scan# 2010
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:250 Resp: 556013
Ion Ratio Lower Upper
250 100
248 64.6 51.0 76.4
252 65.4 50.8 76.2





#77
 Carbazole
 Concen: 32.077 ng/ul
 RT: 17.851 min Scan# 2493
 Delta R.T. -0.006 min
 Lab File: BP014154.D
 Acq: 20 Mar 2023 23:02




Instrument :
 BNA_P
 ClientSampleId :
 SLCS502




Tgt Ion:167 Resp: 1802539
 Ion Ratio Lower Upper
 167 100
 166 21.9 17.5 26.3
 139 13.4 10.5 15.7

Manual Integrations
 APPROVED



Reviewed By :Christian Giraldo 03/21/2023
 Supervised By :Jagrut Upadhyay 03/21/2023




#78
 Di-n-butylphthalate
 Concen: 33.481 ng/ul
 RT: 18.380 min Scan# 2583
 Delta R.T. -0.006 min
 Lab File: BP014154.D
 Acq: 20 Mar 2023 23:02




Tgt Ion:149 Resp: 2351369
 Ion Ratio Lower Upper
 149 100
 150 9.3 7.4 11.2
 104 5.5 4.6 6.8





#80
Fluoranthene
Concen: 32.704 ng/u1
RT: 19.445 min Scan# 21
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion:202 Resp: 2495526
Ion Ratio Lower Upper
202 100
101 9.4 6.6 10.0
100 7.5 5.0 7.4




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#82
Pyrene
Concen: 32.499 ng/u1
RT: 19.798 min Scan# 2824
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:202 Resp: 2586988
Ion Ratio Lower Upper
202 100
101 11.2 7.8 11.8
100 9.1 6.5 9.7





#83

Butylbenzylphthalate

Concen: 35.562 ng/ul

RT: 20.651 min Scan# 2970

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02

Instrument :

BNA_P

ClientSampleId :

SLCS502

Tgt Ion:149 Resp: 1141420

Ion Ratio Lower Upper

149 100

91 76.8 57.9 86.9





206 21.9 17.9 26.9

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/21/2023

Supervised By :Jagrut Upadhyay 03/21/2023



#84

3,3'-Dichlorobenzidine

Concen: 28.529 ng/ul

RT: 21.439 min Scan# 3103

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02




Tgt Ion:252 Resp: 886486


Ion Ratio Lower Upper

252 100

254 64.7 51.3 76.9

126 9.2 6.9 10.3





#85

Benzo(a)anthracene

Concen: 31.639 ng/ul

RT: 21.510 min Scan# 3116

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02

Instrument :

BNA_P

ClientSampleId :

SLCS502

Tgt Ion:228 Resp: 2701215

Ion Ratio Lower Upper

228 100

229 20.0 15.5 23.3





226 27.4 22.1 33.1

Manual Integrations

APPROVED

Reviewed By :Christian Giraldo 03/21/2023

Supervised By :Jagrut Upadhyay 03/21/2023



#86

Bis(2-ethylhexyl)phthalate

Concen: 35.803 ng/ul

RT: 21.404 min Scan# 3097

Delta R.T. -0.006 min

Lab File: BP014154.D

Acq: 20 Mar 2023 23:02




Tgt Ion:149 Resp: 1718040


Ion Ratio Lower Upper

149 100

167 27.7 22.3 33.5


279 5.0 4.3 6.5





#87
Chrysene
Concen: 31.278 ng/ul
RT: 21.568 min Scan# 3125
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502



Tgt Ion:228 Resp: 2544958
Ion Ratio Lower Upper
228 100
226 30.0 24.2 36.2
229 19.4 15.7 23.5




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#89
Di-n-octyl phthalate
Concen: 38.526 ng/ul
RT: 22.368 min Scan# 3261
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Tgt Ion:149 Resp: 3032101









#90
Benzo(b)fluoranthene
Concen: 32.413 ng/ul
RT: 23.274 min Scan# 3416
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02

Instrument :
BNA_P
ClientSampleId :
SLCS502

Tgt Ion:252 Resp: 2693913
Ion Ratio Lower Upper
252 100
253 21.9 17.3 25.9
125 7.9 5.4 8.2




Manual Integrations
APPROVED


Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023



#91
Benzo(k)fluoranthene
Concen: 33.343 ng/ul
RT: 23.327 min Scan# 3424
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02


Tgt Ion:252 Resp: 2642540
Ion Ratio Lower Upper
252 100
253 21.9 17.4 26.0
125 7.5 5.5 8.3





#93
Benzo(a)pyrene
Concen: 31.727 ng/ul
RT: 23.939 min Scan# 31
Delta R.T. -0.006 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Instrument :
BNA_P
ClientSampleId :
SLCS502




Tgt Ion:252 Resp: 2293583
Ion Ratio Lower Upper
252 100
253 21.6 17.4 26.2
125 8.2 6.2 9.2

Manual Integrations
APPROVED



Reviewed By :Christian Giraldo 03/21/2023
Supervised By :Jagrut Upadhyay 03/21/2023




#94
Indeno(1,2,3-cd)pyrene
Concen: 27.608 ng/ul
RT: 26.703 min Scan# 3998
Delta R.T. -0.012 min
Lab File: BP014154.D
Acq: 20 Mar 2023 23:02




Tgt Ion:276 Resp: 2642005
Ion Ratio Lower Upper
276 100
138 18.4 13.1 19.7
277 25.5 20.2 30.2





#95
 Dibenzo(a,h)anthracene
 Concen: 27.787 ng/ul
 RT: 26.715 min Scan# 4002
 Delta R.T. -0.018 min
 Lab File: BP014154.D
 Acq: 20 Mar 2023 23:02




Instrument :
 BNA_P
 ClientSampleId :
 SLCS502



Tgt Ion:278 Resp: 2210169
 Ion Ratio Lower Upper
 278 100
 139 12.1 8.7 13.1
 279 23.3 19.0 28.4

Manual Integrations
 APPROVED

Reviewed By :Christian Giraldo 03/21/2023
 Supervised By :Jagrut Upadhyay 03/21/2023



#96
 Benzo(g,h,i)perylene
 Concen: 27.044 ng/ul
 RT: 27.521 min Scan# 4137
 Delta R.T. -0.018 min
 Lab File: BP014154.D
 Acq: 20 Mar 2023 23:02

Tgt Ion:276 Resp: 2068610
 Ion Ratio Lower Upper
 276 100
 138 16.2 11.9 17.9
 277 24.3 18.7 28.1

